

Reference excerpt 02-037

Source: arXiv:1604.05177v1 | License: CC BY 4.0 | Word window: 1367-1626

Figure 3: Examples of CSA and ECG trace are shown. The waves a, x, v e y are correctly detected by an algorithm. The P, R e T events of the ECG trace are indicated. Figure 4: Examples of CSA and ECG trace are shown. The x and v e y are not correctly detected. Correct waves are indicated. The P, R e T events of the ECG are indicated. 3 DISCUSSION This note describes a methodology to extrapolate some inter - esting quantitative parameters from the IJV CSA trace, whic h is an analogue of the JVP. There are two important aspects that must be considered: i) the contouring of the IJV on each sonogram of the video-clip must be as accurate as possible. This means that the algorithm presented in [Sisini et al.(20 15)] must be improved to achieve an accuracy close to 100 %, while the current one and about 95 %. ii) The wave detection algo- rithm must be improved. In fact it is not always sucient to identify the waves, while they can be identied by the analys is of tracks performed by a human expert. This paper has not discussed the clinical use of these parameters. ACKNOWLEDGEMENT The author wants to thank Giacomo Gadda, Valentina Tavoni and Valentina Sisini to have kindly review the manuscript. REFERENCES [Applefeld(1990)] Applefeld MM. The Jugular Venous Pres- sure and Pulse Contour. In: SourceClinical Methods: The History, Physical, and Laboratory Examinations. Boston: Butterworths, 1990: Chapter 19. [Beulen et al.(2011)] Beulen, B. et al. Toward noninvasive blood pressure assessment in