D. Fit for the intended use

Criterion	Parameter	Estimate [CI]	CI level	H_0	Decision
Test 1 (accuracy)	alpha (mM)	0.980 [0.880, 1.084]	95 %	0	rejected
Test 2 (precision)	beta	0.056 [0.043, 0.070]	95 %	0	rejected
Test 3 (agreement)	alpha (mM)	-0.665 [-0.881, -0.473]	97.5 %	0	rejected
Test 3 (agreement)	beta	1.105 [1.071, 1.142]	97.5 %	1	rejected
Limits of agreement	vs acceptable limits	[-1.0, 1.0] mM			outside