

Calibration Certificate

ISO/IEC 17025:2017 · Accredited Calibration Laboratory

Instrument

Certificate Number	CAL-2026-DIM-04471
Description	Digital Caliper — Absolute Digimatic Caliper
Manufacturer	Mitutoyo
Model	500-196-30 (Series 500)
Serial Number	21B-7745099
Measuring Range	0 to 150 mm (0 to 6 in)
Resolution	0.01 mm

Calibration

Calibration Date	15 January 2026
Calibration Due Date	15 January 2027
Calibration Interval	12 months
Condition as received	In tolerance

Traceability

The measurement results are metrologically traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST), USA, using NIST-traceable grade 0 gauge blocks.

Measurement Uncertainty

The expanded measurement uncertainty is ± 0.015 mm, reported at a coverage factor $k = 2$, corresponding to a confidence level of approximately 95 percent.

Results

Nominal (mm)	Measured (mm)	Deviation (mm)	Tolerance (mm)
0.00	0.00	0.00	± 0.02
25.00	25.01	+0.01	± 0.02
50.00	50.00	0.00	± 0.02
100.00	100.01	+0.01	± 0.03

150.00	149.99	-0.01	±0.03
--------	--------	-------	-------

Reference standard: NIST-traceable gauge block set, grade 0. Calibration procedure SOP-DIM-01.
Ambient 20 ± 1 C, 50 percent RH. This certificate shall not be reproduced except in full. DEMO SAMPLE
— not a real calibration record.