

A Newcomer's Guide to Functional Near Infrared Spectroscopy Experiments

Rand K. Almajidy, Kunal Mankodiya[✉], Member, IEEE, Mohammadreza Abtahi, and Ulrich G. Hofmann[✉]

(Methodological Review)

Abstract—This review presents a practical primer for functional near-infrared spectroscopy (fNIRS) with respect to technology, experimentation, and analysis software. Its purpose is to jump-start interested practitioners considering utilizing a non-invasive, versatile, nevertheless challenging window into the brain using optical methods. We briefly recapitulate relevant anatomical and optical foundations and give a short historical overview. We describe competing types of illumination (trans-illumination, reflectance, and differential reflectance) and data collection methods (continuous wave, time domain and frequency domain). Basic components (light sources, detection, and recording components) of fNIRS systems are presented. Advantages and limitations of fNIRS techniques are offered, followed by a list of very practical recommendations for its use. A variety of experimental and clinical studies with fNIRS are sampled, shedding light on many brain-related ailments. Finally, we describe and discuss a number of freely available analysis and presentation packages suited for data analysis. In conclusion, we recommend fNIRS due to its ever-growing body of clinical applications, state-of-the-art neuroimaging technique and manageable hardware requirements. It can be safely concluded that fNIRS adds a new arrow to the quiver of neuro-medical examinations due to both its great versatility and limited costs.

Index Terms—Functional near infrared spectroscopy, fNIRS, neuroimaging.

I. INTRODUCTION

THE brain is undoubtedly one of the most complex structures known to humankind, as evidenced by its sheer numbers of neurons (ca. 10^{11}), supported in the cortex by about four times as many glial cells [1], and building some 10^{14}

Manuscript received October 11, 2018; revised February 18, 2019 and August 7, 2019; accepted August 28, 2019. Date of publication October 17, 2019; date of current version January 20, 2020. This work was supported in part by the National Science Foundation under Grants 1565962 and 1539068, and in part by the German Research Foundations Cluster of Excellence Brainlinks-Braintools (EXC 1086). (Corresponding author: Kunal Mankodiya.)

R. K. Almajidy and U. G. Hofmann are with the Neuroelectronic Systems, Faculty of Medicine, Department of Neurosurgery, Medical Center, University of Freiburg, 79085 Freiburg im Breisgau, Germany (e-mail: rand.almajidy@klinikum.uni-freiburg.de; ulrich.hofmann@coregen.uni-freiburg.de).

K. Mankodiya and M. Abtahi are with the Wearable Biosensing Lab, Department of Electrical, Computer, and Biomedical Engineering, University of Rhode Island, Kingston, RI 02881 USA (e-mail: kunalm@uri.edu; mabtahi2012@uri.edu).

Digital Object Identifier 10.1109/RBME.2019.2944351

synaptic connections [2]. As such, grasping the inner workings and functions of the human brain is among the most profound and far-reaching challenges of our time. This quest for understanding promises new treatments for brain disorders, fundamental discoveries about the brain's functions, and impactful applications spanning from neuro-medicine and live brain monitoring to new communication devices. Unfortunately, the progress is slow, and there are many open questions, not least because of a lack of unobtrusive, high resolution and fast measurement systems for natural environments. This leads to oversimplifications of the brain's workings. For example, it is a very common misconception to view the brain as simply a collection of neurons, ignoring the essential roles of both glial cells and blood supply in the brain. It is estimated that almost every neuron has its own nourishing capillary, altogether constituting a 400-mile supply infrastructure [3].

As neurons do not maintain any substantial provisions of oxygen or glucose, an increase in neural activity due to computational workload has to be followed by an increase in blood supply by vessels [4], most likely triggered by chemical signaling from the neurons themselves [5]. However, the true relationship between local neural activity and the resulting adaptations in cerebral hemodynamics, called neurovascular coupling (NVC), is not fully understood. Most investigations into NVC employ expensive, bulky, stationary functional magnetic resonance imaging (fMRI) devices with limited time resolution. Despite the superior spatial resolution offered by fMRI, its high cost, low temporal resolution, and limited mobility represent a challenge for many researchers.

Thankfully, this situation might improve with the emergence of functional near infrared spectroscopy (fNIRS) systems which provide a portable and less costly imaging modality for cerebral hemodynamics [6]. Similar to the BOLD [Blood-Oxygen-Level Dependent] signal—the hallmark of fMRI, fNIRS data relies on NVC. However, fNIRS spatial resolution is limited when compared to that of fMRI signals. As the acronym suggests, fNIRS uses near-infrared light of wavelengths longer than visible at 750 nm–1200 nm and benefits from the particular optical properties of tissue regarding this low energy radiation [6].

II. fNIRS PRINCIPLES AND THEORY

Different brain imaging techniques measure changes in tissues' physical or chemical properties, including during brain

activity. These changes are then translated in accordance with prior knowledge of the tissues' properties principles of the measurement techniques into data that reflects changes in brain activity.

fMRI BOLD signals employ changes in the blood's magnetic susceptibility during neural activity to measure changes in brain activity [4]. fNIRS on the other hand, monitors changes in optical properties of tissues, primarily blood's absorbance, during neural activity to measure that activity. fNIRS employs light in the NIR range for that purpose. Hence its beneficial for fNIRS researchers to understand the light propagation principles and the tissues' optical properties that govern fNIRS.

Light propagation depends on the light's wavelength and the medium's optical properties which govern incident light's reflection, scattering, and absorption. Absorption depends on the medium's chemical constitution [6], whereas scattering (considered as a deviation from a straight trajectory), is influenced by many parameters such as wavelength and particulate consistency [7]. Reflection, on the other hand, depends on the incident angle between the light and tissue, and on the materials optical density [6].

Absorbed light is dissipated as heat in the absorber medium and its molecular makeup determines the specific wavelength at which maximal absorption occurs [8]. The most important chromophores, or chemical groups absorbing light at specific wavelengths [9], in healthy perfused tissue are oxygenated hemoglobin HbO₂, deoxygenated hemoglobin Hb, their sum - total hemoglobin HbT [10], and Cytochrome c oxidase [11], [12]. These concentrations change over time and with oxygen concentration [13].

Near infrared light displays advantageous propagation characteristics in biological tissue, with limited absorbance by water or relevant chromophores in tissue ("Optical Window"). Light above 1200 nm is predominantly absorbed by the tissue's water content [6].

Absorption is quantified by the molar extinction coefficient a as a function of wavelength and shows to what extent the chromophore absorbs light at that wavelength. It results in a unique absorption spectrum for each chromophore [13]. The prominent Cytochrome c oxidase (Caa3 in Fig. 1) is not used as an indicator for tissue oxygenation, as it is a mitochondrial enzyme representing intracellular oxygenation whose concentration relies on factors other than changes in oxygen [14]. Instead, Hb and HbO₂ concentrations are of primary interest in tissue monitoring.

The computations translating NIR photons collected from the body surface to information about tissue activation depend upon the optical properties of these illuminated tissues. The following paragraphs will introduce some of these optical properties with the description on how to deduce the diffusion paths of NIR photons through tissue, and how to estimate the concentration changes of Hb and HbO₂ using these properties.

A. Beer-Lambert Law

As both absorption and scattering contribute to light attenuation, both parameters should be considered in NIRS. The Beer-Lambert law relates light attenuation by absorption to




Fig. 1. Absorption spectra of Hb, HbO₂, H₂O and other chromophores in NIR range (redrawn after Murkin and Arango 2009 [11]). The isosbestic point of the Hb/HbO₂ absorption spectrum is circled in white within the NIR optical window.

chromophore concentration:

$$A = -\log_{10} \left(\frac{I}{I_o} \right) = a * c * d$$

The attenuation A of incident light is given by the logarithmic ratio of the intensity of the received light (I) to the intensity of the source light (I_o). This equals the product of the molar extinction coefficient a , the molar chromophore concentration c , and the distance between the light source and the detector d .

The second major factor in light attenuation is scattering, where a photon's trajectory is changed by an interaction with matter without substantial energy loss. Scattering is the dominant mechanism of light propagation in biological tissue. Mie-scattering (where the scatterer's dimension is similar to the incident wavelength) is weakly wavelength dependent. The human head is composed of many different layers with unique densities and thicknesses, resulting in many different scattering paths for NIR light. Thus, skin, bone and cerebral matter must be treated carefully in simulations [10].

A photon crossing through a medium containing a uniform distribution of identical scatterers may be scattered away from its straight path with a probability $P_s(z)$ over a distance z . This probability is characterized by its scattering coefficient μ_s , the inverse of the distance a photon may cover without being deflected in 1/cm. The inverse of μ_s can be interpreted as the scattering free mean path length m_{fps} , or the average distance a photon travels before scattering events.

To correct for a tissue's anisotropic scattering properties, μ_s must be corrected to the reduced scattering coefficient μ'_s , which considers the anisotropy factor g [9], [15]:

$$\mu'_s = \mu_s (1 - g)$$

Typical reduced scattering coefficients for grey and white matter in the brain are 11.8 1/cm and 11.1 1/cm at the 760 nm

and 830 nm wavelengths respectively, that are generally used in fNIRS.

Since NIR light photons suffer more from scattering than absorption in body tissues, NIR photon diffusion through the body can be described (and simulated) as a random walk with a step size of $1/\mu\text{s}^2$ [cm] [16].

B. Modified Beer-Lambert Law

As photons do not travel the distance from source to detector in a straight line, but instead follow a random path and thus travel greater distances than d , this increased true path distance is introduced as the differential path length (DP). To modify the Beer-Lambert law, the differential path length factor (DPF) is introduced [17] as follows:

$$A = a * c * d * DPF + G$$

In the modified Beer-Lambert law, the attenuation is not linearly related to the extinction coefficient because of the unknown term G , which includes the effect of the shape of the optodes and the scattering factor. Therefore, it is not possible to calculate exact chromophore concentrations with the modified Beer-Lambert law. However, by assuming G to be constant for all chromophores, it is possible to eliminate G from the equations and calculate changes in the chromophore's concentrations [18]. Those can be found by assuming that d and DPF are constant over the experimental time frame.

Fig. 1 depicts the NIR sweet spot of low water absorbance between 700 nm and 950 nm, with Hb and HbO₂ spectra crossing at an isosbestic point around 805 nm. The unknowns from the equation can be eliminated by solving the Beer-Lambert law for two (or more) measurement wavelengths on either side of the isosbestic point. This helps to actually find changes in Hb and HbO₂ concentrations [13], [18], [19]. As we detect small changes in attenuation for both wavelengths

$$\Delta A_{\lambda_1} = a_{Hb}^{\lambda_1} \cdot L \cdot [Hb] + a_{HbO_2}^{\lambda_1} \cdot L \cdot [HbO_2]$$

$$\Delta A_{\lambda_2} = a_{Hb}^{\lambda_2} \cdot L \cdot [Hb] + a_{HbO_2}^{\lambda_2} \cdot L \cdot [HbO_2]$$

with L representing the total mean path length $d * DPF$, a representing the respective extinction coefficients, and the respective concentrations denoted as $[HbX]$, we can solve for the concentrations [20] :

$$[HbO_2] = \frac{a_{HbO_2}^{\lambda_2} \cdot \Delta A_{\lambda_1} - a_{Hb}^{\lambda_1} \cdot \Delta A_{\lambda_2}}{L \cdot (a_{HbO_2}^{\lambda_1} \cdot a_{Hb}^{\lambda_2} - a_{HbO_2}^{\lambda_2} \cdot a_{Hb}^{\lambda_1})}$$

$$[Hb] = \frac{a_{HbO_2}^{\lambda_1} \cdot \Delta A_{\lambda_2} - a_{Hb}^{\lambda_2} \cdot \Delta A_{\lambda_1}}{L \cdot (a_{HbO_2}^{\lambda_1} \cdot a_{Hb}^{\lambda_2} - a_{HbO_2}^{\lambda_2} \cdot a_{Hb}^{\lambda_1})}$$

Additional wavelengths may be used to measure the concentration of other chromophores such as Cytochrome c oxidase and water or to improve the accuracy of Hb and HbO₂ concentration measurements [14], [18]. It is feasible to determine NIR wavelengths that can minimize errors in calculations carried out using these equations, introduced by the assumptions above [19].

III. CHRONOLOGY OF fNIRS EVOLUTION

Glenn Millikan's attempt to measure oxygen concentration in well-perfused muscle (with an Oximeter) in the 1940s [21] is considered the origin of optical sensing methods [22], [23]. Frans Jöbsis presented one of his first efforts to measure blood oxygenation levels and its variation in a cat's brain using trans-illumination spectroscopy in 1977 [6]. He explained the relative transparency of brain tissues to NIR light and demonstrated the feasibility of monitoring changes in the brain's Hb oxygenation using NIRS [6]. These experiments and his subsequent research [24] made him the founder of *in vivo* NIRS.

In the 1980s, Marco Ferrari started to measure brain oxygenation and its changes in animals. His results further confirmed that NIR light can efficiently detect blood oxygenation changes [25], [26]. In 1985, Ferrari carried out experiments to monitor blood oxygenation changes in human adults using custom-made NIRS instruments [27]. These experiments and those by Brazy and his colleagues (including Jöbsis) to monitor preterm infants cerebral oxygenation [28] represent the first successful applications of NIRS in human patients. In 1986, Ferrari and colleagues presented more cerebrovascular measurements from neonates [29] and cerebrovascular patients. The data showed the effect of carotid artery compression on regional cerebral blood volume and oxygenation [30]. The same time frame witnessed the first quantitative data showing HbO₂, Hb and HbT changes collected from sick infants' cerebral blood by David Delpy and his colleagues (including M Cope). They employed a custom-made, four wavelength trans-illumination NIRS system to monitor oxygenation level changes [31]. Their findings paved the way for NIRS' use as a bedside cerebral oxygenation monitor. Their 1988 experiments provided the hemoglobin absorption spectra at different NIR wavelengths, facilitating the quantification of NIRS data collected from the brain [12] and estimation of the optical path length of NIR light through the rat brain [17]. They also presented a description of their system [32], which served as the base design for the first commercial NIRS system produced by Hamamatsu Photonics K.K. (Hamamatsu City, Japan) in 1989.

Delpy and his co-workers efforts to accurately calculate the optical path length for NIR photons with time-of-flight measurements [17] were extended by Duncan and colleagues, who collected precise differential path length factor (DPF) values, the absolute path lengths divided by the distances between sources and detectors, using phase resolved spectroscopy from 100 adults and 35 newborn infants heads. Their findings indicated differences between infants and adults, between males and females, and between wavelengths [33]. It was not until 1993 that the first human fNIRS systems measurements were published. These experiments utilized single-channel fNIRS systems and included the work of Hoshi [34], Chance [35], Villringer [36], Kato [37] and Okada [38]. Hoshi's data showed an increase in HbO₂ and a decrease in Hb in the relevant area during brain activation from a cognitive task (solving an arithmetic problem). This change is associated with an increase in cerebral blood flow and was more prominent in younger subjects than in adults [34]. Chance and colleagues interpreted the variation in




Fig. 2. A chronology showing some of the early experiments and developments contributing to the evolution of fNIRS.

blood oxygen concentration in a relatively limited illuminated tissue (a banana shape) as a measure of brain activation during a similar cognitive task (problem solving). They concluded that NIRS can monitor localized brain activity [35].

In the year of 1993, Villringer and colleagues used a NIRS setup to assess hemodynamic changes in the brain during cognitive tasks and visual stimulation. Their results demonstrated that NIRS can indeed record changes in brain activity and not only hemodynamic changes in the skin [36]. Kato and colleagues investigated HbO₂ and Hb changes during visual stimulation [37]. Okada and colleagues (including Hoshi and Tamura) documented differences due to handedness and gender [38]. They also published the first clinical use of fNIRS with schizophrenic patients in 1997 [39].

Gratton and colleagues tested the feasibility of employing NIRS for optical scanning. They provided the first evidence of the indirect relationship between changes in the brain's optical properties and neuronal activity. They showed the feasibility of measurement in reflectance mode by demonstrating the interaction of near infrared light shined from a strong light source with chromophores. The interaction happened on the long, random path of the light through tissue. A sensitive photodetector has detected some backscattered photons [40]. These efforts supported the emergence of the first multi-channel fNIRS system (for details see the review by Ferrari and Quaresima [41]).

Currently, fNIRS is a very useful neuroimaging technique because of its lower cost and greater portability than fMRI or

PET. It has found its place in clinical and research settings monitoring cerebral functionality related to vision [42], hearing [43], speech [44], motor tasks [45], learning [46], and emotional stimuli [47].

NIRS signals are highly correlated with regional cerebral blood flow and thus cost-effectively augment PET or fMRI's BOLD measurements [48], [49]. They can therefore shed light on the coupling of hemodynamic responses with neuronal activity [50], [51], as revealed by electroencephalography (EEG) [52].

NIRS performance is influenced by handedness and gender [38], and in particular by aging [53] due to changes in the optical properties of scalp tissues [54] and decreased brain activation [55]. The increasing popularity of fNIRS is owed to its portability, its moderate spatial and temporal resolution, its ease of use, and its ability to scan slowly moving human subjects. All these advantages have encouraged researchers to utilize it individually or along with other modalities such as fMRI [8], [56]–[58] or EEG [59]–[64].

IV. DIFFERENT TYPES OF NIRS

The different applications of NIRS systems require a basic understanding of their principles. There are three main types of NIRS systems: I) Continuous Wave (CW), II) Time Domain and III) Frequency Domain spectrometers (see Fig. 3). Each type has its strengths and weaknesses. Therefore, researchers




Fig. 3. Principles of three different NIRS types.

must design their experiments in accordance with system characteristics.

- **Continuous Wave NIRS:** CW-NIRS is the oldest and the most common commercially used NIRS system (reviewed by [65]). This type of device uses multiple wavelength sources and measures the attenuation of light (see Fig. 3) with a photodiode or a photodetector [66]. Compared to other types of NIRS, CW-NIRS has advantages in simplicity, size, weight, and cost. However, it is very difficult to separate attenuation from absorption and scattering, and some systems have a small light penetration depth due to short source-detector distance (SDD) [66].
- **Time Domain NIRS:** Also known as time-resolved or time-of-flight NIRS systems, a solid-state laser is usually used to provide very short but powerful pulses. Light attenuation is measured by very sensitive special cameras or even single photon counters sorting them based on their arrival time [66] (reviewed by [67]). This type of NIRS has the advantage of higher accuracy and spatial resolution [68], [69], but it is limited by the system's bulkiness and higher cost [70].
- **Frequency Domain NIRS:** Also known as frequency-resolved or intensity modulated NIRS systems, an LED, laser diode [68], [71], [72], or white light source usually provides input light. These systems measure the attenuation, phase shift and modulation depth of the light with respect to the systems' incident light [70], [73]. They exploit the linear relation between the optical path length and a phase shift for the frequencies <200 MHz [73], [74]. A gain-modulated area detector or a photon counting device is used to take measurements [66], [75].

V. NIRS ILLUMINATION MODES (FIGURE 4)

- 1) **Trans-illumination:** This mode is applicable to newborn infants [32]. Due to changes in the optical properties of scalp tissues as a function of aging and the increased head size, this mode is not used with adults.
- 2) **Reflectance:** This mode is utilized in most current NIRS devices [70], [76]. In reflectance mode, the penetration depth of NIRS is estimated at around 1/3 of the SDD [9].
- 3) **Differential reflectance:** More than one NIR detector (or source) is utilized to measure the difference between extra and intra cranial light paths [76].

VI. NIRS SYSTEMS

The major factors controlling the efficiency of NIRS are 1) the type of NIR source and detector, 2) the efficiency of NIR transmission into/collection from the tissue, and 3) the accuracy of tissue optical property coefficients and models used to calculate HbO_2 and Hb [78].

Previously, NIRS penetration's depth was limited to 3 mm of the skull [79]. However, current instrumentation allows the light to reach up to 1–2 cm in depth [9], depending on several factors including NIR light radiant energy [9], the optical properties of the head beneath the NIRS optodes (NIR source/detector), the SDD [80], and the detector area [81].

Although increasing SDD is believed to increase the penetration depth [82], data quality deteriorates with increasing SDD beyond specific limits [83].

Hence, SDDs between 2–4cm [79], [81], [83], [84] are usually employed for NIRS systems. In the implementations with longer SDDs, small detectors are associated with unstable DPF. Hence, detectors must be chosen in accordance with SDDs [81]. There are currently efforts underway to im-




Fig. 4. Different NIRS operation modes: trans-illuminance, reflectance and differential reflectance (redrawn after [76], [77]).




Fig. 5. Multiparametric comparison of fNIRS (blue), EEG (red), fMRI (yellow), PET (purple) and MEG (green). (a) Depth resolution, relation to neural activity, intrusivity, temporal resolution, spatial resolution, estimated system costs, and running costs. (b) Handling, size, head fixture, ease of long term studies, subjects' mobility, environment, safety, and portability.

prove fidelity in both light sources and detectors to achieve timing precision, permitting zero source-detector distances (OSD) and thus improving the localization of hemodynamic responses [67].

NIRS systems contain the following main components:

- **NIR Light Source:** In most cases, the light source is either a light emitting diode (LED) [85]–[89] or a laser diode [34], [48], [90]–[92] with 670 and 890nm [88], 730, 805, and 850nm [93] or 760 and 850 nm (widely used combinations by developers and commercial systems) [65], [85] wavelengths. LEDs are often preferred for safety reasons [94]. Picosecond lasers are used in experimental OSD systems [67].
- **NIR Detector:** Photodiodes [85], [90], [95]–[97], avalanche photodiodes [87], [88], [92], [98] or photo-multiplier tubes (PMT) [32], [54], [65], [99], are usually utilized as NIRS detectors. They exhibit low wavelength selectivity and thus caution must be taken to block or avoid ambient light.
- **Control and Data Collection Electronics:** NIRS sources and detectors must be controlled by sophisticated electronic circuits. Data collected by NIRS detectors are either amplified and saved on the same hardware or transmitted tethered [85] or wirelessly [86], [87], [96], [100] to another electronic circuit or a computer where

further amplification, noise reduction and signal analysis is performed.

- **NIR Light Transfer Module:** NIR light is shone directly to the scalp from the NIR source [85], [97] or conveyed by optic fibers [101]. The reflected NIR light is either received from the head directly by the NIR detector [97], [100] or guided via optic fibers to the NIR detector [85], [88], [102], [103].

VII. ADVANTAGES

Portable, Low Power and Low Cost: A 16 channel (=16 dual sources and 2 detectors) NIRS setup can be powered with a single 3.6 V–1000 mAh battery [100]. fNIRS devices can be designed to be portable for employment with freely moving subjects. NIRS can be utilized at bed sides [104], in an emergency situation, or in an ambulance [89], [105]–[107]. A basic system may cost around \$10.000 [65], with lower operational costs than MRI [9].

Non-invasive and Safe: LEDs and even laser diodes can stay well below the critical heat deposition of 0.2 W/cm^2 (at 630 nm) to 0.4 W/cm^2 (at 850 nm) [9] known to cause pain or heat damage to the skin [105], [108], [109].

Easy Preparation and Setup: No special skin adhesive is required to attach optodes to the scalp. Optodes in most

systems are reusable and last for long periods over many measurements [110], [111]. Optodes' cleaning after employment is generally easy but may depend on the manufacturers.

Motion Artifacts: Motion artifacts are less pronounced for minimized fNIRS systems where NIRS optodes and the controlling circuits are in close proximity and both attached to the body [107] than in fNIRS systems employing optical fibers. The fibers may shift position during vigorous motion affecting optical coupling [112].

SNR and Temporal and Spatial Resolution: fNIRS sampling rate may exceed 25 Hz/channel versus 1 Hz for fMRI [113]. In other words, fMRI provides brain images at the rate of 1 frame/second while fNIRS could provide 25 images per second. Therefore, the temporal resolution of fNIRS is considerably higher than that of fMRI, but slightly lower than that of EEG, as illustrated in Fig. 5(a). However, the spatial resolution of fMRI is far greater than that of fNIRS. The spatial resolution of fNIRS is also slightly lower than that of EEG [114]. This is depicted in Fig. 5(a) as well. Cui *et al.* (2011) performed a detailed comparison of fNIRS and fMRI signals in the temporal and spatial domains. They reported a weaker signal to noise ratio (SNR) in fNIRS than in fMRI. Although the SNR is reported to be weaker, the signals are highly correlated. In the spatial domain, they reported that the banana-shaped path of the photons is strongly correlated to the BOLD signal [115].

Application in Special Populations: fNIRS is feasible for patients with implanted devices. For example, no interference in the fNIRS optodes have been detected in patients with pacemakers [116]. fNIRS is more convenient for young and claustrophobic patients as compared to fMRI [9], [117]. Patient acceptance is better in fNIRS as the narrow "tube" of fMRI is avoided [107].

Response compared to fMRI: Minati *et al.* (2011) recorded NIRS and fMRI simultaneously in event-related visual stimulation. They measured inter-subject coefficients of variation (CVs) for the response peak amplitude and reported considerably larger CVs for NIRS than in fMRI. The inter-subject CVs for response latencies and intra-subject CVs for response amplitudes are reported to be comparable in their study [118].

VIII. LIMITATIONS

Susceptibility to Ambient Light: As the light transfer module always contains a minimal air gap between transducer and skin or glass fiber and skin, it is challenging to avoid ambient light influencing measurements [94], [102]. As such, the placement of optodes on the head to send and receive light in the proper angle is important.

Shallow Penetration: fNIRS cannot reach the deeper areas of the brain with its shallow penetration depth of around 1–3 cm in the cortex [110]. As fNIRS in reflectance mode depends on photons scattered towards the sensor, the number of photons decays exponentially with SDD. This cannot be easily compensated for with higher flux as damage to the skin must be avoided at all costs. Consequently, hemodynamic responses from deeper brain structures may not be measured with a simple fNIRS device [113].




Fig. 6. Example of the author's spring loaded fNIRS (a) source, (b) detector, and (c) exemplary montage of optodes and comparable EEG channels map.




Fig. 7. NIR light path between source and detector [129].

Low Temporal Resolution (compared to EEG): Although fNIRS measurement is more rapid than fMRI, its temporal resolution is lower than that of EEG [119], [120] which displays between 1 msec [104] and 100 msec [121] time constants. Consequently, utilizing fNIRS in applications such as brain-computer interfaces (BCI) requires longer task periods [107], [111], [122]. However, this may be inadequate to monitor delays in activity between brain areas [123].

Low Spatial Resolution (compared to fMRI): fNIRS' spatial resolution is quite limited to about 1cm [57], [113], compared to fMRI's millimeter voxel sizes [104], [110]. Even with the picturesque simplification of a "banana-shaped" light path between source and sensor, it is hard to talk about a single "spatial resolution," given that absorption acts in an integrative manner along each individual photon's path.

Noise, Artifacts, and Interference: fNIRS offers sometimes noisy channels with a small bandwidth, so in line with the Shannon-Hartley theorem, the reported information transfer rate is low at about 4 bits/min [110], [124], [125]. As fNIRS is an optical method, the presence of hair -especially dark hair- in the region of interest (ROI) may block light and reduce signal strength both entering and exiting the skull [9]. A longer preparation period may be necessary to ensure minimal hair presence below optodes [106]. Hence, pre-experimental preparation time and signal strength depend on the ROI. Although the signals are not affected by muscle artifacts from body motion, signal quality may still be negatively affected by head movement [89], [112]. This may cause fluctuations in the efficiency of light transfer. Noise due to various physiological oscillations around 0.1 Hz Mayer waves are reported, which are caused




Fig. 8. NIRS data visualization and analysis employing different software (a) 2D visualization of brain activation by employing HOMER2, (b) 3D visualization of brain activation, (c) NIRS data filtering, and (d) blood oxygenation change calculation utilizing NIRS lab.

by slow changes in blood pressure [107]. Respiratory oscillations (0.2–0.5 Hz) [125] and heartbeat artifacts (1–1.5 Hz) [107] were found as well. Extracranial activity is reflected in NIRS data [79], [126]–[129]. This contribution may have a strong impact on data accuracy in some cases.

Participant Discomfort in Long-term Use: An easy way to minimize ambient light artifacts, optodes placements, and optimize coupling in the light transfer module is to press the working ends of glass fibers or LEDs into the skin. However, this becomes uncomfortable after a while, causing stress or headaches which can affect experimental trials [89], [102], [130]–[132].

IX. RECOMMENDATIONS

Spring-Loaded Optodes: One of the challenges with the fNIRS optodes and cap is to handle the hair. As shown in Fig. 6(a) and (b) [9], [94], [133], a spring-loaded mechanism have been developed and implemented to accomplish: 1) parting the hair away from the light path and 2) sustaining secure pressure between the optode and skull [89], [106].

Optode Shielding and Special Caps: Studies have shown that ambient light has a significant influence on the performance of fNIRS [134]. It is generally a good practice to shield optodes from ambient light with dark caps. For example, 3D printing and laser cutting technologies are used to design optode capsules made of dark materials encapsulating optodes to reduce the influence of ambient light [132]. In addition to dark shields for the optodes, fNIRS caps are also covered with a black overcap to further reduce ambient light. Employing special caps [130], [131] and/or secondary caps to hold the optodes can support

secure optode-skin coupling and minimize ambient light [89] and motion artifacts [112].

Safety: The long-term use of fNIRS may elevate the temperature at the contact point of the light source and the scalp [135]. In general, commercial fNIRS systems have to pass safety or regulatory standards. However, when employing a laser light source, care must be taken to prevent eye or skin injury [94]. Study participants should be communicated with regarding safety limits of the fNIRS device in use. Laboratory designed fNIRS systems must meet the requirements of IEC80601-2-71:2015, that is they must be designed to regulate the basic safety and essential performance of fNIRS equipment.

Signal Quality: To reduce extracerebral or superficial influences, optodes with different SDD are employed. Data collected from the short SDDs will indicate superficial activity (Fig. 7) which can be then be isolated from deeper brain activity by means of proper modeling [89]. It is always important to implement appropriate approaches to filtering, noise reduction [9], [110], [112] and channel rejection for channels with weak/extremely noisy signals.

X. fNIRS APPLICATIONS

As fNIRS is a rather mature technology that remains open to new developments and creative implementations, it has led to an ever-increasing field of applications. Naturally, they are all based on the hemodynamic response of the brain under a wealth of conditions and research paradigms. These experiments are at least augmenting, and in some cases replacing, the use of more expensive, stationary imaging modalities. It is therefore used in functional connectivity and cognitive neuroscience experi-

TABLE I
SELECTED STUDIES USING fNIRS TECHNOLOGY FOR VARIOUS APPLICATIONS

Application	Neuroimaging modality	Citation (year)
Language studies	fNIRS	Lei <i>et al.</i> (2018) [136], Watanabe <i>et al.</i> (2016) [137], Takahashi <i>et al.</i> (2015) [138], Rossi <i>et al.</i> (2012, a review) [120], Quaresima et.al (2012, a review) [139], Minagawa-Kawai <i>et al.</i> (2011) [140].
	fNIRS-EEG	Walloon <i>et al.</i> (2012, a review) [62].
Brain functional connectivity	fNIRS	Bu <i>et al.</i> (2018) [141], Vergotte <i>et al.</i> (2018) [142], Racz <i>et al.</i> (2017) [143], Gallagher <i>et al.</i> (2016) [144], Wang <i>et al.</i> (2016) [145], Medvedev <i>et al.</i> (2011) [146]. Resting-state: Wang <i>et al.</i> (2017) [147], Zhang <i>et al.</i> (2011) [148], Sasai <i>et al.</i> (2011) [149].
	fMRI-fNIRS	Sasai <i>et al.</i> (2012) [150], Duan <i>et al.</i> (2012) [151].
Psychiatry	fNIRS	Ohi <i>et al.</i> (2017) [152], Lin <i>et al.</i> (2017) [153], Okada <i>et al.</i> (2016) [154], Ehliis <i>et al.</i> (2014, a review) [155], Matsuzawa <i>et al.</i> (2012) [156]. Schizophrenia: Luo <i>et al.</i> (2018) [157], Kumar <i>et al.</i> (2017, a review) [158], Nod <i>et al.</i> (2017) [159], Koike <i>et al.</i> (2013, a review) [160], Chou <i>et al.</i> (2017) [161], (2014) [162]. Depression: Nishizawa <i>et al.</i> (2019) [163], Kondo <i>et al.</i> (2018) [164], Fu <i>et al.</i> (2018) [165], Hirano <i>et al.</i> (2017) [166], Kawano <i>et al.</i> (2016) [167], Zhang <i>et al.</i> (2015, a review) [168].
	fNIRS-EEG	Epilepsy: Sannagowdara <i>et al.</i> (2018) [169], Bourel-Ponchel <i>et al.</i> (2017) [170], Manoochehri <i>et al.</i> (2017) [171], Modir <i>et al.</i> (2017) [172], Peng <i>et al.</i> (2016, a review) [173].
Rehabilitation	fNIRS	Neuro: Mihara and Miyai (2016, a review) [174], Balconi (2016, a review) [175], van Dokkum <i>et al.</i> (2015) [176]. Motor: Bae <i>et al.</i> (2017) [177], Chang <i>et al.</i> (2014) [178], Rea <i>et al.</i> (2014) [179], Lin <i>et al.</i> (2009, a review) [114].
	fNIRS-EEG	Yamamoto <i>et al.</i> (2018) [180].
Anesthetic depth	fNIRS	Liang <i>et al.</i> (2016) [181], Srensen (2016) [182], Hernandez-Meza <i>et al</i> (2015, a review) [183], Leon-Dominguez <i>et al.</i> (2014) [184].
Aging studies	fNIRS	Cognitive: Agbangla <i>et al.</i> (2017, a review) [185], Li <i>et al.</i> (2018, a review) [186], Uemura <i>et al.</i> (2016) [187]. Sensory: Lin <i>et al.</i> (2017) [188]. Discourse comprehension: Martin <i>et al.</i> (2018) [189]. Anxiety: Adorni <i>et al.</i> (2018) [190]. Microvascular dysfunction: Rosenberry <i>et al.</i> (2018) [191].
Cognitive neuroscience	fNIRS	Takeda <i>et al.</i> (2017) [192], Causse <i>et al.</i> (2017) [193], Ozawa and Hiraki (2017) [194], Keshmiri <i>et al.</i> (2017) [195], Fishburn <i>et al.</i> (2014) [196], Byun <i>et al.</i> (2014) [197], Cutini <i>et al.</i> (2012) [101].
	fNIRS-EEG	Liu <i>et al.</i> (2017) [198], Omurtag <i>et al.</i> (2017) [59].
BOLD signal	fMRI-fNIRS	Emir <i>et al.</i> (2008) [199], Schroeter <i>et al.</i> (2006) [200], Steinbrink <i>et al.</i> (2006, a review) [8].
Brain-computer interface (BCI)	fNIRS	Naseer <i>et al.</i> (2015, a review) [107], Qureshi <i>et al.</i> (2017) [201], Shin <i>et al.</i> (2016) [202], Chaudhary <i>et al.</i> (2015, a review) [104], Shin and Jeong(2014) [125].
	fNIRS-EEG	Shin <i>et al.</i> (2018) [203], Hong <i>et al.</i> (2018, a review) [204], Min <i>et al.</i> (2010, a review) [111], Tomita <i>et al.</i> (2014) [205].
Brain motor activity	fNIRS	Herold <i>et al.</i> (2017, a review) [206], Abdalmalak <i>et al.</i> (2017) [207], Nishiyori <i>et al.</i> (2016) [208], Iso <i>et al.</i> (2016) [209], Drenckhahn <i>et al.</i> (2015) [123].
Driving research	fNIRS	Liu <i>et al.</i> (2016, a review) [210], Sturman <i>et al.</i> (2018) [211].

ments, as well as in neurological diagnostics or rehabilitation and neuroimaging and, not the least, as communication modality. **Table I** lists a deliberately incomplete list of applications from the literature.

XI. ANALYSIS SOFTWARE

No matter what fNIRS hardware is used to perform the desired experiments, whether custom built or commercial, post-processing and data analysis are a huge part of any researcher's workload. Rather than re-inventing the wheel, it is very much

worth the effort to gain proficiency and even improve on existing software tools.

The factors a researcher needs to consider when choosing fNIRS software are a) compatibility with the OS and fNIRS data format, b) the range, speed, and accuracy of the fNIRS data processing it offers, c) the software price, d) the software language, which can expedite or hinder understanding of the software depending on the researcher's experience, e) the software's extensibility and customizability, and f) software-specific features or advantages. **Table II** summarizes some currently available fNIRS software.

TABLE II
A LIST OF EXISTING fNIRS SOFTWARE

Package name	Available from	Features	Price	Software and OS compatibility	Data format	Advantages
EasyTopo 2.0 [212]	https://sites.google.com/site/fenghuatian/software/easytopo	Diffuse optical topography algorithms. It visualizes the data which underwent angular interpolation, hence provides a more realistic representation of the data [213].	Free	MATLAB toolbox	MATLAB file .mat	Good computational efficiency
Functional Connectivity Analysis Tool for near-infrared spectroscopy data (FC-NIRS) [214]	https://www.nitrc.org/projects/fcnirs/	Functional connectivity calculation, visualization and network analysis. Signal quality control and batch processing are feasible. The package includes HOMER [215], thereby it allows data preprocessing. It also includes an fMRI based network analysis toolbox named GRaph thEoreTical Network Analysis (GRETNA) [216].	Free	MATLAB based package	fNIRS data with .csv or .nirs (HOMER2 [215])format	Fast processing with an improved accuracy is to be expected
Fieldtrip [217]	http://www.fieldtriptoolbox.org/download/	The software offers to create optode layouts, data preprocessing and artifact correction.	Free, Open source	MATLAB toolbox	Artinis [218] NIRS data .oxy3, .oxy4, and XML files	The structure of the package allows the users to optimize or extend its functionality according to their requirements [217]. It can process EEG and MEG.
fNIRS Optodes' Location Decider (fOLD) [219]	https://github.com/nirx/fOLD-public	It facilitates positioning the NIRS optodes efficiently in accordance with the brain region of interest (ROI).	Free	MATLAB package or Windows standalone executable	.nii NIFTI [220], and .img (ANALYZE 7.5 [221])	Default parameters are available for head tissue segmentation. Thus may be employed without loading subject-specific data.
Functional optical signal analysis (FOSA) [222]	http://www.ucl.ac.uk/medphys/research/borl/nirs/current-projects/fosa	The user can perform fNIRS- optical topography data processing, and statistical analysis by employing SPM	Free, Open source	MATLAB based package		
Hemodynamic Evoked Response (HOMER2) [215]	http://homer-fnirs.org/ , https://www.nitrc.org/projects/homer2	One of the most widely utilized NIRS analysis toolboxes. Tools to analyze the brains hemodynamic changes, the optodes settings (wavelengths, geometry, etc.), and brain activation imaging (Fig. 8(a)). The latter two are realized by the AtlasViewer [223] a software that utilizes forward modeling.	Free, Open source	MATLAB package or Windows standalone executable	Many of the current commercial fNIRS systems data formats can be saved in a format to be imported	Batch processing and grand averaging capabilities. The former facilitates faster and more convenient processing of the data while the latter allows better understanding of the experiments' outcome.
Imperial College near infrared spectroscopy neuroimaging analysis (ICNNA) [224]	http://hamlyn.doc.ic.ac.uk/icnna/	The software allows limited data processing and visualization but offers graph theory-based connectivity analysis, statistical analysis and manifold-based topological analysis.	Free for academic use	toolbox developed under MATLAB	only supports [225] HITACHI ETG-4000 [226]	Great analytical flexibility with focus on the experiment.
Monte Carlo eXtreme(MCX) [227]	http://mcx.space/	The software exploits the high computational speed provided by modern Graphics Processing Units (GPU) for a parallelized computation of NIR photons propagation simulations in 3D turbid media.	Free	written in CUDA programming language and nightly builds for Windows, MacOS, Linux		It claims to reach up to 400 times the computational speed of a single-threaded CPU [228].

TABLE II
CONTINUED

Package name	Available from	Features	Price	Software and OS compatibility	Data format	Advantages
NIRS analysis package(NAP) [229]	https://sites.google.com/site/tomerfekete2/	Offers systematic noise and motion artifact reduction by various preprocessing methods. Statistical analysis of data by employing SPM and data visualization.	Free, Open source	MATLAB	MATLAB format .mat or HITACHI excel format	
NIRFASTSlicer [230]	http://www.dartmouth.edu/nir/nirfast/	multi-modal optical tomography software [231]. It allows optical computation and modeling of light propagation (in different wavelengths and luminescence), 3D image segmentation, 3D data visualization, and image reconstruction.	Free, open source	MATLAB package or Windows standalone executable	Capable of importing data from different imaging systems	Users can integrate a customized version of 3DSlicer.
NIRSlab	https://www.nitrc.org/projects/fnirs_downstate/ or https://nirx.net/nirlab-1/	The package offers analysis and processing of NIRS. The user can create an experiment-specific optode map (Fig. 6c), data preprocessing (Fig. 8c and d), artifact removal, employing SPM to extract data dynamic features [232], batch processing, grand averaging, connectivity analysis and brain activation visualization in 2D (Fig. 8b), 3D and in video format with the task displayed together with the localized brain activity.	Free [233] or included with NIRx [234] systems	Package of .p compiled MATLAB files or Windows standalone executable	data collected by NIRx-fNIRS systems)	Offers batch processing, and reducing the size of data files.
NIRS-Statistical Parameter Mapping(NIRS-SPM)	from the Bio Imaging Signal Processing lab (http://bispl.weebly.com/nirs-spm.html#/) and NITRC (https://www.nitrc.org/projects/nirs_spm/)	Statistically analyzing the data. It allows the removal of global trends and computation of high-resolution maps of Hb, HbO ₂ changes in different brain areas of fNIRS data [235] as well as cerebral metabolic rate of oxygen (CMRO ₂) estimation from NIRS and fMRI [236].	Free, open source	Software based on Matlab and statistical parameter mapping (SPM [237]) compatible with any computer with MATLAB	The software can process fNIRS data from at least 9 different commercial fNIRS systems [http://bispl.weebly.com/nirs-spm.html#/]	The software is based on a general linear model (GLM), and Sun's tube formula / Lipschitz-Killing curvature (LKC) based expected Euler characteristics [235], [238].
Nirstorm [239]	https://github.com/Nirstorm/nirstorm	A Brainstorm [177] plugin for fNIRS data analysis. Data Tools to filter, reduce signal noise and perform some statistical analysis on single data sets or in group form.	Free, open source	Developed mainly with MATLAB (and Java) Windows, MacOS, Linux	.nirs data in a HOMER-format.	Brainstorm can analyze EEG, EMG, and ECoG data.
Platform for optical topography analysis tools (POTATO) [240]	https://www.hitechglobal.com/global/products/ind_solutions/ict/human/brain/ot/analysis/platform.html	Users can preprocess, analyze, and visualize fNIRS data by employing this software in either Normal mode, where the analysis steps and settings are fixed, or Research mode.	Free for academic use	MATLAB based package		The experienced researchers can utilize different data processing recipes according to their data and preferred settings.
Toast++ [241]	http://www.toastplusplus.org	NIR visualization in diffuse optical tomography (DOT). For image reconstruction and NIR light propagation simulations, various methods are employed such as finite element, sparse linear algebra, and others for solving the forward and inverse problems [241].	Open source software	C++software suite or precompiled binary for Linux, Windows, and Mac OS.		Interfaces for PYTHON and MATLAB are included in which exploits these packages' wide range of functions and allows researchers with different programming knowledge to benefit from this software.

XII. CONCLUSION

Although near-infrared spectroscopy is established and has widespread applications in the non-destructive testing of agricultural, pharmaceutical or textile products [242], its more exciting applications deal with its ability to provide a versatile window into the processes of the human brain. fNIRS technology has come a long way from single optode systems capable of limited spectroscopy to multi-channel miniaturized wireless

systems, easily deployed in natural settings and no longer limited by bulky lab systems. The relative simplicity of setting up and utilizing these systems to gain useful data from elegant experiments is further supported by a range of high-powered software suites. They form a crucial link between experiments and understanding and build the backbone of many current publications. fNIRS technology and methods are thus gaining ground in the health sciences and proving capable in more than intra-ICU monitoring applications. Due to its versatility and its increased

coverage of the human cortex, fNIRS has found its way into basic neuroscience, shedding light on the most important activation patterns and connectivities in the cortex—on the very processes of being human. Most recently, a strong interest in combining fNIRS with other modalities such as EEG and fMRI with objectives ranging from validation to data fusion to brain-computer interfaces has grown, allowing fNIRS to widen our non-invasive and minimally obtrusive windows to the human brain.

The future prospects of fNIRS are promising, not least because the number of commercially available fNIRS systems has grown rapidly in the last decade. Nevertheless, further improvements in fidelity, sensors and analysis methods will expand its applications to true bedside and emergency health care as well.

ACKNOWLEDGMENT

The authors would like to thank the reviewers for their helpful suggestions. We convey sincere appreciation to Alyssa Zisk for proofreading.

REFERENCES

- [1] F. A. C. Azevedo *et al.*, “Equal numbers of neuronal and nonneuronal cells make the human brain an isometrically scaled-up primate brain,” *J. Comparative Neurology*, vol. 513, no. 5, pp. 532–541, Apr. 2009.
- [2] B. Pakkenberg *et al.*, “Aging and the human neocortex,” *Exp. Gerontol.*, vol. 38, no. 1–2, pp. 95–99, Feb. 2003.
- [3] B. V. Zlokovic, “The blood-brain barrier in health and chronic neurodegenerative disorders,” *Neuron*, vol. 57, no. 2, pp. 178–201, Jan. 2008.
- [4] N. K. Logothetis, “The neural basis of the blood-oxygen-level-dependent functional magnetic resonance imaging signal,” *Philos. Trans. Roy. Soc. London B., Biol. Sci.*, vol. 357, no. 1424, pp. 1003–1037, Aug. 2002.
- [5] H. Uhlirova *et al.*, “Cell type specificity of neurovascular coupling in cerebral cortex,” *Elife*, vol. 5, May 2016, Art. no. e14315.
- [6] F. F. Jöbsis, “Noninvasive, infrared monitoring of cerebral and myocardial oxygen sufficiency and circulatory parameters,” *Science*, vol. 198, no. 4323, pp. 1264–1267, Dec. 1977.
- [7] S. J. Matcher, M. Cope, and D. T. Delpy, “In vivo measurements of the wavelength dependence of tissue-scattering coefficients between 760 and 900 nm measured with time-resolved spectroscopy,” *Appl. Opt.*, vol. 36, no. 1, pp. 386–396, Jan. 1997.
- [8] J. Steinbrink, A. Villringer, F. Kempf, D. Haux, S. Boden, and H. Obrig, “Illuminating the BOLD signal: Combined fMRI-fNIRS studies,” *Magn. Reson. Imag.*, vol. 24, no. 4, pp. 495–505, May 2006.
- [9] G. Strangman, D. A. Boas, and J. P. Sutton, “Non-invasive neuroimaging using near-infrared light,” *Biol. Psychiatry*, vol. 52, no. 7, pp. 679–693, Oct. 2002.
- [10] C. Elwell, “A practical users guide to near infrared spectroscopy,” Hamamatsu Photonics KK, London, U.K., 1995.
- [11] J. M. Murkin and M. Arango, “Near-infrared spectroscopy as an index of brain and tissue oxygenation,” *Brit. J. Anaesthesia*, vol. 103, Suppl 1, pp. i3–i3, Dec. 2009.
- [12] S. Wray, M. Cope, D. T. Delpy, J. S. Wyatt, and E. O. Reynolds, “Characterization of the near infrared absorption spectra of cytochrome aa3 and haemoglobin for the non-invasive monitoring of cerebral oxygenation,” *Biochimica Biophysica Acta*, vol. 933, no. 1, pp. 184–192, Mar. 1988.
- [13] A. Pellicer and M. del C. Bravo, “Near-infrared spectroscopy: A methodology-focused review,” *Semin. Fetal Neonatal Med.*, vol. 16, no. 1, pp. 42–49, Feb. 2011.
- [14] H. R. Heekeren *et al.*, “Noninvasive assessment of changes in cytochrome-c oxidase oxidation in human subjects during visual stimulation,” *J. Cerebral Blood Flow Metabolism*, vol. 19, no. 6, pp. 592–603, Jun. 1999.
- [15] F. Martelli, S. Del Bianco, A. Ismaelli, and G. Zaccanti, *Light Propagation Through Biological Tissue and Other Diffusive Media: Theory, Solutions, and Software*. Bellingham, WA, USA: SPIE, 2009.
- [16] “Optical properties,” ECE 532, 3. Accessed: Feb. 13, 2019. [Online]. Available: <https://omlc.org/classroom/ece532/class3/musp.html>
- [17] D. T. Delpy, M. Cope, P. van der Zee, S. Arridge, S. Wray, and J. Wyatt, “Estimation of optical pathlength through tissue from direct time of flight measurement,” *Phys. Med. Biol.*, vol. 33, no. 12, pp. 1433–1442, Dec. 1988.
- [18] S. J. Matcher and C. E. Cooper, “Absolute quantification of deoxyhaemoglobin concentration in tissue near infrared spectroscopy,” *Phys. Med. Biol.*, vol. 39, no. 8, pp. 1295–1312, Aug. 1994.
- [19] L. Kocsis, P. Herman, and A. Eke, “The modified Beer–Lambert law revisited,” *Phys. Med. Biol.*, vol. 51, no. 5, pp. N91–N98, Mar. 2006.
- [20] H. Ayaz, P. A. Shewokis, A. Curtin, M. Izzetoglu, K. Izzetoglu, and B. Onaral, “Using MazeSuite and functional near infrared spectroscopy to study learning in spatial navigation,” *J. Vis. Exp.*, no. 56, Oct. 2011. [Online]. Available: <https://www.jove.com/video/3443>
- [21] G. A. Millikan, “The oximeter, an instrument for measuring continuously the oxygen saturation of arterial blood in man,” *Rev. Sci. Instrum.*, vol. 13, no. 10, pp. 434–444, Oct. 1942.
- [22] B. Chance, “Optical method,” *Annu. Rev. Biophys. Biophys. Chem.*, vol. 20, pp. 1–28, 1991.
- [23] J. W. Severinghaus and P. B. Astrup, “History of blood gas analysis. VI. Oximetry,” *J. Clin. Monit.*, vol. 2, no. 4, pp. 270–288, Oct. 1986.
- [24] F. F. Jo Bsis-Vandervliet, “Discovery of the near-infrared window into the body and the early development of near-infrared spectroscopy,” *J. Biomed. Opt.*, vol. 4, no. 4, pp. 392–396, Oct. 1999.
- [25] M. Ferrari *et al.*, “Non invasive infrared monitoring of tissue oxygenation and circulatory parameters,” in XII World Congress of Angiology, Athens, Greece, Sep. 7–12, 1980, p. 663.
- [26] I. Giannini, M. Ferrari, A. Carpi, and P. Fasella, “Rat brain monitoring by near-infrared spectroscopy: An assessment of possible clinical significance,” *Physiological Chem. Phys.*, vol. 14, no. 3, pp. 295–305, 1982.
- [27] M. Ferrari, I. Giannini, G. Sideri, and E. Zanette, “Continuous non invasive monitoring of human brain by near infrared spectroscopy,” *Advances Exp. Med. Biol.*, vol. 191, pp. 873–882, 1985.
- [28] J. E. Brazy, D. V. Lewis, M. H. Mitnick, and F. F. Jöbsis vander Vliet, “Noninvasive monitoring of cerebral oxygenation in preterm infants: Preliminary observations,” *Pediatrics*, vol. 75, no. 2, pp. 217–225, Feb. 1985.
- [29] M. Ferrari *et al.*, “Cerebral blood volume and hemoglobin oxygen saturation monitoring in neonatal brain by near IR spectroscopy,” *Advances Exp. Med. Biol.*, vol. 200, pp. 203–211, 1986.
- [30] M. Ferrari, E. Zanette, I. Giannini, G. Sideri, C. Fieschi, and A. Carpi, “Effects of carotid artery compression test on regional cerebral blood volume, hemoglobin oxygen saturation and cytochrome-c-oxidase redox level in cerebrovascular patients,” *Advances Exp. Med. Biol.*, vol. 200, pp. 213–221, 1986.
- [31] J. S. Wyatt, M. Cope, D. T. Delpy, S. Wray, and E. O. Reynolds, “Quantification of cerebral oxygenation and haemodynamics in sick newborn infants by near infrared spectro-photometry,” *Lancet*, vol. 2, no. 8515, pp. 1063–1066, Nov. 1986.
- [32] M. Cope and D. T. Delpy, “System for long-term measurement of cerebral blood and tissue oxygenation on newborn infants by near infra-red transillumination,” *Med. Biol. Eng. Comput.*, vol. 26, no. 3, pp. 289–294, May 1988.
- [33] A. Duncan *et al.*, “Optical pathlength measurements on adult head, calf and forearm and the head of the newborn infant using phase resolved optical spectroscopy,” *Phys. Med. Biol.*, vol. 40, no. 2, pp. 295–304, Feb. 1995.
- [34] Y. Hoshi and M. Tamura, “Detection of dynamic changes in cerebral oxygenation coupled to neuronal function during mental work in man,” *Neuroscience Lett.*, vol. 150, no. 1, pp. 5–8, Feb. 1993.
- [35] B. Chance, Z. Zhuang, C. UnAh, C. Alter, and L. Lipton, “Cognition-activated low-frequency modulation of light absorption in human brain,” *Proc. Nat. Acad. Sci. USA*, vol. 90, no. 8, pp. 3770–3774, Apr. 1993.
- [36] A. Villringer, J. Planck, C. Hock, L. Schleinkofer, and U. Dirnagl, “Near infrared spectroscopy (NIRS): A new tool to study hemodynamic changes during activation of brain function in human adults,” *Neuroscience Lett.*, vol. 154, no. 1–2, pp. 101–104, May 1993.
- [37] T. Kato, A. Kamei, S. Takashima, and T. Ozaki, “Human visual cortical function during photic stimulation monitoring by means of near-infrared spectroscopy,” *J. Cerebral Blood Flow Metabolism*, vol. 13, no. 3, pp. 516–520, May 1993.
- [38] F. Okada, Y. Tokumitsu, Y. Hoshi, and M. Tamura, “Gender- and handedness-related differences of forebrain oxygenation and hemodynamics,” *Brain Res.*, vol. 601, no. 1–2, pp. 337–342, Jan. 1993.
- [39] M. Tamura, Y. Hoshi, and F. Okada, “Localized near-infrared spectroscopy and functional optical imaging of brain activity,” *Philos. Trans. Royal Soc. London B., Biol. Sci.*, vol. 352, no. 1354, pp. 737–742, Jun. 1997.

- [40] G. Gratton, J. S. Maier, M. Fabiani, W. W. Mantulin, and E. Gratton, "Feasibility of intra-cranial near-infrared optical scanning," *Psychophysiology*, vol. 31, no. 2, pp. 211–215, Mar. 1994.
- [41] M. Ferrari and V. Quaresima, "A brief review on the history of human functional near-infrared spectroscopy (fNIRS) development and fields of application," *Neuroimage*, vol. 63, no. 2, pp. 921–935, Nov. 2012.
- [42] M. J. Herrmann *et al.*, "Enhancement of activity of the primary visual cortex during processing of emotional stimuli as measured with event-related functional near-infrared spectroscopy and event-related potentials," *Human Brain Mapping*, vol. 29, no. 1, pp. 28–35, Jan. 2008.
- [43] P. Zaramella *et al.*, "Brain auditory activation measured by near-infrared spectroscopy (NIRS) in neonates," *Pediatric Res.*, vol. 49, no. 2, pp. 213–219, Feb. 2001.
- [44] A. F. Cannestra, I. Wartenburger, H. Obrig, A. Villringer, and A. W. Toga, "Functional assessment of Broca's area using near infrared spectroscopy in humans," *Neuroreport*, vol. 14, no. 15, pp. 1961–1965, Oct. 2003.
- [45] D. R. Leff *et al.*, "Assessment of the cerebral cortex during motor task behaviours in adults: A systematic review of functional near infrared spectroscopy (fNIRS) studies," *Neuroimage*, vol. 54, no. 4, pp. 2922–2936, Feb. 2011.
- [46] J. León-Carrión *et al.*, "Efficient learning produces spontaneous neural repetition suppression in prefrontal cortex," *Behav. Brain Res.*, vol. 208, no. 2, pp. 502–508, Apr. 2010.
- [47] J. Leon-Carrión *et al.*, "Differential time course and intensity of PFC activation for men and women in response to emotional stimuli: A functional near-infrared spectroscopy (fNIRS) study," *Neuroscience Lett.*, vol. 403, no. 1–2, pp. 90–95, Jul. 2006.
- [48] C. Hock *et al.*, "Decrease in parietal cerebral hemoglobin oxygenation during performance of a verbal fluency task in patients with Alzheimer's disease monitored by means of near-infrared spectroscopy (NIRS)—correlation with simultaneous rCBF-PET measurements," *Brain Res.*, vol. 755, no. 2, pp. 293–303, May 1997.
- [49] T. J. Huppert, R. D. Hoge, S. G. Diamond, M. A. Franceschini, and D. A. Boas, "A temporal comparison of BOLD, ASL, and NIRS hemodynamic responses to motor stimuli in adult humans," *Neuroimage*, vol. 29, no. 2, pp. 368–382, Jan. 2006.
- [50] B. Chance *et al.*, "A novel method for fast imaging of brain function, non-invasively, with light," *Opt. Express*, vol. 2, no. 10, pp. 411–423, May 1998.
- [51] R. B. Buxton, E. C. Wong, and L. R. Frank, "Dynamics of blood flow and oxygenation changes during brain activation: The balloon model," *Magn. Reson. Med.*, vol. 39, no. 6, pp. 855–864, Jun. 1998.
- [52] M. Teplan, "Fundamentals of EEG measurement," *Meas. Sci. Rev.*, vol. 2, pp. 1–11, 2002.
- [53] A. Vermeij, A. S. S. Meel-van den Abeelen, R. P. C. Kessels, A. H. E. A. van Beek, and J. A. H. R. Claassen, "Very-low-frequency oscillations of cerebral hemodynamics and blood pressure are affected by aging and cognitive load," *Neuroimage*, vol. 85 Pt 1, pp. 608–615, Jan. 2014.
- [54] A. Duncan *et al.*, "Measurement of cranial optical path length as a function of age using phase resolved near infrared spectroscopy," *Pediatric Res.*, vol. 39, no. 5, pp. 889–894, May 1996.
- [55] M. J. Herrmann, A. Walter, A. C. Ehli, and A. J. Fallgatter, "Cerebral oxygenation changes in the prefrontal cortex: Effects of age and gender," *Neurobiology Aging*, vol. 27, no. 6, pp. 888–894, Jun. 2006.
- [56] H. Toyoda *et al.*, "Source of nonlinearity of the BOLD response revealed by simultaneous fMRI and NIRS," *Neuroimage*, vol. 39, no. 3, pp. 997–1013, Feb. 2008.
- [57] A. Kleinschmidt *et al.*, "Simultaneous recording of cerebral blood oxygenation changes during human brain activation by magnetic resonance imaging and near-infrared spectroscopy," *J. Cerebral Blood Flow Metabolism*, vol. 16, pp. 817–826, 1996.
- [58] M. Fabiani *et al.*, "Neurovascular coupling in normal aging: A combined optical, ERP and fMRI study," *Neuroimage*, vol. 85 Pt 1, pp. 592–607, Jan. 2014.
- [59] A. Omurtag, H. Aghajani, and H. O. Keles, "Decoding human mental states by whole-head EEG+fNIRS during category fluency task performance," *J. Neural Eng.*, vol. 14, no. 6, 2017, Art. no. 066003.
- [60] F. Wallois, A. Patil, C. Héberlé, and R. Grebe, "EEG-NIRS in epilepsy in children and neonates," *Neurophysiology Clin.*, vol. 40, no. 5–6, pp. 281–292, Dec. 2010.
- [61] A. Machado *et al.*, "Detection of hemodynamic responses to epileptic activity using simultaneous electro-encephalography (EEG)/near infra red spectroscopy (NIRS) acquisitions," *Neuroimage*, vol. 56, no. 1, pp. 114–125, May 2011.
- [62] F. Wallois, M. Mahmoudzadeh, A. Patil, and R. Grebe, "Usefulness of simultaneous EEG-NIRS recording in language studies," *Brain Lang.*, vol. 121, no. 2, pp. 110–123, May 2012.
- [63] A. Gallagher *et al.*, "Non-invasive pre-surgical investigation of a 10 year-old epileptic boy using simultaneous EEG-NIRS," *Seizure*, vol. 17, no. 6, pp. 576–582, Sep. 2008.
- [64] J. Furusho, A. Suzuki, Y. Takakusa, F. Kawaguchi, N. Ichikawa, and T. Kato, "Simultaneous study of interictal EEG and near-infrared spectroscopy in a boy with epilepsy," *Int. Congr. Ser.*, vol. 1232, pp. 673–676, Apr. 2002.
- [65] F. Scholkmann *et al.*, "A review on continuous wave functional near-infrared spectroscopy and imaging instrumentation and methodology," *Neuroimage*, vol. 85 Pt 1, pp. 6–27, Jan. 2014.
- [66] D. T. Delpy and M. Cope, "Quantification in tissue near-infrared spectroscopy," *Philos. Trans. Roy. Soc. B, Biol. Sci.*, vol. 352, no. 1354, pp. 649–659, Jun. 1997.
- [67] A. Torricelli *et al.*, "Time domain functional NIRS imaging for human brain mapping," *Neuroimage*, vol. 85 Pt 1, pp. 28–50, Jan. 2014.
- [68] M. Wolf *et al.*, "Functional frequency-domain near-infrared spectroscopy detects fast neuronal signal in the motor cortex," *Neuroimage*, vol. 17, no. 4, pp. 1868–1875, Dec. 2002.
- [69] S. A. Carp, P. Farzam, N. Redes, D. M. Hueber, and M. A. Franceschini, "Combined multi-distance frequency domain and diffuse correlation spectroscopy system with simultaneous data acquisition and real-time analysis," *Biomed. Opt. Express*, vol. 8, no. 9, pp. 3993–4006, Sep. 2017.
- [70] A. Bakker, B. Smith, P. Ainslie, and K. Smith, "Near-infrared spectroscopy," in *Applied Aspects of Ultrasonography in Humans*, P. Ainslie, Ed., Rijeka, Croatia: InTech, 2012.
- [71] M. Wolf *et al.*, "Absolute frequency-domain pulse oximetry of the brain: Methodology and measurements," *Advances Exp. Med. Biol.*, vol. 530, pp. 61–73, 2003.
- [72] M. Wolf *et al.*, "Fast cerebral functional signal in the 100-ms range detected in the visual cortex by frequency-domain near-infrared spectrophotometry," *Psychophysiology*, vol. 40, no. 4, pp. 521–528, Jul. 2003.
- [73] Y. Hoshi, "Functional near-infrared optical imaging: Utility and limitations in human brain mapping," *Psychophysiology*, vol. 40, no. 4, pp. 511–520, Jul. 2003.
- [74] S. R. Arridge, M. Cope, and D. T. Delpy, "The theoretical basis for the determination of optical pathlengths in tissue: Temporal and frequency analysis," *Phys. Med. Biol.*, vol. 37, pp. 1531–1560, 1992.
- [75] B. Chance, M. B. Maris, J. Sorge, and M. Z. Zhang, "Phase modulation system for dual wavelength difference spectroscopy of hemoglobin deoxygenation in tissues," *Proc. SPIE*, vol. 1204, pp. 481–491, 1990.
- [76] J. A. Wahr, K. K. Tremper, S. Samra, and D. T. Delpy, "Near-infrared spectroscopy: Theory and applications," *J. Cardiothoracic Vascular Anesthesia*, vol. 10, no. 3, pp. 406–418, Apr. 1996.
- [77] M. Hiraoka *et al.*, "A Monte Carlo investigation of optical pathlength in inhomogeneous tissue and its application to near-infrared spectroscopy," *Phys. Med. Biol.*, vol. 38, no. 12, pp. 1859–1876, Dec. 1993.
- [78] P. Rolfe, "In vivo near-infrared spectroscopy," *Annu. Rev. Biomed. Eng.*, vol. 2, pp. 715–754, 2000.
- [79] M. Firbank, E. Okada, and D. T. Delpy, "A theoretical study of the signal contribution of regions of the adult head to near-infrared spectroscopy studies of visual evoked responses," *Neuroimage*, vol. 8, no. 1, pp. 69–78, Jul. 1998.
- [80] W. Cui, C. Kumar, and B. Chance, "Experimental study of migration depth for the photons measured at sample surface," *Proc. SPIE*, vol. 1431, pp. 180–191, 1991.
- [81] L. Wang, H. Ayaz, M. Izzetoglu, and B. Onural, "Evaluation of light detector surface area for functional near infrared spectroscopy," *Comput. Biol. Med.*, vol. 89, pp. 68–75, Oct. 2017.
- [82] T. J. Germon, P. D. Evans, N. J. Barnett, P. Wall, A. R. Manara, and R. J. Nelson, "Cerebral near infrared spectroscopy: Emitter-detector separation must be increased," *Brit. J. Anaesthesia*, vol. 82, no. 6, pp. 831–837, Jun. 1999.
- [83] V. O. Korhonen *et al.*, "Light propagation in NIR spectroscopy of the human brain," *IEEE J. Sel. Topics Quantum Electron.*, vol. 20, no. 2, pp. 289–298, Mar. 2014.
- [84] A. C. Ehli, M. J. Herrmann, A. Wagener, and A. J. Fallgatter, "Multi-channel near-infrared spectroscopy detects specific inferior-frontal activation during incongruent stroop trials," *Biol. Psychol.*, vol. 69, no. 3, pp. 315–331, Jul. 2005.

- [85] S. K. Piper *et al.*, "A wearable multi-channel fNIRS system for brain imaging in freely moving subjects," *Neuroimage*, vol. 85 Pt 1, pp. 64–71, Jan. 2014.
- [86] G. Yurtseven, A. Bozkurt, F. Kepics, K. Pourrezaei, and A. Devaraj, "Pocket PC based wireless continuous wave near infrared spectroscopy system for functional imaging of human brain," in *Proc. 25th Annual Int. Conf. IEEE Eng. Medicine and Biology Soc.*, 2003, pp. 3435–3437.
- [87] E. Lareau, F. Lesage, P. Pouliot, D. Nguyen, J. Le Lan, and M. Sawan, "Multichannel wearable system dedicated for simultaneous electroencephalography/near-infrared spectroscopy real-time data acquisitions," *J. Biomed. Opt.*, vol. 16, no. 9, Sep. 2011, Art. no. 096014.
- [88] G. Bauernfeind, R. Leeb, S. C. Wriessnegger, and G. Pfurtscheller, "Development, set-up and first results for a one-channel near-infrared spectroscopy system," *Biomed. Techn.*, vol. 53, no. 1, pp. 36–43, Feb. 2008.
- [89] S. M. Coyle, T. E. Ward, and C. M. Markham, "Brain-computer interface using a simplified functional near-infrared spectroscopy system," *J. Neural Eng.*, vol. 4, no. 3, pp. 219–226, Sep. 2007.
- [90] E. Watanabe, A. Maki, F. Kawaguchi, Y. Yamashita, H. Koizumi, and Y. Mayanagi, "Noninvasive cerebral blood volume measurement during seizures using multichannel near infrared spectroscopic topography," *J. Biomed. Opt.*, vol. 5, no. 3, pp. 287–290, Jul. 2000.
- [91] R. Bokiniec *et al.*, "Assessment of brain oxygenation in term and preterm neonates using near infrared spectroscopy," *Advances Med. Sci.*, vol. 57, no. 2, pp. 348–355, 2012.
- [92] M. A. Franceschini, S. Fantini, J. H. Thompson, J. P. Culver, and D. A. Boas, "Hemodynamic evoked response of the sensorimotor cortex measured noninvasively with near-infrared optical imaging," *Psychophysiology*, vol. 40, no. 4, pp. 548–560, Jul. 2003.
- [93] S. C. Bunce, M. Izetoglu, K. Izetoglu, B. Onaral, and K. Pourrezaei, "Functional near-infrared spectroscopy," *IEEE Eng. Med. Biol. Mag.*, vol. 25, no. 4, pp. 54–62, Aug. 2006.
- [94] F. Orihueta-Espina, D. R. Leff, D. R. C. James, A. W. Darzi, and G. Z. Yang, "Quality control and assurance in functional near infrared spectroscopy (fNIRS) experimentation," *Phys. Med. Biol.*, vol. 55, no. 13, pp. 3701–3724, Jul. 2010.
- [95] T. Li, F. Zhong, B. Pan, Z. Li, C. Huang, and Z. Deng, "A brief review of OPT101 sensor application in near-infrared spectroscopy instrumentation for intensive care unit clinics," *Sensors Basel*, vol. 17, no. 8, Jul. 2017, Art. no. E1701.
- [96] T. Muehlemann, D. Haensse, and M. Wolf, "Wireless miniaturized in-vivo near infrared imaging," *Opt. Express*, vol. 16, no. 14, pp. 10323–10330, Jul. 2008.
- [97] I.-Y. Son and B. Yazici, "Near infrared imaging and spectroscopy for brain activity monitoring," in *Advances in Sensing With Security Applications*, J. Byrnes and G. Ostheimer, Eds., Norwell, MA, USA: Kluwer, 2006, pp. 341–372.
- [98] N. F. Watson, C. Dodrill, D. Farrell, M. D. Holmes, and J. W. Miller, "Determination of language dominance with near-infrared spectroscopy: Comparison with the intracarotid amobarbital procedure," *Seizure*, vol. 13, no. 6, pp. 399–402, Sep. 2004.
- [99] S. D. Power, A. Kushki, and T. Chau, "Intersession consistency of single-trial classification of the prefrontal response to mental arithmetic and the no-control state by NIRS," *PLoS One*, vol. 7, no. 7, Jul. 2012, Art. no. e37791.
- [100] J. Safaie, R. Grebe, H. A. Moghaddam, and F. Wallois, "Wireless distributed acquisition system for near infrared spectroscopy—WDA-NIRS," *J. Innovative Opt. Health Sci.*, vol. 06, no. 03, Jul. 2013, Art. no. 1350019.
- [101] S. Cutini, F. Scarpa, P. Scatturin, R. Dell'Acqua, and M. Zorzi, "Number-space interactions in the human parietal cortex: Enlightening the SNARC effect with functional near-infrared spectroscopy," *Cerebral Cortex*, vol. 24, no. 2, pp. 444–451, Feb. 2014.
- [102] S. Coyle, C. Markham, W. Lanigan, and T. Ward, "A mechanical mounting system for functional near-infrared spectroscopy brain imaging studies," *Proc. SPIE*, vol. 5826, pp. 618–627, 2005.
- [103] M. A. Franceschini, V. Toronov, M. Filiaci, E. Gratton, and S. Fantini, "On-line optical imaging of the human brain with 160-ms temporal resolution," *Opt. Express*, vol. 6, no. 3, pp. 49–57, Jan. 2000.
- [104] U. Chaudhary, N. Birbaumer, and M. R. Curado, "Brain-machine interface (BMI) in paralysis," *Ann. Phys. Rehabil. Med.*, vol. 58, no. 1, pp. 9–13, Feb. 2015.
- [105] A. Villringer and B. Chance, "Non-invasive optical spectroscopy and imaging of human brain function," *Trends Neuroscience*, vol. 20, no. 10, pp. 435–442, Oct. 1997.
- [106] I. M. Kopton and P. Kenning, "Near-infrared spectroscopy (NIRS) as a new tool for neuroeconomic research," *Frontiers Human Neuroscience*, vol. 8, pp. 1–13, Aug. 2014, Art. no. 549.
- [107] N. Naseer and K.-S. Hong, "fNIRS-based brain-computer interfaces: A review," *Frontiers Human Neuroscience*, vol. 9, pp. 1–15, Jan. 2015, Art. no. 3.
- [108] Y. Ito, R. P. Kennan, E. Watanabe, and H. Koizumi, "Assessment of heating effects in skin during continuous wave near infrared spectroscopy," *J. Biomed. Opt.*, vol. 5, no. 4, pp. 383–390, Oct. 2000.
- [109] C. J. Soraghan, T. E. Ward, F. Matthews, and C. Markham, "Optical safety assessment of a near-infrared brain-computer interface," in *Proc. IET Irish Signals Syst. Conf.*, 2008, pp. 174–179.
- [110] L. F. Nicolas-Alonso and J. Gomez-Gil, "Brain computer interfaces: A review," *Sensors Basel*, vol. 12, no. 2, pp. 1211–1279, Jan. 2012.
- [111] B.-K. Min, M. J. Marzelli, and S.-S. Yoo, "Neuroimaging-based approaches in the brain-computer interface," *Trends Biotechnology*, vol. 28, no. 11, pp. 552–560, Nov. 2010.
- [112] R. J. Cooper *et al.*, "A systematic comparison of motion artifact correction techniques for functional near-infrared spectroscopy," *Frontiers Neuroscience*, vol. 6, pp. 1–10, Oct. 2012, Art. no. 147.
- [113] F. Biessmann, S. Plis, F. C. Meinecke, T. Eichele, and K.-R. Müller, "Analysis of multi-modal neuroimaging data," *IEEE Rev. Biomed. Eng.*, vol. 4, pp. 26–58, Oct. 2011. [Online]. Available: <https://ieeexplore.ieee.org/document/6035960>
- [114] P. Y. Lin, S. I. Lin, T. Penney, and J. J. Chen, "Applications of near infrared spectroscopy and imaging for motor rehabilitation in stroke patients," *J. Med. Biol. Eng.*, vol. 29, no. 5, pp. 210–221, 2009.
- [115] X. Cui, S. Bray, D. M. Bryant, G. H. Glover, and A. L. Reiss, "A quantitative comparison of NIRS and fMRI across multiple cognitive tasks," *Neuroimage*, vol. 54, no. 4, pp. 2808–2821, Feb. 2011.
- [116] B. He *et al.*, "Grand challenges in mapping the human brain: NSF workshop re-port," *IEEE Trans Biomed Eng*, vol. 60, no. 11, pp. 2983–2992, Nov. 2013.
- [117] R. R. Edelman and S. Warach, "Magnetic resonance imaging," *New England J. Med.*, vol. 328, no. 10, pp. 708–716, Mar. 1993.
- [118] L. Minati, E. Visani, N. G. Dowell, N. Medford, and H. D. Critchley, "Variability comparison of simultaneous brain near-infrared spectroscopy and functional magnetic resonance imaging during visual stimulation," *J. Med. Eng. Technol.*, vol. 35, no. 6–7, pp. 370–376, Oct. 2011.
- [119] X.-S. Hu, K.-S. Hong, and S. S. Ge, "fNIRS-based online deception decoding," *J. Neural Eng.*, vol. 9, no. 2, Apr. 2012, Art. no. 026012.
- [120] S. Rossi, S. Telkemeyer, I. Wartenburger, and H. Obrig, "Shedding light on words and sentences: Near-infrared spectroscopy in language research," *Brain Lang.*, vol. 121, no. 2, pp. 152–163, May 2012.
- [121] R. P. Kennan, S. G. Horowitz, A. Maki, Y. Yamashita, H. Koizumi, and J. C. Gore, "Simultaneous recording of event-related auditory odd-ball response using transcranial near infrared optical topography and surface EEG," *Neuroimage*, vol. 16, no. 3 Pt 1, pp. 587–592, Jul. 2002.
- [122] M. Strait, C. Canning, and M. Scheutz, "Limitations of NIRS-based BCI for realistic Applications in human-computer interaction," in *Proc. 5th Int. Brain-Comput. Interface Meeting*, 2013, pp. 2–3.
- [123] C. Drenckhahn, S. P. Koch, J. Dümmler, M. Kohl-Bareis, J. Steinbrink, and J. P. Dreier, "A validation study of the use of near-infrared spectroscopy imaging in primary and secondary motor areas of the human brain," *Epilepsy Behav.*, vol. 49, pp. 118–125, Aug. 2015.
- [124] S. D. Power, A. Kushki, and T. Chau, "Towards a system-paced near-infrared spectroscopy brain-computer interface: Differentiating pre-frontal activity due to mental arithmetic and mental singing from the no-control state," *J. Neural Eng.*, vol. 8, no. 6, Dec. 2011, Art. no. 066004.
- [125] J. Shin and J. Jeong, "Multiclass classification of hemodynamic responses for performance improvement of functional near-infrared spectroscopy-based brain-computer interface," *J. Biomed. Opt.*, vol. 19, no. 6, Jun. 2014, Art. no. 067009.
- [126] A. Jelzow, H. Wabnitz, I. Tachtsidis, E. Kirilina, R. Brühl, and R. Macdonald, "Separation of superficial and cerebral hemodynamics using a single distance time-domain NIRS measurement," *Biomed. Opt. Express*, vol. 5, no. 5, pp. 1465–1482, May 2014.
- [127] E. Okada, "The effect of superficial tissue of the head on spatial sensitivity profiles for near infrared spectroscopy and imaging," *Opt. Rev.*, vol. 7, no. 5, pp. 375–382, Sep. 2000.
- [128] L. Gagnon, M. A. Yücel, D. A. Boas, and R. J. Cooper, "Further improvement in reducing superficial contamination in NIRS using double short separation measurements," *Neuroimage*, vol. 85 Pt 1, pp. 127–135, Jan. 2014.

- [129] N. M. Gregg, B. R. White, B. W. Zeff, A. J. Berger, and J. P. Culver, "Brain specificity of diffuse optical imaging: Improvements from superficial signal regression and tomography," *Frontiers Neuroenergetics*, vol. 2, pp. 1–8, Jul. 2010, Art. no. 14.
- [130] A. Kassab, J. L. Lan, P. Vannasing, and M. Sawan, "Functional near-infrared spectroscopy caps for brain activity monitoring: A review," *Appl. Opt.*, vol. 54, no. 3, pp. 576–586, Jan. 2015.
- [131] P. Giacometti and S. G. Diamond, "Compliant head probe for positioning electroencephalography electrodes and near-infrared spectroscopy optodes," *J. Biomed. Opt.*, vol. 18, no. 2, Feb. 2013, Art. no. 27005.
- [132] M. J. Saikia and K. Mankodiya, "3D-printed human-centered design of fNIRS optode for the portable neuroimaging," *Proc. SPIE*, vol. 10870, 2019, Art. no. 108700Z.
- [133] R. K. Almajid, K. S. Le, and U. G. Hofmann, "Novel near infrared sensors for hybrid BCI applications," *Proc. SPIE*, vol. 9536, 2015, Art. no. 95361H.
- [134] B. Kovalenko, M. Roskosky, B. A. Freedman, and M. S. Shuler, "Effect of ambient light on near infrared spectroscopy," *J. Trauma Treat.*, vol. 4, no. 3, pp. 258–263, Jun. 2015.
- [135] A. Bozkurt and B. Onaral, "Safety assessment of near infrared light emitting diodes for diffuse optical measurements," *Biomed. Eng. Online*, vol. 3, no. 1, pp. 9–19, 2004.
- [136] M. Lei, T. Miyoshi, Y. Niwa, I. Dan, and H. Sato, "Comprehension-dependent cortical activation during speech comprehension tasks with multiple languages: Functional near-infrared spectroscopy study," *Jpn. Psychol. Res.*, vol. 60, no. 4, pp. 300–310, Oct. 2018.
- [137] K. Watanabe, H. Tanaka, K. Takahashi, Y. Niimura, K. Watanabe, and Y. Kurihara, "NIRS-based language learning BCI system," *IEEE Sensors J.*, vol. 16, no. 8, pp. 2726–2734, Apr. 2016.
- [138] K. Takahashi, K. Watanabe, T. Kaburagi, H. Tanaka, K. Watanabe, and Y. Kurihara, "A novel NIRS index to evaluate brain activity in prefrontal regions while listening to first and second languages for long time periods," *Int. J. Med., Health, Biomed., Pharmaceutical Eng.*, vol. 9, no. 3, pp. 235–241, 2015.
- [139] V. Quaresima, S. Bisconti, and M. Ferrari, "A brief review on the use of functional near-infrared spectroscopy (fNIRS) for language imaging studies in human newborns and adults," *Brain Lang.*, vol. 121, no. 2, pp. 79–89, May 2012.
- [140] Y. Minagawa-Kawai, H. van der Lely, F. Ramus, Y. Sato, R. Mazuka, and E. Dupoux, "Optical brain imaging reveals general auditory and language-specific processing in early infant development," *Cerebral Cortex*, vol. 21, no. 2, pp. 254–261, Feb. 2011.
- [141] L. Bu *et al.*, "Alteration in brain functional and effective connectivity in subjects with hypertension," *Frontiers Physiol.*, vol. 9, pp. 1–12, May 2018, Art. no. 669.
- [142] G. Vergote, S. Perrey, M. Muthuraman, S. Janaqi, and K. Torre, "Concurrent changes of brain functional connectivity and motor variability when adapting to task constraints," *Frontiers Physiol.*, vol. 9, pp. 1–14, 909, Jul. 2018, Art. no. 909.
- [143] F. S. Racz, P. Mukli, Z. Nagy, and A. Eke, "Increased prefrontal cortex connectivity during cognitive challenge assessed by fNIRS imaging," *Biomed. Opt. Express*, vol. 8, no. 8, pp. 3842–3855, Aug. 2017.
- [144] A. Gallagher, J. Tremblay, and P. Vannasing, "Language mapping in children using resting-state functional connectivity: Comparison with a task-based approach," *J. Biomed. Opt.*, vol. 21, no. 12, Dec. 2016, Art. no. 125006.
- [145] B. Wang, M. Zhang, L. Bu, L. Xu, W. Wang, and Z. Li, "Posture-related changes in brain functional connectivity as assessed by wavelet phase coherence of NIRS signals in elderly subjects," *Behav. Brain Res.*, vol. 312, pp. 238–245, Oct. 2016.
- [146] A. V. Medvedev, J. M. Kainerstorfer, S. V. Borisov, and J. VanMeter, "Functional connectivity in the prefrontal cortex measured by near-infrared spectroscopy during ultrarapid object recognition," *J. Biomed. Opt.*, vol. 16, no. 1, Feb. 2011, Art. no. 016008.
- [147] J. Wang, Q. Dong, and H. Niu, "The minimum resting-state fNIRS imaging duration for accurate and stable mapping of brain connectivity network in children," *Sci. Rep.*, vol. 7, no. 1, Jul. 2017, Art. no. 6461.
- [148] H. Zhang, L. Duan, Y.-J. Zhang, C.-M. Lu, H. Liu, and C.-Z. Zhu, "Test-retest assessment of independent component analysis-derived resting-state functional connectivity based on functional near-infrared spectroscopy," *Neuroimage*, vol. 55, no. 2, pp. 607–615, Mar. 2011.
- [149] S. Sasai, F. Homae, H. Watanabe, and G. Taga, "Frequency-specific functional connectivity in the brain during resting state revealed by NIRS," *Neuroimage*, vol. 56, no. 1, pp. 252–257, May 2011.
- [150] S. Sasai *et al.*, "A NIRS-fMRI study of resting state network," *Neuroimage*, vol. 63, no. 1, pp. 179–193, Oct. 2012.
- [151] L. Duan, Y.-J. Zhang, and C.-Z. Zhu, "Quantitative comparison of resting-state functional connectivity derived from fNIRS and fMRI: A simultaneous recording study," *Neuroimage*, vol. 60, no. 4, pp. 2008–2018, May 2012.
- [152] K. Ohi *et al.*, "Impact of familial loading on prefrontal activation in major psychiatric disorders: A near-infrared spectroscopy (NIRS) study," *Sci. Rep.*, vol. 7, Mar. 2017, Art. no. 44268.
- [153] X. Lin, S. Xu, H. F.-H. Ieong, and Z. Yuan, "Optical mapping of prefrontal activity in pathological gamblers," *Appl. Opt.*, vol. 56, no. 21, pp. 5948–5953, Jul. 2017.
- [154] N. Okada *et al.*, "Characterizing prefrontal cortical activity during inhibition task in methamphetamine-associated psychosis versus schizophrenia: A multi-channel near-infrared spectroscopy study," *Addiction Biol.*, vol. 21, no. 2, pp. 489–503, Mar. 2016.
- [155] A.-C. Ehlis, S. Schneider, T. Dresler, and A. J. Fallgatter, "Application of functional near-infrared spectroscopy in psychiatry," *Neuroimage*, vol. 85 Pt 1, pp. 478–488, Jan. 2014.
- [156] D. Matsuzawa *et al.*, "Correlation of pre-frontal activity measured by near-infrared spectroscopy (NIRS) with mood, BDNF geno-type and serum BDNF level in healthy individuals," *Open J. Psychiatry*, vol. 02, no. 03, pp. 194–203, 2012.
- [157] X. Luo *et al.*, "Prefrontal cortex dysfunction during verbal fluency task after atypical antipsychotic treatment in schizophrenia: A near-infrared spectroscopy imaging study," *Neuroscience Lett.*, vol. 686, pp. 101–105, Nov. 2018.
- [158] V. Kumar, V. Shivakumar, H. Chhabra, A. Bose, G. Venkatasubramanian, and B. N. Gangadhar, "Functional near infra-red spectroscopy (fNIRS) in schizophrenia: A review," *Asian J. Psychiatry*, vol. 27, pp. 18–31, Jun. 2017.
- [159] T. Noda, K. Nakagome, S. Setoyama, and E. Matsushima, "Working memory and pre-frontal/temporal hemodynamic responses during post-task period in patients with schizophrenia: A multi-channel near-infrared spectroscopy study," *J. Psychiatric Res.*, vol. 95, pp. 288–298, Sep. 2017.
- [160] S. Koike, Y. Nishimura, R. Takizawa, N. Yahata, and K. Kasai, "Near-infrared spectroscopy in schizophrenia: A possible biomarker for predicting clinical outcome and treatment response," *Frontiers Psychiatry*, vol. 4, pp. 1–12, Nov. 2013, Art. no. 145.
- [161] P.-H. Chou, W.-H. Lin, W.-R. Li, C.-M. Huang, and C.-W. Sun, "Reduced language lateralization in first episode schizophrenia: A near infrared spectroscopy study," *Prog. Neuropsychopharmacol Biol. Psychiatry*, vol. 78, pp. 96–104, Aug. 2017.
- [162] P.-H. Chou *et al.*, "Distinct effects of duration of untreated psychosis on brain cortical activities in different treatment phases of schizophrenia: A multi-channel near-infrared spectroscopy study," *Prog. Neuropsychopharmacol Biol. Psychiatry*, vol. 49, pp. 63–69, Mar. 2014.
- [163] Y. Nishizawa *et al.*, "fNIRS assessment during an emotional stroop task among patients with depression: Replication and extension," *Psychiatry Investigation*, vol. 16, no. 1, pp. 80–86, Jan. 2019.
- [164] A. Kondo *et al.*, "Characteristics of oxygenated hemoglobin concentration change during pleasant and unpleasant image-recall tasks in patients with depression: Comparison with healthy subjects," *Psychiatry Clin. Neurosciences*, vol. 72, May 2018, pp. 611–622.
- [165] L. Fu *et al.*, "Reduced prefrontal activation during the tower of London and verbal fluency task in patients with bipolar depression: A multi-channel NIRS study," *Frontiers Psychiatry*, vol. 9, pp. 1–10, May 2018, Art. no. 214.
- [166] J. Hirano *et al.*, "Frontal and temporal cortical functional recovery after electro-convulsive therapy for depression: A longitudinal functional near-infrared spectroscopy study," *J. Psychiatric Res.*, vol. 91, pp. 26–35, Feb. 2017.
- [167] M. Kawano *et al.*, "Correlation between frontal lobe oxyhemoglobin and severity of depression assessed using near-infrared spectroscopy," *J. Affect. Disorders*, vol. 205, pp. 154–158, Nov. 2016.
- [168] H. Zhang *et al.*, "Near-infrared spectroscopy for examination of prefrontal activation during cognitive tasks in patients with major depressive disorder: A meta-analysis of observational studies," *Psychiatry Clin. Neurosciences*, vol. 69, no. 1, pp. 22–33, Jan. 2015.
- [169] K. Sannagowdara *et al.*, "S187. Cerebral oxygen saturation and cytochrome oxidase redox state in children with epilepsy: A pilot study - MULTICHANNEL NIRS for epilepsy seizure detection," *Clin. Neurophysiology*, vol. 129, May 2018, Art. no. e212.
- [170] E. Bourel-Ponchel, M. Mahmoudzadeh, A. Delignières, P. Berquin, and F. Wallois, "Non-invasive, multimodal analysis of cortical activity, blood volume and neurovascular coupling in infantile spasms using EEG-fNIRS monitoring," *Neuroimage Clin.*, vol. 15, pp. 359–366, May 2017.

- [171] M. Manoochehri, M. Mahmoudzadeh, E. Bourel-Ponchel, and F. Wallois, "Cortical light scattering during interictal epileptic spikes in frontal lobe epilepsy in children: A fast optical signal and electroencephalographic study," *Epilepsia*, vol. 58, no. 12, pp. 2064–2072, Oct. 2017.
- [172] A. Modir, M. A. Khalilzadeh, and A. Gorji, "Detection of focal epileptic seizure using NIRS signal based on discrete wavelet transform," *Int. Clin. Neuroscience J.*, vol. 4, no. 4, pp. 134–139, Oct. 2017.
- [173] K. Peng, P. Pouliot, F. Lesage, and D. K. Nguyen, "Multichannel continuous electroencephalography-functional near-infrared spectroscopy recording of focal seizures and interictal epileptiform discharges in human epilepsy: A review," *Neurophotonics*, vol. 3, no. 3, Jul. 2016, Art. no. 031402.
- [174] M. Mihara and I. Miyai, "Review of functional near-infrared spectroscopy in neurorehabilitation," *Neurophotonics*, vol. 3, no. 3, Jul. 2016, Art. no. 031414.
- [175] M. Balconi, "Brain plasticity and rehabilitation by using near-infrared spectroscopy," *Neuropsychological Trends*, vol. 19, no. 19, pp. 71–82, Apr. 2016.
- [176] L. E. H. van Dokkum, T. Ward, and I. Laffont, "Brain computer interfaces for neurorehabilitation—Its current status as a rehabilitation strategy post-stroke," *Ann. Phys. Rehabil. Med.*, vol. 58, no. 1, pp. 3–8, Feb. 2015.
- [177] S. J. Bae, S. H. Jang, J. P. Seo, and P. H. Chang, "The optimal speed for cortical activation of passive wrist movements performed by a rehabilitation robot: A functional NIRS study," *Frontiers Human Neuroscience*, vol. 11, pp. 1–7, Apr. 2017, Art. no. 194.
- [178] P.-H. Chang *et al.*, "The cortical activation pattern by a rehabilitation robotic hand: A functional NIRS study," *Frontiers Human Neuroscience*, vol. 8, pp. 1–7, Feb. 2014, Art. no. 49.
- [179] M. Rea *et al.*, "Lower limb movement preparation in chronic stroke: A pilot study toward an fNIRS-BCI for gait rehabilitation," *Neurorehabilitation Neural Repair*, vol. 28, no. 6, pp. 564–575, Jul. 2014.
- [180] I. Yamamoto, M. Matsui, T. Higashi, N. Iso, K. Hachisuka, and A. Hachisuka, "Wrist rehabilitation robot system and its effectiveness for patients," *Sensors Mater.*, vol. 30, no. 8, Aug. 2018, Art. no. 1825.
- [181] Z. Liang *et al.*, "Design of multichannel functional near-infrared spectroscopy system with application to propofol and sevoflurane anesthesia monitoring," *Neurophotonics*, vol. 3, no. 4, Oct. 2016, Art. no. 045001.
- [182] H. Sørensen, "Near infrared spectroscopy evaluated cerebral oxygenation during anesthesia," *Danish Med. J.*, vol. 63, no. 12, Dec. 2016, Art. no. B5318.
- [183] G. Hernandez-Meza, M. Izzetoglu, M. Osbakken, M. Green, and K. Izzetoglu, "Near-infrared spectroscopy for the evaluation of anesthetic depth," *Biomed Res. Int.*, vol. 2015, Oct. 2015, Art. no. 939418.
- [184] U. Leon-Dominguez *et al.*, "Molecular concentration of deoxyHb in human prefrontal cortex predicts the emergence and suppression of consciousness," *Neuroimage*, vol. 85 Pt 1, pp. 616–625, Jan. 2014.
- [185] N. F. Agbangla, M. Audiffren, and C. T. Albinet, "Use of near-infrared spectroscopy in the investigation of brain activation during cognitive aging: A systematic review of an emerging area of research," *Ageing Res. Rev.*, vol. 38, pp. 52–66, Sep. 2017.
- [186] K. Z. H. Li, L. Bherer, A. Mirelman, I. Maidan, and J. M. Hausdorff, "Cognitive involvement in balance, gait and dual-tasking in aging: A focused review from a neuroscience of aging perspective," *Frontiers Neurology*, vol. 9, pp. 1–13, Oct. 2018, Art. no. 913.
- [187] K. Uemura *et al.*, "Reduced prefrontal oxygenation in mild cognitive impairment during memory retrieval," *Int. J. Geriatric Psychiatry*, vol. 31, no. 6, pp. 583–591, Jun. 2016.
- [188] C.-C. Lin, J. W. Barker, P. J. Sparto, J. M. Furman, and T. J. Huppert, "Functional near-infrared spectroscopy (fNIRS) brain imaging of multi-sensory integration during computerized dynamic posturography in middle-aged and older adults," *Exp. Brain Res.*, vol. 235, no. 4, pp. 1247–1256, Feb. 2017.
- [189] C.-O. Martin *et al.*, "Narrative discourse in young and older adults: Behavioral and NIRS analyses," *Frontiers Aging Neuroscience*, vol. 10, pp. 1–13, Mar. 2018, Art. no. 69.
- [190] R. Adorni, A. Brugnera, A. Gatti, G. A. Tasca, K. Sakatani, and A. Compare, "Psycho-physiological responses to stress related to anxiety in healthy aging," *J. Psychophysiol.*, vol. 33, no. 3, pp. 188–197, Jul. 2019.
- [191] R. Rosenberry *et al.*, "Age-related microvascular dysfunction: Novel insight from near-infrared spectroscopy," *Exp. Physiol.*, vol. 103, no. 2, pp. 190–200, Feb. 2018.
- [192] T. Takeda *et al.*, "PFC blood oxygenation changes in four different cognitive tasks," *Advances Exp. Med. Biol.*, vol. 977, pp. 199–204, 2017.
- [193] M. Causse, Z. Chua, V. Peysakhovich, N. Del Campo, and N. Matton, "Mental workload and neural efficiency quantified in the prefrontal cortex using fNIRS," *Sci. Rep.*, vol. 7, no. 1, Jul. 2017, Art. no. 5222.
- [194] S. Ozawa and K. Hiraki, "Distraction decreases prefrontal oxygenation: A NIRS study," *Brain Cognition*, vol. 113, pp. 155–163, Feb. 2017.
- [195] S. Keshmiri, H. Sumioka, R. Yamazaki, and H. Ishiguro, "A Non-parametric approach to the overall estimate of cognitive load using NIRS time series," *Frontiers Human Neuroscience*, vol. 11, pp. 1–14, Feb. 2017, Art. no. 15.
- [196] F. A. Fishburn, M. E. Norr, A. V. Medvedev, and C. J. Vaidya, "Sensitivity of fNIRS to cognitive state and load," *Frontiers Human Neuroscience*, vol. 8, pp. 1–11, Feb. 2014, Art. no. 76.
- [197] K. Byun, K. Hyodo, K. Suwabe, S. Kujach, M. Kato, and H. Soya, "Possible influences of exercise-intensity-dependent increases in non-cortical hemodynamic variables on NIRS-based neuroimaging analysis during cognitive tasks: Technical note," *J. Exercise Nutrition Biochem.*, vol. 18, no. 4, pp. 327–332, Dec. 2014.
- [198] Y. Liu, H. Ayaz, and P. A. Shewokis, "Multisubject 'Learning' for mental workload classification using concurrent EEG, fNIRS, and physiological measures," *Frontiers Human Neuroscience*, vol. 11, pp. 1–14, Jul. 2017, Art. no. 389.
- [199] U. E. Emir, C. Ozturk, and A. Akin, "Multimodal investigation of fMRI and fNIRS derived breath hold BOLD signals with an expanded balloon model," *Physiological Meas.*, vol. 29, no. 1, pp. 49–63, Jan. 2008.
- [200] M. L. Schroeter, T. Kupka, T. Mildner, K. Uluda, and D. Y. von Cramon, "Investigating the post-stimulus undershoot of the BOLD signal—A simultaneous fMRI and fNIRS study," *Neuroimage*, vol. 30, no. 2, pp. 349–358, Apr. 2006.
- [201] N. K. Qureshi, N. Naseer, F. M. Noori, H. Nazeer, R. A. Khan, and S. Saleem, "Enhancing classification performance of functional near-infrared spectroscopy-brain-computer interface using adaptive estimation of general linear model coefficients," *Frontiers Neurorobotics*, vol. 11, pp. 1–10, Jul. 2017, Art. no. 33.
- [202] J. Shin, K.-R. Müller, and H.-J. Hwang, "Near-infrared spectroscopy (NIRS)-based eyes-closed brain-computer interface (BCI) using pre-frontal cortex activation due to mental arithmetic," *Sci. Rep.*, vol. 6, Nov. 2016, Art. no. 36203.
- [203] J. Shin, J. Kwon, and C.-H. Im, "A ternary hybrid EEG-NIRS brain-computer interface for the classification of brain activation patterns during mental arithmetic, motor imagery, and idle state," *Frontiers Neuroinformatics*, vol. 12, pp. 1–9, Feb. 2018, Art. no. 5.
- [204] K.-S. Hong, M. J. Khan, and M. J. Hong, "Feature extraction and classification methods for hybrid fNIRS-EEG brain-computer interfaces," *Frontiers Human Neuroscience*, vol. 12, pp. 1–25, Jun. 2018, Art. no. 246.
- [205] Y. Tomita, F.-B. Vialatte, G. Dreyfus, Y. Mitsukura, H. Bakardjian, and A. Cichocki, "Bimodal BCI using simultaneously NIRS and EEG," *IEEE Trans. Biomed. Eng.*, vol. 61, no. 4, pp. 1274–1284, Apr. 2014.
- [206] F. Herold, P. Wiegel, F. Scholkmann, A. Thiers, D. Hamacher, and L. Schega, "Functional near-infrared spectroscopy in movement science: A systematic review on cortical activity in postural and walking tasks," *Neurophotonics*, vol. 4, no. 4, Oct. 2017, Art. no. 041403.
- [207] A. Abdalmalak *et al.*, "Can time-resolved NIRS provide the sensitivity to detect brain activity during motor imagery consistently?" *Biomed. Opt. Express*, vol. 8, no. 4, pp. 2162–2172, Apr. 2017.
- [208] R. Nishiyori, S. Bisconti, and B. Ulrich, "Motor cortex activity during functional motor skills: An fNIRS Study," *Brain Topography*, vol. 29, no. 1, pp. 42–55, Jan. 2016.
- [209] N. Iso *et al.*, "Monitoring local regional hemodynamic signal changes during motor execution and motor imagery using near-infrared spectroscopy," *Frontiers Physiol.*, vol. 6, pp. 1–10, Jan. 2016, Art. no. 416.
- [210] T. Liu, M. Pelowski, C. Pang, Y. Zhou, and J. Cai, "Near-infrared spectroscopy as a tool for driving research," *Ergonomics*, vol. 59, no. 3, pp. 368–379, Mar. 2016.
- [211] D. Sturman, M. W. Wiggins, J. C. Auton, and S. Loft, "Cue utilization predicts resource allocation during rail control simulations: A NIRS study," *Proc. Human Factors Ergonom. Soc. Annu. Meeting*, vol. 62, no. 1, pp. 726–730, Sep. 2018.
- [212] F. Tian *et al.*, "Test-retest assessment of cortical activation induced by repetitive transcranial magnetic stimulation with brain atlas-guided optical topography," *J. Biomed. Opt.*, vol. 17, no. 11, Nov. 2012, Art. no. 116020.
- [213] F. Tian, Z.-J. Lin, and H. Liu, "EasyTopo: A toolbox for rapid diffuse optical topography based on a standard template of brain atlas," *Proc. SPIE*, vol. 8578, 2013, Art. no. 85782J.
- [214] J. Xu *et al.*, "FC-NIRS: A functional connectivity analysis tool for near-infrared spectroscopy data," *Biomed. Res. Int.*, vol. 2015, Oct. 2015, Art. no. 248724.

- [215] T. J. Huppert, S. G. Diamond, M. A. Franceschini, and D. A. Boas, "HomER: A review of time-series analysis methods for near-infrared spectroscopy of the brain," *Appl. Opt.*, vol. 48, no. 10, pp. D280–D298, Apr. 2009.
- [216] J. Wang, X. Wang, M. Xia, X. Liao, A. Evans, and Y. He, "Corrigendum: GRETNA: A graph theoretical network analysis toolbox for imaging connectomics," *Frontiers Human Neuroscience*, vol. 9, pp. 1–16, Aug. 2015, Art. no. 458.
- [217] R. Oostenveld, P. Fries, E. Maris, and J.-M. Schoffelen, "FieldTrip: Open source software for advanced analysis of MEG, EEG, and invasive electrophysiological data," *Comput. Intell. Neuroscience*, vol. 2011, 2011, Art. no. 156869.
- [218] "Artinis medical systems—fNIRS devices—NIRS devices-home," Accessed: Feb. 13, 2019. [Online]. Available: <https://www.artinis.com/>
- [219] G. A. Zimeo Morais, J. B. Balardin, and J. R. Sato, "fNIRS optodes' location decider (FOLD): A toolbox for probe arrangement guided by brain regions-of-interest," *Sci. Rep.*, vol. 8, no. 1, Feb. 2018, Art. no. 3341.
- [220] "NIFTI-1 data format—Neuroimaging informatics technology initiative," Accessed: Feb. 13, 2019. [Online]. Available: <https://nifti.nimh.nih.gov/nifti-1/>
- [221] "TPC - Analyze 7.5 image format," Accessed: Feb. 13, 2019. [Online]. Available: http://www.turkupetcentre.net/petanalysis/format_image_analyze.html. Accessed on: Feb. 13, 2019.
- [222] P. H. Koh *et al.*, "Functional optical signal analysis: A software tool for near-infrared spectroscopy data processing incorporating statistical parametric mapping," *J. Biomed. Opt.*, vol. 12, no. 6, Dec. 2007, Art. no. 064010.
- [223] C. M. Aasted *et al.*, "Anatomical guidance for functional near-infrared spectroscopy: AtlasViewer tutorial," *Neurophotonics*, vol. 2, no. 2, Apr. 2015, Art. no. 020801.
- [224] F. Orihuela-Espina, D. R. Leff, D. R. C. James, A. W. Darzi, and G.-Z. Yang, "Imperial college near infrared spectroscopy neuroimaging analysis framework," *Neurophotonics*, vol. 5, no. 01, pp. 1–11, Sep. 2017, Art. no. 011011.
- [225] F. Orihuela-Espina, "ICNA/MENA user guide," Imperial College London, p. 19, Jul. 2010. [Online]. Available: <http://hamlyn.doc.ic.ac.uk/icnna/ICNA-MENAUUserGuide.pdf>. Accessed on: Feb. 13, 2019.
- [226] "Hitachi medical systems: Optical topography," Accessed: Feb. 13, 2019. [Online]. Available: <http://www.hitachi-medical-systems.eu/products-and-services/optical-topography.html>
- [227] Q. Fang and D. A. Boas, "Monte Carlo simulation of photon migration in 3D turbid media accelerated by graphics processing units," *Opt. Express*, vol. 17, no. 22, pp. 20178–20190, Oct. 2009.
- [228] "Monte Carlo eXtreme—The photon player," Accessed: Feb. 13, 2019. [Online]. Available: <http://mcx.space/>
- [229] T. Fekete, D. Rubin, J. M. Carlson, and L. R. Mujica-Parodi, "The NIRS analysis package: Noise reduction and statistical inference," *PLoS One*, vol. 6, no. 9, Sep. 2011, Art. no. e24322.
- [230] M. Jermyn *et al.*, "Fast segmentation and high-quality three-dimensional volume mesh creation from medical images for diffuse optical tomography," *J. Biomed. Opt.*, vol. 18, no. 8, Aug. 2013, Art. no. 86007.
- [231] H. Dehghani *et al.*, "Near infrared optical tomography using NIRFAST: Algorithm for numerical model and image reconstruction," *Commun. Numer. Methods Eng.*, vol. 25, no. 6, pp. 711–732, Aug. 2008.
- [232] Y. Xu, H. L. Graber, and R. L. Barbour, "nirsLAB: A computing environment for fNIRS neuroimaging data analysis," in *Proc. Biomed. Opt.*, Washington, DC, USA, 2014, Paper BM3A.1.
- [233] "NITRC: fNIRS data analysis environment: Tool/Resource info," Accessed: Feb. 13, 2019. [Online]. Available: https://www.nitrc.org/projects/fnirs_downstate/
- [234] "nirsLAB—fNIRS Systems—NIRS Devices—NIRx," Accessed: Feb. 13, 2019. [Online]. Available: <https://nirx.net/nirlab-1/>
- [235] J. C. Ye, S. Tak, K. E. Jang, J. Jung, and J. Jang, "NIRS-SPM: Statistical parametric mapping for near-infrared spectroscopy," *Neuroimage*, vol. 44, no. 2, pp. 428–447, Jan. 2009.
- [236] S. Tak, J. Jang, K. Lee, and J. C. Ye, "Quantification of CMRO(2) without hypercapnia using simultaneous near-infrared spectroscopy and fMRI measurements," *Phys. Med. Biol.*, vol. 55, no. 11, pp. 3249–3269, Jun. 2010.
- [237] "SPM - Statistical parametric mapping," Accessed: Feb. 13, 2019. [Online]. Available: <https://www.fil.ion.ucl.ac.uk/spm/>
- [238] H. Li, S. Tak, and J. C. Ye, "Lipschitz-Killing curvature based expected Euler characteristics for p-value correction in fNIRS," *J. Neuroscience Methods*, vol. 204, no. 1, pp. 61–67, Feb. 2012.
- [239] "Home · Nirstorm/nirstorm Wiki · GitHub," Accessed: Feb. 13, 2019. [Online]. Available: <https://github.com/Nirstorm/nirstorm/wiki>
- [240] S. Sutoko *et al.*, "Tutorial on platform for optical topography analysis tools," *Neurophotonics*, vol. 3, no. 1, Jan. 2016, Art. no. 010801.
- [241] M. Schweiger and S. Arridge, "The toast++ software suite for forward and inverse modeling in optical tomography," *J. Biomed. Opt.*, vol. 19, no. 4, Apr. 2014, Art. no. 040801.
- [242] Y. Ozaki, S. Kawata, H. W. Siebler, and H. M. Heise, *Near-Infrared Spectroscopy: Principles, Instruments, Applications*. Hoboken, NJ, USA: Wiley, 2008.



Rand K. Almajidy received the B.E. degree in biomedical engineering from Baghdad University, Baghdad, Iraq, in 2001, the M.S. degree in medical engineering from Al-Nahrain University, Baghdad, Iraq, in 2006, and the Ph.D. research in the University of Freiburg, Freiburg im Breisgau, Germany. From 2016 to 2017, She was a Postdoctoral Researcher with Biomedical Optics Research Laboratory, University Hospital Zurich, Zurich, Switzerland. Her research interests include brain-computer interface, EEG, near-infrared spectroscopy, medical system design, and data analysis and classification.



Kunal Mankodiya (S'08–M'14) received the B.E. degree in biomedical engineering from Saurashtra University, Rajkot, India, in 2003, the M.S. degree in biomedical engineering, and the Ph.D. degree in computer science from the University of Lübeck, Lübeck, Germany, in 2007 and 2010, respectively. From 2011 to 2014, he was a Postdoctoral Researcher with Intel Science and Technology Center affiliated with Carnegie Mellon University, Pittsburgh, PA, USA.

He is currently an Associate Professor of biomedical engineering and the Director of Wearable Biosensing Lab, Department of Electrical, Computer, and Biomedical Engineering, University of Rhode Island, Kingston, RI, USA. His research interests include wearable systems, smart textiles, Internet-of-Things, and neural engineering.



Mohammadreza Abtahi received the B.Sc. degree in electrical engineering from Sharif University of Technology, Tehran, Iran, in 2012, and the M.Sc. and Ph.D. degrees in electrical engineering from the University of Rhode Island, Kingston, RI, USA, in 2014 and 2018, respectively. He has received a certificate in Neuroscience from the Interdisciplinary Neuroscience Program, University of Rhode Island, in 2018.

He is currently a Research and Development Scientist in medical device industry. His research interests include brain imaging, neural engineering, brain-computer interface, biosignal processing, machine learning, wearable sensors, and medical device development.



Ulrich G. Hofmann received the Diploma in technical physics and the Ph.D. degree in biophysics from the Technical University of Munich, Munich, Germany, in 1993 and 1996, respectively. He was with the Abo Akademi, Finland, and the California Institute of Technology as a Postdoc from 1996 to 1998. From 1999 to 2012, he headed the RG on Biosignal Processing and Neuroengineering with the Institute for Signal Processing, University of Lübeck. From 2012 to 2018, he was the Endowed Peter-Osykpa-

Professor with the Medical Center, University of Freiburg. He is currently a Professor for neuroelectronic interfaces with the Department for General Neurosurgery. His long term research aims to interface wet brains with cold electronics not the least to treat diverse human illnesses and conditions.