



ISO CERTIFIED NO. 21EQCV51
SSI REGD. NO. EM22400721040204
IEC NO. : 0814021514



Shreeji Instruments

SALES • SERVICE • REPAIRING • CALIBRATORS

Manufacturing & Suppliers of : • All types of Survey Instruments (Total Station, Theodolite, Auto Level, GPS) • Soil, Cement, Concrete, Aggregate & Bitumen Equipments • Electronic Weighing Scale • Safety Items • Measuring Instruments • NDT Testing Instruments • Metal Detector • Construction Instruments

CALIBRATION CERTIFICATE

REF NO :- AI-260801

DATE:-27/08/2026

NAME OF PARTY :- CHODHARY CONSTRUCTION CO. LTD.

EQUIPMENT NAME :- AGGREGATE IMPACT VALUE MACHINE

CAPACITY & MAKE :- SHREEJI INSTRUMENTS

SERIAL NO :- SI-AI-9012

NEXT DUE DATE:-26/08/2027

SITE LOCATION :- AHMEDABAD HIGHWAY PROJECT SITE - BLOCK A

SR NO.	TEST PARAMETER	SPECIFIED RANGE	OBSERVED VALUE	REMARKS
01	Weight of Hammer Assembly	13.75 kg $\pm$ 0.25 kg	13.85 kg	PASS
02	Free Fall Height of Hammer	380 mm $\pm$ 5 mm	381.0 mm	PASS
03	Cylindrical Measure Int. Dia.	75 mm $\pm$ 1 mm	75.2 mm	PASS
04	Metal Tamping Rod Length	230 mm $\pm$ 5 mm	232.0 mm	PASS

CALIBRATED BY: YOGESH B JOSHI

- REMARKS: This certificate is valid for 12 months from the date of calibration.
- This certificate refers to the value obtained at the time of calibration.
- NOTE: Equipment meets specified tolerances under IS: 2386 (Part IV).

DETAILS & TRACEABILITY OF MASTER INSTRUMENT			
MASTER INST:	Digital Vernier Caliper & Weight Set		
SERIAL NO:	MC-AI-02	VALID UNTIL:	19/09/2027
RANGE / CAP:	0 - 300 mm / 0 - 20 kg	CALIB CERT:	NABL/AI/26-88
TRACEABILITY:	Calibrated & Traceable to National Standards via NABL L		

[QR CODE]



FOR, SHREEJI INSTRUMENTS

PROPRIETOR

Regd. Office: Shop 9/G, Karnavati Shopping Center, P.D. Pandya College Road, Nr. Burhani Society, B/h. Smruti Mandir, Ghodasar, Ahmedabad-380 050.
Mobile: 99049 04610, 93282 01463, 93771 96244
E-mail: shreejiinstrument83@yahoo.com, info@shreejiinstruments.com
Website: www.shreejiinstruments.com

Workshop : 64-65, Gopinath Industrial Hub, Ambica Tube Mill Compound, Nr. Vatva GIDC Over Bridge, Vatva, Ahmedabad-382445.
Mobile: 8734866224



ISO CERTIFIED NO. 21EQCV51
SSI REGD. NO. EM22400721040204
IEC NO. : 0814021514



Shreeji Instruments

SALES • SERVICE • REPAIRING • CALIBRATORS

Manufacturing & Suppliers of : • All types of Survey Instruments (Total Station, Theodolite, Auto Level, GPS) • Soil, Cement, Concrete, Aggregate & Bitumen Equipments • Electronic Weighing Scale • Safety Items • Measuring Instruments • NDT Testing Instruments • Metal Detector • Construction Instruments

CALIBRATION CERTIFICATE

REF NO :- AL-260802

DATE:-27/08/2026

NAME OF PARTY :- KASALA INFRASTRUCTURE PVT LTD

EQUIPMENT NAME :- AUTO LEVEL (SURVEYING INSTRUMENT)

CAPACITY & MAKE :- SOKKIA / SHREEJI

SERIAL NO :- AL-38491

NEXT DUE DATE:-26/08/2027

SITE LOCATION :- DAHOD METRO RAIL PROJECT

SR NO.	TEST PARAMETER	TOLERANCE / LIMIT	MEASURED VALUE	STATUS
01	Horizontal Line of Sight Error	± 3 arc seconds	1.2 arc sec	OK
02	Compensator Setting Accuracy	± 0.5 arc seconds	0.3 arc sec	OK
03	Telescope Magnification Check	32x Nominal	32x Verified	OK
04	Circular Bubble Sensitivity	8' / 2 mm	8' / 2 mm	OK

CALIBRATED BY: YOGESH B JOSHI

- REMARKS: This certificate is valid for 12 months from the date of calibration.
- This certificate refers to the value obtained at the time of calibration.
- NOTE: Instrument tested for collimation error on standard baseline.

DETAILS & TRACEABILITY OF MASTER INSTRUMENT			
MASTER INST:	Optical Collimator System / Master Level		
SERIAL NO:	CS/AL-08	VALID UNTIL:	15/05/2027
RANGE / CAP:	0 - 360° / 30 m	CALIB CERT:	NABL/AL/2026/102
TRACEABILITY:	Calibrated & Traceable to National Standards via NABL L		

[QR CODE]



FOR, SHREEJI INSTRUMENTS

Y B JOSHI
PROPRIETOR

Regd. Office: Shop 9/G, Karnavati Shopping Center, P.D. Pandya College Road, Nr. Burhani Society, B/h. Smruti Mandir, Ghodasar, Ahmedabad-380 050.
Mobile: 99049 04610, 93282 01463, 93771 96244
E-mail: shreejiinstrument83@yahoo.com, info@shreejiinstruments.com
Website: www.shreejiinstruments.com

Workshop : 64-65, Gopinath Industrial Hub, Ambica Tube Mill Compound, Nr. Vatva GIDC Over Bridge, Vatva, Ahmedabad-382445.
Mobile: 8734866224



ISO CERTIFIED NO. 21EQCV51
SSI REGD. NO. EM22400721040204
IEC NO. : 0814021514



Shreeji Instruments

SALES • SERVICE • REPAIRING • CALIBRATORS

Manufacturing & Suppliers of : • All types of Survey Instruments (Total Station, Theodolite, Auto Level, GPS) • Soil, Cement, Concrete, Aggregate & Bitumen Equipments • Electronic Weighing Scale • Safety Items • Measuring Instruments • NDT Testing Instruments • Metal Detector • Construction Instruments

TEST REPORT FOR CUBE MOULD

(150 MM x 150 MM x 150 MM)

REF NO :- CM-260803

DATE:-27/08/2026

NAME OF PARTY :- SHIVAM CONSTRUCTIONS LTD

EQUIPMENT NAME :- CUBE MOULD (150 MM)

CAPACITY & MAKE :- STANDARD

SERIAL NO :- CCM-01/06

NEXT DUE DATE:-26/08/2027

SITE LOCATION :- GANDHINAGAR SMART CITY PH-3

MOULD NO.	FACE DISTANCE (MM)	HEIGHT (MM)	ANGLE (DEG)	FLATNESS (MM)
Mould 1	150.15	150.10	90.0°	0.02
Mould 2	150.08	150.05	90.0°	0.01
Mould 3	150.20	150.18	89.9°	0.02
Mould 4	150.12	150.08	90.0°	0.01
Mould 5	150.05	150.02	90.0°	0.01
Mould 6	150.18	150.15	90.1°	0.02

CALIBRATED BY: YOGESH B JOSHI

- REMARKS: This report is valid for 12 months from the date of testing.
- All cube moulds conform to IS: 10086 dimensional requirements.

DETAILS & TRACEABILITY OF MASTER INSTRUMENT			
MASTER INST:	Digital Vernier Caliper & Micrometer		
SERIAL NO:	VC-CM-09	VALID UNTIL:	15/07/2027
RANGE / CAP:	0 - 300 mm	CALIB CERT:	NABL/CM/26-211
TRACEABILITY:	Calibrated & Traceable to National Standards via NABL L		

[QR CODE]



FOR, SHREEJI INSTRUMENTS

Y B JOSHI
PROPRIETOR

Regd. Office: Shop 9/G, Karnavati Shopping Center, P.D. Pandya College Road, Nr. Burhani Society, B/h. Smruti Mandir, Ghodasar, Ahmedabad-380 050.
Mobile: 99049 04610, 93282 01463, 93771 96244
E-mail: shreejiinstrument83@yahoo.com, info@shreejiinstruments.com
Website: www.shreejiinstruments.com

Workshop : 64-65, Gopinath Industrial Hub, Ambica Tube Mill Compound, Nr. Vatva GIDC Over Bridge, Vatva, Ahmedabad-382445.
Mobile: 8734866224



ISO CERTIFIED NO. 21EQCV51
SSI REGD. NO. EM22400721040204
IEC NO. : 0814021514



Shreeji Instruments

SALES • SERVICE • REPAIRING • CALIBRATORS

Manufacturing & Suppliers of : • All types of Survey Instruments (Total Station, Theodolite, Auto Level, GPS) • Soil, Cement, Concrete, Aggregate & Bitumen Equipments • Electronic Weighing Scale • Safety Items • Measuring Instruments • NDT Testing Instruments • Metal Detector • Construction Instruments

CALIBRATION CERTIFICATE

REF NO :- CP-260804

DATE:-27/08/2026

NAME OF PARTY :- PATEL GEOTECH & TESTING LAB

EQUIPMENT NAME :- CONE PENETROMETER (LIQUID LIMIT APPARATUS)

CAPACITY & MAKE :- SHREEJI INSTRUMENTS

SERIAL NO :- CP-7712

NEXT DUE DATE:-26/08/2027

SITE LOCATION :- VADODARA INDUSTRIAL PARK

SR NO.	PARAMETER	REQUIRED SPECIFICATION	ACTUAL VALUE	RESULT
01	Cone Angle	$30^{\circ} \pm 1^{\circ}$	30.1°	PASS
02	Cone Surface Finish	Polished Steel	Smooth & Polished	PASS
03	Total Sliding Shaft Weight	148 g to 150 g	149.2 g	PASS
04	Penetration Scale Accuracy	0.1 mm resolution	Verified	PASS

CALIBRATED BY: YOGESH B JOSHI

- REMARKS: This certificate is valid for 12 months from the date of calibration.
- This certificate refers to the value obtained at the time of calibration.

DETAILS & TRACEABILITY OF MASTER INSTRUMENT			
MASTER INST:	Digital Vernier Caliper & Digital Timer		
SERIAL NO:	MC-CP-05	VALID UNTIL:	04/06/2027
RANGE / CAP:	0 - 50 mm / 0 - 60 sec	CALIB CERT:	NABL/CP/26-67
TRACEABILITY:	Calibrated & Traceable to National Standards via NABL L		

[QR CODE]



FOR, SHREEJI INSTRUMENTS

Y B JOSHI
PROPRIETOR

Regd. Office: Shop 9/G, Karnavati Shopping Center, P.D. Pandya College Road, Nr. Burhani Society, B/h. Smruti Mandir, Ghodasar, Ahmedabad-380 050.
Mobile: 99049 04610, 93282 01463, 93771 96244
E-mail: shreejiinstrument83@yahoo.com, info@shreejiinstruments.com
Website: www.shreejiinstruments.com

Workshop : 64-65, Gopinath Industrial Hub, Ambica Tube Mill Compound, Nr. Vatva GIDC Over Bridge, Vatva, Ahmedabad-382445.
Mobile: 8734866224



ISO CERTIFIED NO. 21EQCV51
SSI REGD. NO. EM22400721040204
IEC NO. : 0814021514



Shreeji Instruments

SALES • SERVICE • REPAIRING • CALIBRATORS

Manufacturing & Suppliers of : • All types of Survey Instruments (Total Station, Theodolite, Auto Level, GPS) • Soil, Cement, Concrete, Aggregate & Bitumen Equipments • Electronic Weighing Scale • Safety Items • Measuring Instruments • NDT Testing Instruments • Metal Detector • Construction Instruments

CALIBRATION CERTIFICATE

REF NO :- CC-260805

DATE:-27/08/2026

NAME OF PARTY :- MEGHA ENGINEERING & INFRA

EQUIPMENT NAME :- CORE CUTTER WITH STEEL RAMMER

CAPACITY & MAKE :- ASC

SERIAL NO :- CC-1092

NEXT DUE DATE:-26/08/2027

SITE LOCATION :- AHMEDABAD BULLET TRAIN SITE

SR NO.	COMPONENT	SPECIFIED DIMENSION	ACTUAL VALUE	RESULT
01	Cutter Internal Diameter	100.0 mm $\pm$ 0.25 mm	100.12 mm	OK
02	Cutter Internal Height	130.0 mm $\pm$ 0.25 mm	130.08 mm	OK
03	Calculated Volume	1020.5 cm ³ $\pm$ 5 cm ³	1021.4 cm ³	OK
04	Rammer Mass	4.75 kg to 5.00 kg	4.85 kg	OK

CALIBRATED BY: YOGESH B JOSHI

- REMARKS: This certificate is valid for 12 months from the date of calibration.
- This certificate refers to the value obtained at the time of calibration.

DETAILS & TRACEABILITY OF MASTER INSTRUMENT			
MASTER INST:	Digital Vernier Caliper & Precision Balance		
SERIAL NO:	MC-CC-07	VALID UNTIL:	08/04/2027
RANGE / CAP:	0 - 300 mm / 0 - 15 kg	CALIB CERT:	NABL/CC/26-53
TRACEABILITY:	Calibrated & Traceable to National Standards via NABL L		

[QR CODE]



FOR, SHREEJI INSTRUMENTS

Y B JOSHI
PROPRIETOR

Regd. Office: Shop 9/G, Karnavati Shopping Center, P.D. Pandya College Road, Nr. Burhani Society, B/h. Smruti Mandir, Ghodasar, Ahmedabad-380 050.
Mobile: 99049 04610, 93282 01463, 93771 96244
E-mail: shreejiinstrument83@yahoo.com, info@shreejiinstruments.com
Website: www.shreejiinstruments.com

Workshop : 64-65, Gopinath Industrial Hub, Ambica Tube Mill Compound, Nr. Vatva GIDC Over Bridge, Vatva, Ahmedabad-382445.
Mobile: 8734866224



ISO CERTIFIED NO. 21EQCV51
SSI REGD. NO. EM22400721040204
IEC NO. : 0814021514



Shreeji Instruments

SALES • SERVICE • REPAIRING • CALIBRATORS

Manufacturing & Suppliers of : • All types of Survey Instruments (Total Station, Theodolite, Auto Level, GPS) • Soil, Cement, Concrete, Aggregate & Bitumen Equipments • Electronic Weighing Scale • Safety Items • Measuring Instruments • NDT Testing Instruments • Metal Detector • Construction Instruments

CALIBRATION CERTIFICATE

REF NO :- CTM-260806

DATE:-27/08/2026

NAME OF PARTY :- KASALA CONSTRUCTION PVT LTD

EQUIPMENT NAME :- CUBE TESTING MACHINE (1000 KN - ELECTRICAL)

CAPACITY & MAKE :- SHREEJI INSTRUMENTS

SERIAL NO :- SI-26824

NEXT DUE DATE:-26/08/2027

SITE LOCATION :- DAHOD SITE BLOCK-B

CALIBRATION INSTRUMENT 1000KN | PROVING RING NO: 1000KN 065 IS 4169:2014 | CALIBRATED BY: NCCBM

DEFLECTION (DIV)	LOAD IN KN	1st SET (KN)	2nd SET (KN)	3rd SET (KN)	AVERAGE (KN)
79.1	100	100	100	100	100.0
155.2	200	200	200	200	200.0
232.3	300	300	300	300	300.0
308.1	400	400	400	400	400.0
384.4	500	500	500	500	500.0
460.1	600	600	600	600	600.0
536.7	700	705	705	705	705.0
613.4	800	805	805	805	805.0
689.7	900	905	905	905	905.0
766.1	1000	1000	1000	1000	1000.0

CALIBRATION BY :- YOGESH BHAI

- REMARKS: This certificate is valid for 12 months from the date of calibration.
- This certificate refers to the value obtained at the time of calibration.

DETAILS & TRACEABILITY OF MASTER INSTRUMENT

MASTER INST: Master Proving Ring / Load Cell (2000 kN)
SERIAL NO: PR-2000/05
RANGE / CAP: 0 - 2000 kN
TRACEABILITY: Calibrated & Traceable to NPL India via NABL Accredited
VALID UNTIL: 10/08/2027
CALIB CERT: NPL/CTM/26-501

[QR CODE]



FOR, SHREEJI INSTRUMENTS

Y B Tani
PROPRIETOR

Regd. Office: Shop 9/G, Karnavati Shopping Center, P.D. Pandya College Road, Nr. Burhani Society, B/h. Smruti Mandir, Ghodasar, Ahmedabad-380 050.
Mobile: 99049 04610, 93282 01463, 93771 96244
E-mail: shreejiinstrument83@yahoo.com, info@shreejiinstruments.com
Website: www.shreejiinstruments.com

Workshop : 64-65, Gopinath Industrial Hub, Ambica Tube Mill Compound, Nr. Vatva GIDC Over Bridge, Vatva, Ahmedabad-382445.
Mobile: 8734866224



ISO CERTIFIED NO. 21EQCV51
SSI REGD. NO. EM22400721040204
IEC NO. : 0814021514



Shreeji Instruments

SALES • SERVICE • REPAIRING • CALIBRATORS

Manufacturing & Suppliers of : • All types of Survey Instruments (Total Station, Theodolite, Auto Level, GPS) • Soil, Cement, Concrete, Aggregate & Bitumen Equipments • Electronic Weighing Scale • Safety Items • Measuring Instruments • NDT Testing Instruments • Metal Detector • Construction Instruments

TEST REPORT FOR CUBE MOULD

(70.6 MM x 70.6 MM x 70.6 MM)

REF NO :- CM-260807

DATE:-27/08/2026

NAME OF PARTY :- L&T CONSTRUCTION LTD

EQUIPMENT NAME :- CUBE MOULD (70.6 MM)

CAPACITY & MAKE :- ASC

SERIAL NO :- CM-70-01/03

NEXT DUE DATE:-26/08/2027

SITE LOCATION :- SURAT METRO RAIL CORRIDOR-1

MOULD ID	FACE DISTANCE (MM)	HEIGHT (MM)	PERPENDICULARITY	RESULT
CM-70-01	70.65	70.62	90.0°	PASS
CM-70-02	70.60	70.58	90.0°	PASS
CM-70-03	70.68	70.65	89.9°	PASS

CALIBRATED BY: YOGESH B JOSHI

- REMARKS: This report is valid for 12 months from the date of testing.
- Cube moulds conform to IS: 10086 tolerances.

DETAILS & TRACEABILITY OF MASTER INSTRUMENT

MASTER INST: Digital Vernier Caliper & Micrometer
SERIAL NO: VC-CM-12
RANGE / CAP: 0 - 300 mm
TRACEABILITY: Calibrated & Traceable to National Standards via NABL L

VALID UNTIL: 15/07/2027
CALIB CERT: NABL/CM/26-210

[QR CODE]



FOR, SHREEJI INSTRUMENTS

Y B Joshi
PROPRIETOR

Regd. Office: Shop 9/G, Karnavati Shopping Center, P.D. Pandya College Road, Nr. Burhani Society, B/h. Smruti Mandir, Ghodasar, Ahmedabad-380 050.
Mobile: 99049 04610, 93282 01463, 93771 96244
E-mail: shreejiinstrument83@yahoo.com, info@shreejiinstruments.com
Website: www.shreejiinstruments.com

Workshop : 64-65, Gopinath Industrial Hub, Ambica Tube Mill Compound, Nr. Vatva GIDC Over Bridge, Vatva, Ahmedabad-382445.
Mobile: 8734866224



ISO CERTIFIED NO. 21EQCV51
SSI REGD. NO. EM22400721040204
IEC NO. : 0814021514



Shreeji Instruments

SALES • SERVICE • REPAIRING • CALIBRATORS

Manufacturing & Suppliers of : • All types of Survey Instruments (Total Station, Theodolite, Auto Level, GPS) • Soil, Cement, Concrete, Aggregate & Bitumen Equipments • Electronic Weighing Scale • Safety Items • Measuring Instruments • NDT Testing Instruments • Metal Detector • Construction Instruments

CALIBRATION CERTIFICATE

REF NO :- DT-260808

DATE:-27/08/2026

NAME OF PARTY :- INFRA STRUCTURE TESTING LAB

EQUIPMENT NAME :- DIGITAL THERMOMETER WITH PROBE

CAPACITY & MAKE :- MESTECH

SERIAL NO :- DT-9904

NEXT DUE DATE:-26/08/2027

SITE LOCATION :- AHMEDABAD LAB

MASTER TEMP (°C)	UUT READING (°C)	DEVIATION (°C)	UNCERTAINTY (°C)	STATUS
0.0	0.1	+0.1	± 0.2	OK
50.0	50.2	+0.2	± 0.2	OK
100.0	100.1	+0.1	± 0.2	OK
150.0	149.8	-0.2	± 0.2	OK
200.0	200.3	+0.3	± 0.3	OK

CALIBRATED BY: YOGESH B JOSHI

- REMARKS: This certificate is valid for 12 months from the date of calibration.
- This certificate refers to the value obtained at the time of calibration.

DETAILS & TRACEABILITY OF MASTER INSTRUMENT

MASTER INST: Master Temperature Calibrator / RTD Sensor
SERIAL NO: TC-RTD-11 VALID UNTIL: 12/04/2027
RANGE / CAP: -50°C to +300°C CALIB CERT: NABL/TEMP/26-109
TRACEABILITY: Calibrated & Traceable to NABL Accredited Calibration L

[QR CODE]



FOR, SHREEJI INSTRUMENTS

Y B JOSHI
PROPRIETOR

Regd. Office: Shop 9/G, Karnavati Shopping Center, P.D. Pandya College Road, Nr. Burhani Society, B/h. Smruti Mandir, Ghodasar, Ahmedabad-380 050.
Mobile: 99049 04610, 93282 01463, 93771 96244
E-mail: shreejiinstrument83@yahoo.com, info@shreejiinstruments.com
Website: www.shreejiinstruments.com

Workshop : 64-65, Gopinath Industrial Hub, Ambica Tube Mill Compound, Nr. Vatva GIDC Over Bridge, Vatva, Ahmedabad-382445.
Mobile: 8734866224



ISO CERTIFIED NO. 21EQCV51
SSI REGD. NO. EM22400721040204
IEC NO. : 0814021514



Shreeji Instruments

SALES • SERVICE • REPAIRING • CALIBRATORS

Manufacturing & Suppliers of : • All types of Survey Instruments (Total Station, Theodolite, Auto Level, GPS) • Soil, Cement, Concrete, Aggregate & Bitumen Equipments • Electronic Weighing Scale • Safety Items • Measuring Instruments • NDT Testing Instruments • Metal Detector • Construction Instruments

CALIBRATION CERTIFICATE

REF NO :- EG-260809

DATE:-27/08/2026

NAME OF PARTY :- HINDUSTAN CONSTRUCTION CO.

EQUIPMENT NAME :- ELONGATION GAUGE FOR AGGREGATES

CAPACITY & MAKE :- STANDARD

SERIAL NO :- EG-5521

NEXT DUE DATE:-26/08/2027

SITE LOCATION :- RAJKOT DUAL CARRIAGEWAY SITE

SLOT NO.	SIEVE SIZE FRACTION (MM)	STANDARD GAP (MM)	MEASURED GAP (MM)	RESULT
01	63.0 to 50.0	101.7	101.75	PASS
02	50.0 to 40.0	81.0	81.08	PASS
03	40.0 to 31.5	64.3	64.35	PASS
04	25.0 to 20.0	40.5	40.52	PASS
05	20.0 to 12.5	29.2	29.25	PASS
06	12.5 to 10.0	20.2	20.24	PASS

CALIBRATED BY: YOGESH B JOSHI

- REMARKS: This certificate is valid for 12 months from the date of calibration.
- This certificate refers to the value obtained at the time of calibration.

DETAILS & TRACEABILITY OF MASTER INSTRUMENT			
MASTER INST:	Digital Vernier Caliper	VALID UNTIL:	21/08/2027
SERIAL NO:	VC-EG-01	CALIB CERT:	NABL/EG/26-14
RANGE / CAP:	0 - 200 mm		
TRACEABILITY:	Calibrated & Traceable to National Standards via NABL L		

[QR CODE]



FOR, SHREEJI INSTRUMENTS

Y B JOSHI
PROPRIETOR

Regd. Office: Shop 9/G, Karnavati Shopping Center, P.D. Pandya College Road, Nr. Burhani Society, B/h. Smruti Mandir, Ghodasar, Ahmedabad-380 050.
Mobile: 99049 04610, 93282 01463, 93771 96244
E-mail: shreejiinstrument83@yahoo.com, info@shreejiinstruments.com
Website: www.shreejiinstruments.com

Workshop : 64-65, Gopinath Industrial Hub, Ambica Tube Mill Compound, Nr. Vatva GIDC Over Bridge, Vatva, Ahmedabad-382445.
Mobile: 8734866224



ISO CERTIFIED NO. 21EQCV51
SSI REGD. NO. EM22400721040204
IEC NO. : 0814021514



Shreeji Instruments

SALES • SERVICE • REPAIRING • CALIBRATORS

Manufacturing & Suppliers of : • All types of Survey Instruments (Total Station, Theodolite, Auto Level, GPS) • Soil, Cement, Concrete, Aggregate & Bitumen Equipments • Electronic Weighing Scale • Safety Items • Measuring Instruments • NDT Testing Instruments • Metal Detector • Construction Instruments

CALIBRATION CERTIFICATE

REF NO :- FG-260810

DATE:-27/08/2026

NAME OF PARTY :- DINESHCHANDRA R AGRAWAL INFRA

EQUIPMENT NAME :- FLAKINESS GAUGE FOR AGGREGATES

CAPACITY & MAKE :- ASC

SERIAL NO :- FG-3312

NEXT DUE DATE:-26/08/2027

SITE LOCATION :- BHARUCH EXPRESSWAY SITE

SLOT NO.	SIEVE FRACTION (MM)	SPECIFIED WIDTH (MM)	MEASURED WIDTH (MM)	STATUS
01	63.0 to 50.0	33.90	33.92	OK
02	50.0 to 40.0	27.00	27.04	OK
03	40.0 to 31.5	21.45	21.48	OK
04	25.0 to 20.0	13.50	13.52	OK
05	20.0 to 12.5	9.75	9.78	OK
06	12.5 to 10.0	6.75	6.77	OK

CALIBRATED BY: YOGESH B JOSHI

- REMARKS: This certificate is valid for 12 months from the date of calibration.
- This certificate refers to the value obtained at the time of calibration.

DETAILS & TRACEABILITY OF MASTER INSTRUMENT

MASTER INST: Digital Vernier Caliper
SERIAL NO: VC-FG-02
RANGE / CAP: 0 - 200 mm
TRACEABILITY: Calibrated & Traceable to National Standards via NABL L

VALID UNTIL: 21/08/2027
CALIB CERT: NABL/FG/26-15

[QR CODE]



FOR, SHREEJI INSTRUMENTS

PROPRIETOR

Regd. Office: Shop 9/G, Karnavati Shopping Center, P.D. Pandya College Road, Nr. Burhani Society, B/h. Smruti Mandir, Ghodasar, Ahmedabad-380 050.
Mobile: 99049 04610, 93282 01463, 93771 96244
E-mail: shreejiinstrument83@yahoo.com, info@shreejiinstruments.com
Website: www.shreejiinstruments.com

Workshop : 64-65, Gopinath Industrial Hub, Ambica Tube Mill Compound, Nr. Vatva GIDC Over Bridge, Vatva, Ahmedabad-382445.
Mobile: 8734866224



ISO CERTIFIED NO. 21EQCV51
SSI REGD. NO. EM22400721040204
IEC NO. : 0814021514



Shreeji Instruments

SALES • SERVICE • REPAIRING • CALIBRATORS

Manufacturing & Suppliers of : • All types of Survey Instruments (Total Station, Theodolite, Auto Level, GPS) • Soil, Cement, Concrete, Aggregate & Bitumen Equipments • Electronic Weighing Scale • Safety Items • Measuring Instruments • NDT Testing Instruments • Metal Detector • Construction Instruments

COMPREHENSIVE LABORATORY CALIBRATION REPORT

REF NO :- FL-260811

DATE:-27/08/2026

NAME OF PARTY :- NATIONAL HIGHWAYS AUTHORITY OF INDIA

EQUIPMENT NAME :- COMBINED TESTING EQUIPMENT INDEX

CAPACITY & MAKE :- MULTIPLE MAKERS

SERIAL NO :- LAB-FULL-2026

NEXT DUE DATE:-26/08/2027

SITE LOCATION :- CENTRAL TESTING LAB - AHMEDABAD

SR NO.	EQUIPMENT NAME	SERIAL NO	CALIBRATION STATUS	VALIDITY
01	Cube Testing Machine 2000KN	CTM-2000-881	CALIBRATED	26/08/2027
02	Rapid Moisture Meter	RM-260803	CALIBRATED	26/08/2027
03	Digital Thermometer	DT-9904	CALIBRATED	26/08/2027
04	Test Sieves Set (Brass)	TS-200-88	CALIBRATED	26/08/2027
05	Weight Balance 30kg	WB-30K-12	CALIBRATED	26/08/2027

CALIBRATED BY: YOGESH B JOSHI

- REMARKS: This master report certifies calibration for all listed laboratory equipment.
- All calibrations performed using NABL traceable reference standards.

DETAILS & TRACEABILITY OF MASTER INSTRUMENT			
MASTER INST:	Master Reference Standards (Proving Ring, Caliper, T		
SERIAL NO:	ML-REF-SET-01	VALID UNTIL:	26/08/2027
RANGE / CAP:	Multi-Parameter Calibration Standards	CALIB CERT:	NABL/FL/26-1000
TRACEABILITY:	All Reference Standards Calibrated & Traceable to NPL /		

[QR CODE]



FOR, SHREEJI INSTRUMENTS

PROPRIETOR

Regd. Office: Shop 9/G, Karnavati Shopping Center, P.D. Pandya College Road, Nr. Burhani Society, B/h. Smruti Mandir, Ghodasar, Ahmedabad-380 050.
Mobile: 99049 04610, 93282 01463, 93771 96244
E-mail: shreejiinstrument83@yahoo.com, info@shreejiinstruments.com
Website: www.shreejiinstruments.com

Workshop : 64-65, Gopinath Industrial Hub, Ambica Tube Mill Compound, Nr. Vatva GIDC Over Bridge, Vatva, Ahmedabad-382445.
Mobile: 8734866224



ISO CERTIFIED NO. 21EQCV51
SSI REGD. NO. EM22400721040204
IEC NO. : 0814021514



Shreeji Instruments

SALES • SERVICE • REPAIRING • CALIBRATORS

Manufacturing & Suppliers of : • All types of Survey Instruments (Total Station, Theodolite, Auto Level, GPS) • Soil, Cement, Concrete, Aggregate & Bitumen Equipments • Electronic Weighing Scale • Safety Items • Measuring Instruments • NDT Testing Instruments • Metal Detector • Construction Instruments

CALIBRATION CERTIFICATE

REF NO :- GEN-260812

DATE:-27/08/2026

NAME OF PARTY :- ADANI ROAD TRANSPORT LTD

EQUIPMENT NAME :- GENERAL TESTING APPARATUS

CAPACITY & MAKE :- SHREEJI INSTRUMENTS

SERIAL NO :- GEN-5541

NEXT DUE DATE:-26/08/2027

SITE LOCATION :- MUNDRA PORT SITE

SR NO.	PARAMETER	SPECIFIED LIMIT	OBSERVED VALUE	RESULT
01	Dimension Check	100.0 mm $\pm$ 0.5 mm	100.1 mm	PASS
02	Mass Verification	500.0 g $\pm$ 1.0 g	500.3 g	PASS
03	Operational Test	Smooth Functioning	Satisfactory	PASS

CALIBRATED BY: YOGESH B JOSHI

- REMARKS: This certificate is valid for 12 months from the date of calibration.
- This certificate refers to the value obtained at the time of calibration.

DETAILS & TRACEABILITY OF MASTER INSTRUMENT			
MASTER INST:	Digital Vernier Caliper & Master Weights		
SERIAL NO:	GEN-M-01	VALID UNTIL:	25/08/2027
RANGE / CAP:	0 - 300 mm / 0 - 50 kg	CALIB CERT:	NABL/GEN/26-500
TRACEABILITY:	Calibrated & Traceable to National Standards via NABL L		

[QR CODE]



FOR, SHREEJI INSTRUMENTS

Y B Joshi
PROPRIETOR

Regd. Office: Shop 9/G, Karnavati Shopping Center, P.D. Pandya College Road, Nr. Burhani Society, B/h. Smruti Mandir, Ghodasar, Ahmedabad-380 050.
Mobile: 99049 04610, 93282 01463, 93771 96244
E-mail: shreejiinstrument83@yahoo.com, info@shreejiinstruments.com
Website: www.shreejiinstruments.com

Workshop : 64-65, Gopinath Industrial Hub, Ambica Tube Mill Compound, Nr. Vatva GIDC Over Bridge, Vatva, Ahmedabad-382445.
Mobile: 8734866224



ISO CERTIFIED NO. 21EQCV51
SSI REGD. NO. EM22400721040204
IEC NO. : 0814021514



Shreeji Instruments

SALES • SERVICE • REPAIRING • CALIBRATORS

Manufacturing & Suppliers of : • All types of Survey Instruments (Total Station, Theodolite, Auto Level, GPS) • Soil, Cement, Concrete, Aggregate & Bitumen Equipments • Electronic Weighing Scale • Safety Items • Measuring Instruments • NDT Testing Instruments • Metal Detector • Construction Instruments

CALIBRATION CERTIFICATE

REF NO :- HYD-260813

DATE:-27/08/2026

NAME OF PARTY :- GEOCON TESTING & RESEARCH LAB

EQUIPMENT NAME :- SOIL HYDROMETER (TYPE 151H)

CAPACITY & MAKE :- STANLEY

SERIAL NO :- HYD-151H-04

NEXT DUE DATE:-26/08/2027

SITE LOCATION :- VADODARA GEOTECH LAB

NOMINAL DENSITY (g/ml)	SCALE READING	CORRECTION FACTOR	CORRECTED READING	STATUS
0.995	0.9950	+0.0000	0.9950	OK
1.000	1.0002	-0.0002	1.0000	OK
1.010	1.0101	-0.0001	1.0100	OK
1.020	1.0203	-0.0003	1.0200	OK
1.030	1.0298	+0.0002	1.0300	OK

CALIBRATED BY: YOGESH B JOSHI

- REMARKS: This certificate is valid for 12 months from the date of calibration.
- Hydrometer scale corrections determined at 27°C reference water temperature.

DETAILS & TRACEABILITY OF MASTER INSTRUMENT

MASTER INST: Master Standard Reference Hydrometer & Digital Therm
SERIAL NO: RH-HYD-01 VALID UNTIL: 02/10/2027
RANGE / CAP: 0.995 - 1.050 g/ml CALIB CERT: NABL/HYD/26-441
TRACEABILITY: Calibrated & Traceable to National Standards via NABL L

[QR CODE]



FOR, SHREEJI INSTRUMENTS

Y B JOSHI
PROPRIETOR

Regd. Office: Shop 9/G, Karnavati Shopping Center, P.D. Pandya College Road, Nr. Burhani Society, B/h. Smruti Mandir, Ghodasar, Ahmedabad-380 050.
Mobile: 99049 04610, 93282 01463, 93771 96244
E-mail: shreejiinstrument83@yahoo.com, info@shreejiinstruments.com
Website: www.shreejiinstruments.com

Workshop : 64-65, Gopinath Industrial Hub, Ambica Tube Mill Compound, Nr. Vatva GIDC Over Bridge, Vatva, Ahmedabad-382445.
Mobile: 8734866224



ISO CERTIFIED NO. 21EQCV51
SSI REGD. NO. EM22400721040204
IEC NO. : 0814021514

TEST REPORT FOR ISI CUBE MOULD

(150 MM x 150 MM x 150 MM - ISI MARKED)

REF NO :- ISI-260814

DATE:-27/08/2026

NAME OF PARTY :- WELSPUN ENTERPRISES LTD

EQUIPMENT NAME :- ISI CUBE MOULD (150 MM)

CAPACITY & MAKE :- ISI BRAND

SERIAL NO :- ISI-CM-901/904

NEXT DUE DATE:-26/08/2027

SITE LOCATION :- KUTCH INDUSTRIAL SITE

MOULD SERIAL	INSIDE DIA/LEN (MM)	HEIGHT (MM)	DIAGONAL (MM)	RESULT
ISI-901	150.10	150.05	212.1	PASS
ISI-902	150.15	150.10	212.2	PASS
ISI-903	150.08	150.02	212.0	PASS
ISI-904	150.12	150.08	212.1	PASS

CALIBRATED BY: YOGESH B JOSHI

- REMARKS: This test report is valid for 12 months from the date of testing.
- Moulds meet all standard tolerances prescribed under IS: 10086.

DETAILS & TRACEABILITY OF MASTER INSTRUMENT			
MASTER INST:	Digital Vernier Caliper	VALID UNTIL:	15/07/2027
SERIAL NO:	VC-ISI-03	CALIB CERT:	NABL/ICM/26-212
RANGE / CAP:	0 - 300 mm		
TRACEABILITY:	Calibrated & Traceable to National Standards via NABL L		

[QR CODE]



FOR, SHREEJI INSTRUMENTS

Y B JOSHI
PROPRIETOR

Regd. Office: Shop 9/G, Karnavati Shopping Center, P.D. Pandya College Road, Nr. Burhani Society, B/h. Smruti Mandir, Ghodasar, Ahmedabad-380 050.
Mobile: 99049 04610, 93282 01463, 93771 96244
E-mail: shreejiinstrument83@yahoo.com, info@shreejiinstruments.com
Website: www.shreejiinstruments.com

Workshop : 64-65, Gopinath Industrial Hub, Ambica Tube Mill Compound, Nr. Vatva GIDC Over Bridge, Vatva, Ahmedabad-382445.
Mobile: 8734866224



ISO CERTIFIED NO. 21EQCV51
SSI REGD. NO. EM22400721040204
IEC NO. : 0814021514



Shreeji Instruments

SALES • SERVICE • REPAIRING • CALIBRATORS

Manufacturing & Suppliers of : • All types of Survey Instruments (Total Station, Theodolite, Auto Level, GPS) • Soil, Cement, Concrete, Aggregate & Bitumen Equipments • Electronic Weighing Scale • Safety Items • Measuring Instruments • NDT Testing Instruments • Metal Detector • Construction Instruments

CALIBRATION CERTIFICATE

REF NO :- MC-260815

DATE:-27/08/2026

NAME OF PARTY :- GUJARAT WATER SUPPLY BOARD

EQUIPMENT NAME :- GRADUATED MEASURING CYLINDER (1000 ML)

CAPACITY & MAKE :- BOROSIL

SERIAL NO :- MC-1000-88

NEXT DUE DATE:-26/08/27

SITE LOCATION :- CENTRAL WATER LAB - RAJKOT

NOMINAL VOLUME (ML)	ACTUAL MEASURED VOL (ML)	ERROR (ML)	TOLERANCE (ML)	STATUS
100.0	100.1	+0.1	± 0.5	OK
250.0	250.2	+0.2	± 1.0	OK
500.0	499.8	-0.2	± 2.0	OK
1000.0	1000.4	+0.4	± 5.0	OK

CALIBRATED BY: YOGESH B JOSHI

- REMARKS: This certificate is valid for 12 months from the date of calibration.
- Volumetric calibration performed with distilled water at 27°C.

DETAILS & TRACEABILITY OF MASTER INSTRUMENT			
MASTER INST:	Precision Analytical Balance & Volumetric Standard		
SERIAL NO:	AB-MC-04	VALID UNTIL:	17/11/2027
RANGE / CAP:	0 - 1000 ml	CALIB CERT:	NABL/MC/26-309
TRACEABILITY:	Calibrated & Traceable to National Standards via NABL L		

[QR CODE]



FOR, SHREEJI INSTRUMENTS

Y B JOSHI
PROPRIETOR

Regd. Office: Shop 9/G, Karnavati Shopping Center, P.D. Pandya College Road, Nr. Burhani Society, B/h. Smruti Mandir, Ghodasar, Ahmedabad-380 050.
Mobile: 99049 04610, 93282 01463, 93771 96244
E-mail: shreejiinstrument83@yahoo.com, info@shreejiinstruments.com
Website: www.shreejiinstruments.com

Workshop : 64-65, Gopinath Industrial Hub, Ambica Tube Mill Compound, Nr. Vatva GIDC Over Bridge, Vatva, Ahmedabad-382445.
Mobile: 8734866224



ISO CERTIFIED NO. 21EQCV51
SSI REGD. NO. EM22400721040204
IEC NO. : 0814021514



Shreeji Instruments

SALES • SERVICE • REPAIRING • CALIBRATORS

Manufacturing & Suppliers of : • All types of Survey Instruments (Total Station, Theodolite, Auto Level, GPS) • Soil, Cement, Concrete, Aggregate & Bitumen Equipments • Electronic Weighing Scale • Safety Items • Measuring Instruments • NDT Testing Instruments • Metal Detector • Construction Instruments

CALIBRATION CERTIFICATE

REF NO :- OV-260816

DATE:-27/08/2026

NAME OF PARTY :- RELIANCE INDUSTRIES LTD

EQUIPMENT NAME :- LABORATORY HOT AIR OVEN

CAPACITY & MAKE :- SHREEJI INSTRUMENTS

SERIAL NO :- OV-500-12

NEXT DUE DATE:-26/08/2027

SITE LOCATION :- JAMNAGAR REFINERY TESTING LAB

SET TEMP (°C)	CENTER TEMP (°C)	TOP CORNER (°C)	BOTTOM CORNER (°C)	UNIFORMITY
105.0	105.2	105.6	104.8	± 0.8 °C
150.0	150.1	150.8	149.5	± 1.3 °C
200.0	200.4	201.2	199.6	± 1.6 °C

CALIBRATED BY: YOGESH B JOSHI

- REMARKS: This certificate is valid for 12 months from the date of calibration.
- Temperature mapping performed across 5 internal spatial locations.

DETAILS & TRACEABILITY OF MASTER INSTRUMENT

MASTER INST: Master Temperature Data Logger & Thermocouple Sensor
SERIAL NO: TH-LOG-03 VALID UNTIL: 18/07/2027
RANGE / CAP: 0°C to +300°C CALIB CERT: NABL/OVEN/26-44
TRACEABILITY: Calibrated & Traceable to NABL Accredited Calibration L

[QR CODE]



FOR, SHREEJI INSTRUMENTS

Y B Joshi
PROPRIETOR

Regd. Office: Shop 9/G, Karnavati Shopping Center, P.D. Pandya College Road,
Nr. Burhani Society, B/h. Smruti Mandir, Ghodasar, Ahmedabad-380 050.
Mobile: 99049 04610, 93282 01463, 93771 96244
E-mail: shreejiinstrument83@yahoo.com, info@shreejiinstruments.com
Website: www.shreejiinstruments.com

Workshop : 64-65, Gopinath Industrial Hub,
Ambica Tube Mill Compound,
Nr. Vatva GIDC Over Bridge,
Vatva, Ahmedabad-382445.
Mobile: 8734866224



ISO CERTIFIED NO. 21EQCV51
SSI REGD. NO. EM22400721040204
IEC NO. : 0814021514



Shreeji Instruments

SALES • SERVICE • REPAIRING • CALIBRATORS

Manufacturing & Suppliers of : • All types of Survey Instruments (Total Station, Theodolite, Auto Level, GPS) • Soil, Cement, Concrete, Aggregate & Bitumen Equipments • Electronic Weighing Scale • Safety Items • Measuring Instruments • NDT Testing Instruments • Metal Detector • Construction Instruments

CALIBRATION CERTIFICATE

REF NO :- PH-260817

DATE:-27/08/2026

NAME OF PARTY :- TORRENT PHARMACEUTICALS LTD

EQUIPMENT NAME :- DIGITAL PH METER WITH ELECTRODE

CAPACITY & MAKE :- ELICO

SERIAL NO :- PH-8841

NEXT DUE DATE:-26/08/2027

SITE LOCATION :- AHMEDABAD QUALITY CONTROL LAB

BUFFER VALUE (pH)	MEASURED pH	DEVIATION (pH)	SLOPE EFFICIENCY	RESULT
4.01	4.02	+0.01	98.5%	PASS
7.00	7.00	0.00	98.5%	PASS
9.20	9.18	-0.02	98.5%	PASS

CALIBRATED BY: YOGESH B JOSHI

- REMARKS: This certificate is valid for 12 months from the date of calibration.
- Calibrated using NIST traceable certified buffer solutions at 25°C.

DETAILS & TRACEABILITY OF MASTER INSTRUMENT			
MASTER INST:	Certified Standard pH Buffer Solutions (pH 4.01, 7.0)		
SERIAL NO:	BUF-NIST-2026	VALID UNTIL:	05/09/2027
RANGE / CAP:	0 - 14 pH	CALIB CERT:	NABL/PH/26-78
TRACEABILITY:	Traceable to NIST / NPL Standards via NABL Accredited L		

[QR CODE]



FOR, SHREEJI INSTRUMENTS

Y B JOSHI
PROPRIETOR

Regd. Office: Shop 9/G, Karnavati Shopping Center, P.D. Pandya College Road, Nr. Burhani Society, B/h. Smruti Mandir, Ghodasar, Ahmedabad-380 050.
Mobile: 99049 04610, 93282 01463, 93771 96244
E-mail: shreejiinstrument83@yahoo.com, info@shreejiinstruments.com
Website: www.shreejiinstruments.com

Workshop : 64-65, Gopinath Industrial Hub, Ambica Tube Mill Compound, Nr. Vatva GIDC Over Bridge, Vatva, Ahmedabad-382445.
Mobile: 8734866224



ISO CERTIFIED NO. 21EQCV51
SSI REGD. NO. EM22400721040204
IEC NO. : 0814021514



Shreeji Instruments

SALES • SERVICE • REPAIRING • CALIBRATORS

Manufacturing & Suppliers of : • All types of Survey Instruments (Total Station, Theodolite, Auto Level, GPS) • Soil, Cement, Concrete, Aggregate & Bitumen Equipments • Electronic Weighing Scale • Safety Items • Measuring Instruments • NDT Testing Instruments • Metal Detector • Construction Instruments

CALIBRATION CERTIFICATE

REF NO :- PYC-260818

DATE:-27/08/2026

NAME OF PARTY :- CADILA HEALTHCARE TESTING DIVISION

EQUIPMENT NAME :- PYCNOMETER BOTTLE WITH CONE

CAPACITY & MAKE :- ASC

SERIAL NO :- PYC-1KG-09

NEXT DUE DATE:-26/08/2027

SITE LOCATION :- SANAND INDUSTRIAL ESTATE

PARAMETER	SPECIFIED VALUE	MEASURED VALUE	STATUS
Mass of Empty Pycnometer Bottle	650 g $\pm$ 20 g	654.2 g	PASS
Volume Capacity of Bottle	1000 ml Nominal	1002.5 ml	PASS
Cone Top Opening Diameter	6.0 mm $\pm$ 0.5 mm	6.1 mm	PASS

CALIBRATED BY: YOGESH B JOSHI

- REMARKS: This certificate is valid for 12 months from the date of calibration.
- This certificate refers to the value obtained at the time of calibration.

DETAILS & TRACEABILITY OF MASTER INSTRUMENT			
MASTER INST:	Precision Balance & Digital Vernier Caliper		
SERIAL NO:	PB-PYC-03	VALID UNTIL:	28/05/2027
RANGE / CAP:	0 - 2000 g / 0 - 300 mm	CALIB CERT:	NABL/PYC/26-92
TRACEABILITY:	Calibrated & Traceable to National Standards via NABL L		

[QR CODE]



FOR, SHREEJI INSTRUMENTS

Y B JOSHI
PROPRIETOR

Regd. Office: Shop 9/G, Karnavati Shopping Center, P.D. Pandya College Road, Nr. Burhani Society, B/h. Smruti Mandir, Ghodasar, Ahmedabad-380 050.
Mobile: 99049 04610, 93282 01463, 93771 96244
E-mail: shreejiinstrument83@yahoo.com, info@shreejiinstruments.com
Website: www.shreejiinstruments.com

Workshop : 64-65, Gopinath Industrial Hub, Ambica Tube Mill Compound, Nr. Vatva GIDC Over Bridge, Vatva, Ahmedabad-382445.
Mobile: 8734866224



ISO CERTIFIED NO. 21EQCV51
SSI REGD. NO. EM22400721040204
IEC NO. : 0814021514



Shreeji Instruments

SALES • SERVICE • REPAIRING • CALIBRATORS

Manufacturing & Suppliers of : • All types of Survey Instruments (Total Station, Theodolite, Auto Level, GPS) • Soil, Cement, Concrete, Aggregate & Bitumen Equipments • Electronic Weighing Scale • Safety Items • Measuring Instruments • NDT Testing Instruments • Metal Detector • Construction Instruments

CALIBRATION CERTIFICATE

TABLE TO CONVERT RAPID MOISTURE METER READING TO PERCENTAGE MOISTURE CONTENT ON DRY WEIGHT BASIS.

REF NO :- RM-260803

DATE:-27/08/2026

NAME OF PARTY :- CHODHARY CONSTRUCTION CO. LTD.

EQUIPMENT NAME :- RAPID MOISTURE METER (0 - 50 %)

CAPACITY & MAKE :- SHREEJI INSTRUMENTS

SERIAL NO :- SI-260803

NEXT DUE DATE:-26/08/2027

SITE LOCATION :- HIGHWAY PROJECT SITE BLOCK C

Rapid Moisture %	Dry Weight Basis %	Rapid Moisture %	Dry Weight Basis %
01	1	26	27
02	2	27	28
03	3	28	29
04	4	29	30
05	5	30	32
06	6	31	33
07	7	32	34
08	8	33	35
09	9	34	36
10	10	35	37
11	12	36	38
12	13	37	39
13	14	38	40
14	15	39	41
15	16	40	42
16	17	41	43
17	18	42	44
18	19	43	45
19	20	44	46
20	21	45	47
21	22	46	48
22	23	47	49
23	24	48	50
24	25	49	51
25	26	50	52

CALIBRATED BY: YOGESH B JOSHI

- REMARKS: This certificate is valid for 12 months from the date of calibration.
- This certificate refers to the value obtained at the time of calibration.
- NOTE: This calibration report refers to 'Oven Drying Method'.

[QR CODE]

FOR, SHREEJI INSTRUMENTS

Y. B. Joshi
PROPRIETOR

DETAILS & TRACEABILITY OF MASTER INSTRUMENT

MASTER INST: Standard Pressure Gauge & Precision Balance
SERIAL NO: PG-RM-06 VALID UNTIL: 13/03/2027
RANGE / CAP: 0 - 50 % CALIB CERT: NABL/RM/26-177
TRACEABILITY: Calibrated & Traceable to National Standards via NABL L

Re: Damani Society, D.M. Shrihari Mandir, Ahmedabad-380 050.
Mobile: 99049 04610, 93282 01463, 93771 96244
E-mail: shreejiinstrument83@yahoo.com, info@shreejiinstruments.com
Website: www.shreejiinstruments.com

Workshop : 64-65, Gopinath Industrial Hub,
Ambica Tube Mill Compound,
Nr. Vatva GIDC Over Bridge,
Vatva, Ahmedabad-382445.
Mobile: 8734866224



ISO CERTIFIED NO. 21EQCV51
SSI REGD. NO. EM22400721040204
IEC NO. : 0814021514



Shreeji Instruments

SALES • SERVICE • REPAIRING • CALIBRATORS

Manufacturing & Suppliers of : • All types of Survey Instruments (Total Station, Theodolite, Auto Level, GPS) • Soil, Cement, Concrete, Aggregate & Bitumen Equipments • Electronic Weighing Scale • Safety Items • Measuring Instruments • NDT Testing Instruments • Metal Detector • Construction Instruments

CALIBRATION CERTIFICATE

REF NO :- SPC-260820

DATE:-27/08/2026

NAME OF PARTY :- GAMMON INDIA LIMITED

EQUIPMENT NAME :- SAND POURING CYLINDER (100 MM)

CAPACITY & MAKE :- STANDARD

SERIAL NO :- SPC-100-55

NEXT DUE DATE:-26/08/2027

SITE LOCATION :- VADODARA HIGHWAY SITE

PARAMETER	SPECIFIED VALUE	OBSERVED VALUE	RESULT
Volume of Calibrating Can	1000 cm ³ ± 10 cm ³	1002 cm ³	PASS
Mean Mass of Sand in Cone	450 g ± 15 g	452 g	PASS
Internal Diameter of Cylinder	100 mm ± 1 mm	100.2 mm	PASS

CALIBRATED BY: YOGESH B JOSHI

- REMARKS: This certificate is valid for 12 months from the date of calibration.
- Calibrated using standard Ottawa sand fraction 600 ¼ m to 300 ¼ m.

DETAILS & TRACEABILITY OF MASTER INSTRUMENT			
MASTER INST:	Digital Vernier Caliper & Precision Balance		
SERIAL NO:	VC-SPC-08	VALID UNTIL:	09/01/2027
RANGE / CAP:	0 - 300 mm / 0 - 10 kg	CALIB CERT:	NABL/SPC/26-61
TRACEABILITY:	Calibrated & Traceable to National Standards via NABL L		

[QR CODE]



FOR, SHREEJI INSTRUMENTS

Y B JOSHI
PROPRIETOR

Regd. Office: Shop 9/G, Karnavati Shopping Center, P.D. Pandya College Road, Nr. Burhani Society, B/h. Smruti Mandir, Ghodasar, Ahmedabad-380 050.
Mobile: 99049 04610, 93282 01463, 93771 96244
E-mail: shreejiinstrument83@yahoo.com, info@shreejiinstruments.com
Website: www.shreejiinstruments.com

Workshop : 64-65, Gopinath Industrial Hub, Ambica Tube Mill Compound, Nr. Vatva GIDC Over Bridge, Vatva, Ahmedabad-382445.
Mobile: 8734866224



ISO CERTIFIED NO. 21EQCV51
SSI REGD. NO. EM22400721040204
IEC NO. : 0814021514



Shreeji Instruments

SALES • SERVICE • REPAIRING • CALIBRATORS

Manufacturing & Suppliers of : • All types of Survey Instruments (Total Station, Theodolite, Auto Level, GPS) • Soil, Cement, Concrete, Aggregate & Bitumen Equipments • Electronic Weighing Scale • Safety Items • Measuring Instruments • NDT Testing Instruments • Metal Detector • Construction Instruments

TEST SIEVES CALIBRATION CERTIFICATE

REF NO :- TS-260821

DATE:-27/08/2026

NAME OF PARTY :- ULTRATECH CEMENT TESTING DIVISION

EQUIPMENT NAME :- BRASS TEST SIEVES (200 MM DIA)

CAPACITY & MAKE :- STANDARD

SERIAL NO :- TS-200-88

NEXT DUE DATE:-26/08/2027

SITE LOCATION :- JAFFRABAD CEMENT PLANT

SR NO.	MAKE	SIEVE TYPE	SIEVE APERTURE SIZE	RESULT
01	STANDARD	Brass 200 MM DIA	4.75 MM	OK
02	STANDARD	Brass 200 MM DIA	2.36 MM	OK
03	STANDARD	Brass 200 MM DIA	1.18 MM	OK
04	STANDARD	Brass 200 MM DIA	600 MICRON	OK
05	STANDARD	Brass 200 MM DIA	300 MICRON	OK
06	STANDARD	Brass 200 MM DIA	150 MICRON	OK
07	STANDARD	Brass 200 MM DIA	75 MICRON	OK

CALIBRATED BY: YOGESH B JOSHI

- REMARKS: This certificate is valid for 12 months from the date of calibration.
- Sieve aperture measurements verified using optical projector traceable to NPL.

DETAILS & TRACEABILITY OF MASTER INSTRUMENT

MASTER INST: Digital Vernier Caliper
SERIAL NO: ACCUPLUS/13-200
RANGE / CAP: 0 - 200 mm
TRACEABILITY: Calibrated & Traceable to National Standard through IDE
VALID UNTIL: 03/07/2026
CALIB CERT: 62

[QR CODE]



FOR, SHREEJI INSTRUMENTS

Y B JOSHI
PROPRIETOR

Regd. Office: Shop 9/G, Karnavati Shopping Center, P.D. Pandya College Road, Nr. Burhani Society, B/h. Smruti Mandir, Ghodasar, Ahmedabad-380 050.
Mobile: 99049 04610, 93282 01463, 93771 96244
E-mail: shreejiinstrument83@yahoo.com, info@shreejiinstruments.com
Website: www.shreejiinstruments.com

Workshop : 64-65, Gopinath Industrial Hub, Ambica Tube Mill Compound, Nr. Vatva GIDC Over Bridge, Vatva, Ahmedabad-382445.
Mobile: 8734866224



ISO CERTIFIED NO. 21EQCV51
SSI REGD. NO. EM22400721040204
IEC NO. : 0814021514



Shreeji Instruments

SALES • SERVICE • REPAIRING • CALIBRATORS

Manufacturing & Suppliers of : • All types of Survey Instruments (Total Station, Theodolite, Auto Level, GPS) • Soil, Cement, Concrete, Aggregate & Bitumen Equipments • Electronic Weighing Scale • Safety Items • Measuring Instruments • NDT Testing Instruments • Metal Detector • Construction Instruments

SLUMCONE CALIBRATION CERTIFICATE

Calibration laboratory certifies that the instrument has been inspected, tested, and calibrated in accordance with documented procedures using measuring and test equipment traceable to international standards.

From the below test this is to certify that the Slump test apparatus is meeting necessary requirements as per IS: 7320-1974 within the permissible limit.

REF NO :- SC-260802

DATE:-27/08/2026

NAME OF PARTY :- CHODHARY CONSTRUCTION CO. LTD.

EQUIPMENT NAME :- SLUMP CONE APPARATUS

CAPACITY & MAKE :- SHREEJI INSTRUMENTS

SERIAL NO :- SC-260802

NEXT DUE DATE:-26/08/2027

SITE LOCATION :- 300 MW HYBRID NTPC BLOCK-2 LAKADIYA

PARAMETER	AS PER IS (MM)	ACTUAL MEASURED (AVG. OF THREE)
TOP DIA	100 + 3.0 - 1.5	100.10
BOTTOM DIA	200 + 3.0 - 1.5	200.80
HEIGHT	300 + 1.5 - 1.5	300.40

CALIBRATED BY: YOGESH B JOSHI

- REMARKS: This certificate is valid for 12 months from the date of calibration.
- This certificate refers to the value obtained at the time of calibration.

DETAILS & TRACEABILITY OF MASTER INSTRUMENT

MASTER INST:	Digital Vernier Caliper & Steel Rule	VALID UNTIL:	14/02/2027
SERIAL NO:	VC-SC-10	CALIB CERT:	NABL/SC/26-83
RANGE / CAP:	0 - 300 mm		
TRACEABILITY:	Calibrated & Traceable to National Standards via NABL L		

[QR CODE]



FOR, SHREEJI INSTRUMENTS

Y B JOSHI
PROPRIETOR

Regd. Office: Shop 9/G, Karnavati Shopping Center, P.D. Pandya College Road, Nr. Burhani Society, B/h. Smruti Mandir, Ghodasar, Ahmedabad-380 050.
Mobile: 99049 04610, 93282 01463, 93771 96244
E-mail: shreejiinstrument83@yahoo.com, info@shreejiinstruments.com
Website: www.shreejiinstruments.com

Workshop : 64-65, Gopinath Industrial Hub, Ambica Tube Mill Compound, Nr. Vatva GIDC Over Bridge, Vatva, Ahmedabad-382445.
Mobile: 8734866224



ISO CERTIFIED NO. 21EQCV51
SSI REGD. NO. EM22400721040204
IEC NO. : 0814021514



Shreeji Instruments

SALES • SERVICE • REPAIRING • CALIBRATORS

Manufacturing & Suppliers of : • All types of Survey Instruments (Total Station, Theodolite, Auto Level, GPS) • Soil, Cement, Concrete, Aggregate & Bitumen Equipments • Electronic Weighing Scale • Safety Items • Measuring Instruments • NDT Testing Instruments • Metal Detector • Construction Instruments

CALIBRATION CERTIFICATE

REF NO :- TS-260823

DATE:-27/08/2026

NAME OF PARTY :- LARSEN & TOUBRO INFRASTRUCTURE

EQUIPMENT NAME :- TOTAL STATION (SURVEYING INSTRUMENT)

CAPACITY & MAKE :- LEICA / SHREEJI

SERIAL NO :- TS-77401

NEXT DUE DATE:-26/08/2027

SITE LOCATION :- AHMEDABAD GANDHINAGAR METRO PH-2

PARAMETER	NOMINAL TOLERANCE	MEASURED VALUE	STATUS
Horizontal Angle Accuracy	2 arc seconds	1.5 arc sec	PASS
Vertical Angle Accuracy	2 arc seconds	1.8 arc sec	PASS
EDM Distance Accuracy	2 mm + 2 ppm	1.2 mm	PASS
Laser Plummet Alignment	Center dot offset < 0.5 mm	0.2 mm	PASS

CALIBRATED BY: YOGESH B JOSHI

- REMARKS: This certificate is valid for 12 months from the date of calibration.
- Calibrated using multi-baseline optical test bench.

DETAILS & TRACEABILITY OF MASTER INSTRUMENT			
MASTER INST:	Optical Collimator Bench & Precision Polygon Mirror		
SERIAL NO:	TS-CAL-04	VALID UNTIL:	20/06/2027
RANGE / CAP:	0 - 360° / 3000 m	CALIB CERT:	NABL/TS/2026/84
TRACEABILITY:	Calibrated & Traceable to National Standards via NABL L		

[QR CODE]



FOR, SHREEJI INSTRUMENTS

PROPRIETOR

Regd. Office: Shop 9/G, Karnavati Shopping Center, P.D. Pandya College Road, Nr. Burhani Society, B/h. Smruti Mandir, Ghodasar, Ahmedabad-380 050.
Mobile: 99049 04610, 93282 01463, 93771 96244
E-mail: shreejiinstrument83@yahoo.com, info@shreejiinstruments.com
Website: www.shreejiinstruments.com

Workshop : 64-65, Gopinath Industrial Hub, Ambica Tube Mill Compound, Nr. Vatva GIDC Over Bridge, Vatva, Ahmedabad-382445.
Mobile: 8734866224



ISO CERTIFIED NO. 21EQCV51
SSI REGD. NO. EM22400721040204
IEC NO. : 0814021514



Shreeji Instruments

SALES • SERVICE • REPAIRING • CALIBRATORS

Manufacturing & Suppliers of : • All types of Survey Instruments (Total Station, Theodolite, Auto Level, GPS) • Soil, Cement, Concrete, Aggregate & Bitumen Equipments • Electronic Weighing Scale • Safety Items • Measuring Instruments • NDT Testing Instruments • Metal Detector • Construction Instruments

CALIBRATION CERTIFICATE

REF NO :- VC-260824

DATE:-27/08/2026

NAME OF PARTY :- TATA MOTORS SANAND PLANT

EQUIPMENT NAME :- VERNIER CALIPER (0 - 300 MM)

CAPACITY & MAKE :- MITUTOYO

SERIAL NO :- VC-300-44

NEXT DUE DATE:-26/08/2027

SITE LOCATION :- SANAND AUTOMOTIVE DIVISION

NOMINAL SIZE (MM)	MEASURED VALUE (MM)	ERROR (MM)	UNCERTAINTY (MM)	STATUS
0.00	0.00	0.00	± 0.01	OK
50.00	50.01	+0.01	± 0.01	OK
100.00	100.01	+0.01	± 0.01	OK
150.00	149.99	-0.01	± 0.01	OK
200.00	200.02	+0.02	± 0.01	OK
250.00	250.01	+0.01	± 0.01	OK
300.00	300.02	+0.02	± 0.01	OK

CALIBRATED BY: YOGESH B JOSHI

- REMARKS: This certificate is valid for 12 months from the date of calibration.
- Slip gauge reference standards traceable to NPL India.

DETAILS & TRACEABILITY OF MASTER INSTRUMENT			
MASTER INST:	Gauge Block Set Grade 0 (Slip Gauges)		
SERIAL NO:	GBS-01/88	VALID UNTIL:	14/03/2027
RANGE / CAP:	0.5 mm - 100 mm	CALIB CERT:	NABL/VC/26-301
TRACEABILITY:	Calibrated & Traceable to NPL India via NABL Accredited		

[QR CODE]



FOR, SHREEJI INSTRUMENTS

PROPRIETOR

Regd. Office: Shop 9/G, Karnavati Shopping Center, P.D. Pandya College Road, Nr. Burhani Society, B/h. Smruti Mandir, Ghodasar, Ahmedabad-380 050.
Mobile: 99049 04610, 93282 01463, 93771 96244
E-mail: shreejiinstrument83@yahoo.com, info@shreejiinstruments.com
Website: www.shreejiinstruments.com

Workshop : 64-65, Gopinath Industrial Hub, Ambica Tube Mill Compound, Nr. Vatva GIDC Over Bridge, Vatva, Ahmedabad-382445.
Mobile: 8734866224



ISO CERTIFIED NO. 21EQCV51
SSI REGD. NO. EM22400721040204
IEC NO. : 0814021514



Shreeji Instruments

SALES • SERVICE • REPAIRING • CALIBRATORS

Manufacturing & Suppliers of : • All types of Survey Instruments (Total Station, Theodolite, Auto Level, GPS) • Soil, Cement, Concrete, Aggregate & Bitumen Equipments • Electronic Weighing Scale • Safety Items • Measuring Instruments • NDT Testing Instruments • Metal Detector • Construction Instruments

CALIBRATION CERTIFICATE

REF NO :- WB-260825

DATE:-27/08/2026

NAME OF PARTY :- AMUL DAIRY QUALITY TESTING DIVISION

EQUIPMENT NAME :- THERMOSTATIC WATER BATH

CAPACITY & MAKE :- SHREEJI INSTRUMENTS

SERIAL NO :- WB-50L-03

NEXT DUE DATE:-26/08/2027

SITE LOCATION :- ANAND MAIN PLANT LAB

SET TEMP (°C)	MEASURED TEMP (°C)	DEVIATION (°C)	STABILITY	STATUS
37.0	37.1	+0.1	± 0.2 °C	OK
60.0	60.2	+0.2	± 0.3 °C	OK
100.0	99.8	-0.2	± 0.4 °C	OK

CALIBRATED BY: YOGESH B JOSHI

- REMARKS: This certificate is valid for 12 months from the date of calibration.
- Water bath temperature stability verified over 60-minute duration.

DETAILS & TRACEABILITY OF MASTER INSTRUMENT			
MASTER INST:	Master Digital Thermometer & Temperature Sensor		
SERIAL NO:	MDT-WB-09	VALID UNTIL:	11/05/2027
RANGE / CAP:	0°C to 100°C	CALIB CERT:	NABL/WB/26-118
TRACEABILITY:	Calibrated & Traceable to NABL Accredited Calibration L		

[QR CODE]



FOR, SHREEJI INSTRUMENTS

Y B JOSHI
PROPRIETOR

Regd. Office: Shop 9/G, Karnavati Shopping Center, P.D. Pandya College Road, Nr. Burhani Society, B/h. Smruti Mandir, Ghodasar, Ahmedabad-380 050.
Mobile: 99049 04610, 93282 01463, 93771 96244
E-mail: shreejiinstrument83@yahoo.com, info@shreejiinstruments.com
Website: www.shreejiinstruments.com

Workshop : 64-65, Gopinath Industrial Hub, Ambica Tube Mill Compound, Nr. Vatva GIDC Over Bridge, Vatva, Ahmedabad-382445.
Mobile: 8734866224



ISO CERTIFIED NO. 21EQCV51
SSI REGD. NO. EM22400721040204
IEC NO. : 0814021514



Shreeji Instruments

SALES • SERVICE • REPAIRING • CALIBRATORS

Manufacturing & Suppliers of : • All types of Survey Instruments (Total Station, Theodolite, Auto Level, GPS) • Soil, Cement, Concrete, Aggregate & Bitumen Equipments • Electronic Weighing Scale • Safety Items • Measuring Instruments • NDT Testing Instruments • Metal Detector • Construction Instruments

CALIBRATION CERTIFICATE

REF NO :- WB-260826

DATE:-27/08/2026

NAME OF PARTY :- ACC CEMENT BATCHING PLANT

EQUIPMENT NAME :- WEIGH BATCHER (500 KG CAPACITY)

CAPACITY & MAKE :- SHREEJI INSTRUMENTS

SERIAL NO :- WB-500K-07

NEXT DUE DATE:-26/08/2027

SITE LOCATION :- AHMEDABAD BATCHING PLANT SITE

APPLIED LOAD (KG)	OBSERVED DISPLAY (KG)	ERROR (KG)	PERMISSIBLE ERROR	RESULT
0.0	0.0	0.0	± 0.5 kg	PASS
100.0	100.2	+0.2	± 0.5 kg	PASS
200.0	200.1	+0.1	± 0.5 kg	PASS
300.0	299.8	-0.2	± 0.5 kg	PASS
400.0	400.3	+0.3	± 0.5 kg	PASS
500.0	500.2	+0.2	± 0.5 kg	PASS

CALIBRATED BY: YOGESH B JOSHI

- REMARKS: This certificate is valid for 12 months from the date of calibration.
- Calibrated using M1 class standard cast iron test weights.

DETAILS & TRACEABILITY OF MASTER INSTRUMENT

MASTER INST: Standard M1 Class Test Weights
SERIAL NO: M1-TW-500
RANGE / CAP: 1 kg - 500 kg
TRACEABILITY: Calibrated & Traceable to NABL Accredited Calibration L

VALID UNTIL: 30/01/2027
CALIB CERT: NABL/WBC/26-412

[QR CODE]



FOR, SHREEJI INSTRUMENTS

Y B JOSHI
PROPRIETOR

Regd. Office: Shop 9/G, Karnavati Shopping Center, P.D. Pandya College Road, Nr. Burhani Society, B/h. Smruti Mandir, Ghodasar, Ahmedabad-380 050.
Mobile: 99049 04610, 93282 01463, 93771 96244
E-mail: shreejiinstrument83@yahoo.com, info@shreejiinstruments.com
Website: www.shreejiinstruments.com

Workshop : 64-65, Gopinath Industrial Hub, Ambica Tube Mill Compound, Nr. Vatva GIDC Over Bridge, Vatva, Ahmedabad-382445.
Mobile: 8734866224



Shreeji Instruments

SALES • SERVICE • REPAIRING • CALIBRATORS

Manufacturing & Suppliers of : • All types of Survey Instruments (Total Station, Theodolite, Auto Level, GPS) • Soil, Cement, Concrete, Aggregate & Bitumen Equipments • Electronic Weighing Scale • Safety Items • Measuring Instruments • NDT Testing Instruments • Metal Detector • Construction Instruments

ISO CERTIFIED NO. 21EQCV51
SSI REGD. NO. EM22400721040204
IEC NO. : 0814021514

CALIBRATION CERTIFICATE

REF NO :- WB-260805

DATE:-25/08/2026

NAME OF PARTY :- CHODHARY CONSTRUCTION

EQUIPMENT NAME :- WEIGHT BALANCE (30 KG)

CAPACITY & MAKE :- REPTECH

SERIAL NO :- SI-260805

NEXT DUE DATE:-24/08/2027

SITE LOCATION :- 300 MW HYBRID NTPC BLOCK-2 LAKADIYA

SR.NO	APPLIED WEIGHT IN (KG)	ACTUAL VALUE IN (g/kg)	UNCERTAINTY AT 95% C.L. (k=2)
01	0.05	0.050	±0.6 g
02	0.1	0.100	±0.6 g
03	0.5	0.501	±6 g
04	1.0	0.999	±6 g
05	2.0	1.998	±6 g
06	5.0	4.998	±6 g
07	10.0	9.996	±6 g
08	20.0	19.995	±6 g
09	30.0	29.995	±6 g

CALIBRATED BY: YOGESH B JOSHI

- REMARKS: This certificate is valid for 12 months from the date of calibration.
- This certificate refers to the value obtained at the time of calibration.
- NOTE: Uncertainty is calculated at a confidence level of 95% (k=2).

DETAILS & TRACEABILITY OF MASTER INSTRUMENT

MASTER INST: Standard E2 / F1 Class Weights Set

SERIAL NO: SW-E2-500

VALID UNTIL: 22/02/2027

RANGE / CAP: 1 mg - 30 kg

CALIB CERT: RRSL/WB/26-905

TRACEABILITY: Calibrated & Traceable to NPL / Regional Reference Stan

[QR CODE]



FOR, SHREEJI INSTRUMENTS

Y B JOSHI
PROPRIETOR

Regd. Office: Shop 9/G, Karnavati Shopping Center, P.D. Pandya College Road,
Nr. Burhani Society, B/h. Smruti Mandir, Ghodasar, Ahmedabad-380 050.

Mobile: 99049 04610, 93282 01463, 93771 96244

E-mail: shreejiinstrument83@yahoo.com, info@shreejiinstruments.com

Website: www.shreejiinstruments.com

Workshop : 64-65, Gopinath Industrial Hub,
Ambica Tube Mill Compound,
Nr. Vatva GIDC Over Bridge,
Vatva, Ahmedabad-382445.

Mobile: 8734866224