

MEETING 5: Full System Testing

Date: Thursday, July 15, 2026 Time: 12:00 PM - 2:00 PM
Venue: Level 5, COCIS Attendees: All team members present

Agenda

- Full system assembly review
- Testing results
- Bug fixes
- Documentation review

Discussion

- Gideon assembled the complete system on the IV stand with the printed bracket
- Gideon and Zahra tested the full system together
- Tested LED responses for Normal, Low, and Critical states - All worked
- Following Precious's research findings, the team agreed to remove the servo clamp mechanism from the design, as it was found to be unnecessary and potentially unsafe for the intended use case
- Tested ThingSpeak data upload - Data flowing properly
- Tested app fetching and displaying data - Successful
- Joshua fixed a buffer bug that was causing data loss
- Precious updated the Technical Report and SRS to reflect the removal of the servo clamp
- Rahim added features to the app based on feedback

Decisions Made

- Remove servo clamp mechanism from the system based on Precious's research recommendation
- Add EEPROM for buffer persistence
- Finalize app push notifications
- Complete remaining documentation

Action Items

Task	Assigned To	Status
Fix buffer bug	Joshua	Done
Remove servo mechanism from system	Gideon & Zahra	Done
Test app with live data	Zahra	Done
Complete documentation	Precious	In progress