

# 4D Dynamic Contrast-Enhanced Breast CT: Evaluation of quantitative accuracy

Juan J. Pautasso, Mikhail Mikerov, Liselot Goris, Koen Michielsen and Ioannis Sechopoulos

Advanced X-ray Tomographic Imaging (AXTI) Lab  
Department of Medical Imaging  
Radboud University Medical Center

---

# Disclosures

This work was funded by European Research Council (ERC);  
European Union's Horizon 2020,  
Grant/Award Number: 864929

*Ioannis Sechopoulos* has research agreements with  
Siemens Healthcare, Canon Medical Systems,  
ScreenPoint Medical, Sectra Benelux, Volpara Healthcare,  
Lunit, a speaker agreement with Siemens Healthcare,  
and is a Scientific Advisory Board member of Koning Corp.




---

# BREAST4D




European Research Council  
Established by the European Commission





**Treatment**



# Breast cancer heterogeneity





---

# 4-dimensional dynamic contrast-enhanced dedicated breast computed tomography





---


# 4-dimensional dynamic contrast-enhanced dedicated breast computed tomography



SUBJECT 061  
19780813  
00061

## Transverse view

Looking down  
from head




W/L: 456 / 40 Type: MIP Slab: 3.0mm Zoom: 71%

SUBJECT 061  
19780813  
00061

## Coronal view

Looking from  
anterior




W/L: 456 / 40 Type: MIP Slab: 3.0mm Zoom: 71%

BILATERAL STUDY, [Right](D)  
20211119  
Right

SUBJECT 061  
19780813  
00061


## Sagittal view

Looking from  
left



W/L: 456 / 40 Type: MIP Slab: 3.0mm Zoom: 71%


SUBJECT 061  
19780813  
00061



W/L: 456 / 40 Type: MIP Zoom: 63% Full

BILATERAL STUDY, [Right](D)  
20211119  
Right

BILATERAL STUDY, [Right](D)  
20211119  
Right




**X-ray scatter**




**Image quality**  
(quantitative information)


## DEEP LEARNING CORRECTED



## UNCORRECTED



WW/WL: 0.01/0.03  $\text{mm}^{-1}$




Ref.: Pautasso, J. J., Caballo, M., Mikerov, M., Boone, J. M., Michielsen, K., & Sechopoulos, I. (2023). Deep learning for x-ray scatter correction in dedicated breast CT. *Medical physics*, 50(4), 2022-2036.

---

# 4-dimensional dynamic **contrast-enhanced** dedicated breast computed tomography

## Optimal settings for imaging





---

# 4-dimensional dynamic contrast-enhanced dedicated breast computed tomography

# 4D DCE-bCT protocol


Contrast Enhanced Contrast Timer (sec)  Description **IWBI 2024**

Contrast R 


Pre-contrast Post-contrast 1 Post-contrast 2 Post-contrast 3

Prep Ends: 0 secs (Inject) 100ml, 2ml/s

Pulse Width (ms):  Scan Duration (s):  Total Pulses:  Pulses / Rev:




**Pre-contrast**




**Post-contrast**


# 4D DCE-bCT protocol





# 4D DCE-bCT protocol



# Mean glandular dose in 4D DCE-bCT protocol










In





Out




Center


 = 0.25 mm









# Post-contrast analysis (100-second Intervals)



Post-contrast 1




Post-contrast 2




Post-contrast 3


# Post-contrast analysis (100-second Intervals)



Post-contrast 1




Post-contrast 2




Post-contrast 3


# Post-contrast analysis (100-second Intervals)



Post-contrast 1



Post-contrast 2




Post-contrast 3

---

# Limitations




**Timing of peak  
capture**



**Fluid bolus behavior**



# Conclusion



**4D DCE-bCT has the potential to provide quantitatively accurate estimates of iodine concentration**