2260B Series

360W and 720W Multi-Range Programmable DC Power Supplies

Verification Procedure

Rev. A / March 2014



Series 2260B

360W and 720W Multi-Range Programmable DC Power Supplies

VERIFICATION PROCEDURE



This manual contains proprietary information, which is protected by copyright. All rights are reserved. No part of this manual may be photocopied, reproduced or translated to another language without prior written consent of the Keithley company.

The information in this manual was correct at the time of printing. However, Keithley continues to improve products and reserves the right to change specifications, equipment, and maintenance procedures at any time without notice.

VERIFICATION

Test equipment

- 1. DMMs (2) with minimum performance of 5- ½ digit resolution and 0.012% DC volt accuracy
- 2. DC Load or pure load
- 3. Oscilloscope with minimun bandwidth of 100MHz
- 4. AC Power Source
- 5. Insulation resistance/Hipot tester
- 7. 200A Shunt
- 8. 400A Shunt
- 9. PC (personal computer)

Sonai computer)	Test conditions	Limits	Notes
			Power off
		D C 300 V _ 201V132	1 Ower on
Resistance		DC500V>30MO	
		DC300 V _ 30 W 32	
Withstanding		DC 2kV 1MIN < 4mA	Power on
_		De Zity Timit (IIII)	i ower on
, 5.000			
	3.Test AC line and Ground		
Overshoot	1. Set the current to the	≤3.3V	CV mode.
Verification	maximum.		
	2. Set the voltage to 3.0V		
	3. Set the DSO to single mode.		
	4. Set the DSO vertical range to		
	0.5V/Div.		
	5. Turn the output on and see		
	that the maximum voltage is		
	≤3.3V.		
<u> </u>	•		CV mode.
•		, ·	
Accuracy			
		•	
		•	
	setting is within mints.		
i			
Voltage	1 Turn the power on	With less than 8hr warmun	I CV mode
Voltage Measurement	Turn the power on. Set the current to maximum	With less than 8hr warmup	CV mode.
Measurement	2. Set the current to maximum.	period:	CV mode.
_	2. Set the current to maximum.3. Set voltage to:	period: ≤0.1%+10mV	CV mode.
Measurement	2. Set the current to maximum.3. Set voltage to:31.50V for 2260B-30XX	period: ≤0.1%+10mV With more than 8hr warmup	CV mode.
Measurement	2. Set the current to maximum. 3. Set voltage to: 31.50V for 2260B-30XX 84.00V for 2260B-80XX.	period: ≤0.1%+10mV With more than 8hr warmup period:	CV mode.
Measurement	2. Set the current to maximum. 3. Set voltage to: 31.50V for 2260B-30XX 84.00V for 2260B-80XX. 4. Turn the output on.	period: ≤0.1%+10mV With more than 8hr warmup period: 30Volt models:	CV mode.
Measurement	 2. Set the current to maximum. 3. Set voltage to: 31.50V for 2260B-30XX 84.00V for 2260B-80XX. 4. Turn the output on. 5. Verify that the difference 	period: ≤0.1%+10mV With more than 8hr warmup period: 30Volt models: ≤0.1%+25mV	CV mode.
Measurement	 2. Set the current to maximum. 3. Set voltage to: 31.50V for 2260B-30XX 84.00V for 2260B-80XX. 4. Turn the output on. 5. Verify that the difference between the voltage on the 	period: ≤0.1%+10mV With more than 8hr warmup period: 30Volt models: ≤0.1%+25mV 80Volt models:	CV mode.
Measurement	 2. Set the current to maximum. 3. Set voltage to: 31.50V for 2260B-30XX 84.00V for 2260B-80XX. 4. Turn the output on. 5. Verify that the difference 	period: ≤0.1%+10mV With more than 8hr warmup period: 30Volt models: ≤0.1%+25mV	CV mode.
	Test Insulation Resistance Withstanding Voltage Test Overshoot	Test Test conditions Insulation	Test Test conditions Limits Insulation + output terminal and ground, - output terminal and ground. DC500V≥20MΩ AC Neutral and ground, AC line and ground. DC500V≥30MΩ Withstanding Voltage Test 1. short-circuit + output terminal and ground 2. short-circuit AC Neutral and AC line 3.Test AC line and Ground DC 2kV 1MIN < 4mA

1

4	Voltage display accuracy	 Turn the power on. Set the current to max Set the votlage to any value. Turn the output on. Verify that the reading on the DMM and the display are within limits. Note: Verificaiton procedure applies to all voltages between 3V to rating voltage. 	≤0.1%+2 digits.	CV mode.
5	OVP verification	1. Hold the function key and turn the power on. This will enter the "Power On Configuration" settings. F-90 will be displayed on screen when you have successfully entered the "Power on Configuration" settings. 2. Set F-95 to 1 (Power switch trip disable) Procedure: Turn the voltage knob until F-95 is shown. Turn the current knob until 1 is shown. Press the voltage knob to save the setting. 3. Turn the power off. 4. Turn the power off. 4. Turn the power on. 5. Press the OVP/OCP key and set the OVP value. 6. Press the Set key and then set the output voltage to a value lower than the OVP voltage level. 7. Turn the output on. 8. Progressively increase the output voltage until the OVP is tripped. 9. Verify the OVP funciton over the entire OVP range.	≤1%	CV mode OVP setting range 2260B-30XX: 3V~33V 2260B-80XX: 8V~88V
6	Load Regulation (Load effect)	1. Set the current to the maximum. 2. Set the voltage to: 2260B 30XX:10V, 30V, 2260B 80XX: 26.7V, 80V. (30-36&80-13:360W, 30-72&80-27:720W) 4. Set the DC load to CC mode and turn on. 5. Measure the ouptut voltage on the DMM. (record the voltage from the sense terminals.) 6. Turn the load off. 7. Record the output voltage again. Calculate the difference in voltage output between full load and no load. Verify that the voltage ouptut difference is within limits.	≤ 0.05%+5mV.	CV Mode
7	Line Regulation (Source Effect)	 Set the line input voltage to 90V. Follow steps 1-5 of item 6 to record the output voltage. 	≤.05%+3mV	CV Mode

	I	2 Channel II	<u> </u>	
		3. Change the line input voltage		
		from 90V to 265V.		
		4. Record the output voltage.		
		Verify that the output voltage doesn't differ more than		
		limits.		
8	Voltage Ripple and	1. Follow steps 1-4 of item 6.	RIPPLE: mVrms	CV Mode.
0	Noise	2. Set the scope BW limit to	30-36&80-13:≤7	CV Mode.
	IVOISC	20MHz.	30-72&80-27:≤11	
		3. Use the scope to measure	30-72&80-27.511	
		from the "Sense" terminals.	NOISE: mVn n	
		4. Confirm that the voltage ripple	NOISE: mVp-p 30-36&80-13:≤60	
		and noise meets the	30-72&80-27:≤80	
		specifications.	30-72&80-2780	
		5. Use a pure DC load if		
		necessary for the noise		
		measurement.		
9	Current	1. Turn the power on.	With less than 8hr warmup	CC mode
	Programming	2. Set the voltage to the	period:	
	Accuraccy	maximum setting.	$30-36: \le 0.1\% +30 \text{ mA}$	
		3. Set the current to the	30-72: ≦ 0.1% +60 mA	
		following values:	$80-13: \le 0.1\% + 10 \text{ mA}$	
		2260B 30-36: 37.80A	80-27: ≦ 0.1% +30 mA	
		2260B 30-72: 75.60A 2260B 80-13: 14.18A	With more than 8hr warmup	
		2260B 80-27: 28.36A	period:	
		4. Turn the output on.	30-36: ≦ 0.1% +48 mA	
		5. Check the difference between	30-72: ≦ 0.1% +96 mA	
		the setting current and the	80-13: ≦0.1% +17 mA	
		measured current across the	80-27: ≤ 0.1% +44 mA	
		shunt resistor (DMM voltage		
		across shunt/shunt R =		
		measured current).		
		(200A shunt)		
10	Current	1. Follow steps 1-4 of item 9.	With less than 8hr warmup	
	Measurement	2. Verify that the difference	period:	
	Accuraccy	between the current calculated	30-36: ≤ 0.1% +30 mA	
	(readback accuracy)	across the shunt resistor to that	30-72: ≦ 0.1% +60 mA	
		of the remote reading on the PC	80-13: ≤ 0.1% +10 mA	
		is within limits.	80-27: ≤ 0.1% +30 mA	
			With more than 8hr warmup	
			period:	
			$30-36: \le 0.1\% +48 \text{ mA}$	
			$30-72: \le 0.1\% + 96 \text{ mA}$	
			80-13: ≤0.1% +17 mA	
			80-27: ≤ 0.1% +44 mA	
11	Front Panel Display	1. Follow steps 1-4 of item 9.	$30-36$: $\leq 0.1\% + 4$ digits	CC Mode
	Accuracy	2. Calculate the difference	$30-72: \le 0.1\% + 7 \text{ digits}$	CV mode
		between the measured current	$80-13: \le 0.1\% + 2 \text{ digits}$	(with load)
		across the shunt resistor and the displayed value.	$80-27: \le 0.1\% + 4 \text{ digits}$	
12	OCP function check	1. Hold the function key and	≤1%	CC mode
'-	JCI IMICLION CHECK	turn the power on.	_ 170	OCP SET
		This will enter the "Power On		range =
		Configuration"settings.		10%~110
		F-90 will be displayed on screen		rating range.
		when you have successfully		
		entered the "Power on		
		Configuration" settings.		
		2. Set F-95 to 1 (Power switch		

	•	1	·	1
		trip disable) Procedure: Turn the voltage knob until F-95 is shown.Turn		
		the current knob until 1 is		
		shown. Press the voltage knob to		
		save the setting.		
		3. Turn the power off.		
		4. Turn the power on.		
		5. Press the OVP/OCP key and set the OCP value.		
		6. Press the Set key and set the		
		output current less than the		
		OCP value.		
		7. Turn the output on.		
		8. Progressively increase the		
		output current until the OCP is		
		tripped. 9. Verify the OCP funciton over		
		the entire OCP range.		
13	Current load effect	1. Set the voltage to:	≤0.1%+5mA	CC mode
		2260B 30XX: 10V, 30V:		
		2260B 80XX: 26.7V, 80V.		
		2. Set thecurrent values to any		
		value.		
		3. Turn the load on.4. Set the DC load to CR mode.		
		5. Note the power rating of the		
		power supply.		
		6. Turn the electronic load on.		
		7. Modify the DC load to make		
		2260B output change from CV to		
		CC.		
		8. Short the DC load.		
		9. Compare the difference in		
		output current before/after the DC load was shorted. (From the		
		DMM current measurement)		
14	Current source	1. Follow steps 1-4 of item 9.	≦0.1%+5mA	CC mode
	effect	2. Vary the line input voltage		
		from 90V to 265V.		
		3. Confirm that the output		
		current does not change more than limits		
15	Ripple Current	1. Set the voltage to:	RIPPLE: mVrms	CC mode
	measurement	2260B 30XX: 10V, 30V.	30-36:≤72	
		2260B 80XX: 26.7V, 80V.	30-72:≤144	
		2. Set the current any value.	80-13:≤27	
		3. Turn the output on.	80-27:≤54	
		4. Set the DC load to CR mode.	Unit: mArms	
		5. Note the power rating of the		
		power supply. 6. Turn the load on.		
		7. Modify the current on the		
		2260B until the mode changes		
		from CV to CC mode.		
		8. Calculate DCA by measuring		
		the DCV across the shunt		
		resistor with the DMM.		
		9.Measure the voltage (DCV) from the sense terminals and		
		calculate R by R=DCV/DCA.		
L	1		I.	İ

		10 C		<u> </u>
		10. Set the DMM to ACV		
		measurement (DMM connected		
		to the sense terminals).		
		11. Record the AC voltage on the		
		DMM.		
		12. Calculate the ACA value .		
		(ACA=ACV/R)		
16	Recovery Time	1. Use an electronic load to test	< 1ms	CV mode.
		the dynamic response. (time		
		taken for the voltage to recover		
		from a 50% change in load		
		current. (50% to 100%)).		
		2. Set the DC load to CC mode.		
17	Rise Time Full Load	1. Set thevoltage to:	≤50mS	CV mode.
		2260B 30XX: 10V, 30V:		
		2260B 80XX: 26.7V, 80V		
		2. Set the current to the		
		maximum.		
		3. Turn the output on.		
		4. Note the power rating of the		
		power supply.		
		5. The load should be set to CR		
		mode.		
		6. Use the DSO to measure from		
		the sense terminals.		
		7. Turn the DC load on.		
		8. Turn the output on again.		
		9. Verify the 10%~90% ouptut		
		rise time for the waveform.		
18	Fall Time Full Load	1. Follow steps 1-7 for item 17.	≤50mS	CV mode.
	Tun Time Tun Loud	2. Turn the output off.	_301113	ev mode.
		3. measure the 10%~90% fall		
		time of the output with full load.		
19	Fall Time No Load	1. Set the voltage to the rating	≤500mS	CV mode.
	Tun Time IVO Loud	setting.	_3001113	ev mode.
		2. make a load go from On→Off.		
		3. Verify the 10%~90% fall time		
		for the output with no load.		
20	EXT Voltage ,	1. Hold the function key and	≤ 8 hrs warm up:	CV mode.
20	Voltage Out	turn the power on.	≤0.5%	CV mode.
	Tollage Out	This will enter the "Power On	_ *****	
		Configuration"settings.	>8hrs warm up:	
		F-90 will be displayed on screen	30XX: ≤1.5%±15mV	
		when you have successfully	80XX: ≤1.5%±40mV	
		entered the "Power on		
		Configuration" settings.		
		2. Set F-90 to 1 (CV control – Ext		
		`		
		voltage). Do this by turning the current knob until 1 is shown.		
		Press the voltage knob to save		
		the setting.		
		3. The external voltage should be		
		connected as per the user		
		manual.		
		4. Turn the power off.		
		5. Turn the power on.		
		6. Set the external power supply		
		voltage from 0.1V to 10V.		
		7. Make sure you have an		
		external DMM to monitor the		
İ		actual output value from the		

		external power supply.]	
		8. Make sure the relative voltage		
		output is within limits.		
21	Ext R	1. Hold the function key and	≤ 8 hrs warm up:	CV mode.
	Voltage out	turn the power on.	≤1.5%	
		This will enter the "Power On	>8hrs warm up:	
		Configuration" settings. F-90 will be displayed on screen	30XX: ≤1.5%±15mV	
		when you have successfully	80XX: ≤1.5%±40mV	
		entered the "Power on		
		Configuration" settings.		
		2. Set F-90 to 2 (Ext-R ∠ − CV		
		control)		
		Turn the current knob until 2 is		
		shown. Press the voltage knob to		
		save the setting.		
		3. The external resistance should		
		be connected as per the user		
		manual.		
		4. Turn the power off.		
		5. Turn the power on.		
		6. Set the external resistance		
		from $0.1k\Omega$ to $10k\Omega$.		
		7. Make sure the relative voltage		
22	Fut DA	output is within limits.	0 has 110 mg 1101	CV mode.
22	Ext R	1. Follow the steps for item 21 except set F-90 to 3 (Ext-R\(\sigma -	≤ 8 hrs warm up:	Cv mode.
	Voltage out	CV control)	≤1.5%	
		CV control)	>8hrs warm up:	
			$30XX: \le 1.5\% \pm 15mV$	
23	VMON	1 Hold the function key and	80XX: ≤ 1.5%±40mV	CV Mode.
23	VIVION	1. Hold the function key and turn the power on.	≤1%	CV Mode.
		This will enter the "Power On		
		Configuration" settings.		
		F-90 will be displayed on screen		
		when you have successfully		
		entered the "Power on		
		Configuration" settings.		
		2. Set F-90 to 0 (panel control)		
		to disable the external control		
		configuration. Do this by turning		
		the current knob until 0 is		
		shown. Press the voltage knob to		
		save the setting. 3. The DMM should be		
		connected as per the user		
		manual.		
		4. Turn the power off.		
		5. Turn the power on.		
		6. The monitor output voltage		
		ranges from 1% to 100% of		
		fullscale.		
		7. Make sure the relative voltage		
		output is within limits.		
24	EXT Voltage,	1. Hold the function key and	≤ 8 hrs warm up:	CC mode.
	Current Out	turn the power on.	≤1%	
		This will enter the "Power On	>8hrs warm up:	
		Configuration"settings.	30-36: ≤1%±18mA	
		F-90 will be displayed on screen	30-72: ≤1%±36mA	
		when you have successfully	80-13: ≤1%±6mA	
1		entered the "Power on	80-27: ≤1%±13mA	

		Configuration" settings. 2. Turn the Voltage knob until F- 91 is displayed. Set F-91 to 1 (CC control – Ext voltage) by turning the current knob until 1 is shown. Press the voltage knob to save the setting. 3. The external voltage should be connected as per the user manual. 4. Turn the power off. 5. Turn the power on. 6. Set the external power supply voltage from 0.1V to 10V. 7. Make sure you have an external DMM to monitor the actual output value from the external power supply. 8. Make sure the relative current output is within limits.		
25	Ext R Current out	 Hold the function key and turn the power on. This will enter the "Power On Configuration" settings. F-90 will be displayed on screen when you have successfully entered the "Power on Configuration" settings. Turn the Voltage knob until F-91 is displayed. Fe F-91 to 2 (Ext-R CC control) by turning the current knob until 2 is shown. Press the voltage knob to save the setting. The external resistance should be connected as per the user manual. Turn the power off. Turn the power on. Set the external resistance from 0.1kΩ to 10kΩ. Make sure the relative current outtput is within limits. 	≤ 8 hrs warm up: ≤1.5% >8hrs warm up: 30-36: ≤1.5%±18mA 30-72: ≤1.5%±36mA 80-13: ≤1.5%±6mA 80-27: ≤1.5%±13mA	CC mode.
26	Ext R \(\sum \) Voltage out	1. Follow the steps for item 21 except set F-91 to 3 (Ext-R CC control)	≤ 8 hrs warm up: ≤1.5% >8hrs warm up: 30-36: ≤1.5%±18mA 30-72: ≤1.5%±36mA 80-13: ≤1.5%±6mA 80-27: ≤1.5%±13mA	CC mode.
27	IMON	1. Hold the function key and turn the power on. This will enter the "Power On Configuration" settings. F-90 will be displayed on screen when you have successfully entered the "Power on Configuration" settings. 2. Turn the Voltage knob until F-91 is displayed. Set F-91 to 0 (panel control) by	≤1%	CC Mode.

is shown. Press the voltage knob to save the setting. 3. The DMM should be connected as per the user manual. 4. Turn the power off. 5. Turn the power on. 6. The monitor output voltage ranges from 1% to 100% of fullscale. 7. Make sure the relative voltage output is within limits. 28 Series Operation Verification 1. Hold the function key and turn the power on.	
3. The DMM should be connected as per the user manual. 4. Turn the power off. 5. Turn the power on. 6. The monitor output voltage ranges from 1% to 100% of fullscale. 7. Make sure the relative voltage output is within limits. 28 Series Operation Verification 1. Hold the function key and turn the power on.	
connected as per the user manual. 4. Turn the power off. 5. Turn the power on. 6. The monitor output voltage ranges from 1% to 100% of fullscale. 7. Make sure the relative voltage output is within limits. 28 Series Operation Verification 1. Hold the function key and turn the power on.	
manual. 4. Turn the power off. 5. Turn the power on. 6. The monitor output voltage ranges from 1% to 100% of fullscale. 7. Make sure the relative voltage output is within limits. 28 Series Operation Verification 1. Hold the function key and turn the power on.	
4. Turn the power off. 5. Turn the power on. 6. The monitor output voltage ranges from 1% to 100% of fullscale. 7. Make sure the relative voltage output is within limits. 28 Series Operation Verification 1. Hold the function key and turn the power on.	
5. Turn the power on. 6. The monitor output voltage ranges from 1% to 100% of fullscale. 7. Make sure the relative voltage output is within limits. 28 Series Operation Verification 1. Hold the function key and turn the power on.	
6. The monitor output voltage ranges from 1% to 100% of fullscale. 7. Make sure the relative voltage output is within limits. 28 Series Operation Verification Verification CV Monitor Output voltage ranges from 1% to 100% of fullscale. 7. Make sure the relative voltage output is within limits. CV Monitor Output voltage ranges from 1% to 100% of fullscale. 7. Make sure the relative voltage output is within limits.	
ranges from 1% to 100% of fullscale. 7. Make sure the relative voltage output is within limits. 28 Series Operation Verification turn the power on. CV Moreover 1000 turn the power on.	
7. Make sure the relative voltage output is within limits. 28 Series Operation Verification turn the power on. CV Months	
output is within limits. 28 Series Operation 1. Hold the function key and Verification turn the power on.	
28 Series Operation 1. Hold the function key and Verification turn the power on.	
Verification turn the power on.	odo
	oue.
This will enter the "Power On	
Configuration"settings.	
F-90 will be displayed on screen	
when you have successfully entered the "Power on	
Configuration" settings.	
2. Turn the Voltage knob until F-	
93 is displayed.	
Set F-93 to 0 for the master unit,	
and to 4 for the slave units. Do this by turning the current knob	
until 0 or 4 is shown. Press the	
voltage knob to save the setting.	
3. The units should be	
connected in series as per the	
user manual. 4. Series connection: Master	
+terminal to slave -terminal.	
5. Connect the DMM across the	
Master+ and Slave – terminals.	
6. Set the voltage from the Master unit and verify the series	
operation.	
7. Verify the series operation	
with the bleeder control set to	
Off & On on the master unit.	
(bleeder control for Slave units is fixed to On.)	
29 Parallel Operation 1. Hold the function key and CC Mo	ode.
Verification turn the power on.	
This will enter the "Power On	
Configuration"settings. F-90 will be displayed on screen	
when you have successfully	
entered the "Power on	
Configuration" settings.	
2. Turn the Voltage knob until F- 93 is displayed.	
Set F-93 to 2 for the master unit,	
and to 3 for the slave units. Do	
this by turning the current knob	
until 2 or 3 is shown. Press the	
voltage knob to save the setting. 3. The units should be	
connected in parallel as per the	

		user manual.	
		4. From Master +terminal to	
		slave +terminal, Master –	
		terminal to slave -terminal.	
		5. Connect the DMM across the	
		Master+ and Master – terminal.	
		6. Set the current from the	
		Master unit and verify the	
		parallel operation.	
		7. Verify the parallel operation	
		with the bleeder control set to	
		Off & On on the master unit.	
		(bleeder control for Slave units is	
		fixed to off.)	
30	Interface	1. Lan Verification. Use the LAN	
	Verification	interface to remotely check the	
		system information (model	
		number, serial number,	
		firmware, Mac address etc.).	
		2. GPIB Verification. Use the	
		optional GPIB-USB adapter and	
		the NI Measurement and	
		Automation Explorer to verifiy	
		the GPIB functionality.	
		3. Confirm that the USB port is	
		functioning. The display should	
		show "USB ON"	

Specifications are subject to change without notice.

All Keithley trademarks and trade names are the property of Keithley Instruments, Inc.

All other trademarks and trade names are the property of their respective companies.

Keithley Instruments, Inc.

Corporate Headquarters • 28775 Aurora Road • Cleveland, Ohio 44139 • 440-248-0400 • Fax: 440-248-6168 • 1-888-KEITHLEY • www.keithley.com

