
Questions about IntelliGen NCS network protocol

Karl Anderson <Karl.Anderson@entegris.com>
To: Justus Languell <justus@tamu.edu>

Fri, Sep 25, 2026 at 11:33 AM

This Message Is From an External Sender

This message came from outside your organization.

Hi Justus,

I briefly looked at your questions. One I think I can answer. Some are going beyond my basic programming skills.

I would like to see if I can connect you with someone from our firmware/software Engineering team.

What days next week would you be available for a Teams session?

3. Does the COM1 on the NCS correspond to logical serial number 0 or 1?

COM1 is logical serial number 0. COM7 is logical serial number 7.

I am going by the formula found in the attached template. The formula should be in the IntelliGen NCS User Guide.

Please let me know if there are more questions.

Karl Anderson

Manager - Market and Technical Solutions

Advanced Purity Solutions Division

InVue Sensing & Control

IntelliGen Dispense Systems

M +1 763 370 8167



entegris.com



101 Peavey Road
Chaska, MN 55318 USA



From: Justus Languell <justus@tamu.edu>
Sent: Thursday, September 24, 2026 8:00 AM
To: Karl Anderson <Karl.Anderson@entegris.com>
Subject: Questions about IntelliGen NCS network protocol

You don't often get email from justus@tamu.edu. [Learn why this is important](#)

External Email: This email originated from outside of the organization. Do not click links or open attachments unless you recognize the sender and know the content is safe.

[Quoted text hidden]

ENTEGRIS PROPRIETARY AND CONFIDENTIAL – INTERNAL



SVID List for Stratus, template 18Mar2021.xlsx
19K