



Entegris RULV/ULV/LV/MV/HMV/HV20 SVIDs

*Generated at Fri Nov 8 12:52:46 2024
For firmware version V1011.1536*

DESCRIPTION:

SVIDs for the IntelliGen RULV, ULV, LV, MV, HMV, and HV20 generation dispense systems.

This document is auto generated directly from the source code of MMI/FIRMWARE.

Entegris Confidential Not For Commercial Use

Entegris RULV/ULV/LV/MV/HMV/HV20 SVIDs	1
SVIDs	3
Appendix	10
SVIDs	10

SVIDs

SVID	Size	Type	Data	Description
1	24	STRING	SerialNumbers.Pump[24]	Pump Serial Number
2	24	STRING	SerialNumbers.FillMotor[24]	Fill Motor Serial Number
3	24	STRING	SerialNumbers .DispenseMotor[24]	Dispense Motor Serial Number
4	24	STRING	SerialNumbers.FillPressure[24]	Fill Pressure Serial Number
5	24	STRING	SerialNumbers .DispensePressure[24]	Dispense Pressure Serial Number
6	24	STRING	SerialNumbers.MainBoard[24]	Main PCB Serial Number
7	24	STRING	SerialNumbers .InternalBoard[24]	Internal PCB Serial Number
8	24	STRING	SerialNumbers .InterfaceBoard[24]	Interface PCB Serial Number
10	24	STRING	Version.Model[24]	Model Name
11	48	STRING	Version.Version[48]	Firmware Version String
12	24	STRING	SystemNames.CustomerName[24]	Customer Name
13	24	STRING	SystemNames.FabName[24]	Fab Name
14	24	STRING	SystemNames.ToolName[24]	Tool Name
15	24	STRING	SystemNames.CoaterName[24]	Coater Name
16	24	STRING	SystemNames.NozzleName[24]	Nozzle Name
17	24	STRING	SystemNames.ResistName[24]	Resist Name
18	24	STRING	SystemNames.PumpName[24]	Pump Name
19	24	STRING	SystemNames.FluidName[24]	Fluid Name/Type
20	24	STRING	Recipe.RecipeTitle[24]	Description
21	24	STRING	DispenseInfo.JobInfo[24]	Job Information
257	2	INT16	RtStatus.Status	System Status
258	2	INT16	UtilityStatus.HomingCurrent_mA	Peak homing current of DC motor
259	2	INT16	UtilityStatus.reserved1	reserved1
260	2	INT16	UtilityStatus .DisCurrent_UnCal_mA	Dispense Motor Current
261	2	INT16	UtilityStatus .DispenseCurrent_mA	Dispense Motor Current
263	2	INT16	FillMotor.Position_uL	Fill Position in uL
264	2	INT16	Recharge.AutoVentEnable	Auto-Vent Enable
265	2	INT16	Recharge.VentFreqEnable	Vent Frequency Enable

266	2	INT16	Recipe.OpenAdjust	Outlet Valve Open Adj.
267	2	INT16	Recipe.CloseAdjust	Outlet Valve Close Adj.
268	2	INT16	Recipe.VolumeCalibration	Volume Calibration
269	2	INT16	PressureStatus .Dis_pressure_cPSI	dispense calibrated cPSI
270	2	INT16	PressureStatus .Fill_pressure_cPSI	Fill calibrated cPSI
275	2	INT16	DispenseInfo.ReadyPressure	Ready - Ready Pressure
276	2	INT16	DispenseInfo.PeakPressure	Dispense - Maximum Pressure
277	2	INT16	DispenseInfo .ReferencePeakPressure	Reference Dispense - Maximum Pressure
278	2	INT16	DispenseInfo.AveragePressure	Dispense - Average Pressure
279	2	INT16	DispenseInfo .ReferenceAveragePressure	Reference Dispense - Average Pressure
280	2	INT16	DispenseInfo.CutoffPressure	Dispense - Cutoff Pressure
281	2	INT16	DispenseInfo .ReferenceCutoffPressure	Reference Dispense - Cutoff Pressure
282	2	INT16	DispenseInfo.MaxFiltrationPrs	Filtration - Maximum Upstream Pressure
283	2	INT16	DispenseInfo .ReferenceMaxFiltrationPrs	Reference Filtration - Maximum Upstream Pressure
284	2	INT16	DispenseInfo.MinFillPrs	Fill - Minimum Fill Pressure
285	2	INT16	DispenseInfo .ReferenceMinFillPrs	Reference Fill - Minimum Fill Pressure
286	2	INT16	SkipstepInfo.StepMismatch_uL	StepMismatch uL
287	2	INT16	UtilityStatus.N2Sensor_cPSI	N2 Sensor
288	2	INT16	UtilityStatus.VacSensor_cPSI	Vacuum Sensor
289	2	INT16	DispenseInfo .ReferenceReadyPressure	Reference Ready - Ready Pressure
290	2	INT16	DispenseInfo.AverageCurrent	Dispense - Average Motor Current
291	2	INT16	DispenseInfo .ReferenceAverageCurrent	Reference Dispense - Average Motor Current
292	2	INT16	DispenseInfo.PredictedCurrent	Dispense - Predicted Motor Current
293	2	INT16	DispenseInfo .ReferencePredictedCurrent	Reference Dispense - Predicted Motor Current
294	2	INT16	DispenseInfo .AverageDeltaPressure	Filtration - Average Delta Pressure
295	2	INT16	DispenseInfo .ReferenceAverageDeltaPressure	Reference Filtration - Average Delta Pressure
296	2	INT16	DispenseInfo.FillMotorMisMatch	Cycle - Fill Motor Mismatch

297	2	INT16	DispenseInfo .ReferenceFillMotorMismatch	Cycle Dispense - Fill Motor Mismatch
298	2	INT16	UtilityStatus .TempFillSensor_cC	Fill Temperature
299	2	INT16	UtilityStatus.TempDisSensor_cC	Dispense Temperature
300	2	INT16	UtilityStatus.N2SensorTemp_cC	N2 Sensor Temp
301	2	INT16	UtilityStatus.VacSensorTemp_cC	Vacuum Sensor Temp
302	2	INT16	RtStatus.OfflineState	OfflineState
513	4	FLOAT	PressureStatus .Dis_pressure_PSI	dispense calibrated PSI
514	4	FLOAT	PressureStatus .Fill_pressure_PSI	Fill calibrated PSI
769	2	UINT16	FirmwareInfo.ProductQualifier	Qualifer for ProductCode to be valid
770	2	UINT16	FirmwareInfo.Comms_Version	Communications protocol version
771	2	UINT16	FirmwareInfo.MajorVersion	Major Version Number
772	2	UINT16	FirmwareInfo.ModuleType	Interface Module Type
773	2	UINT16	FirmwareInfo.MinorVersion	Firmware Minor Version
774	2	UINT16	FirmwareInfo.ProductCode	Product ID Code
775	2	UINT16	SystemParameters.Address	Serial Address
776	2	UINT16	SystemParameters .FluidViscosity_cP	Fluid Viscosity
778	2	UINT16	SystemParameters .ReadyPressureEnable	Ready Pressure Enable
780	2	UINT16	SystemParameters.AutoRecEnable	Auto-Recirculation
781	2	UINT16	SystemParameters .AutoRecVolume_uL	Auto-Recirc. Volume
782	2	UINT16	SystemParameters .AutoRecRate_uLS	Auto-Recirc. Rate
783	2	UINT16	SystemParameters .MaintenanceModeTimeoutEnable	Maint. Mode Timeout
784	2	UINT16	SystemParameters .MaintenanceModeTimeout_min	Maint. Mode Timeout
795	2	UINT16	RtStatus.LastRecipeNum	Last Dispensed Recipe Num
796	2	UINT16	RtStatus.LastDispenseVol_uL	Last Dispensed Total volume
797	2	UINT16	RtStatus .LastDispenseConfirmation_per	Last Dispensed confirmation value
799	2	UINT16	RtStatus.CriticalErrors	counts Critical System Errors after they are logged
800	2	UINT16	RtStatus.Errors	counts System Errors after they are logged

801	2	UINT16	RtStatus.Warnings	counts System Warnings after they are logged
802	2	UINT16	RtStatus.Infos	counts System infos after they are logged
803	2	UINT16	RtStatus.TotalAlarms	counts total uncleared alarms after they are logged
808	2	UINT16	DisMotor.Position_uL	Dispense Position in uL
809	2	UINT16	SystemParameters .SerialTriggerEnable	Serial Trigger Enable
810	2	UINT16	SystemParameters.reserved8	reserved
811	2	UINT16	SystemParameters .VentOnlyVolume_uL	Vent Only Volume
812	2	UINT16	SystemParameters .VentOnlyRate_uLS	Vent Only Rate
813	2	UINT16	SystemParameters .PreFiltrationAlarmLevel	Pre-Filtration Alarm Level
814	2	UINT16	SystemParameters .FiltrationAlarmLevel	Filtration Alarm Level
815	2	UINT16	SystemParameters .UtilityAlarmLevel	Utility Alarm Level
816	2	UINT16	Recharge.VentFreq	Vent Frequency
817	2	UINT16	Recharge.FreqVentVolume	Vent Volume (Frequency)
818	2	UINT16	Recharge.FreqVentRate	Vent Rate (Frequency)
819	2	UINT16	Recharge.PurgeVolume_uL	Purge Volume
820	2	UINT16	Recharge.PurgeRate	Purge Rate
821	2	UINT16	Recharge.FillRate	Fill Rate
822	2	UINT16	Recharge.FiltrationRate	Filtration Rate
823	2	UINT16	Recharge .VariableDispenseVolume	Max. Dispense Volume
824	2	UINT16	Recipe.RecipeNum	Recipe number
825	2	UINT16	Recipe.DisRampUpTime	Ramp Up Time
826	2	UINT16	Recipe.DispenseVolume_1	Segment 1 Volume
827	2	UINT16	Recipe.DispenseRate_1	Segment 1 Rate
828	2	UINT16	Recipe.DispenseVolume_2	Segment 2 Volume
829	2	UINT16	Recipe.DispenseRate_2	Segment 2 Rate
830	2	UINT16	Recipe.DispenseVolume_3	Segment 3 Volume
831	2	UINT16	Recipe.DispenseRate_3	Segment 3 Rate
832	2	UINT16	Recipe.DispenseVolume_4	Segment 4 Volume
833	2	UINT16	Recipe.DispenseRate_4	Segment 4 Rate

834	2	UINT16	Recipe.DispenseVolume_5	Segment 5 Volume
835	2	UINT16	Recipe.DispenseRate_5	Segment 5 Rate
836	2	UINT16	Recipe.DisRampDownTime	Ramp Down Time
837	2	UINT16	Recipe.PreDispenseRecipe	Pre-Dispense Recipe
838	2	UINT16	DispenseInfo .ReferenceCycleTime	Reference Cycle - Total Time
839	2	UINT16	DispenseInfo.CycleTime	Cycle - Total Time
840	2	UINT16	DispenseInfo.FillTime	Fill - Total Time
847	2	UINT16	DispenseInfo.RecipeNum	Recipe number
848	2	UINT16	DispenseInfo.SlotID	Slot ID
849	2	UINT16	DispenseInfo.Confirmation	Dispense - Pressure Profile Compare
850	2	UINT16	DispenseInfo .ReferenceConfirmation	Reference Dispense - Pressure Profile Compare
851	2	UINT16	DispenseInfo.AverageRate	Dispense - Average Motor Rate
852	2	UINT16	DispenseInfo .ReferenceAverageRate	Reference Dispense - Average Motor Rate
853	2	UINT16	DispenseInfo.TotalVolume	Dispense - Total Motor Volume
854	2	UINT16	DispenseInfo .ReferenceTotalVolume	Reference Dispense - Total Motor Volume
855	2	UINT16	DispenseInfo.AirDetect	Dispense - Air Detect Volume
856	2	UINT16	DispenseInfo .ReferenceAirDetect	Reference Dispense - Air Detect Volume
858	2	UINT16	DispenseInfo.FlowConfirmation	Dispense - Flow Profile Compare
859	2	UINT16	DispenseInfo .ReferenceFlowConfirmation	Reference Dispense - Flow Profile Compare
860	2	UINT16	DispenseInfo.ReferenceFillTime	Reference Fill - Total Time
861	2	UINT16	DispenseInfo.FiltrationRate	Filtration - Average Motor Rate
862	2	UINT16	DispenseInfo .ReferenceFiltrationRate	Reference Filtration - Average Motor Rate
863	2	UINT16	DispenseInfo.FiltrationTime	Filtration - Total Time
864	2	UINT16	DispenseInfo .ReferenceFiltrationTime	Reference Filtration - Total Time
866	2	UINT16	DispenseInfo.AverageFillRate	Fill - Average Motor Rate
867	2	UINT16	DispenseInfo .ReferenceAverageFillRate	Reference Fill - Average Motor Rate
868	2	UINT16	DispenseInfo.VentTime	Vent - Total Time
869	2	UINT16	DispenseInfo.ReferenceVentTime	Reference Vent - Total Time

870	2	UINT16	DispenseInfo .ConfirmationMiniStyle	Dispense - Pres Profile Cmp (IGMINI style)
871	2	UINT16	DispenseInfo .ReferenceConfirmationMiniStyle	Reference Dispense - Pres Profile Cmp (IGMINI style)
1025	4	ULONG	TimeNow.Time	UTC seconds from 1 Jan 1970
1026	4	ULONG	CycleCounters.CycleCounts	Total Cycles
1027	4	ULONG	CycleCounters.PowerCycles	Total Power Cycles
1028	4	ULONG	CycleCounters.DispenseVolume	Total Dispensed Volume (L)
1029	4	ULONG	CycleCounters.FiltratedVolume	Total Filtrated Volume (L)
1030	4	ULONG	CycleCounters.VentedVolume	Total Vented Volume (L)
1031	4	ULONG	CycleCounters.PurgedVolume	Total Purged Volume (L)
1032	4	ULONG	CycleCounters.FilledVolume	Total Filled Volume (L)
1033	4	ULONG	CycleCounters .CycleCounts_reset_A	Resettable A Cycles
1034	4	ULONG	CycleCounters .PowerCycles_reset_A	Resettable A Power Cycles
1035	4	ULONG	CycleCounters .DispenseVolume_reset_A	Resettable A Dispensed Volume (L)
1036	4	ULONG	CycleCounters .FiltratedVolume_reset_A	Resettable A Filtrated Volume (L)
1037	4	ULONG	CycleCounters .VentedVolume_reset_A	Resettable A Vented Volume (L)
1038	4	ULONG	CycleCounters .PurgedVolume_reset_A	Resettable A Purged Volume (L)
1039	4	ULONG	CycleCounters .FilledVolume_reset_A	Resettable A Filled Volume (L)
1040	4	ULONG	CycleCounters .CycleCounts_reset_B	Resettable B Cycles
1041	4	ULONG	CycleCounters .PowerCycles_reset_B	Resettable B Power Cycles
1042	4	ULONG	CycleCounters .DispenseVolume_reset_B	Resettable B Dispensed Volume (L)
1043	4	ULONG	CycleCounters .FiltratedVolume_reset_B	Resettable B Filtrated Volume (L)
1044	4	ULONG	CycleCounters .VentedVolume_reset_B	Resettable B Vented Volume (L)
1045	4	ULONG	CycleCounters .PurgedVolume_reset_B	Resettable B Purged Volume (L)
1046	4	ULONG	CycleCounters .FilledVolume_reset_B	Resettable B Filled Volume (L)

1047	4	ULONG	SystemParameters .FilterCounterLimit	Filter Cycle Alarm Limit
1048	4	ULONG	RtStatus.NewestAlarmNum	Number of newest alarm
1049	4	ULONG	RtStatus.NewestRecordNum	Number of newest record
1056	4	ULONG	RtStatus.LastDispenseCycle	Last Dispensed Cycle
1057	4	ULONG	RtStatus.LastClearedRecord	# of the last cleared record
1058	4	ULONG	RtStatus.ProfileCycle	Most recent profile cycle
1059	4	ULONG	SystemParameters.Time	Last saved
1060	4	ULONG	SystemParameters.Cycle	Cycle when written
1061	4	ULONG	Recharge.Time	Last saved
1062	4	ULONG	Recharge.Cycle	Cycle when written
1063	4	ULONG	Recipe.Time	Last saved
1064	4	ULONG	Recipe.Cycle	Cycle when written
1069	4	ULONG	DispenseInfo.Cycle	Cycle Number
1070	4	ULONG	DispenseInfo.ReferenceCycle	Reference Cycle Number
1071	4	ULONG	DispenseInfo.Time	Time of dispense (trigger)
1072	4	ULONG	SkipstepInfo.Cycle	Cycle for this information
1285	4	LONG	SystemParameters .ReadyPressureSetpoint_mPSI	RDY Pressure Setpoint
1286	4	LONG	SystemParameters .AutoRecTimer_min	Auto-Recirc. Timer
1293	4	LONG	PressureStatus .Dis_pressure_mPSI	dispense calibrated mPSI
1294	4	LONG	PressureStatus .Fill_pressure_mPSI	Fill calibrated mPSI
1295	4	LONG	Recharge .FiltrationPressureSetpoint	Filtration psi Setpoint

Appendix

SVIDs

The SVID command allows the pump to be queried for specific data values in command. A list of System Variable ID values are sent as UINT16 data in the command. The command will return the requested data in the same order as it was requested. Note that different data types have different sizes (in bytes) as defined below:

'INT16' : 2
'UINT16' : 2,
'LONG' : 4,
'ULONG' : 4,
'FLOAT' : 4, IEEE 32 bit floating point
'STRING' : 2 bytes (USHORT) is the size of the returned string in
bytes followed the string of bytes.

The actual SVID values may change for different firmware revisions so the list is published with new firmware versions.