Gamma Ray Spectroscopy of an Unknown Radioactive Source

Kellie McGuire^{1,*} and Ronny Nguyen¹

¹Department of Physics, University of New Hampshire
(Dated: September 29, 2018)

We identified an unknown radioactive source by measuring its gamma emission spectrum using a NaI scintillator with photomultiplier. To calibrate the experiment we measured the spectra of two known radioactive sources and mapped their energy photopeaks to the channels of a multichannel analyzer. This enabled us to identify the unknown source as Bismuth-207. We used this same experimental setup to measure the mass absorption coefficient of lead at 609 keV, the results of which are consistent with published values.

BACKGROUND AND MOTIVATION

Gamma ray spectroscopy using a NaI detector is a standard method for identifying radioactive substances. [CITE]

Decay of radioactive materials.—Radioactive elements are unstable nuclei that spontaneously decay to lower-energy states, emitting their excess energy in the form of particles. The rate of this decay is governed by quantum mechanics and is probabilistic. The lifetime of a radioactive sample is characterized by its half-life: the time necessary for a collection of identical radionuclides to reduce to half their initial number. Given a large enough sample, one can observe this decay directly in the lab.

In general, a radioactive nucleus can undergo many different decays before reaching its final, stable state. These decays are governed by conservation laws, such that only certain decay schemes can occur for a given radionuclide [1]. Further, each possible decay occurs with a definite probability. Radioactive materials can thus be identified by the products of their decay, which act as the signature of the radioactive sample.

Scintillator method.—Scintillating materials are commonly used as radioactivity detectors because they emit visible light when struck by charged particles. This light can either be observed directly or translated into energy information, as is done in this experiment.

In the case of gamma ray detection, the incident photon must transfer its energy to a charged particle (electron) before it can be detected by the scintillator. This occurs via three main mechanisms: the photoelectic effect, Compton scattering, and pair production. [CITE] Whereas in the photoelectric effect and pair production the gamma ray is completely absorbed, in Compton scattering the photon deposits only part of its energy. The recoiling photon is then left to either interact with another electron or escape the detector altogether. As a result, energy resolution is lost in the region where Compton scattering is dominant. A NaI detector is well-suited for gamma ray spectroscopy because its photoelectric and pair production cross-sections are much larger than its Compton scattering cross-section [2].

EXPERIMENTAL SETUP

A NaI crystal was used to detect gamma rays emitted by the radioactive samples. The NaI crystal was coupled to a [make/model]photomultiplier, which amplified the resulting photoelectric current through a series of high-voltage dynodes. The amplified current was then converted to a voltage and directed through a [make/model]spectroscopic amplifier.

The signal was observed on an oscilloscope, and the amplification was increased until it fell just below saturation. This signal was then directed to a multichannel analyzer (MCA), which sorted each pulse (decay event) into a channel according to its amplitude (energy). The emission spectra were then observed and recorded using GammaVision software.




FIG. 1. Experimental setup for gamma ray spectroscopy. A NaI crystal detects gamma emissions from the radioactive source and re-emits them as visible photons. The photons travel to the photomultiplier tube (PMT), where they produce a signal that is amplified by the spectroscopic analyzer before reaching the multichannel analyzer (MCA), which sorts each pulse according its energy.

MEASUREMENTS

All measurements were performed in Lab 317 in the University of New Hampshire's DeMeritt Hall.

Gamma spectra were recorded on two separate occasions. The second data were necessary because the first data included an overly attenuated spectra that was inadequate for performing the lead attenuation measurement.

For data set 1, seven spectra were recorded: two of known radioactive sources (Cesium-137 and Radium-226), two of unidentified sources, one with a lead attenuator, and two with no source present (background).

For data set 1, each radioactive source was positioned 20 ± 0.5 cm in front of the NaI detector "hot end." This distance was chosen to optimize the emission cross-section at the detector, permitting the NaI crystal to "see" the maximum number of decay events without exceeding the response time of the detector. The signals from the photomultiplier were amplified by a factor of 500. All spectra were recorded for 200 seconds (real time), except for the second background measurement, which ran for 300 seconds (real time).

An unidentified source labeled "Ore B" was measured, and a visual comparison to the Radium-226 spectrum revealed that it was very likely Radium-226 (Fig. 2). A measurement was also made with a different source, labeled "No. 3."




FIG. 2. A visual comparison of Ore B with $^{226}\mathrm{Ra}$ suggests that they are the same radionuclide.

For data set 2, six spectra were recorded: Cesium-137 and Radium-226, one background, sample No. 3, and two Radium-226 with lead attenuator. Each source was placed 20 cm in front of the detector. The signals were amplified by a factor of 1000 and recorded for 180 seconds (live time). A different lead absorber (thicknesses 1.25 ± 0.25 mm and 7 ± 0.5 mm) was used for each of the lead attenuator measurements. For both measurements, the lead absorber was placed halfway between the Radium-226 sample and detector hot end.

Calibration—Prior to identifying sample No. 3 and measuring the attenuation coefficient of lead, we calibrated the spectra by mapping the photopeak energies of the two known sources to the channels of the MCA. Because our data were gathered on two separate occasions, it was necessary to perform separate calibrations for each data set. The calibration results for data set 1

are presented here, with results from data set 2 given in the Appendix.

We performed the calibration by first subtracting the recorded background spectrum from the Cesium-137 and Radium-226 spectra and fitting the local background to an exponential plus constant. We then fit the most prominent energy photopeaks to a Gaussian,

$$Ae^{\frac{-(x-X)^2}{2\sigma^2}},\tag{1}$$

where X represents the MCA channel number of the photopeak max, σ is the width parameter, and A is the amplitude of the photopeak.

The photopeak energies for Cesium-137 and Radium-226 were obtained from [3] and [4], and their errors were taken to be zero. The resulting fit parameters, given in Table I, were used to determine a linear relation between the gamma energies and the channels of the MCA.




FIG. 3. The most prominent photopeaks of ¹³⁷Ce and ²²⁶Ra were fitted to a Gaussian after removing global and local background noise from the spectra. The photopeak energies were obtained from tables of known decays.

TABLE I. Photopeak Fit Parameters.

Energy (keV)	ADC Channel No.
1120	561±16
661.64	313±10
609.312	286 ± 10
351.932	153±6
295.224	126 ± 5
241.997	101±4
187.1	76±3

Errors on the linear fit parameters were calculated using the following relations from [5]

$$\sigma_b = \sigma_y \sqrt{\frac{\sum_{i=1}^N x^2}{\Delta}},\tag{2}$$




FIG. 4. A plot of known photopeak energies and MCA channel numbers, obtained from fitting the photopeaks to Gaussian curves, shows a strong linear relationship in the measured energy range.

$$\sigma_m = \sigma_y \sqrt{\frac{N}{\Lambda}},\tag{3}$$

$$\sigma_y = \sqrt{\frac{1}{N-2} \sum_{i=1}^{N} (y_i - b - mx_i)^2},$$
 (4)

$$\Delta = N \sum_{i=1}^{N} x^2 - (\sum_{i=1}^{N} x)^2.$$
 (5)

Our results indicate an offset of 53 ± 5.9 keV for our energy axis, with a resolution of 1.92 ± 0.02 keV per channel, or a relative resolution of $6.0\pm0.32\%$, which we calculated from

$$\frac{1}{n}\left(\sum_{i=1}^{n} \frac{2m\sigma_i}{E_i}\right) \times 100,\tag{6}$$

where σ is the width parameter of the peak, E is the photopeak energy, and n is the number of photopeaks measured.

RESULTS

Identity of unknown sample—The unidentified sample was plotted on a calibrated energy axis and its photopeaks fitted to Gaussians. The energies were found to be $560\pm9~\mathrm{keV}$ and $1070\pm10~\mathrm{keV}$, as shown in Fig. 6.

To identify the source, we used the Nuclear Data Center at KAERI's "Table of Gamma Rays" [6] to narrow the

search to radionuclides with prominent photopeaks that matched our results. Bismuth-207 was the only result that fit our search that contained no additional prominent gamma photopeaks, which we would not have been able to account for in our data. Our data also show a large x-ray spike at approximately 75 keV, consistent with the decay spectrum of Bismuth-207.

Bismuth-207 decays via three major gamma emissions: 570 keV (97.74 I γ), 1.064 MeV (74.5 I γ), and 1.770 MeV(6.87 I γ), where I γ is the number of counts per 100 decays [7]. It's unclear whether the 1.770 MeV photopeak is present in our data. A longer acquisition time would likely help resolve this less-intense photopeak.




FIG. 5. Gamma spectra with energy calibration.

Lead attenuation of Radium-226—As mentioned in the Measurements section, it was necessary to obtain a second set of spectra, as the first data did not contain enough information to measure the absorption coefficient of lead. The calibration results for data set 2 are given in the Appendix.

Lead absorbers of two different thicknesses (1.25 mm and 7 mm) were used to attenuate the gamma spectra of Radium-226.

A background spectrum was subtracted from the Radium-226 spectra, which were then fitted to an exponential term plus constant to remove any remaining background. The 609 keV photopeaks of Radium-226 were then fitted to a Gaussian.

The total area under each Gaussian was taken to be the intensity of the corresponding photopeak, the results of which are presented in Table II and in Figs. 6 and ??. Errors on the counts were calculated using one standard deviation errors of the fit parameters, A and σ .

We used these results to calculate the absorption coefficient for lead, using

$$I(x) = I_0 e^{-\mu x},\tag{7}$$




FIG. 6. Lead attenuation of the 609 keV photopeak of ²²⁶Ra, showing the total number of counts under each curve.

TABLE II. Lead attenuation of 609 keV gamma rays.

Source	I (count)
²²⁶ Ra unattenuated	$14,800\pm100$
With 1.25 mm absorber	$12,400\pm100$
With 7 mm absorber	$7,200\pm200$

where I and I_0 are the attenuated and unattenuated counts under the curves, respectively, x is the thickness of the absorber, and μ is the attenuation coefficient. Our results show $1.0 \pm 0.1 \text{cm}^{-1}$ using the 1.25mm absorber, and $1.4 \pm 0.6 \text{cm}^{-1}$ using the 7mm absorber.

We calculated the mass attenuation from lead, using $\frac{\mu}{\rho}$, where ρ is the mass density of lead, taken to be 11.34 g/cm³. Our results show 0.12 ± 0.05 cm³/g for the 1.25 mm absorber and 0.09 ± 0.009 cm²/g for the 7 mm absorber. To calculate the error in our results, we considered uncertainties in the counts under each curve, as well as in the thickness of the absorbers. The result from the 1.25 mm absorber is in close agreement with 0.1248 cm²/g for 600 keV radiation, found in [8]. The result from the 7 mm absorber is slightly higher than this value. This shouldn't be regarded as an inconsistency; we were attenuating gammas at a higher

energy level (609 keV) than [8], and mass attenuation is a function of incident energy. Further, we note our value of 0.09 ± 0.009 cm²/g falls within 0.1248-0.0887 ²/g, for 600 keV and 800 keV, respectively.

CONCLUSIONS AND REMARKS

(Acknowledgements)

APPENDIX

Energy Calibration of Data Set 2—

* kjm1042@wildcats.unh.edu

- Long E. PHYS 705/805 Nuclear Physics Workshop Parts I/II. (Presentation). Durham, NH: Univ. of New Hampshire. 2018.
- [2] Leo WR. Techniques for Nuclear and Particle Physics Experiments: A How-to Approach. 1994.
- [3] Radionuclide Data Sheet: Cesium 137. Spectrum Techniques, LLC. [The data are available at http://www.spectrumtechniques.com/wp-content/uploads/2016/12/Cesium-137-Information-Sheet.pdf]
 - https://www.cpp.edu/pbsiegel/bio431/genergies.html
- [4] Gamma-Ray Energies and Intensities, ²²⁶Radium. Radiochemistry Society. [The data are available at www.radiochemistry.org/periodictable/gamma_spectra/pdf /ra226.pdf]
- [5] Taylor JR. An Introduction to Error Analysis: The Study of Uncertainties in Physical Measurements. 1997.
- [6] Nuclear Data Center at KAERI, "Table of Gamma Rays," [The data are available at http://atom.kaeri.re.kr/]
- [7] Gamma-Ray Energies and Intensities, ²⁰⁷Bismuth. Radiochemistry Society. [The data are available at https://www.radiochemistry.org/periodictable/gamma_spectra/pdf/bi207.pdf]
- [8] Hubbell JH, Seltzer SM, Tables of X-Ray Mass Attenuation Coefficients and Mass Energy-Absorption Coefficients 1 keV to 20 MeV for Elements Z=1 to 92 and 48 Additional Subtances of Dosimetric Interest. NIST. 1995. [The data are available at https://physics.nist.gov/PhysRefData/XrayMassCoef/ElemTab/z82.html]




FIG. 7. Energy calibration results for data set 2. The relative resolution of the detector was $8.1\pm0.3\%$.