

# Conceptual Design for a BSL-3 Suitable Stage-Top Incubator



**Prepared by:**

Joab Gray Kloppers

**Prepared for:**

Mandy Mason

Robyn Verrinder

Justin Pead

University of Cape Town

February 2, 2026

# Contents

|          |                                              |           |
|----------|----------------------------------------------|-----------|
| <b>1</b> | <b>Introduction</b>                          | <b>1</b>  |
| 1.1      | Background and Summary . . . . .             | 1         |
| 1.2      | Scope and Limitations . . . . .              | 1         |
| <b>2</b> | <b>Research</b>                              | <b>2</b>  |
| 2.1      | Zeis/Pecon . . . . .                         | 2         |
| 2.2      | Ibidi . . . . .                              | 3         |
| 2.3      | Tokai-Hit . . . . .                          | 3         |
| 2.4      | Oko-Lab . . . . .                            | 4         |
| <b>3</b> | <b>Requirements Analysis</b>                 | <b>5</b>  |
| 3.1      | Traceability Analysis . . . . .              | 5         |
| 3.1.1    | Traceability Analysis 1 . . . . .            | 5         |
| 3.1.2    | Traceability Analysis 2 . . . . .            | 5         |
| 3.2      | Requirements . . . . .                       | 5         |
| 3.3      | Specifications . . . . .                     | 5         |
| 3.4      | Acceptance Testing Procedures . . . . .      | 6         |
| <b>4</b> | <b>Subsystem Design</b>                      | <b>7</b>  |
| 4.1      | Design Decisions . . . . .                   | 7         |
| 4.2      | Final Design . . . . .                       | 7         |
| 4.3      | Failure Management . . . . .                 | 7         |
| 4.4      | System Integration and Interfacing . . . . . | 8         |
| <b>5</b> | <b>Discussion and Results</b>                | <b>9</b>  |
| <b>6</b> | <b>Conclusion</b>                            | <b>10</b> |
| 6.1      | Recommendations . . . . .                    | 10        |
|          | <b>Bibliography</b>                          | <b>11</b> |

# Chapter 1

## Introduction

### 1.1 Background and Summary

Long term Live-Cell imaging is a very useful research tool when determining long term behavior of cell cultures in various situations. Commercial solutions are unable to provide adequate containment for the use within a Bio-Safety Level 3 (BSL-3) and often require the use of a separate containment enclosure, which introduces additional costs and complexity.

This project focuses on the conceptual design of a low-cost Stage-Top Incubator that is capable of being utilized within a BSL-3 laboratory, without the requirement for additional containment.

### 1.2 Scope and Limitations

This project is limited to the following:

- Preliminary Research
- Temperature Control
- Compatibility with Multi-Well microtiter plates

This project is limited to compatibility with a BSL-3 Lab and does not consider higher containment requirements. The environmental control aspects of the incubator are limited to just temperature control for the preliminary conceptual design. The initial design is limited to compatibility with a multi-well microtiter plate.

# Chapter 2

## Research

The conceptual design began with assessing existing commercial products to identify the key environmental control specifications required for the incubator to function correctly. Additionally, this was useful in identifying why the commercial products are incompatible with a BSL-3 lab when used without a separate external biological containment mechanism.

### 2.1 Zeis/Pecon

The Zeis/Pecon stage top incubator utilities four modules to control the environmental conditions within the incubator stage.

The Pecon Temperature Module S regulates the temperature of the incubator. It has 4 electrical connections to the Incubator lid, base, microscope objective and humidifier device. Additionally, it directly connects to the Gas regulating units to provide temperature control of the gas flow mixture itself.

The Pecon CO<sub>2</sub> Module S attaches to the temperature module, and provides CO<sub>2</sub> concentration control to the airflow passed into the incubation chamber. It requires a pressurised CO<sub>2</sub> supply in order to regulate CO<sub>2</sub> concentration.

The Pecon O<sub>2</sub> Module S attaches to the CO<sub>2</sub> module and provides O<sub>2</sub> concentration control to the airflow passed into the incubation chamber. It requires a pressurised N<sub>2</sub> supply in order to displace the O<sub>2</sub> in the airflow.

The Pecon Heating Device Humidity S connects in series with the airflow tube to the incubator stage. This module works by bubbling the airflow through heated sterile water in order to raise the relative humidity (RH) to approximately 100%.

The incubator stage includes a heated glass topped lid and a heated insert which accepts mounts for the different dishes used in microscopy. The heated insert supports perfusion experiments utilising two ports which allow tubes to be run to the culture.

The main reason why the Zeis/Pecon stage top incubator system is not suitable for long term imaging without an external containment enclosure is due to the method of gas flow through the incubator. The design of the Incubator stage is such that the inflowing gas mixture flows over the culture media and then exhausts passively through the opening on the bottom and through the sides. This means that the cell culture is not completely sealed inside the incubator which can result in contamination.




Figure 2.1: Diagram Showing the Airflow Path Within an Ibidi Multi-Well Incubator Chamber [1]

## 2.2 Ibidi

The Ibidi stage top incubator functions similarly to the Zeis/Pecon module, however it makes use of only 3 modules to maintain environmental conditions.

The Heating System Silver Line module regulates the temperature of the incubation chamber stage and the microscope. The heating of the gas mixture is subjected to the gas regulator module.

The Gas Incubation System comes in two different models, the CO<sub>2</sub> - Silver Line and the CO<sub>2</sub>/O<sub>2</sub> - Silver Line. The difference between these two systems is the former is only capable of regulating CO<sub>2</sub> concentration, where as the later has the ability to regulate O<sub>2</sub> concentration as well as CO<sub>2</sub> concentration. The gas incubation system is responsible for heating the gas mixture and the humidifying column to the appropriate temperature.

The Humidifying system works by mixing airflow directly from the Mixing module and from the humidifying column, which bubbles a separate airflow through heated water to obtain approximately 100% RH air. The gas mixer module then controls the mix ratio of dry gas to 100% RH gas to obtain a controlled humidity flowing into the incubation chamber.

The Ibidi incubation chamber comes in two versions, one version supports culture mediums with Petri dishes and slides, the second version supports multi-well plates. The Petri dish/Slide version of the chamber supports perfusion experiments using two ports which allow tubes to be run to the culture. Both versions support direct medium temperature monitoring which allow for more accurate temperature control by directly measuring the temperature of the culture medium.

The Ibidi system functions in a very similar fashion to the Pecon system, with the addition of being able to control the relative humidity in the chamber, rather than a forced approximate 100% relative humidity. This system still poses the same problem wherein the airflow is designed to passively exhaust from the incubation chamber, as shown in Figure 2.1, and therefore does not have a seal that meets BSL-3 containment standards.

## 2.3 Tokai-Hit

The Tokai-Hit system functions similarly to the Pecon and Ibidi systems however it utilises a single module to regulate the environmental conditions for temperature and CO<sub>2</sub>. Humidity is controlled by

a heated water bath inside the chamber, and O<sub>2</sub> control is achieved by utilising an optional module which then overrides all gas control.

The Tokai-Hit STXG/STXF temperature and CO<sub>2</sub> controller regulates both temperature and CO<sub>2</sub> concentration in the incubation chamber. The controller is capable of regulating the temperature of the top-lid heater, water bath heater, stage heater and the microscope objective heater. Additionally it is capable of regulating the airflow into the chamber depending on the model and air supply.

The incubator chamber uses a single design that is compatible with all major culture dishes including Petri dishes, slides and multi-well plates. This chamber supports perfusion experiments, drug administration and direct culture temperature measurement via access ports designed into the lid.

This system still poses the same issue as the aforementioned chambers, which prevents this chamber from forming a completely sealed environment suitable for a BSL-3 lab.

## 2.4 Oko-Lab

I should probably explore this option i just have not fully looked at method. test what is it doing why not running

The Oko-Lab H301-T-UNIT-BL-PLUS [2] unit is responsible for maintaining the thermal conditions within the chamber enclosure. It operates by monitoring and controlling the temperature of the chamber base, lid and optionally the microscope objective. This in turn heats the sample to the desired temperature. This heating unit is capable of sample temperature feedback utilising a thin thermocouple that can be placed in distilled water near the sample to obtain more accurate sample temperature control. Otherwise the unit is capable of using pre-sets/calibrations to infer sample temperature based on enclosure temperatures. The heating unit can be operated via a computer using available software or using optional touch screen accessory. The heating unit is capable of interfacing with the Oko-Lab Bold Line Gas Controllers.

## Chapter 3

# Requirements Analysis

### 3.1 Traceability Analysis

Show how the requirements, specifications and testing procedures all link, [Table 3.1](#) is provided.

Table 3.1: Requirements Traceability Matrix

| # | Requirements | Specifications | Acceptance Test |
|---|--------------|----------------|-----------------|
| 1 | UR01         | SP01           | AT01            |
| 2 | FR01         | SP02,SP05      | AT03            |
|   |              |                |                 |

#### 3.1.1 Traceability Analysis 1

#### 3.1.2 Traceability Analysis 2

### 3.2 Requirements

The requirements for a micromouse power module are described in [Table 3.2](#).

Table 3.2: User and functional requirements of the power subsystem.

| Requirement ID | Description |
|----------------|-------------|
| UR01           |             |
| FR01           |             |
|                |             |
|                |             |

### 3.3 Specifications

The specifications, refined from the requirements in [Table 3.2](#), for the micromouse power module are described in [Table 3.3](#).

Table 3.3: Specifications of the sensing subsystem derived from the requirements in [Table 3.2](#).

| Specification ID | Description |
|------------------|-------------|
| SP01             | User        |
|                  |             |
|                  |             |

### 3.4 Acceptance Testing Procedures

Table 3.4: Subsystem acceptance tests

| ID   | Description | Testing Procedure                                           | Pass/Fail Criteria |
|------|-------------|-------------------------------------------------------------|--------------------|
| AT01 | Powers on   | <ul style="list-style-type: none"><li>•</li><li>•</li></ul> |                    |
|      |             | <ul style="list-style-type: none"><li>•</li><li>•</li></ul> |                    |

# Chapter 4

## Subsystem Design

### 4.1 Design Decisions

### 4.2 Final Design

The following design...



Figure 4.1: Schematic

### 4.3 Failure Management

Table 4.1: CAPTION

| Name | Description |
|------|-------------|
|      |             |




Figure 4.2: PCB

## 4.4 System Integration and Interfacing

To integrate the subsystem with the rest of the system ....

Table 4.2: Interfacing specifications

| Interface | Description                                                     | Pins/Output                                                                                                                                                                                                                                                                                                        |
|-----------|-----------------------------------------------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| I001      | SparkFun 9DoF IMU Breakout board to STM for data transfer (SPI) | <ul style="list-style-type: none"> <li>• MISO: Breakout MISO* to STM PB14</li> <li>• MOSI: Breakout MOSI* to STM PB15</li> <li>• SCLK: Breakout SCLK* to STM PB13</li> <li>• CS: Breakout CS* to STM PB12 (GPIO)</li> <li>• Power: Breakout 1V8-5V5* to STM 3V</li> <li>• GND: Breakout GND* to STM GND</li> </ul> |
| I002      | STM to FTDI for data transfer to computer (UART)                | <ul style="list-style-type: none"> <li>• STM PA2(RX) to FTDI RXD</li> <li>• STM PA3(TX) to FTDI TXD</li> </ul>                                                                                                                                                                                                     |
| I003      | FTDI to computer for data transfer from STM                     | <ul style="list-style-type: none"> <li>• Assuming the FTDI is connected to a Micro B USB port (which can be used to connect to a USB cable to connect to a computer), output via cable</li> </ul>                                                                                                                  |

## Chapter 5

# Discussion and Results

# Chapter 6

## Conclusion

Lorem ipsum dolor sit amet, consectetur adipiscing elit. Ut purus elit, vestibulum ut, placerat ac, adipiscing vitae, felis. Curabitur dictum gravida mauris. Nam arcu libero, nonummy eget, consectetur id, vulputate a, magna. Donec vehicula augue eu neque. Pellentesque habitant morbi tristique senectus et netus et malesuada fames ac turpis egestas. Mauris ut leo. Cras viverra metus rhoncus sem. Nulla et lectus vestibulum urna fringilla ultrices. Phasellus eu tellus sit amet tortor gravida placerat. Integer sapien est, iaculis in, pretium quis, viverra ac, nunc. Praesent eget sem vel leo ultrices bibendum. Aenean faucibus. Morbi dolor nulla, malesuada eu, pulvinar at, mollis ac, nulla. Curabitur auctor semper nulla. Donec varius orci eget risus. Duis nibh mi, congue eu, accumsan eleifend, sagittis quis, diam. Duis eget orci sit amet orci dignissim rutrum.

### 6.1 Recommendations

Lorem ipsum dolor sit amet, consectetur adipiscing elit. Ut purus elit, vestibulum ut, placerat ac, adipiscing vitae, felis. Curabitur dictum gravida mauris. Nam arcu libero, nonummy eget, consectetur id, vulputate a, magna. Donec vehicula augue eu neque. Pellentesque habitant morbi tristique senectus et netus et malesuada fames ac turpis egestas. Mauris ut leo. Cras viverra metus rhoncus sem. Nulla et lectus vestibulum urna fringilla ultrices. Phasellus eu tellus sit amet tortor gravida placerat. Integer sapien est, iaculis in, pretium quis, viverra ac, nunc. Praesent eget sem vel leo ultrices bibendum. Aenean faucibus. Morbi dolor nulla, malesuada eu, pulvinar at, mollis ac, nulla. Curabitur auctor semper nulla. Donec varius orci eget risus. Duis nibh mi, congue eu, accumsan eleifend, sagittis quis, diam. Duis eget orci sit amet orci dignissim rutrum.

# Bibliography

- [1] ibidi, “ibidi Heating System Multiwell Plate - Silver Line Manual.” [Online]. Available: [https://ibidi.com/img/cms/downloads/in/IN\\_12150\\_HeatingSystem\\_Multiwell\\_SL.pdf](https://ibidi.com/img/cms/downloads/in/IN_12150_HeatingSystem_Multiwell_SL.pdf)
- [2] Oko-Lab, “H301-T-UNIT-BL-PLUS\_manual.” [Online]. Available: [https://www.oko-lab.com/images/okaImages/PDF-files/H301-T-UNIT-BL-PLUS\\_Manual.pdf](https://www.oko-lab.com/images/okaImages/PDF-files/H301-T-UNIT-BL-PLUS_Manual.pdf)