METAFuN Tau MR-PET Session Participant Info Sheet

Date: _____ Start Time: _____ End Time: _____

ADC ID: _____ Scan ID: _____ Behavior Staff (mRem): _____

Scanner Staff (mRem): _____

Before Participant arrival, Research staff should:

- ☐ Check *Behavioral Laptop* is fully charged
  - If not, allow computer to charge during the NIH Toolbox portion of assessments
- ☐ Check *NIH Toolbox iPad and Bluetooth Keyboard* are charged
- ☐ Check *Sennheiser HD 280 Pro Headphones*, *Headphone Jack Adapter*, and *tape measure* are with testing materials
- ☐ Check for proper USB drive to export data from laptop and iPad are with testing materials
- ☐ Make sure iPad is Connected to Bluetooth and WiFi, Do Not Disturb is turned on, Brightness is set to the maximum level, and all other apps are closed
- ☐ Sync Dosimeter to IntsaDose App (Either on mobile device or Sync Station at WTI)

MR Tech and ADRC staff will welcome the participant and administer PET tracer.

- ☐ Participant must be finished with assessments ***no more than 60 minutes after tracer injection time*** to allow for proper PET data collection

Pre-Scan Vitals

Heart Rate: _____ (BPM) O2 Sat: _____ (%)

BP: _____ (mmHg) Resp. Rate: _____ (breaths/min)

Time of PET Tracer Injection (HH:MM:SS)

_____

UPTAKE ROOM #: _____

IMAGING TRACER: _____

Meet Participant In Tracer Uptake Room:

Before beginning behavioral tasks:

- Introduce yourself to the participant and inform them about the ongoing research project.
- If the participant is wearing heavy eye makeup, we can supply materials to remove it as to not interfere with eyetracking software or vision
- If wearing glasses:
  - Glasses ON: Instructions, Corrected Near Visual Acuity, and Stroop Task
    - **PERFORM NEAR VISUAL ACUITY WITH AND WITHOUT GLASSES**
  - Glasses OFF: Uncorrected Near Visual Acuity and Sensory Attention Task
- Contacts may remain ON for all tasks
- If wearing hearing aids:
  - Hearing Aids ON: Instructions & Corrected Hearing Acuity
    - **PERFORM HEARING ACUITY WITH AND WITHOUT HEARING AIDS**
  - Hearing Aids OFF: Uncorrected Hearing Acuity, Words-In-Noise and Sensory Attention Task

"Have you had caffeine or caffeinated beverages today?" If yes, how many and when? _____	**YES NO**
"Are you wearing contacts or glasses?"	**YES NO**
"Do you wear hearing aids?"	**YES NO**

NIH Toolbox Tasks

****** READ INSTRUCTIONS EXACTLY AS THEY APPEAR FROM NIH TOOLBOX ******

**** NIH TOOLBOX AUTOMATICALLY ADJUSTS BRIGHTNESS & VOLUME! DO NOT ADJUST THEM!! ****

Prepare Subject Profile in NIH Toolbox App

- **iPad Passcode: 1719 // NIH Toolbox Passcode: 17191719**

☐ Turn On Bluetooth Keyboard and check it is connected to the iPad in Bluetooth Settings
- Create new participant and save: "**Participants**" → "**+ New Participant**"

☐ **Identifier = METAFUN/ADC Participant ID**, **Date of Birth** (Age will auto populate), **Handedness** (R/L)
- Assign METAFUN Cognitive Battery to Participant:

"**+ New Assessment**" → "**Add Tools**" → "**Batteries**" → *INSERT ASSESSMENT* → "**Save**" → "**Save**"

☐ **If YES to Glasses, ADD:**
- **Task Title: "Corrected Vision" → "METAFUN Prescan Behavioral Assessments - Visual Correction"**

☐ **If YES to Hearing Aids, ADD:**
- **Task Title: "Corrected Hearing" → "METAFUN Prescan Behavioral Assessments - Hearing Correction"**

☐ **Once Corrected Assessments are set, ADD:**
- **Task Title: "Uncorrected Assessments" → "METAFUN Prescan Behavioral Assessments"**

- Navigate to Assessments tab → Open Drop Down Menu for task to run → "**Start Assessments**"
  (Order: Corrected Vision → Corrected Hearing → Uncorrected Assessments)

NOTE: If the screen accidentally locks, re-enter the passwords and NIH Toolbox will continue the assessment from that exact point.

☐ **Hearing Acuity Task(s)** | ☐ Corrected | ☐ Uncorrected

- For Task Setup: Select "**BOTH EARS**" and Headphones: "**SENNHEISER HD 280 PRO**"
- ***Instructions: "Are the headphones comfortable? Do you have any trouble hearing me right now?"***
  - If Yes, participants may remove headphones for verbal instructions before beginning the task
- Ensure you are close enough to the participant so you are not pulling the headphones cord
- Ensure you are seated or standing comfortably in a position where the participant cannot see the iPad screen
- **TASK: Participants will tap the screen when a beep is heard in their ear. Each ear is run separately.**
- ***Repeat instructions exactly as prompted by NIH Toolbox On-Screen Script***
- **Volume is automatically adjusted. DO NOT TOUCH**

☐ **Words-in-Noise Task**

- **TASK: Participants will be told to repeat a specific word. If the person correctly says the word, Select it on screen. Once a set of 5 words is completed, select "Next" to advance.**
- ***Repeat Instructions exactly as prompted by NIH Toolbox On-Screen Script***
- **Volume is automatically adjusted. DO NOT TOUCH**
- **Only select words that the participant reports hearing**
- Once complete, remove headphones, unplug from adapter, and plug headphones in to the behavioral laptop
- Once tasks are complete, assessments will show as "Completed" and assessment report PDFs will be available

☐ **Visual Acuity Task(s)** | ☐ Corrected | ☐ Uncorrected

- Instruct participant to sit up straight and place tape measure at the nasion (approximately the midpoint of eye brows)
  - Ask participant to hold tape measure in place while you measure **16 inches**
- Remove iPad from the case and instruct the participant to hold the iPad at an arms distance away
- **TASK: Participants will say the letter they see on screen. Press that letter on the keyboard and then press SPACE to advance to the next trial.**
- ***Repeat instructions exactly as prompted by NIH Toolbox On-Screen Script***
- **Only press the letter they report, and then press "space bar" to proceed to next trial**
  - ***If a participant is struggling, prompt with: "Please give me your best guess!"***
- Once complete plug headphones into the iPad headphone adapter

Configure Sensory Attention Task Starting Values

- On the NIH Toolbox iPad, Select the "**Participants**" panel and select the current participant. Select "**Assessments**" → "**Uncorrected Assessments**" → In the Results Section, Open "**Assessment Report PDF**" and go to **Page 2**
- Circle the corresponding start values in the table

Static Visual Acuity Snellen: _____

Words-in-Noise Right Threshold: _____

Words-in-Noise Left Threshold: _____

Prepare Behavioral Laptop Tasks

- **Laptop Password: p*at***

Visual Snellen Acuity	Starting Visual Weber (%)
20/20 or Better	40%
20/20.1 – 20/40	80%
20/40.1 – 20/60	120%
20/60.1 – 20/80	160%
20/80.1 or Worse	200%

Words-In-Noise Threshold (Better of 2 Ears)	Starting Auditory Weber (%)
3.99 or Lower	200%
4.00 – 7.99	250%
8.00 – 11.99	300%
12.00 – 15.99	350%
16.00 or Greater	400%

☐ **CHECK LAPTOP VOLUME = 20%**

- Open PsychoPy application → "**View**" → "**Show Coder**"
- In Coder → "**File**" → "**Open**" → **"METAFUN_StroopClassic.py"** & **"METAFUN_SAT_FinalVersion.py"** Path: **"Desktop/METAFUN_Tasks"**
  - Click "**Green Play Button**" to begin experiment

Behavioral Laptop Tasks

☐ **Present STROOP TASK INSTRUCTIONS: "StroopTaskDemo.pptx" or Stroop Instructions Video**

- Open PsychoPy Coder → Open **METAFUN_StroopClassic.py** Tab → Click "**Green Play Button**"

☐ **Stroop Color Naming Task (METAFUN_StroopClassic.py)**

- **Check: Participant = SubjectID (ADC####), "Session" = 001**
- ***Instructions: "We will now begin the next set of behavioral tasks. Are you ready to begin?"***
  - If necessary, adjust positioning with a pillow placed on the participant's lap and click "OK" when ready

- Open PsychoPy Coder → Open **METAFUN_SAT_FinalVersion_Gap.py** Tab → Click "**Green Play Button**"

☐ **Sensory Attention Task (METAFUN_SAT_FinalVersion_Gap.py)**

- **"Participant ID" = ADC####, "Starting Modality" = (VISUAL) or (AUDITORY), "AuditoryStimulus" = WhiteNoise _RippleRate, "Output Audio Device" = Headphones (2- RealTek(R) Audio)**

START VISUAL WF% = _____

START AUDITORY WF% = _____

- ***Instructions: "We will now begin the final set of behavioral tasks. Are you ready to begin?"***
  - If necessary, adjust positioning with a pillow placed on the participant's lap and click "OK" when ready

How well could you see the pictures? *0 = "I could not see anything"; 10 = "Perfectly clear"*	_____
How would you rate your focus on the pictures? *0 = "None at all"; 10 = "Totally Focusing"*	_____
How well could you hear the sounds? *0 = "I could not hear anything"; 10 = "Perfectly clear"*	_____
How would you rate your focus on the sounds? *0 = "None at all"; 10 = "Totally Focusing"*	_____
How would you rate your wakefulness? *0 = "Completely Asleep"; 10 = "Wide awake and alert"*	_____

Post-Behavioral Task Completion

☐ Confirm all data is properly saved on iPad (**Assessment reports**) & Laptop (**"StroopClassic_data" and "MultimodalSwitch_data"**)

- Open "/Desktop/METAFUN_Tasks/MultimodalSwitchData" and **open the subject's .xlsx sheet**. Locate:
    - **Column AQ: (vis_weber_mean_last5_pct)**
    - **Column AS: (aud_weber_mean_last5_pct)**

END VISUAL WF% = _____

END AUDITORY WF% = _____

Behavior Stop Time (HH:MM) _____

---------- Research Staff Escort Participant to PET/MR Console Room ----------

In Console Room:

- On the MR Console Computer, select "New Patient" in the top left corner.
- **ENSURE NOVA 32 CHANNEL HEAD COIL WITH MIRROR IS CONNECTED**

☐ Enter **PATIENT ID (ADC ID), HEIGHT, WEIGHT, & DATE OF BIRTH (MM-DD-YYYY)** from the screening form

☐ Enter **Exam Description: "Project: METAFUN"** (*This is critical for XNAT transfer routing)

- **CIRCLE** the protocol used: **"WTI PET/MR" → "Adult" → "HEAD" → "Enable Filter for PET/MR"** → Select&Start

1.49 - PET/MR VISLAB Alzheimer's F-18 AV1451 V2 06/08/26	1.70 - PET/MR VISLAB CLARITI MK6240 WITH TASK	1.85 - VISSCHER ADRC MMS TASK 06/11/2026 FINAL VERSION
(ADRC: AV1451 + Task)	(CLARiTI: MK6240 + Task)	(Task Sequence Only)

Preparing Eye Tracking

- Check Eye tracking cables are plugged into ports 107 and 108
- Check that the orange fiber optic data cable and power source are plugged into the camera

☐ Connect the capture card from the EyeLink computer to the right USB hub of the capture laptop. Unplug the blue USB cord from the NNL tablet, and connect to the left USB hub of the capture laptop.

☐ On the capture laptop, open OBS and delete all scenes except "NNL and EyeLink Capture" and delete all sources.

☐ Add EyeLink Capture source (+ sign). Select "Video Capture Device" and select "AFN_Cap Video". Select OK. Add NNL tablet source (+ sign). Select "Video Capture Device" and select "USB Video". Select OK.

☐ Set up the Eyelink Window capture and maximize it so it covers the left side of the window. Set up the NNL window capture, decrease size by ⅓ and bring the window to the bottom right corner.

☐ Select "Start Recording".

Putting Participant in scanner (Technologist):

- Walk participant into scanning room and instruct them to sit down on the bed facing the door
- Demonstrate the answer paddles and instruct them on how to use them
- Demonstrate the emergency squeeze ball and to only use it in a true emergency
- Give the participant earplugs and ask them to insert them into each ear
- Ask participant to lie on the scanner bed and guide them to position into head coil **without hitting the mirror**
- Place knee cushion onto scanner bed under participant's legs (ask them to raise their legs)
- Slide the head coil forward and lock it into position. Place ear padding inside the head coil
- Give participant Emergency squeeze ball and answer paddles
- Cover the participant with a sheet or blanket
- Ensure optimal head placement and proper cushioning

☐ **When aligning the participant on the table with the light marker/laser, be sure that the laser crosshair is aligned to the vertical midpoint and horizontal midpoint of the mirror. Ensure the participant's head is in the center of the head coil.**

Scan Start Time (HH:MM) _____

Setup: Nordic NeuroLabs Tablet

NOTE: Tablet Setup and Localizer can and should be completed SIMULTANEOUSLY

- Open /**Desktop/psychopy_projects/4_30_versions/randomize_image**
  - ☐ **Go to prev_aud.txt, and type in the Final Auditory Weber % from the out of scanner xlsx**
  - ☐ **Go to prev_vis.txt, and type in the Final Visual Weber % from the out of scanner xlsx**
  - ☐ **Go to prev_sub.txt, and type in the participant ID: ADC####**
- Open PsychoPy, Select "**View**" → "**Show Builder**" → Open **"METAFUN_SAT_InScanner_Gap.psyexp"**

☐ **Confirm NNL Tablet Volume Set to 100**

- Open "**/Desktop/FixationCross.png"**

MRI Assessments

1) Structural Scans: Localizer (27sec)

- ***"How are you doing? Can you see all 4 corners of the TV and the White Plus Sign? For this first scan, just relax and try to keep your head as still as possible."***

☐ Double Click **"3-Plane Localizer"** → **"Save Rx"** → **"Scan"**

- **QA CHECK: Is the participant positioned accurately in the bore? Click on the completed scan and scroll slices**

MRI Technologist will initiate PET Acquisition: ***"PET BRAIN"*** → ***"MRAC1: PET BRAIN"*** → ***"ZTE MRAC GEM HNU 1: PET BRAIN"***

PET Start Time (HH:MM:SS) _____

- Once PET collection has begun, prepare remaining scans in simultaneous window:

☐ **Single Click Sequence → Click "Setup" → Confirm Bounding Boxes → Click "Save Rx" → Continue until all scans during the Simultaneous Window are of Status = "RxD"**

☐ **QA Check: Confirm all BOLD sequences have Anatomical Reference: AC/PC & Slices = 72**

☐ **Power ON both NNL sync boxes, Plug the headphone jack into the NNL tablet, and Re-Confirm Volume**

☐ ***Instructions: "We are going to test your answer buttons now. Please press the left button with your pointer finger? Please press the right button with your pointer finger?"***

- **QA CHECK: From left to right on lower sync box, LEFT POINTER = Light 2 and RIGHT POINTER = Light 3. Must reset box after testing**

2) Anatomical Practice & T1 BRAVO Scan (3min 37sec)

SCAN QUALITY: (1=Poor, 4=Best) _____

ADC ID _____ Session = "999"

Start Aud. % _____ Start Vis. % _____

☐ Single Click "**3D_Sag_T1_BRAVO_1mm_Isotropic**"
(sequence will be highlighted yellow in the Task Window)

- **POWER ON NNL Sync Boxes**
- ***In Builder Window, ensure the experiment toggle is set to RUN and not Pilot*** → **Press Run (Green play button)**
- ***"How are you doing? Did you hear that beep?"***
- If no beep played, ensure steps above were completed to properly set up audio file presentation, audio jack is properly plugged into the tablet, and Sound Output on the tablet is set to Headphones.
- ***"For the next scan, you will practice the task with the images and sounds. Remember that BLUE means focus on the images. Press your left answer button if the FIRST image is THINNER, or your right answer button if the SECOND image is THINNER. YELLOW means focus on the sounds. Press the left answer button if the FIRST sound is FASTER or the right answer button if the SECOND sound is FASTER. This will be about 3 minutes. Are you ready to begin?***

_____

☐ **When you are 100% ready to begin the scan, press: *** "Space Bar" on the NNL tablet keyboard & Click "Scan" simultaneously! *****

☐ **QA CHECK: Sync box volume count will begin to increase by 1. First colored fixation cross appears on Volume 9**

- **POWER OFF NNL Sync Boxes**

END PRACTICE AUDITORY SCORE:

NR ___ Correct _____ WF% _____

END PRACTICE VISUAL SCORE:

NR ___ Correct _____ WF% _____

During the last task (T1 Practice), on a scale from 0-10…

How well could you see the pictures? *0 = "I could not see anything"; 10 = "Perfectly clear"*	_____
How would you rate your focus on the pictures? *0 = "None at all"; 10 = "Totally Focusing"*	_____
How well could you hear the sounds? *0 = "I could not hear anything"; 10 = "Perfectly clear"*	_____
How would you rate your focus on the sounds? *0 = "None at all"; 10 = "Totally Focusing"*	_____
How would you rate your wakefulness? *0 = "Completely Asleep"; 10 = "Wide awake and alert"*	_____

3) Functional Scans: Task Run 1 (5min 45sec)

SCAN QUALITY: (1=Poor, 4=Best) _____

☐ On Console, Single Click **"Ax_Task_Run1_HB2x3_2200TR _2_1iso" → In Dropdown Menu: "Prep Scan"**

- **POWER ON NNL Sync Boxes when Prep is complete**
- **Press Run, Confirm + Log Session Info, and Click OK**
- ***"How are you doing? Did you hear the beep?"***

ADC ID _____ Session = "001"

Start Aud. % _____ Start Vis. % _____

- ***"For the next scan, we will have you perform the Task Run #1 out of 3 with the images and sounds. Remember that BLUE means focus on the images. Press your left answer button if the first image is THINNER or press your right answer button if the SECOND image is thinner. YELLOW means focus on the sounds. Press your left answer button if the first sound is FASTER or press your right answer button if the SECOND sound is faster. Are you ready?"***

☐ **Press Space at the instruction screen on the NNL Tablet**
QA CHECK: Screen should display white fixation cross and SyncBox should be holding at frame 1 out of 400

☐ **Press flashing green Start Scan Button (or click "Scan" on the MR Console)**
QA CHECK: Sync box volume count should be increasing. The colored fixation cross should appear on Volume 7

- **POWER OFF NNL Sync Boxes**

END RUN 001 AUDITORY SCORE:

NR ___ Correct _____ WF% _____

END RUN 001 VISUAL SCORE:

NR ___ Correct _____ WF% _____

During the last task (Run 001), on a scale from 0-10…

How well could you see the pictures? *0 = "I could not see anything"; 10 = "Perfectly clear"*	_____
How would you rate your focus on the pictures? *0 = "None at all"; 10 = "Totally Focusing"*	_____
How well could you hear the sounds? *0 = "I could not hear anything"; 10 = "Perfectly clear"*	_____
How would you rate your focus on the sounds? *0 = "None at all"; 10 = "Totally Focusing"*	_____
How would you rate your wakefulness? *0 = "Completely Asleep"; 10 = "Wide awake and alert"*	_____

4,5) First Set: EPI Gradients (20sec x2)

- ☐ On Console, Single Click **"ALT_ON_1_Spin_Echo_EPI" → "Scan"**
- ☐ On Console, Single Click **"ALT_OFF_1_Spin_Echo_EPI" → "Scan"**

6) Functional Scans: Task Run 2 (5min 45sec)

SCAN QUALITY: (1=Poor, 4=Best) _____

ADC ID _____ Session = "002"

Start Aud. % _____ Start Vis. % _____

- ☐ On Console, Single Click **"Ax_Task_Run2_HB2x3_2200TR_2_1iso" → In Dropdown Menu: "Prep Scan"**
- **POWER ON NNL Sync Boxes when Prep is complete**
- **Press Run, Confirm + Log Session Info, and Click OK**
- ***"How are you doing? Did you hear the beep?"***
- ***"For the next scan, we will have you perform the Task Run #2 out of 3 with the images and sounds. Remember that blue means focus on the images, and yellow means focus on the sounds. Are you ready to begin?"***

- ☐ **Press Space at the instruction screen on the NNL Tablet**
  QA CHECK: Screen should display white fixation cross and SyncBox should be holding at frame 1 out of 400
- ☐ **Press flashing green Start Scan Button (or click "Scan" on the MR Console)**
  QA CHECK: Sync box volume count should be increasing. The colored fixation cross should appear on Volume 7
- **POWER OFF NNL Sync Boxes**

END RUN 002 AUDITORY SCORE:

NR ___ Correct _____ WF% _____

END RUN 002 VISUAL SCORE:

NR ___ Correct _____ WF% _____

During the last task (Run 002), on a scale from 0-10…

How well could you see the pictures? *0 = "I could not see anything"; 10 = "Perfectly clear"*	_____
How would you rate your focus on the pictures? *0 = "None at all"; 10 = "Totally Focusing"*	_____
How well could you hear the sounds? *0 = "I could not hear anything"; 10 = "Perfectly clear"*	_____
How would you rate your focus on the sounds? *0 = "None at all"; 10 = "Totally Focusing"*	_____
How would you rate your wakefulness? *0 = "Completely Asleep"; 10 = "Wide awake and alert"*	_____

7) Functional Scans: Task Run 3 (5min 45sec)

SCAN QUALITY: (1=Poor, 4=Best) _____

ADC ID _____ Session = "003"

Start Aud. % _____ Start Vis. % _____

☐ On Console, Single Click **"Ax_Task_Run3_HB2x3_2200TR_2_1iso" → In Dropdown Menu: "Prep Scan"**

- **POWER ON NNL Sync Boxes when Prep is complete**
- **Press Run, Confirm + Log Session Info, and Click OK**
- ***"How are you doing? Did you hear the beep?"***
- ***"For the next scan, we will have you perform the Task Run #3 out of 3 with the images and sounds. Remember that blue means focus on the images, and yellow means focus on the sounds. Are you ready to begin?"***

☐ **Press Space at the instruction screen on the NNL Tablet**
QA CHECK: Screen should display white fixation cross and SyncBox should be holding at frame 1 out of 400

☐ **Press flashing green Start Scan Button (or click "Scan" on the MR Console)**
QA CHECK: Sync box volume count should be increasing. The colored fixation cross should appear on Volume 7

- **POWER OFF NNL Sync Boxes**

END RUN 003 AUDITORY SCORE:

NR ___ Correct _____ WF% _____

END RUN 003 VISUAL SCORE:

NR ___ Correct _____ WF% _____

During the last task (Run 003), on a scale from 0-10…

How well could you see the pictures? *0 = "I could not see anything"; 10 = "Perfectly clear"*	_____
How would you rate your focus on the pictures? *0 = "None at all"; 10 = "Totally Focusing"*	_____
How well could you hear the sounds? *0 = "I could not hear anything"; 10 = "Perfectly clear"*	_____
How would you rate your focus on the sounds? *0 = "None at all"; 10 = "Totally Focusing"*	_____
How would you rate your wakefulness? *0 = "Completely Asleep"; 10 = "Wide awake and alert"*	_____

8,9) Second Set EPI Gradients (20sec x2)

- ☐ On Console, Single Click **"ALT_ON_2_Spin_Echo_EPI" → "Scan"**
- ☐ On Console, Single Click **"ALT_OFF_2_Spin_Echo_EPI" → "Scan"**

10) Functional Scans: Post-Task Rest (7min 20sec)

REST SCAN QUALITY: (1=Poor, 4=Best): ____

- **Close PsychoPy and Re-Open FixationCross.png**
- ***"Great Job! For the next few scans, you won't need to do anything. Just keep your eyes open, look at the plus sign, and don't think of anything in particular"***
- ☐ On Console, Single Click **"Ax_Rest1_HB2x3_2200TR_2_1iso" → "Scan"**

11) Structural Scans: Sagittal 3D T2 (5min 53sec)

T2 SCAN QUALITY: (1=Poor, 4=Best) ____

- During the scan, use the ReadyViewer window to QC previous scan qualities.
- ***"How are you doing? For the last scan, you won't need to do anything, just try to keep still. It'll be about 6 minutes."***
- ☐ On Console, Single Click **"SAG_3D_T2" → "Scan"**

(CLARITI Only) Post-Scan Vitals	**Scan Stop Time (HH:MM:SS)**
Heart Rate: _____ (beat/min) O2 Sat: _____ (%) BP: _____ (mmHg) Respiratory Rate: _____ (breath/min)	_____

☐ ***XNAT Data Transfer:***

- Once all scanning is complete, Select **"End Exam"** → **"Successfully Completed All Scans"** → **"End"** → **"Accept"**

TOTAL IMAGES/DICOMS: _____

- In the Protocol Manager Window, scroll to the top and find the scanning protocol that matches the ADC Session ID
- Select the Scan, and then in the Series Window below, Select all the Individual Scans Series
- With all scans selected, on the bottom right of the screen, click **"Tools"** → **"De-identification"**
  - **This can NOT be interrupted and will take approximately 20 minutes to complete**.
  - **If there are scans scheduled immediately after, this will have to wait to be completed at a later time**
- Once De-Identification is complete, at the scroll to the bottom of Protocol Window (upper panel) and find anonymized protocol, which will be prepended as: **"ANON(ADC-SessionID)"**
- Click the anonymized scan in the upper window and select all of the scans in the lower series window
- Once all scans are collected selected, to initiate data transfer, select **"XNAT"** at the bottom of the screen. In the manager tab on the right, a message will display stating the XNAT transfer has been successfully initiated
- **This process can be completed in the background and takes approximately 2 hours to fully upload**

☐ ***Hard-Drive Data Transfer with dumpDB:***

- Insert VISLAB 2TB hard drive into the PET MR Computer
- In the protocol transfer panel on the MRI console (3rd button on the top bar), select the drop down menu and select "Open Command Window". In the command window, type in "nautilus". This will bring up the file browser
- On the left side of the file browser, right click the "VISLAB" drive, and select "Mount Drive"
- Close out of file browser and command window, open a new command window, and cd to the root directory. Type in "cd /run/media/sdc/VISLAB" and press Enter.
- To begin transfer, highlight all anonymized scans in the lower series window you want to transfer
- With all scans selected, type into the command window "dumpDB SERIES #####" (The 5 digit number must match the scanner session ID). Files will begin to transfer on the the hard drive
- **This process takes approximately 50 minutes to 1 hour.**

☐ ***Hard-Drive Data Transfer with EasyExport:***

- Insert VISLAB 2TB hard drive into the PET MR Computer
- In the protocol transfer panel on the MRI console (3rd button on the top bar), select the drop down menu and select "Open Command Window". In the command window, type in "nautilus". This will bring up the file browser
- On the left side of the file browser, right click the "VISLAB" drive, and select "Mount Drive"
- Close out of file browser and command window, open a new command window, and cd to the root directory. Type in "cd EasyExport" and press Enter.
- To begin transfer, highlight all anonymized scans in the lower series window you want to transfer
- With all scans selected, type into the command window "python easy_export.py -r /run/media/sdc/VISLAB " and press enter. Files will begin to transfer to hard drive
- An error **will** occur. **THIS IS OKAY**. The error occurs because the script will attempt to compress the entire VISLAB drive as a tar.gz, and it does not have permission to do so. However, this only occurs after transferring all files.
  - This has not been removed as it would require amending code by Gaurav Rathi and/or Alex Smith.
  - A folder on the 2TB hard drive will show up as e#####, containing all of the files.
- In the nautilus window, right-click the newly created folder, and select compress as a .zip
- **This process takes approximately 3 minutes. However, it can only transfer one scan at a time. The scan must be moved off the hard drive (i.e. deleted from SSD and Recycle Bin) before another one can be added.**