LABEL SPECIFICATION	**DOCUMENT NO: Z311862** **REVISION: B**
TITLE: IFU, LAMINAR FLOW PS & TC REUSABLE	

NOTE: PRINT IFU Z311862, OUTSERT IFU Z311862S, AND EIFU Z311862E SHARE THE SAME CONTENT. FUTURE REVISIONS TO THESE DOCUMENTS MUST BE REVIEWED TOGETHER.

IFU SPECIFICATIONS:

1. **IFU DIMENSIONS:**
   - a. Horizontal: 38.5” ± 1/8”
   - b. Vertical: 19.5” ± 1/8”
   - c. Flat: 19.5” x 38.5” (Tolerance: ± 1/8”)
   - d. Folded: 3.25” x 7.7” (Tolerance: ± 1/8”)
     - i. “Fan” fold to 19.5” X 7.7”
     - ii. “Z” Fold to 6.5” x 7.7”
     - iii. Fold in half to 3.25” x 7.7”
2. **IFU STOCK:** Basis Weight: 35-45 lb. Offset or 16 lb. Bond, Smooth Opaque Finish.
   Thickness range: 0.0030” – 0.0041”
3. **STYLE:** One sheet, two sides printing. Fifteen panels per side.
   Self contained. English faces out.
4. **COLOR OF COPY:**
   All copy prints black ink on white paper. Printing to be clear and legible with no ink smears. No visible damage, loose or attached particles (fibers) are permitted.
5. See attached for artwork layout (Not to Scale)
6. Manufacturer specific information (i.e. Manufacture Part number) if on the label would need to be verified and confirmed by the Manufacturing site.
7. **POINT OF USE**: Irvine, CA (Updating)
8. **DATA SHEET LAYOUT**:

Front	English (EN)	Danish (DA)	Dutch (NL)	Finnish (FI)	French (FR)
	German (DE)	Italian (IT)	Norwegian (NO)	Portuguese (PT)	Spanish (ES)
	Swedish (SV)	Estonian (ET)	Ukrainian (UK)	Arabic (AR)	(Blank)

Back	Polish (PL)	Russian (RU)	Bulgarian (BG)	Slovenian (SL)	Czech (CS)
	Hungarian (HU)	Latvian (LV)	Lithuanian (LT)	Slovak (SK)	Romanian (RO)
	Turkish (TR)	Croatian (HR)	Greek (EL)	Serbian (SR)	(Blank)

CHANGE SUMMARY

REVISION	DESCRIPTION OF CHANGE
A	Initial release per C0194204
B	Amended language related to cleaning and sterilization procedures, and for compliance to EU MDR, per C0233633

Reusable LAMINAR™ Flow Irrigation Sleeve and Test Chamber

Z311862 B
09/2026
1 of 1

EN

- REF OPOS19L Reusable LAMINAR™ Flow Irrigation Sleeve and Test Chamber, 19 Gauge
- REF OPOS20L Reusable LAMINAR™ Flow Irrigation Sleeve and Test Chamber, 20 Gauge
- REF OPOS21L Reusable LAMINAR™ Flow Irrigation Sleeve and Test Chamber, 21 Gauge
- REF OPOHF20L Reusable LAMINAR™ High Flow Irrigation Sleeve and Test Chamber, 20 Gauge
- REF OPOHF21L Reusable LAMINAR™ High Flow Irrigation Sleeve and Test Chamber, 21 Gauge

℞ only

Description:
The Reusable LAMINAR™ Flow Irrigation Sleeve and Test Chamber are used in conjunction with compatible Johnson & Johnson Surgical Vision, Inc. Phacoemulsification Handpieces. The irrigation sleeve channels irrigating solution across the shaft of the handpiece tip to allow the solution to enter the eye. The solution will also cool the tip during the phacoemulsification procedure. The irrigating solution exits the irrigation sleeve via two small holes located at the tip of the sleeve and maintains intraocular pressure. The test chamber provides a protective cover for the phaco handpiece tip prior to use. The test chamber serves as a receptacle for the balanced salt solution during the prime cycle and subsequently keeps the phaco tip wet throughout the handpiece tuning cycle. During verification of irrigation/aspiration balance, the tip of the test chamber simulates the intraocular chamber. The Irrigation Sleeve and Test Chamber are designed as reusable products for reuse up to 20 times and have been sterilized by gamma radiation.

Use only phacoemulsification accessories that have been appropriately validated and approved for use with the phacoemulsification system that is to be used. Prior to use, refer to the phacoemulsification system operator manual for a listing of compatible accessories.

19G = 19 Gauge
20G = 20 Gauge
21G = 21 Gauge

Intended Use / Purpose:
The Irrigation Sleeve is used to deliver solution to the eye and cool the handpiece tip.

Indications for Use:
The Irrigation Sleeve is a device intended to direct irrigation solution across the shaft of a phacoemulsification tip, allowing the solution to enter the eye during ocular surgery. The Test Chamber is intended to verify the correct connectivity and functionality of the irrigation/aspiration function prior to the surgical procedure.

Contraindications / Intended Patient / Intended User:
All relevant information for Contraindications / Intended Patient / Intended User is described in the Phacoemulsification System Operator's Manual. Contraindications include any pre-existing patient condition[s] that the health care provider determines to be a contraindication to cataract surgery. Intended patients include adult patients with cataracts or adult patients undergoing anterior segment ophthalmic surgery (e.g., removal of the crystalline lens). Note: Pregnant, nursing women, or children were not included in the clinical trial[s] conducted. The Johnson & Johnson Surgical Vision, Inc. Phacoemulsification System is intended for use by an ophthalmologist and support staff operating under direction of the surgeon. Any pre-existing patient condition[s] that the health care provider determines to be a contraindication to cataract surgery is considered a contraindication.

Clinical Benefits / Undesirable Side Effects / Adverse Effects:
All relevant information for Clinical Benefits / Undesirable Side Effects / Adverse Effects is described in the Operator's Manual. Clinical benefits of the Johnson & Johnson Surgical Vision, Inc. Phacoemulsification System include the following: rapid visual rehabilitation, early resumption of daily activities, minimal ocular inflammation, minimal postoperative astigmatism, reduced potential for ocular complications. Complications that may be associated with the Johnson & Johnson Surgical Vision, Inc. Phacoemulsification System include broken capsule, corneal burn, infection/inflammation, and corneal edema.

Performance Characteristics:
- Phacoemulsification under irrigation / aspiration
- See the user manual of the phacoemulsification system for phacoemulsification specifications
- Re-use up to 20 times

Product Drawing:



Instructions for Use:
Please refer to the System Owner's Manual for specific information on how to use the Reusable LAMINAR Flow Irrigation Sleeve and Test Chamber. Please refer to the following instructions for how to properly assemble the irrigation sleeve and test chamber onto the phaco handpiece.

Performed by Scrub (Nurse):

Assemble the phaco handpiece and accessories as follows:

1. Inspect and screw the phaco tip onto the end of the phaco handpiece, using care to properly engage the threads. Finger tighten.
2. Place the tip wrench over the phaco tip and over the tip hub near the base of the tip (not over the handpiece horn). Tighten the phaco tip snugly into place (clockwise). Do not under or over tighten. Under tightening will result in a loose tip that will signal "CHECK TIP" diagnostics when phaco power is applied. Over tightening may result in damage to the tip.

CAUTION: NEVER ATTEMPT TO STRAIGHTEN A BENT TIP. THIS WILL PRODUCE A BROKEN TIP WHEN PHACO POWER IS APPLIED.

3. Inspect and carefully place the silicone irrigation sleeve down over the phaco tip and engage the irrigation sleeve threads. Slowly turn the irrigation sleeve clockwise until the end of the phaco tip is exposed out of the end of the irrigation sleeve. The length of the exposed tip should be set according to the surgeon's preference. The angle of the tip bevel should be set at 90 degrees from the two side irrigation holes at the tip of the irrigation sleeve. If improper tightening results in a bent tip, remove and discard.
4. Attach the test chamber to the handpiece during priming of the handpiece and tubing. Please refer to the Owner's Manual for complete instructions on how to perform this test.

Tracking Method:
A 20 use tracking card is provided. This card should be used to track each use of the tip and sleeve. Discard tip and sleeve after 20 uses.

Cleaning:
WARNING: This device has not been validated to withstand automated cleaning. Improper/inadequate cleaning may result in particulate matter adhering to the instrument and exfoliation of particles into the surgical field. In addition, the function and life expectancy of the product may become compromised.
The cleaning procedure has been designed in accordance with: AAMI TIR12 "Designing, testing, and labeling medical devices intended for processing by health care facilities: A guide for device manufacturers". If processing will be performed in a different location from use, care shall be taken to prevent damage by using specific racks, trays or rigid container systems. Deviations from the described steps must be validated by the processor, including the inclusion of automated cleaning.
Prior to cleaning, remove and discard sleeve and test chamber if damaged. The following cleaning procedure is to be performed at the end of each surgical case:

1. With the irrigation sleeve and phaco tip still in place on the handpiece, inject (or flush) 60 cc of distilled water through the irrigation line of the phaco handpiece.
2. Next, inject (or flush) 60 cc of distilled water through the aspiration line of the phaco handpiece. It is helpful to use a female/female connector to attach the syringe to the aspiration line.
3. Carefully remove the phaco irrigation sleeve and phaco tip from the handpiece.
4. The phaco test chamber is cleaned by filling the test chamber with 3mL of distilled water. Pour out the water and repeat two (2) more times.

Packaging:
Wrap the devices in approved sterilization wrap for sterilization. Sterilization trays have not been validated for use with this device.

Sterilization:
CAUTION: DO NOT STERILIZE PRIOR TO PERFORMING THE CLEANING PROCEDURES DESCRIBED ABOVE.
WARNING: Sterility assurance is the responsibility of the user. All non-sterile accessories should be sterilized prior to use. Sterilize the handpiece using a steam sterilization cycle[1]. The sterizilation instructions provided in table that follows, have been validated by Johnson & Johnson Surgical Vision, Inc. as being capable of sterilizing the device for reuse.
It is the responsibility of the processor to ensure that the process is actually performed using the equipment, materials and personnel to achieve the desired result. This requires verification and routine monitoring of the process. Alternate methods that deviate from the following instructions provided should be properly evaluated for effectiveness and potential adverse consequences.

The following sterilization techniques, times and temperatures should be used in order to maximize the performance and life of the reusable products.

Steam Sterilization Cycles:

Sterilizer Type	Sample Configuration	Temperature	Minimum Exposure Time (Minutes)	Minimum Dry Time (Minutes)
Gravity Displacement	Wrapped*	132°C (270°F)	15	25
Gravity Displacement	Wrapped*	135°C (275°F)	10	25
Pulsing Prevacuum	Wrapped*	132°C (270°F)	4	20
Pulsing Prevacuum	Wrapped*	134°C (273°F)	3	20

*Utilize health authority (for example: US FDA) cleared wrap. For an unwrapped configuration, the same parameters as defined for the wrapped configuration may be used.

Note: The above table includes the minimum temperature and time validated to assure sterility. Devices can be sterilized in unwrapped configurations but **Health Authorities in some regions do not accept unwrapped sterilization methods.** Please review the appropriate guidelines, standards and national Health Authorities' guidelines when determining acceptable steam sterilization process parameters for use in each respective country. National Health Authorities may have additional requirements regarding quantity and types of sterilizers required for healthcare facilities, installation requirements, sterilizer capacities, personnel training, process monitoring, reprocessed device recordkeeping, and other requirements relating to reprocessing of reusable medical devices. Perform reprocessing in accordance with all regional, national and local standards (for example, ISO 17665 or ANSI/AAMI ST79 in conjunction with national, regional, or local requirements).

The cleaning methods and sterilization parameters recommended here were validated using qualified equipment and materials in a controlled environment. Due to the potential for the accumulation of particulate and bioburden residues in the sterilizer water reservoirs, it is the responsibility of the surgical facility to properly maintain the sterilization equipment, quality of steam and their associated filters to assure the introduction of contaminant-free steam into the handpiece. It is the responsibility of the surgical facility to conduct appropriate validation(s) for any practice outside of the parameters provided within the instructions for use.

Cooling is not necessary prior to re-assembly; however, caution should be used to prevent burns.

WARNING: The use of gas sterilization is not recommended. Sterility assurance is the responsibility of the user. All non-sterile accessories should be sterilized prior to use.

The components have been designed to withstand the parameters of accepted procedures as outlined in ISO 17665-1 "Sterilization of health care products – Moist heat – Part 1: Requirements for the development, validation and routine control of a sterilization process for medical devices", and ANSI/AAMI ST79 "Comprehensive Guide to Steam Sterilization and Sterility Assurance in Health Care Facilities."

Storage and Handling:
After sterilization, store the components in a safe, clean environment. Keep dry and free of dust. Remember to resterilize prior to each use.

Precautions:
1. It is not advisable to mix Johnson & Johnson Surgical Vision, Inc. products with those of different manufacturers or different systems since designs, materials, mechanics, and constructions are not harmonized. Overall performance may be impaired.
2. Contents are sterile when pouch is sealed and undamaged.

Serious Incident Reporting:
Contact your local representative to report all product complaints and/or adverse events to Johnson & Johnson Surgical Vision, Inc. If the LAMINAR Flow Irrigation Sleeve and Test Chamber are received in an unsatisfactory condition, please contact your local Johnson & Johnson Surgical Vision, Inc. Sales Office immediately. Each component is identified by a lot number, and this should be included in any correspondence or discussions. In the U.S.A. telephone: Toll free 1-877-266-4543.

Serious Incident Reporting (EU):
For a patient/user/third party in the European Union and in countries with identical regulatory regime; if, during the use of this device or as a result of its use, a serious incident has occurred, please report it to Johnson & Johnson Surgical Vision, Inc. and/or its authorised representative and to the competent authority of the Member State in which the user and/or patient is established.

Disposal Information:
Do not dispose with household waste. To prevent environmental risks or endangerments by nonprofessional disposal, the disposal of this product must comply with valid practices and local regulations. Contact your local Johnson & Johnson Surgical Vision, Inc. representative for a disposal location.

Symbol/Title (explanation):

SYMBOL	TITLE (EXPLANATION)
℞ only	Caution: Federal law restricts this device to sale by or on the order of a licensed healthcare practitioner
⚠	Caution
ⓘ	Consult instructions for use or consult electronic instructions for use
⊘	Do not use if package is damaged and consult instructions for use
STERILE R	Sterilized using irradiation
◯	Double sterile barrier system
REF	Catalogue number
LOT	Batch code
UDI	Unique device identifier

🏭	Date of manufacture
⧗	Use-by date
🏭	Manufacturer
EC REP	Authorized representative in the European Community/ European Union
🌐	Importer
▢	Packaging unit
MD	Medical Device
CH REP	Authorized representative in Switzerland
CH	Switzerland
EC	European Community/ European Union

Johnson & Johnson
Surgical Vision, Inc.
31 Technology Drive, Suite 200,
Irvine, CA 92618 USA

Product of Mexico

EC REP
EC
AMO Ireland
Block B
Liffey Valley Office Campus
Quarryvale, Co. Dublin, Ireland

CH REP
CH
AMO Switzerland GmbH
Gubelstrasse 34
6300 Zug
Switzerland

CE 2797

[1] Note: Ensure that the steam sterilizer conforms to regional regulatory requirements, such as ISO 17665, Sterilization of health care products – Moist heat – Requirements for the development, validation and routine control of a sterilization process for medical devices or ANSI/AAMI ST79 Comprehensive guide to steam sterilization and sterility assurance in health care facilities.

LAMINAR flow phaco-spids og irrigationsmanchet til flergangsbrug

XX

Herbruikbare LAMINAR facotip en irrigatiemanchet voor stroming

XX

Uudelleenkäytettävä LAMINAR-virtausfakokärki ja -huuhteluholkki

XX

Embout pour phacoémulsification incurvé et manchon d'irrigation LAMINAR réutilisable

XX

Wiederverwendbare LAMINAR Fluss-Phako-Nadel und Irrigationshülse

XX

TRANSLATION REQUIRED

Punta faco e sleeve di irrigazione LAMINAR Flow riutilizzabili

XX

TRANSLATION REQUIRED

Gjenbrukbar LAMINAR flyt fakospiss og skyllehylse

XX

TRANSLATION REQUIRED

Ponta de facoemulsificação e manga de irrigação de fluxo LAMINAR reutilizáveis

XX

TRANSLATION REQUIRED

Punta de facoemulsificación y funda de irrigación de flujo LAMINAR reutilizables

XX

TRANSLATION REQUIRED

Återanvändbar LAMINAR flödesfakospets och spolhylsa

XX

Korduvkasutatav LAMINAR-i voolu fako-otsak ja loputustoru

XX

Багаторазова прямоточна факоголка LAMINAR і рукав для зрошування

XX

UA.TR.116

رأس عدسة التدفق ومخروط الإرواء من LAMINAR القابلان لإعادة الاستخدام

XX

© Johnson & Johnson and its affiliates 2026

Johnson&Johnson VISION

Końcówka wielokrotnego użytku do fakoemulsyfikacji LAMINAR i dren do wlewów

XX

Zakrzywiona końcówka i dren: 19 G, 30° (nr kat. OPOCR3019R)
Zakrzywiona końcówka i dren: 20 G, 30° (nr kat. OPOCR3020R)
Zakrzywiona końcówka i dren: 20 G, 45° (nr kat. OPOCR4520R)
Zakrzywiona końcówka i dren: 21 G, 30° (nr kat. OPOCR3021R)
Zakrzywiona końcówka i dren: 21 G, 45° (nr kat. OPOCR4521R)
Rozszerzana końcówka i dren: 20 G, 30° (nr kat. OPOF3020R)
Rozszerzana końcówka i dren: 20 G, 45° (nr kat. OPOF4520R)
Zaokrąglona końcówka i dren: 21 G, 0° (nr kat. OPOR0021R)
Zaokrąglona końcówka i dren: 21 G, 15° (nr kat. OPOR1521R)
Zaokrąglona końcówka i dren: 21 G, 30° (nr kat. OPOR3021R)
Zaokrąglona końcówka i dren: 21 G, 45° (nr kat. OPOR4521R)
Zaokrąglona końcówka i dren: 19 G, 0° (nr kat. OPOR0019L)
Zaokrąglona końcówka i dren: 19 G, 15° (nr kat. OPOR1519L)
Zaokrąglona końcówka i dren: 19 G, 30° (nr kat. OPOR3019L)
Zaokrąglona końcówka i dren: 19 G, 45° (nr kat. OPOR4519L)
Zaokrąglona końcówka i dren: 20 G, 0° (nr kat. OPOR0020L)
Zaokrąglona końcówka i dren: 20 G, 15° (nr kat. OPOR1520L)
Zaokrąglona końcówka i dren: 20 G, 30° (nr kat. OPOR3020L)
Zaokrąglona końcówka i dren: 20 G, 45° (nr kat. OPOR4520L)

Opis:
Końcówki wielokrotnego użytku do fakoemulsyfikacji LAMINAR i dreny do wlewów są przeznaczone do stosowania ze wszystkimi aparatami do fakoemulsyfikacji firmy AMO. Końcówka do fakoemulsyfikacji wibruje podłużnie z częstotliwością ultradźwiękową, powodując emulsyfikację obszaru soczewki dotkniętego zaćmą. Pozostałości są usuwane przez odsysanie za pomocą wydrążonej igły (aspiracja). Wynikające z tego zmniejszenie objętości komory przedniej oka jest wyrównywane przez podawany izotoniczny roztwór soli (przepłukiwanie). Kanały drenu do wlewów dla roztworu irygacyjnego, biegnące w trzonie końcówki do fakoemulsyfikacji, pozwalają na wprowadzenie roztworu do oka. Podczas zabiegu fakoemulsyfikacji roztwór chłodzi także igłę. Roztwór irygacyjny wypływa z drenu do wlewów przez dwa małe otwory na końcówce drenu i utrzymuje ciśnienie wewnątrzgałkowe. Końcówka do fakoemulsyfikacji i dren do wlewów zostały zaprojektowane pod kątem wielokrotnego użytku (do 20 razy) i były sterylizowane promieniowaniem gamma.

Wskazania do stosowania:
Końcówki do fakoemulsyfikacji są używane do rozbijania (emulsyfikacji) jądra dotkniętej zaćmą soczewki i usuwania jego pozostałych fragmentów. Dren do wlewów to urządzenie przeznaczone do bezpośredniego wlewu roztworu irygacyjnego przez trzon końcówki do fakoemulsyfikacji, umożliwiające wprowadzenie roztworu do oka podczas zabiegów chirurgii oka.
Końcówka do fakoemulsyfikacji jest przeznaczona do użytku z drenem do wlewów podczas zabiegów chirurgii oka.

Przykładowy rysunek:



Instrukcja użycia:
Szczegółowe instrukcje na temat użytkowania końcówek wielokrotnego użytku do fakoemulsyfikacji LAMINAR i drenów do wlewów można znaleźć w Podręczniku użytkownika systemu. Sposób odpowiedniego montażu końcówki do fakoemulsyfikacji i drenu do wlewów na głowicy do fakoemulsyfikacji opisano w poniższych instrukcjach.

i ich złuszczanie do pola operacyjnego. …

Sterylizacja:

Cykle sterylizacji parowej:

Wyprodukowano w USA.

Проточная головка для факоэмульсификации и ирригационный рукав LAMINAR многоразового использования

XX



Проточен връх за факоемулсификация LAMINAR и иригационен маншон за многократна употреба

XX



Fakoemulzifikacijska konica za laminarni pretok in irigacijski tulec LAMINAR za večkratno uporabo

XX



Opakovaně použitelná průtoková fakoemulzifikační jehla a irigační návlek LAMINAR

XX



Újrafelhasználható LAMINAR áramlásos phaco fej és öblítőhüvely

XX



TRANSLATION REQUIRED

Atkārtoti lietojamais „LAMINAR" plūsmas fako uzgalis un uzmava skalošanai

XX



TRANSLATION REQUIRED

Daugkartinis LAMINAR srauto fakoemulsifikacijos antgalis ir drėkinimo mova

XX



TRANSLATION REQUIRED

Opakované použiteľný prietokový fakoemulzifikačný hrot a výplachová manžeta LAMINAR

XX



TRANSLATION REQUIRED

Vârf de facoemulsificare și manșon de irigare Flux LAMINAR reutilizabile

XX



TRANSLATION REQUIRED

Tekrar Kullanılabilen LAMINAR Akış Fako Ucu ve İrigasyon Kovanı

XX



Protočni nastavak za fakoemulzifikaciju i cjevčica za irigaciju LAMINAR za višekratnu uporabu

XX



Επαναχρησιμοποιήσιμο άκρο φακοθρυψίας και χιτώνιο καταιονισμού ροής LAMINAR

XX



LAMINAR protočni fako vrh i naglavak za ispiranje za višekratnu upotrebu

XX