ILLUSTRATIVE TEXT FOR TESTING - NOT AN OFFICIAL RWANDA FDA DOCUMENT

GUIDELINES ON REGISTRATION OF MEDICAL DEVICES
Version 2026-01

1. Quality management system
The manufacturer must hold an ISO 13485 certificate, valid on the date of submission, whose scope covers the design and manufacture of the device.

2. Declaration of conformity
The declaration of conformity names the device and its model, states its risk class (A, B, C or D), lists the standards applied, and is signed and dated by the manufacturer.

3. Technical file summary
The technical file summary describes the device and its intended use, and includes a summary of the risk management according to ISO 14971 and of the performance or clinical evidence.

4. Instructions for use
The instructions for use state the intended use, the users, contraindications and warnings, and the storage and operating conditions, including temperature limits.