

Calibration

Measured values

pH buffer 1:	TEC_4.000	165.8 mV	18.8 °C
pH buffer 2:	TEC_7.000	3.6 mV	19.1 °C

Calibration data

Slope:	92.8 % / -54.9 mV/pH (a)
Zero point:	pH 7.10 / 4.4 mV (a)
Date and time:	01/28/22 10:17:51

Electrode data

Name:	SCP-pHT-A170MF-3M-DIN-N Batch ID:	B210806001
-------	-----------------------------------	------------

Device information

Device:	TitroLine 7000
Serial number:	10012653
Software version:	1.18.1121.247