

## Calibration

### Measured values

pH buffer 1:	TEC_4.000	165.7 mV	18.4 °C
pH buffer 2:	TEC_7.000	6.7 mV	18.8 °C

### Calibration data

Slope:	91.1 % / -53.9 mV/pH (a)
Zero point:	pH 7.15 / 7.4 mV (a)
Date and time:	01/28/22 10:15:26

### Electrode data

Name:	SCP-pHT-A170MF-3M-DIN-N Batch ID:	B210806001
-------	-----------------------------------	------------

### Device information

Device: TitroLine 7000  
Serial number: 10012653  
Software version: 1.18.1121.247