सी एस आई आर – राष्ट्रीय भौतिक प्रयोगशाला
CSIR-NATIONAL PHYSICAL LABORATORY
(वैज्ञानिक तथा औद्योगिक अनुसंधान परिषद्)
(Council of Scientific and Industrial Research)
(राष्ट्रीय मापिकी संस्थान (एनएमआई), सदस्य बीआईपीएस एवं हस्ताक्षरकर्ता सीआईपीएम - एमआरए)
(National Metrology Institute (NMI), Member BIPM and Signatory CIPM - MRA)
डॉ. के. एस. कृष्णन् मार्ग, नई दिल्ली-110012, भारत
Dr. K. S. Krishnan Marg, New Delhi-110012 INDIA
दूरभाष/Phone : +91-11-4560, 8441, 8610, 9447, फैक्स/Fax : 91-11-4560 8448
ई-मेल/E-mail : cfct@nplindia.org, वेबसाईट/Website : www.nplindia.org

अंशांकन प्रमाण पत्र
CALIBRATION CERTIFICATE

MEASURING CYLINDER

प्रमाण पत्र संख्या / Certificate No.

25050175/D1.01/C-045

दिनांक/ Date	अगले अंशांकन हेतु अनुशंसित तिथि Recommended date for next calibration	पृष्ठ / Page	पृष्ठों की संख्या / No. of Pages
24.07.2025	24.07.2028	1	2

1. **Calibrated for** : Mayuresh Associates LLP
3077A, Bhandup Industrial Estate
Pannalal Compound, LBS Marg, Bhandup West
Mumbai - 400 078, Maharashtra, India

Customer's Ref. No. : Letter , Dated : 16.05.2025

2. **Description and identification of instrument**

Name	:	Measuring Cylinder
Range	:	(100 -10) ml
Least Count	:	1 ml
Make	:	Borosil
Material	:	Borosilicate glass 3.3
S. No.	:	BS309311G
Accuracy class	:	A
Id. No.	:	MA 830

3. **Environmental conditions** : Temperature : (23.0 ± 1.5) °c
Relative humidity : (50 ± 10) %

4. **Standard(s) used with associated Measurement uncertainty** :
1. Precision balance (Sub-Div.#1.01/M/CC2001/03) with uncertainty ± 0.000 5 g .
2. Distilled water.

5. **Metrological Traceability of Standard(s) used** : The standard(s) used for calibration is(are) traceable to the National Standard which realize the units of quantities according to the International System of Units (SI).

6. **Principle/Methodology of Calibration & Calibration Procedure number** : CSIR-NPL Calibration Procedure No. Sub.Div.#1.01/Doc.3/CP#VM/V-01. The Measuring Cylinder has been calibrated for its capacity by gravimetric method.

अंशांकनकर्ता:
Calibrated by:
(D.C. SHARMA)

जाँचकर्ता:
Checked by:
(NIDHI SINGH)

प्रभारी वैज्ञानिक:
Scientist-in-charge:
(NIDHI SINGH)

जारीकर्ता:
Issued by:
डॉ. श्रीनिवास राव रागम
Dr. Srinivasa Rao Ragam
Director's Nominee

सी एस आई आर – राष्ट्रीय भौतिक प्रयोगशाला
CSIR-NATIONAL PHYSICAL LABORATORY
(वैज्ञानिक तथा औद्योगिक अनुसंधान परिषद्)
(Council of Scientific and Industrial Research)
(राष्ट्रीय मापिकी संस्थान (एनएमआई), सदस्य बीआईपीएस एवं हस्ताक्षरकर्ता सीआईपीएम - एमआरए)
(National Metrology Institute (NMI), Member BIPM and Signatory CIPM - MRA)
डॉ. के. एस. कृष्णन् मार्ग, नई दिल्ली-110012, भारत
Dr. K. S. Krishnan Marg, New Delhi-110012 INDIA
दूरभाष/Phone : +91-11-4560, 8441, 8610, 9447, फैक्स/Fax : 91-11-4560 8448
ई-मेल/E-mail : cfct@nplindia.org, वेबसाईट/Website : www.nplindia.org

सीएसआईआर CSIR भारत का नवाचार इंजन The Innovation Engine of India

अंशांकन प्रमाण पत्र
CALIBRATION CERTIFICATE

MEASURING CYLINDER

प्रमाण पत्र संख्या / Certificate No.

25050175/D1.01/C-045

दिनांक/ Date	अगले अंशांकन हेतु अनुशंसित तिथि Recommended date for next calibration	पृष्ठ / Page	पृष्ठों की संख्या / No. of Pages
24.07.2025	24.07.2028	2	2

7. Results :

Denomination (mL)	Volume at 27 °C (mL)	Uncertainty (mL)
100	99.813 1	± 0.010 0
10	9.899 7	± 0.001 0

The reported uncertainty is at coverage factor k = 2.28 which corresponds to a coverage probability of approximately 95% for a normal distribution.

8. Date(s) of Calibration : 17.07.2025

9. Remarks : Nil

अंशांकनकर्ता:
Calibrated by:
(D.C. SHARMA)

जाँचकर्ता:
Checked by:
(NIDHI SINGH)

प्रभारी वैज्ञानिक:
Scientist-in-charge:
(NIDHI SINGH)

जारीकर्ता:
Issued by:
Director's Nominee
डॉ. श्रीनिवास राव रागम
Dr. Srinivasa Rao Ragam