आर - राष्ट्रीय भौतिक प्रयोगशाला
IONAL PHYSICAL LABORATORY
वैज्ञानिक तथा औद्योगिक अनुसंधान परिषद)
ouncil of Scientific and Industrial Research)
की संस्थान (एनएमआई), सदस्य बीआईपीएस एवं हस्ताक्षरकर्ता सीआईपीएम - एमआरए)
Metrology Institute (NMI), Member BIPM and Signatory CIPM - MRA)
डॉ. के. एस. कृष्णन् मार्ग, नई दिल्ली-110012, भारत
Dr. K. S. Krishnan Marg, New Delhi-110012 INDIA
none : +91-11-4560, 8441, 8610, 9447, फैक्स/Fax : 91-11-4560 8448
मल/E-mail : cfct@nplindia.org, वेबसाईट/Website : www.nplindia.org

अंशांकन प्रमाण पत्र
CALIBRATION CERTIFICATE
A SET OF WEIGHTS

प्रमाण पत्र संख्या / Certificate No.

25060371/D1.01/C-054

क/ Date	अगले अंशांकन हेतु अनुशंसित तिथि Recommended date for next calibration	पृष्ठ / Page	पृष्ठों की संख्या / No. of Pa ges
.2025	02.09.2028	1	2

alibrated for : Puniska Injectable Pvt.Ltd.
Plot No.1, Phase - III, Honda Industrial Estate,
Honda Sattari
Goa - 403530, India

Customer's Ref. No. : Letter, Dated : 23.06.2025
From : S-Tech Services , Ponda , Goa

2. **Description and identification of instrument**

200 g to 1g (11 Nos.)	:	Integral knob cylindrical stainless steel weights.
500 mg to 1 mg (12 Nos.)	:	Stainless steel wire weights.
Assumed Density (d)	:	(7950 ± 140) kg/m^3; k = 2
Make	:	Sartorius
Model No.	:	YCS011-522-00
Serial No.	:	32729758
Equipment ID No.	:	Q06WB00300

3. **Environmental conditions** : Temperature : (23.0 ± 1.5) °C
Relative humidity : (50 ± 10) %
[Change in temperature during the calibration was less than ± 0.7 °C per hour]

4. **Standard(s) used with associated Measurement uncertainty** : NPL working standard (Sub-Div.#1.01/M/224(2)/00 & Sub-Div.#1.01/M/60027/55) of mass with uncertainty better than one-third of the reported uncertainty of measurement.

5. **Metrological Traceability of Standard(s) used** : The standard(s) used for calibration is(are) traceable to the National Standard which realize the units of quantities according to the International System of Units (SI).

6. **Principle/Methodology of Calibration & Calibration Procedure number** : CSIR-NPL Calibration Procedure No. Sub-Div.#1.01/Doc.3/ CP#WT/M-03. Method of comparison with the CSIR-NPL working standard(s) using substitution weighing. The reported mass value(s) is(are) the conventional mass value(s) (M_c) related to the true mass value(s) (M_t) by formula : $M_c = M_t \times [1 - 1.2 \times (1/d - 1/8\,000)]$.

अंशांकनकर्ता: Calibrated by:	जाँचकर्ता: Checked by:	प्रभारी वैज्ञानिक: Scientist-in-charge:
(D.C. SHARMA)	(NIDHI SINGH)	(NIDHI SINGH)

आर - राष्ट्रीय भौतिक प्रयोगशाला
TIONAL PHYSICAL LABORATORY
(वैज्ञानिक तथा औद्योगिक अनुसंधान परिषद)
Council of Scientific and Industrial Research)
की संस्थान (एनएमआई), सदस्य बीआईपीएस एवं हस्ताक्षरकर्ता सीआईपीएम - एमआरए)
Metrology Institute (NMI), Member BIPM and Signatory CIPM - MRA)
डॉ. के. एस. कृष्णन् मार्ग, नई दिल्ली-110012, भारत
Dr. K. S. Krishnan Marg, New Delhi-110012 INDIA
none : +91-11-4560, 8441, 8610, 9447, फैक्स/Fax : 91-11-4560 8448
मेल/E-mail : cfct@nplindia.org, वेबसाईट/Website : www.nplindia.org

अंशांकन प्रमाण पत्र
CALIBRATION CERTIFICATE
A SET OF WEIGHTS

प्रमाण पत्र संख्या / Certificate No.

25060371/D1.01/C-054

क/ Date	अगले अंशांकन हेतु अनुशंसित तिथि Recommended date for next calibration	पृष्ठ / Page	पृष्ठों की संख्या / No. of Pages
2.09.2025	02.09.2028	2	2

. Results :

Denomination	Identification No.	Mass Value (g)	Uncertainty (g)
200 g	E2A	200.000 03	± 0.000 10
. 200 g	E29	199.999 94	± 0.000 10
100 g	CT2	99.999 90	± 0.000 05
50 g	C83	49.999 97	± 0.000 03
20 g	CXM	20.000 039	± 0.000 025
. 20 g	CXN	20.000 010	± 0.000 025
10 g	C2G	10.000 012	± 0.000 020
5 g	BX1	5.000 019	± 0.000 016
2 g	CT7	2.000 014	± 0.000 012
. 2 g	CT9	2.000 019	± 0.000 012
1 g	CFL	1.000 005	± 0.000 010
500 mg	-	0.500 001	± 0.000 008
200 mg	-	0.200 005	± 0.000 006
. 200 mg	-	0.200 006	± 0.000 006
100 mg	-	0.100 010	± 0.000 005
50 mg	-	0.050 004	± 0.000 004
20 mg	-	0.020 004	± 0.000 003
. 20 mg	-	0.020 003	± 0.000 003
10 mg	-	0.010 003	± 0.000 003
5 mg	-	0.005 000	± 0.000 003
2 mg	-	0.002 001	± 0.000 003
. 2 mg	-	0.002 002	± 0.000 003
1 mg	-	0.001 003	± 0.000 003

The reported uncertainty is at coverage factor k = 2 which corresponds to a coverage probability of approximately 95 % for a normal distribution.

8. **Date(s) of Calibration** : 23.08.2025

9. **Remarks** :
   1. Mass value(s) of the weight(s) is(are) within the maximum errors permissible in E_2 accuracy class of weights as per OIML R 111-1 : 2004.
   2. Dots(.) are used to distinguish the weights of same nominal value.

अंशांकनकर्ता: Calibrated by:	जाँचकर्ता: Checked by:	प्रभारी वैज्ञानिक: Scientist-in-charge:
(D.C. SHARMA)	(NIDHI SINGH)	(NIDHI SINGH)

जारीकर्ता: