सीं एस आई आर – राष्ट्रीय भौतिक प्रयोगशाला
CSIR-NATIONAL PHYSICAL LABORATORY
(वैज्ञानिक तथा औद्योगिक अनुसंधान परिषद्)
(Council of Scientific and Industrial Research)
(राष्ट्रीय मापिकी संस्थान (एनएमआई), सदस्य बीआईपीएस एवं हस्ताक्षरकर्ता सीआईपीएम - एमआरए)
(National Metrology Institute (NMI), Member BIPM and Signatory CIPM - MRA)
डॉ. के. एस. कृष्णन् मार्ग, नई दिल्ली-110012, भारत
Dr. K. S. Krishnan Marg, New Delhi-110012 INDIA
दूरभाष/Phone : +91-11-4560, 8441, 8610, 9447, फैक्स/Fax : 91-11-4560 8448
ई-मेल/E-mail : cfct@nplindia.org, वेबसाईट/Website : www.nplindia.org

अंशांकन प्रमाण पत्र
CALIBRATION CERTIFICATE
WEIGHTS

प्रमाण पत्र संख्या / Certificate No.

25080613/D1.01/C-140

दिनांक/ Date	अगले अंशांकन हेतु अनुशंसित तिथि Recommended date for next calibration	पृष्ठ / Page	पृष्ठों की संख्या / No. of Pages
26.11.2025	26.11.2028	1	2

1. Calibrated for : IOCL Panipat
Indian Oil Corporation Limited (IOCL)
Refineries Division,
Panipat Refinery
Panipat - 132 140, Haryana, India

Customer's Ref. No. : 31/WI/25-26 Dated : 29.07.2025
From : Weigh India, New Delhi

2. Description and identification of instrument

200 g & 10 g (02 Nos.)	:	Integral knob cylindrical austenitic stainless steel weights.
50 mg (1 No.)	:	Austenitic stainless steel wire weight.
Density (d)	:	(7 950 ± 50) kg/m³; k = 2
Make	:	Weigh India

3. Environmental conditions : Temperature : (23.0 ± 1.5) °C
Relative humidity : (50 ± 10) %
[Change in temperature and relative humidity during the calibration were less than ± 0.3 °C per hour and ± 5.0 % per 4 hours respectively]

4. Standard(s) used with associated uncertainty : CSIR-NPL 1 kg transfer standard [Ś] of mass with uncertainty of ± 0.053 mg; k = 2.

5. Traceability of Standard(s) used : The standards used for calibration is(are) traceable to the National Standard which realize the physical units of mass according to the International System of units (SI).

6. Principle/Methodology of Calibration & Calibration Procedure number : CSIR-NPL Calibration Procedures No. Sub.Div#1.01/Doc.3/CP#WT/M-02. Method of comparison with the CSIR-NPL transfer standard(s) using subdivision weighing. The reported mass value(s) is (are) the conventional mass value(s) (M_c) related to the true mass value (s) (M_t) by formula : $M_c = M_t \times [1 - 1.2 \times (1/d - 1/8\,000)]$.

अंशांकनकर्ता:
Calibrated by:
(D.C. SHARMA)

जाँचकर्ता:
Checked by:
(NIDHI SINGH)

प्रभारी वैज्ञानिक:
Scientist-in-charge:
(NIDHI SINGH)

जारीकर्ता:
Issued by:
डॉ. श्रीनिवास राव रागम
Dr. Srinivasa Rao Ragam

सी एस आई आर - राष्ट्रीय भौतिक प्रयोगशाला
CSIR-NATIONAL PHYSICAL LABORATORY
(वैज्ञानिक तथा औद्योगिक अनुसंधान परिषद्)
(Council of Scientific and Industrial Research)
(राष्ट्रीय मापिकी संस्थान (एनएमआई), सदस्य बीआईपीएस एवं हस्ताक्षरकर्ता सीआईपीएम - एमआरए)
(National Metrology Institute (NMI), Member BIPM and Signatory CIPM - MRA)
डॉ. के. एस. कृष्णन् मार्ग, नई दिल्ली-110012, भारत
Dr. K. S. Krishnan Marg, New Delhi-110012 INDIA
दूरभाष/Phone : +91-11-4560, 8441, 8610, 9447, फैक्स/Fax : 91-11-4560 8448
ई-मेल/E-mail : cfct@nplindia.org, वेबसाईट/Website : www.nplindia.org

अंशांकन प्रमाण पत्र
CALIBRATION CERTIFICATE
WEIGHTS

प्रमाण पत्र संख्या / Certificate No.

25080613/D1.01/C-140

दिनांक/ Date	अगले अंशांकन हेतु अनुशंसित तिथि Recommended date for next calibration	पृष्ठ / Page	पृष्ठों की संख्या / No. of Pages
26.11.2025	26.11.2028	2	2

7. Results :

Denomination	Mass Value (g)	Uncertainty (g)
200 g	200.000 039	± 0.000 030
10 g	10.000 010	± 0.000 006
50 mg	0.050 002	± 0.000 002

The reported uncertainty is at coverage factor k = 2 which corresponds to a coverage probability of approximately 95 % for a normal distribution.

8. Date(s) of Calibration : 28.10.2025

9. Remarks : Mass value(s) of the weight(s) is(are) within the maximum errors permissible in E_1 accuracy class of weights as per OIML R 111-1 : 2004.

Density of weights are provided by customer.

अंशांकनकर्ता:
Calibrated by:
(D.C. SHARMA)

जाँचकर्ता:
Checked by:
(NIDHI SINGH)

प्रभारी वैज्ञानिक:
Scientist-in-charge:
(NIDHI SINGH)

जारीकर्ता:
Issued by:
डॉ. श्रीनिवास राव रागम
Dr. Srinivasa Rao Ragam