


Rear view


J 1	: Supply and anode driver	277-10B40
J 2	: Voltmeter and cathode driver	277-10C40
J 3	: Option: Overload LED driver	277-10G40
J 4	: Not used	
J 5	: Log. amplifier	277-10D40
J 6	: Log. amplifier	277-10D40
J 7	: Input amplifier	277-10E40

etter

J 5 J 6



Since the instrument is adjusted correctly on delivery, adjustment only has to be carried out in case of faulty mode of operation, i.e. when a component has failed and has been replaced.



TEST SET-UP.

1. Connect +24V DC to pin A on J1 (0 Volt to pin B on J1)
2. Feed a signal e.g. 5 kHz sinusoidal to the input terminals on input board (277-10E40) for all channels, i.e. pin A and pin 1 for channel 1, pin B and pin 2 for channel 2 a.s.o. up to pin J and pin 8 for channel 8. Adjust the amplitude of the signal for max. reading on the displays (top of the scales).

ANODE VOLTAGE ADJUST.

1. Remove the instrument from the cabinet.
CAUTION ! Be aware that when supply voltage is supplied, high voltage is on the circuits (250V DC)
2. Set P2 on the supply and anode driver board (277-10B40) to mid position.
3. Turn P1 on the supply and anode driver board (277-10B40) max. CW and then slowly CCW as far as possible without getting a flickering display.
4. Measure the current consumption.
5. Adjust P1 to obtain an increase in current consumption of approx. 10%.

LIGHT INTENSITY ADJUSTMENT.

1. Adjust P2 on the supply and anode driverboard until a current consumption of 450 mA is obtained (or to desired light intensity).

VOLTMETER SENSITIVITY ADJUSTMENT.

1. Connect a voltmeter to the wiper (center pin) on P1 on the voltmeter & cathode driver board (277-10C40). The voltage should be 3V. If not, adjust P1.

LOW LEVEL & REF. LEVEL ADJUSTMENT.

1. Connect a signal of -40dB to input terminals.
2. Adjust the low level potmeter for all channels to obtain reading (-40 on the DIN-scale).
3. Change the signal to ref.level and adjust the ref.level potmeters for all channels to obtain correct reading (0 on DIN-scale).
4. Repeat 1-3.

FALL-BACK TIME ADJUSTMENT.

1. Connect a burst generator e.g. NTP type 507-100 to input terminals.
2. Adjust the fall-back potmeters to obtain recommended fall-back time (1.5 sec. from "0" to "-20" on DIN-scale)