

Problem Statement and Goals

CADENCE

Team #4, CADENCE

Anam Khan
Haniya Kashif
Vivian Nguyen
Kelsey Rough
Hareem Salman

Table 1: Revision History

Date	Developer(s)	Change
2026-09-25	Anam Khan, Haniya Kashif, Vivian Nguyen, Kelsey Rough, Hareem Salman	Initial version of the problem statement and goals: problem, inputs and outputs, stakeholders, environment, goals, stretch goals, custom documents, and reflection

1 Problem Statement

In the laboratory, Doppler ultrasound measurements can be used to study blood flow in arteries such as the brachial and femoral arteries. Previously, the laboratory used a Multigon system to process the audio output from an ultrasound machine and convert the Doppler signal into a voltage signal that could be continuously recorded and analyzed using LabChart.

The existing Multigon system is no longer functional, creating a need for a modern replacement. Without a functioning system, researchers cannot readily convert the Doppler ultrasound output into a continuous measure of MBV using the laboratory's existing recording workflow.

1.1 Problem

We want to provide a system that will receive the Doppler audio output from an ultrasound machine, process the signal, and generate a continuous voltage output that can be recorded by LabChart for subsequent analysis of mean blood velocity.

1.2 Inputs and Outputs

The primary input to the system is the Doppler ultrasound audio signal representing blood flow within an artery.

The primary output is a continuous signal representing mean blood velocity that can be recorded and analyzed using LabChart. The system should also provide sufficient information to allow researchers to verify that the resulting measurements are reasonable and reliable.

1.3 Stakeholders

The primary stakeholders are the researchers and laboratory staff who require MBV measurements for physiological research. The project supervisors and biomedical researchers are also stakeholders, as they will provide domain knowledge, define research requirements, and evaluate the system's results. Ultrasound operators are stakeholders because the system must work with the output produced by the laboratory's ultrasound equipment. Future research participants are indirect stakeholders because the system may eventually be used to obtain measurements during human research studies.

1.4 Environment

The system will operate in a laboratory environment alongside an existing Doppler ultrasound system and LabChart data acquisition software. Researchers will use the system to obtain continuous MBV measurements during physiological experiments. The system will require appropriate hardware for receiving the ultrasound audio output and producing a signal that can be recorded by LabChart.

2 Goals

The primary goal of this project is to develop and validate a functional digital signal processing system that can derive mean blood velocity from the Doppler ultrasound audio output and provide a continuous signal compatible with LabChart.

The system goals are to:

1. Receive and process the audio output from a Doppler ultrasound system in real time.
2. Investigate and implement appropriate signal-processing techniques for extracting information related to blood velocity.
3. Calculate a continuous estimate of mean blood velocity from the processed Doppler signal.
4. Generate an output signal that can be recorded and analyzed using LabChart.

5. Formally validating the system's output against a ground truth, appropriate reference measurements or established methods.
6. Provide documentation describing the system, its operation, and its limitations.

3 Stretch Goals

Additional goals we'd hope to implement in CADENCE include but are not limited to the following: 1. Using the same input signal from the ultrasound, we can apply different signal transformation algorithms to derive metrics other than just MBV, such as Peak systolic velocity (PSV), End-diastolic velocity (EDV), Pulsatility Index (PI), and Resistance Index (RI) to gain further insight into the research participant's cardiovascular health. 2. Alongside the LabChart output, the system displays a real-time spectrogram (a color-coded time-frequency display) to provide a more insightful diagnostic trail for research purposes.

4 Custom Documents (CDs)

For the Fall term, we have chosen to complete a User/Stakeholder/Client Interview Report. Our project is being developed for the Vascular Dynamics Lab, and the researchers and graduate students there are the people who will actually be using our system. Many of the lab's requirements come from how they currently use the Multigon, such as its velocity ranges, offset, calibration signal, and output voltage, and a lot of this knowledge isn't formally documented anywhere. Because nobody on our team has a background in the medical field, interviewing the lab and our supervisors will help us capture these requirements accurately and bridge the knowledge gap between our team and the lab. This will also help us make sure that our system fits into the lab's current requirements and how they hope to use the system.

For the Winter term, we have chosen to complete a Performance Report. Since our system needs to process the Doppler signal and produce a continuous output in real-time, it is essential that the system's output be both precise and accurate. The delay between the incoming Doppler signal and the output voltage recorded in LabChart directly affects whether the lab can use our measurements for research. A performance report will allow us to formally measure and analyze the latency and accuracy of our final system on its hardware, and determine whether or not it meets the project requirements and the lab's needs.

Appendix — Reflection

1. What went well while writing this deliverable?

For this deliverable, we found that identifying the core problem and user needs was straightforward. The Vascular Dynamics Lab had an existing Multigon machine that is broken and discontinued, and they need a replacement. They told us clearly that they were looking for a system that takes the ultrasound's Doppler audio and outputs a continuous mean blood velocity voltage signal for their LabChart software to interpret. The lab has been incredibly helpful in providing us with manuals, test data, and time in the lab to better understand their system and how it works. They also gave us ideas for stretch goals and requirements, and helped us understand their primary user needs. We were also able to get help from Dr. Von Mohrenschildt, who advised us on what information we need to gather about the system's electronics and existing signals, which helped us understand what approach we need to take with the project.

Within our team, we split up the work evenly, each taking on parts of the problem statement, goals, and project research. When it came to determining all of the technical details we needed, we met to research together, each taking on specific elements or interfaces within the system. We also shared responsibilities such as contacting supervisors and the lab to arrange meetings, tracking the administrative tasks we needed to do, and organizing our shared resources as a group.

We found that meeting with the lab and our supervisor early was one of our best decisions. Seeing the equipment in person changed how we wrote the problem statement because instead of describing a generic signal processing tool, we were able to focus it on the specific input and output the lab actually needs. It also helped us identify the lab's needs, along with the information and context we needed early on, which gave us what we needed to accurately plan our goals and development.

2. What pain points did you experience during this deliverable, and how did you resolve them?

The biggest pain point we found for this deliverable was the number of unknowns about the existing system our project aims to replace. Because it's legacy hardware, there's little documentation for operating it, much less technical details such as current and voltage draw, interface impedances, signal formats and phase shifts, data bit depth, and other key details we need to accurately determine our next steps. We addressed a lot of this by going into the lab a couple of times to see the machine, watch some tests being done, and see how the output looks. We haven't resolved all the unknowns yet, so we wrote our goals around what the system needs to deliver rather than around specs we haven't confirmed, and we will be updating these as soon as we have confirmed answers. We're returning to the lab next week to measure the remaining unknowns with a multimeter

and oscilloscope, and we'll update the documents based on what we find. A lot of this project is new to our team, so we've had to do a lot of research and learning as we go, and looking back we recognize that we may have also benefitted from putting together a list of specific measurements to take before our first lab visit instead of after.

One of the other pain points we experienced was the difference in domain knowledge on both sides. The Vascular Dynamics Lab is an amazing resource on the physiology side, but they don't have the technical engineering background that we do. On the other hand, we've had a lot to learn about the medical side of this project and its implications for our design. To bridge this gap, when we visited the lab, we asked them to demonstrate to us how they use the system, and we took notes on how it works, what readings they get, and what may be required for us to provide. We also had a gap between our current technical knowledge and what the project requires. Dr. Von Mohrenschildt was a great help with this. When we met with him, he explained key concepts and helped us see what we need for the project and how to approach it. The main thing we learned is that asking specific questions got us much better answers than general ones, so we plan to come to future meetings with questions prepared ahead of time.

3. How did you and your team adjust the scope of your goals to ensure they are suitable for a Capstone project (not overly ambitious but also of appropriate complexity for a senior design project)?

As our team looked further into the system we need to develop, it became clear that we needed to adjust our goals to make sure they are feasible and suitable for this capstone. Initially, we had ideas of providing other outputs, such as peak systolic velocity, end-diastolic velocity, pulsatility index, and resistance index, along with a real-time spectrogram display. However, from being in the lab, examining the equipment, and speaking with the lab group, it became clear that the core problem is complex enough on its own. Most of these extra metrics can be derived from the same velocity data, but each one would need its own implementation, testing, and verification. That would pull time and effort away from getting the core mean blood velocity output right. The lab also has ways of interpreting these already, so the addition of these would add little to their current research. Because this system will be used for medical research, an accurate and reliable core output matters far more than extra features. We adjusted our goals to prioritize system accuracy and reliability, along with clear documentation and verification of correctness, so the system is fit for research use and easy for the lab to operate. We haven't decided against the additional metrics entirely, and we do have some stretch goals that we aim to complete if the core output is validated with time to spare. We believe this scope is appropriately challenging, because it involves real-time signal processing, hardware interfacing, and validation against real laboratory equipment, which are all relevant to the workforce

and also allow us to apply concepts learned in SFWRENG 3MX3, 3DX4,
and 4AA4.