

## Laboratory Report of Analysis

To: Kenai Watershed Forum  
44129 Sterling Highway  
Soldotna, AK 99669  
(907)260-5449

Report Number: **1194285**

Client Project: **Kenai River Water Quality**

Dear Branden Bornemann,

Enclosed are the results of the analytical services performed under the referenced project for the received samples and associated QC as applicable. The samples are certified to meet the requirements of the National Environmental Laboratory Accreditation Conference Standards. Copies of this report and supporting data will be retained in our files for a period of ten years in the event they are required for future reference. All results are intended to be used in their entirety and SGS is not responsible for use of less than the complete report. Any samples submitted to our laboratory will be retained for a maximum of fourteen (14) days from the date of this report unless other archiving requirements were included in the quote.

If there are any questions about the report or services performed during this project, please call Justin at (907) 562-2343. We will be happy to answer any questions or concerns which you may have.

Thank you for using SGS North America Inc. for your analytical services. We look forward to working with you again on any additional analytical needs.

Sincerely,  
SGS North America Inc.

---

Justin Nelson  
Project Manager  
Justin.Nelson@sgs.com

---

Date

## Case Narrative

SGS Client: **Kenai Watershed Forum**  
 SGS Project: **1194285**  
 Project Name/Site: **Kenai River Water Quality**  
 Project Contact: **Branden Bornemann**

Refer to sample receipt form for information on sample condition.

### **USFWS - Cunningham Park (1194285006) PS**

Calcium, Iron and Magnesium by 200.7 were analyzed by ALS of Kelso, WA.

### **1194397001MSD (1524595) MSD**

4500NO3-F - Nitrate/Nitrite - MSD recovery for Total Nitrite / Nitrate is outside of QC criteria. Refer to LCS for accuracy requirements.

\*QC comments may be associated with the field samples found in this report. When applicable, comments will be applied to associated field samples.

Print Date: 08/20/2019 9:46:01AM

## Laboratory Qualifiers

Enclosed are the analytical results associated with the above work order. The results apply to the samples as received. All results are intended to be used in their entirety and SGS is not responsible for use of less than the complete report. This document is issued by the Company under its General Conditions of Service accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein.

Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. Any unauthorized alteration, forgery or falsification of the context or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

SGS maintains a formal Quality Assurance/Quality Control (QA/QC) program. A copy of our Quality Assurance Plan (QAP), which outlines this program, is available at your request. The laboratory certification numbers are AK00971 (DW Chemistry & Microbiology) & 17-021 (CS) for ADEC and 2944.01 for DOD ELAP/ISO17025 (RCRA methods: 1020B, 1311, 3010A, 3050B, 3520C, 3550C, 5030B, 5035A, 6020A, 7470A, 7471B, 8015C, 8021B, 8082A, 8260C, 8270D, 8270D-SIM, 9040C, 9045D, 9056A, 9060A, AK101 and AK102/103). SGS is only certified for the analytes listed on our Drinking Water Certification, and only those analytes will be reported to the State of Alaska for compliance. Except as specifically noted, all statements and data in this report are in conformance to the provisions set forth by the SGS QAP and, when applicable, other regulatory authorities.

The following descriptors or qualifiers may be found in your report:

*	The analyte has exceeded allowable regulatory or control limits.
!	Surrogate out of control limits.
B	Indicates the analyte is found in a blank associated with the sample.
CCV/CVA/CVB	Continuing Calibration Verification
CCCV/CVC/CVCA/CVCB	Closing Continuing Calibration Verification
CL	Control Limit
DF	Analytical Dilution Factor
DL	Detection Limit (i.e., maximum method detection limit)
E	The analyte result is above the calibrated range.
GT	Greater Than
IB	Instrument Blank
ICV	Initial Calibration Verification
J	The quantitation is an estimation.
LCS(D)	Laboratory Control Spike (Duplicate)
LLQC/LLIQC	Low Level Quantitation Check
LOD	Limit of Detection (i.e., 1/2 of the LOQ)
LOQ	Limit of Quantitation (i.e., reporting or practical quantitation limit)
LT	Less Than
MB	Method Blank
MS(D)	Matrix Spike (Duplicate)
ND	Indicates the analyte is not detected.
RPD	Relative Percent Difference
U	Indicates the analyte was analyzed for but not detected.

Note: Sample summaries which include a result for "Total Solids" have already been adjusted for moisture content. All DRO/RRO analyses are integrated per SOP.

### Sample Summary

<u>Client Sample ID</u>	<u>Lab Sample ID</u>	<u>Collected</u>	<u>Received</u>	<u>Matrix</u>
USFWS - Cunningham Park	1194285001	07/30/2019	07/31/2019	Water (Surface, Eff., Ground)
USFWS - Beaver Creek	1194285002	07/30/2019	07/31/2019	Water (Surface, Eff., Ground)
USFWS - Kenai River	1194285003	07/30/2019	07/31/2019	Water (Surface, Eff., Ground)
USFWS - Pillars	1194285004	07/30/2019	07/31/2019	Water (Surface, Eff., Ground)
USFWS - Poachers Cove	1194285005	07/30/2019	07/31/2019	Water (Surface, Eff., Ground)
USFWS - Cunningham Park	1194285006	07/30/2019	07/31/2019	Water (Surface, Eff., Ground)
USFWS - Beaver Creek	1194285007	07/30/2019	07/31/2019	Water (Surface, Eff., Ground)
USFWS - Kenai River	1194285008	07/30/2019	07/31/2019	Water (Surface, Eff., Ground)
USFWS - Pillars	1194285009	07/30/2019	07/31/2019	Water (Surface, Eff., Ground)
USFWS - Poachers Cove	1194285010	07/30/2019	07/31/2019	Water (Surface, Eff., Ground)
Trip Blank	1194285011	07/30/2019	07/31/2019	Water (Surface, Eff., Ground)

#### Method

EPA 602/624

EP200.8

SM21 4500NO3-F

SM21 4500P-B,E

#### Method Description

602 Aromatics by 624 (W)

Metals in Drinking Water by ICP-MS DISSO

Nitrate/Nitrite Flow injection Pres.

Total Phosphorus (W)

## Detectable Results Summary

Client Sample ID: **USFWS - Cunningham Park**

Lab Sample ID: 1194285001

**Dissolved Metals by ICP/MS**

**Volatile GC/MS**

**Waters Department**

<u>Parameter</u>	<u>Result</u>	<u>Units</u>
Copper	0.428J	ug/L
Toluene	0.370J	ug/L
Total Nitrate/Nitrite-N	0.191J	mg/L
Total Phosphorus	0.0169J	mg/L

Client Sample ID: **USFWS - Beaver Creek**

Lab Sample ID: 1194285002

**Dissolved Metals by ICP/MS**

**Waters Department**

<u>Parameter</u>	<u>Result</u>	<u>Units</u>
Arsenic	4.87J	ug/L
Copper	1.77	ug/L
Lead	3.20	ug/L
Zinc	69.9	ug/L
Total Phosphorus	0.0451	mg/L

Client Sample ID: **USFWS - Kenai River**

Lab Sample ID: 1194285003

**Dissolved Metals by ICP/MS**

**Waters Department**

<u>Parameter</u>	<u>Result</u>	<u>Units</u>
Copper	0.497J	ug/L
Lead	0.697	ug/L
Zinc	10.7	ug/L
Total Nitrate/Nitrite-N	0.180J	mg/L
Total Phosphorus	0.00930J	mg/L

Client Sample ID: **USFWS - Pillars**

Lab Sample ID: 1194285004

**Dissolved Metals by ICP/MS**

**Waters Department**

<u>Parameter</u>	<u>Result</u>	<u>Units</u>
Copper	0.452J	ug/L
Zinc	8.14J	ug/L
Total Nitrate/Nitrite-N	0.185J	mg/L
Total Phosphorus	0.00920J	mg/L

Client Sample ID: **USFWS - Poachers Cove**

Lab Sample ID: 1194285005

**Dissolved Metals by ICP/MS**

**Waters Department**

<u>Parameter</u>	<u>Result</u>	<u>Units</u>
Copper	0.420J	ug/L
Total Nitrate/Nitrite-N	0.182J	mg/L
Total Phosphorus	0.00740J	mg/L

## Results of USFWS - Cunningham Park

Client Sample ID: **USFWS - Cunningham Park**  
 Client Project ID: **Kenai River Water Quality**  
 Lab Sample ID: 1194285001  
 Lab Project ID: 1194285

Collection Date: 07/30/19 09:11  
 Received Date: 07/31/19 16:40  
 Matrix: Water (Surface, Eff., Ground)  
 Solids (%):  
 Location:

## Results by Dissolved Metals by ICP/MS

<u>Parameter</u>	<u>Result Qual</u>	<u>LOQ/CL</u>	<u>DL</u>	<u>Units</u>	<u>DF</u>	<u>Allowable Limits</u>	<u>Date Analyzed</u>
Arsenic	2.50 U	5.00	1.50	ug/L	1		08/05/19 20:58
Cadmium	0.250 U	0.500	0.150	ug/L	1		08/05/19 20:58
Chromium	1.00 U	2.00	0.800	ug/L	1		08/05/19 20:58
Copper	0.428 J	1.00	0.310	ug/L	1		08/05/19 20:58
Lead	0.100 U	0.200	0.0700	ug/L	1		08/05/19 20:58
Zinc	5.00 U	10.0	3.10	ug/L	1		08/05/19 20:58

## Batch Information

Analytical Batch: MMS10582  
 Analytical Method: EP200.8  
 Analyst: DSH  
 Analytical Date/Time: 08/05/19 20:58  
 Container ID: 1194285001-B

Prep Batch: MXX32633  
 Prep Method: E200.2  
 Prep Date/Time: 08/05/19 10:22  
 Prep Initial Wt./Vol.: 20 mL  
 Prep Extract Vol: 50 mL

## Results of USFWS - Cunningham Park

Client Sample ID: **USFWS - Cunningham Park**  
 Client Project ID: **Kenai River Water Quality**  
 Lab Sample ID: 1194285001  
 Lab Project ID: 1194285

Collection Date: 07/30/19 09:11  
 Received Date: 07/31/19 16:40  
 Matrix: Water (Surface, Eff., Ground)  
 Solids (%):  
 Location:

## Results by Volatile GC/MS

<u>Parameter</u>	<u>Result Qual</u>	<u>LOQ/CL</u>	<u>DL</u>	<u>Units</u>	<u>DF</u>	<u>Allowable Limits</u>	<u>Date Analyzed</u>
Benzene	0.200 U	0.400	0.120	ug/L	1		08/05/19 17:38
Ethylbenzene	0.500 U	1.00	0.310	ug/L	1		08/05/19 17:38
o-Xylene	0.500 U	1.00	0.310	ug/L	1		08/05/19 17:38
P & M -Xylene	1.00 U	2.00	0.620	ug/L	1		08/05/19 17:38
Toluene	0.370 J	1.00	0.310	ug/L	1		08/05/19 17:38
Xylenes (total)	1.50 U	3.00	1.00	ug/L	1		08/05/19 17:38
<b>Surrogates</b>							
1,2-Dichloroethane-D4 (surr)	105	81-118		%	1		08/05/19 17:38
4-Bromofluorobenzene (surr)	99.4	85-114		%	1		08/05/19 17:38
Toluene-d8 (surr)	98.4	89-112		%	1		08/05/19 17:38

## Batch Information

Analytical Batch: VMS19267  
 Analytical Method: EPA 602/624  
 Analyst: CMC  
 Analytical Date/Time: 08/05/19 17:38  
 Container ID: 1194285001-C

Prep Batch: VXX34588  
 Prep Method: SW5030B  
 Prep Date/Time: 08/05/19 06:00  
 Prep Initial Wt./Vol.: 5 mL  
 Prep Extract Vol: 5 mL

## Results of USFWS - Cunningham Park

Client Sample ID: **USFWS - Cunningham Park**  
 Client Project ID: **Kenai River Water Quality**  
 Lab Sample ID: 1194285001  
 Lab Project ID: 1194285

Collection Date: 07/30/19 09:11  
 Received Date: 07/31/19 16:40  
 Matrix: Water (Surface, Eff., Ground)  
 Solids (%):  
 Location:

## Results by Waters Department

<u>Parameter</u>	<u>Result Qual</u>	<u>LOQ/CL</u>	<u>DL</u>	<u>Units</u>	<u>DF</u>	<u>Allowable Limits</u>	<u>Date Analyzed</u>
Total Nitrate/Nitrite-N	0.191 J	0.200	0.0500	mg/L	2		08/09/19 15:20

## Batch Information

Analytical Batch: WFI2832  
 Analytical Method: SM21 4500NO3-F  
 Analyst: EWW  
 Analytical Date/Time: 08/09/19 15:20  
 Container ID: 1194285001-A

<u>Parameter</u>	<u>Result Qual</u>	<u>LOQ/CL</u>	<u>DL</u>	<u>Units</u>	<u>DF</u>	<u>Allowable Limits</u>	<u>Date Analyzed</u>
Total Phosphorus	0.0169 J	0.0200	0.00500	mg/L	1		08/15/19 11:53

## Batch Information

Analytical Batch: WDA4629  
 Analytical Method: SM21 4500P-B,E  
 Analyst: DMM  
 Analytical Date/Time: 08/15/19 11:53  
 Container ID: 1194285001-A

Prep Batch: WXX12974  
 Prep Method: SM21 4500P-B,E  
 Prep Date/Time: 08/14/19 21:30  
 Prep Initial Wt./Vol.: 25 mL  
 Prep Extract Vol: 25 mL



## Results of USFWS - Beaver Creek

Client Sample ID: **USFWS - Beaver Creek**  
 Client Project ID: **Kenai River Water Quality**  
 Lab Sample ID: 1194285002  
 Lab Project ID: 1194285

Collection Date: 07/30/19 09:54  
 Received Date: 07/31/19 16:40  
 Matrix: Water (Surface, Eff., Ground)  
 Solids (%):  
 Location:

## Results by Dissolved Metals by ICP/MS

<u>Parameter</u>	<u>Result Qual</u>	<u>LOQ/CL</u>	<u>DL</u>	<u>Units</u>	<u>DF</u>	<u>Allowable Limits</u>	<u>Date Analyzed</u>
Arsenic	4.87 J	5.00	1.50	ug/L	1		08/05/19 21:01
Cadmium	0.250 U	0.500	0.150	ug/L	1		08/05/19 21:01
Chromium	1.00 U	2.00	0.800	ug/L	1		08/05/19 21:01
Copper	1.77	1.00	0.310	ug/L	1		08/05/19 21:01
Lead	3.20	0.200	0.0700	ug/L	1		08/05/19 21:01
Zinc	69.9	10.0	3.10	ug/L	1		08/05/19 21:01

## Batch Information

Analytical Batch: MMS10582  
 Analytical Method: EP200.8  
 Analyst: DSH  
 Analytical Date/Time: 08/05/19 21:01  
 Container ID: 1194285002-B

Prep Batch: MXX32633  
 Prep Method: E200.2  
 Prep Date/Time: 08/05/19 10:22  
 Prep Initial Wt./Vol.: 20 mL  
 Prep Extract Vol: 50 mL

## Results of USFWS - Beaver Creek

Client Sample ID: **USFWS - Beaver Creek**  
 Client Project ID: **Kenai River Water Quality**  
 Lab Sample ID: 1194285002  
 Lab Project ID: 1194285

Collection Date: 07/30/19 09:54  
 Received Date: 07/31/19 16:40  
 Matrix: Water (Surface, Eff., Ground)  
 Solids (%):  
 Location:

## Results by Waters Department

<u>Parameter</u>	<u>Result Qual</u>	<u>LOQ/CL</u>	<u>DL</u>	<u>Units</u>	<u>DF</u>	<u>Allowable Limits</u>	<u>Date Analyzed</u>
Total Nitrate/Nitrite-N	0.100 U	0.200	0.0500	mg/L	2		08/09/19 15:22

## Batch Information

Analytical Batch: WFI2832  
 Analytical Method: SM21 4500NO3-F  
 Analyst: EWW  
 Analytical Date/Time: 08/09/19 15:22  
 Container ID: 1194285002-A

<u>Parameter</u>	<u>Result Qual</u>	<u>LOQ/CL</u>	<u>DL</u>	<u>Units</u>	<u>DF</u>	<u>Allowable Limits</u>	<u>Date Analyzed</u>
Total Phosphorus	0.0451	0.0200	0.00500	mg/L	1		08/15/19 11:56

## Batch Information

Analytical Batch: WDA4629  
 Analytical Method: SM21 4500P-B,E  
 Analyst: DMM  
 Analytical Date/Time: 08/15/19 11:56  
 Container ID: 1194285002-A

Prep Batch: WXX12974  
 Prep Method: SM21 4500P-B,E  
 Prep Date/Time: 08/14/19 21:30  
 Prep Initial Wt./Vol.: 25 mL  
 Prep Extract Vol: 25 mL

## Results of USFWS - Kenai River

Client Sample ID: **USFWS - Kenai River**  
 Client Project ID: **Kenai River Water Quality**  
 Lab Sample ID: 1194285003  
 Lab Project ID: 1194285

Collection Date: 07/30/19 10:27  
 Received Date: 07/31/19 16:40  
 Matrix: Water (Surface, Eff., Ground)  
 Solids (%):  
 Location:

## Results by Dissolved Metals by ICP/MS

<u>Parameter</u>	<u>Result Qual</u>	<u>LOQ/CL</u>	<u>DL</u>	<u>Units</u>	<u>DF</u>	<u>Allowable Limits</u>	<u>Date Analyzed</u>
Arsenic	2.50 U	5.00	1.50	ug/L	1		08/05/19 21:09
Cadmium	0.250 U	0.500	0.150	ug/L	1		08/05/19 21:09
Chromium	1.00 U	2.00	0.800	ug/L	1		08/05/19 21:09
Copper	0.497 J	1.00	0.310	ug/L	1		08/05/19 21:09
Lead	0.697	0.200	0.0700	ug/L	1		08/05/19 21:09
Zinc	10.7	10.0	3.10	ug/L	1		08/05/19 21:09

## Batch Information

Analytical Batch: MMS10582  
 Analytical Method: EP200.8  
 Analyst: DSH  
 Analytical Date/Time: 08/05/19 21:09  
 Container ID: 1194285003-B

Prep Batch: MXX32633  
 Prep Method: E200.2  
 Prep Date/Time: 08/05/19 10:22  
 Prep Initial Wt./Vol.: 20 mL  
 Prep Extract Vol: 50 mL

## Results of USFWS - Kenai River

Client Sample ID: **USFWS - Kenai River**  
 Client Project ID: **Kenai River Water Quality**  
 Lab Sample ID: 1194285003  
 Lab Project ID: 1194285

Collection Date: 07/30/19 10:27  
 Received Date: 07/31/19 16:40  
 Matrix: Water (Surface, Eff., Ground)  
 Solids (%):  
 Location:

## Results by Waters Department

<u>Parameter</u>	<u>Result Qual</u>	<u>LOQ/CL</u>	<u>DL</u>	<u>Units</u>	<u>DF</u>	<u>Allowable Limits</u>	<u>Date Analyzed</u>
Total Nitrate/Nitrite-N	0.180 J	0.200	0.0500	mg/L	2		08/09/19 15:24

## Batch Information

Analytical Batch: WFI2832  
 Analytical Method: SM21 4500NO3-F  
 Analyst: EWW  
 Analytical Date/Time: 08/09/19 15:24  
 Container ID: 1194285003-A

<u>Parameter</u>	<u>Result Qual</u>	<u>LOQ/CL</u>	<u>DL</u>	<u>Units</u>	<u>DF</u>	<u>Allowable Limits</u>	<u>Date Analyzed</u>
Total Phosphorus	0.00930 J	0.0200	0.00500	mg/L	1		08/15/19 11:57

## Batch Information

Analytical Batch: WDA4629  
 Analytical Method: SM21 4500P-B,E  
 Analyst: DMM  
 Analytical Date/Time: 08/15/19 11:57  
 Container ID: 1194285003-A

Prep Batch: WXX12974  
 Prep Method: SM21 4500P-B,E  
 Prep Date/Time: 08/14/19 21:30  
 Prep Initial Wt./Vol.: 25 mL  
 Prep Extract Vol: 25 mL

## Results of USFWS - Pillars

Client Sample ID: **USFWS - Pillars**  
 Client Project ID: **Kenai River Water Quality**  
 Lab Sample ID: 1194285004  
 Lab Project ID: 1194285

Collection Date: 07/30/19 10:44  
 Received Date: 07/31/19 16:40  
 Matrix: Water (Surface, Eff., Ground)  
 Solids (%):  
 Location:

## Results by Dissolved Metals by ICP/MS

<u>Parameter</u>	<u>Result Qual</u>	<u>LOQ/CL</u>	<u>DL</u>	<u>Units</u>	<u>DF</u>	<u>Allowable Limits</u>	<u>Date Analyzed</u>
Arsenic	2.50 U	5.00	1.50	ug/L	1		08/05/19 21:12
Cadmium	0.250 U	0.500	0.150	ug/L	1		08/05/19 21:12
Chromium	1.00 U	2.00	0.800	ug/L	1		08/05/19 21:12
Copper	0.452 J	1.00	0.310	ug/L	1		08/05/19 21:12
Lead	0.100 U	0.200	0.0700	ug/L	1		08/05/19 21:12
Zinc	8.14 J	10.0	3.10	ug/L	1		08/05/19 21:12

## Batch Information

Analytical Batch: MMS10582  
 Analytical Method: EP200.8  
 Analyst: DSH  
 Analytical Date/Time: 08/05/19 21:12  
 Container ID: 1194285004-B

Prep Batch: MXX32633  
 Prep Method: E200.2  
 Prep Date/Time: 08/05/19 10:22  
 Prep Initial Wt./Vol.: 20 mL  
 Prep Extract Vol: 50 mL

## Results of USFWS - Pillars

Client Sample ID: **USFWS - Pillars**  
 Client Project ID: **Kenai River Water Quality**  
 Lab Sample ID: 1194285004  
 Lab Project ID: 1194285

Collection Date: 07/30/19 10:44  
 Received Date: 07/31/19 16:40  
 Matrix: Water (Surface, Eff., Ground)  
 Solids (%):  
 Location:

## Results by Waters Department

<u>Parameter</u>	<u>Result Qual</u>	<u>LOQ/CL</u>	<u>DL</u>	<u>Units</u>	<u>DF</u>	<u>Allowable Limits</u>	<u>Date Analyzed</u>
Total Nitrate/Nitrite-N	0.185 J	0.200	0.0500	mg/L	2		08/09/19 15:25

## Batch Information

Analytical Batch: WFI2832  
 Analytical Method: SM21 4500NO3-F  
 Analyst: EWW  
 Analytical Date/Time: 08/09/19 15:25  
 Container ID: 1194285004-A

<u>Parameter</u>	<u>Result Qual</u>	<u>LOQ/CL</u>	<u>DL</u>	<u>Units</u>	<u>DF</u>	<u>Allowable Limits</u>	<u>Date Analyzed</u>
Total Phosphorus	0.00920 J	0.0200	0.00500	mg/L	1		08/15/19 11:58

## Batch Information

Analytical Batch: WDA4629  
 Analytical Method: SM21 4500P-B,E  
 Analyst: DMM  
 Analytical Date/Time: 08/15/19 11:58  
 Container ID: 1194285004-A

Prep Batch: WXX12974  
 Prep Method: SM21 4500P-B,E  
 Prep Date/Time: 08/14/19 21:30  
 Prep Initial Wt./Vol.: 25 mL  
 Prep Extract Vol: 25 mL

## Results of USFWS - Poachers Cove

Client Sample ID: **USFWS - Poachers Cove**  
 Client Project ID: **Kenai River Water Quality**  
 Lab Sample ID: 1194285005  
 Lab Project ID: 1194285

Collection Date: 07/30/19 11:04  
 Received Date: 07/31/19 16:40  
 Matrix: Water (Surface, Eff., Ground)  
 Solids (%):  
 Location:

## Results by Dissolved Metals by ICP/MS

<u>Parameter</u>	<u>Result Qual</u>	<u>LOQ/CL</u>	<u>DL</u>	<u>Units</u>	<u>DF</u>	<u>Allowable Limits</u>	<u>Date Analyzed</u>
Arsenic	2.50 U	5.00	1.50	ug/L	1		08/05/19 21:18
Cadmium	0.250 U	0.500	0.150	ug/L	1		08/05/19 21:18
Chromium	1.00 U	2.00	0.800	ug/L	1		08/05/19 21:18
Copper	0.420 J	1.00	0.310	ug/L	1		08/05/19 21:18
Lead	0.100 U	0.200	0.0700	ug/L	1		08/05/19 21:18
Zinc	5.00 U	10.0	3.10	ug/L	1		08/05/19 21:18

## Batch Information

Analytical Batch: MMS10582  
 Analytical Method: EP200.8  
 Analyst: DSH  
 Analytical Date/Time: 08/05/19 21:18  
 Container ID: 1194285005-B

Prep Batch: MXX32633  
 Prep Method: E200.2  
 Prep Date/Time: 08/05/19 10:22  
 Prep Initial Wt./Vol.: 20 mL  
 Prep Extract Vol: 50 mL

## Results of USFWS - Poachers Cove

Client Sample ID: **USFWS - Poachers Cove**  
 Client Project ID: **Kenai River Water Quality**  
 Lab Sample ID: 1194285005  
 Lab Project ID: 1194285

Collection Date: 07/30/19 11:04  
 Received Date: 07/31/19 16:40  
 Matrix: Water (Surface, Eff., Ground)  
 Solids (%):  
 Location:

## Results by Waters Department

<u>Parameter</u>	<u>Result Qual</u>	<u>LOQ/CL</u>	<u>DL</u>	<u>Units</u>	<u>DF</u>	<u>Allowable Limits</u>	<u>Date Analyzed</u>
Total Nitrate/Nitrite-N	0.182 J	0.200	0.0500	mg/L	2		08/09/19 15:27

## Batch Information

Analytical Batch: WFI2832  
 Analytical Method: SM21 4500NO3-F  
 Analyst: EWW  
 Analytical Date/Time: 08/09/19 15:27  
 Container ID: 1194285005-A

<u>Parameter</u>	<u>Result Qual</u>	<u>LOQ/CL</u>	<u>DL</u>	<u>Units</u>	<u>DF</u>	<u>Allowable Limits</u>	<u>Date Analyzed</u>
Total Phosphorus	0.00740 J	0.0200	0.00500	mg/L	1		08/15/19 11:59

## Batch Information

Analytical Batch: WDA4629  
 Analytical Method: SM21 4500P-B,E  
 Analyst: DMM  
 Analytical Date/Time: 08/15/19 11:59  
 Container ID: 1194285005-A

Prep Batch: WXX12974  
 Prep Method: SM21 4500P-B,E  
 Prep Date/Time: 08/14/19 21:30  
 Prep Initial Wt./Vol.: 25 mL  
 Prep Extract Vol: 25 mL



## Results of Trip Blank

Client Sample ID: **Trip Blank**  
 Client Project ID: **Kenai River Water Quality**  
 Lab Sample ID: 1194285011  
 Lab Project ID: 1194285

Collection Date: 07/30/19 09:11  
 Received Date: 07/31/19 16:40  
 Matrix: Water (Surface, Eff., Ground)  
 Solids (%):  
 Location:

## Results by Volatile GC/MS

<u>Parameter</u>	<u>Result Qual</u>	<u>LOQ/CL</u>	<u>DL</u>	<u>Units</u>	<u>DF</u>	<u>Allowable Limits</u>	<u>Date Analyzed</u>
Benzene	0.200 U	0.400	0.120	ug/L	1		08/05/19 16:23
Ethylbenzene	0.500 U	1.00	0.310	ug/L	1		08/05/19 16:23
o-Xylene	0.500 U	1.00	0.310	ug/L	1		08/05/19 16:23
P & M -Xylene	1.00 U	2.00	0.620	ug/L	1		08/05/19 16:23
Toluene	0.500 U	1.00	0.310	ug/L	1		08/05/19 16:23
Xylenes (total)	1.50 U	3.00	1.00	ug/L	1		08/05/19 16:23
<b>Surrogates</b>							
1,2-Dichloroethane-D4 (surr)	105	81-118		%	1		08/05/19 16:23
4-Bromofluorobenzene (surr)	98.4	85-114		%	1		08/05/19 16:23
Toluene-d8 (surr)	98.7	89-112		%	1		08/05/19 16:23

## Batch Information

Analytical Batch: VMS19267  
 Analytical Method: EPA 602/624  
 Analyst: CMC  
 Analytical Date/Time: 08/05/19 16:23  
 Container ID: 1194285011-A

Prep Batch: VXX34588  
 Prep Method: SW5030B  
 Prep Date/Time: 08/05/19 06:00  
 Prep Initial Wt./Vol.: 5 mL  
 Prep Extract Vol: 5 mL

## Method Blank

Blank ID: MB for HBN 1797384 [MXX/32633]  
Blank Lab ID: 1523171

Matrix: Water (Surface, Eff., Ground)

QC for Samples:  
1194285001, 1194285002, 1194285003, 1194285004, 1194285005

## Results by EP200.8

<u>Parameter</u>	<u>Results</u>	<u>LOQ/CL</u>	<u>DL</u>	<u>Units</u>
Arsenic	2.50U	5.00	1.50	ug/L
Cadmium	0.250U	0.500	0.150	ug/L
Chromium	1.00U	2.00	0.800	ug/L
Copper	0.500U	1.00	0.310	ug/L
Lead	0.100U	0.200	0.0700	ug/L
Zinc	5.00U	10.0	3.10	ug/L

## Batch Information

Analytical Batch: MMS10583  
Analytical Method: EP200.8  
Instrument: Perkin Elmer Nexlon P5  
Analyst: DSH  
Analytical Date/Time: 8/6/2019 12:42:16PM

Prep Batch: MXX32633  
Prep Method: E200.2  
Prep Date/Time: 8/5/2019 10:22:40AM  
Prep Initial Wt./Vol.: 20 mL  
Prep Extract Vol: 50 mL

Print Date: 08/20/2019 9:46:06AM

## Blank Spike Summary

Blank Spike ID: LCS for HBN 1194285 [MXX32633]

Blank Spike Lab ID: 1523172

Date Analyzed: 08/06/2019 12:45

Matrix: Water (Surface, Eff., Ground)

QC for Samples: 1194285001, 1194285002, 1194285003, 1194285004, 1194285005

## Results by EP200.8

Blank Spike (ug/L)				
Parameter	Spike	Result	Rec (%)	CL
Arsenic	1000	1020	102	( 85-115 )
Cadmium	100	102	102	( 85-115 )
Chromium	400	443	111	( 85-115 )
Copper	1000	1080	108	( 85-115 )
Lead	1000	1050	105	( 85-115 )
Zinc	1000	1100	110	( 85-115 )

## Batch Information

Analytical Batch: **MMS10583**

Analytical Method: **EP200.8**

Instrument: **Perkin Elmer Nexlon P5**

Analyst: **DSH**

Prep Batch: **MXX32633**

Prep Method: **E200.2**

Prep Date/Time: **08/05/2019 10:22**

Spike Init Wt./Vol.: 1000 ug/L Extract Vol: 50 mL

Dupe Init Wt./Vol.: Extract Vol:

## Matrix Spike Summary

Original Sample ID: 1523174  
MS Sample ID: 1523179 MS  
MSD Sample ID:

Analysis Date: 08/05/2019 20:25  
Analysis Date: 08/05/2019 20:28  
Analysis Date:  
Matrix: Water (Surface, Eff., Ground)

QC for Samples: 1194285001, 1194285002

## Results by EP200.8

Parameter	Sample	Matrix Spike (ug/L)			Spike Duplicate (ug/L)			CL	RPD (%)	RPD CL
		Spike	Result	Rec (%)	Spike	Result	Rec (%)			
Arsenic	2.50U	1000	1050	105				70-130		
Cadmium	0.250U	100	105	105				70-130		
Chromium	1.00U	400	428	107				70-130		
Copper	38.2	1000	1090	105				70-130		
Lead	1.40	1000	1110	111				70-130		
Zinc	60.1	1000	1100	104				70-130		

## Batch Information

Analytical Batch: MMS10582  
Analytical Method: EP200.8  
Instrument: Perkin Elmer Nexlon P5  
Analyst: DSH  
Analytical Date/Time: 8/5/2019 8:28:15PM

Prep Batch: MXX32633  
Prep Method: DW Digest for Metals on ICP-MS  
Prep Date/Time: 8/5/2019 10:22:40AM  
Prep Initial Wt./Vol.: 20.00mL  
Prep Extract Vol: 50.00mL

Print Date: 08/20/2019 9:46:09AM

## Matrix Spike Summary

Original Sample ID: 1523180  
MS Sample ID: 1523182 MS  
MSD Sample ID:

Analysis Date: 08/05/2019 21:04  
Analysis Date: 08/05/2019 21:06  
Analysis Date:  
Matrix: Water (Surface, Eff., Ground)

QC for Samples: 1194285001, 1194285002, 1194285003, 1194285004, 1194285005

## Results by EP200.8

Parameter	Sample	Matrix Spike (ug/L)			Spike Duplicate (ug/L)			CL	RPD (%)	RPD CL
		Spike	Result	Rec (%)	Spike	Result	Rec (%)			
Arsenic	25.0	1000	1090	106				70-130		
Cadmium	0.250U	100	104	104				70-130		
Chromium	1.00U	400	433	108				70-130		
Copper	64.3	1000	1110	105				70-130		
Lead	0.953	1000	1110	110				70-130		
Zinc	31.1	1000	1060	103				70-130		

## Batch Information

Analytical Batch: MMS10582  
Analytical Method: EP200.8  
Instrument: Perkin Elmer Nexlon P5  
Analyst: DSH  
Analytical Date/Time: 8/5/2019 9:06:59PM

Prep Batch: MXX32633  
Prep Method: DW Digest for Metals on ICP-MS  
Prep Date/Time: 8/5/2019 10:22:40AM  
Prep Initial Wt./Vol.: 20.00mL  
Prep Extract Vol: 50.00mL

Print Date: 08/20/2019 9:46:09AM

## Method Blank

Blank ID: MB for HBN 1797493 [VXX/34588]  
Blank Lab ID: 1523664

Matrix: Water (Surface, Eff., Ground)

QC for Samples:  
1194285001, 1194285011

## Results by EPA 602/624

<u>Parameter</u>	<u>Results</u>	<u>LOQ/CL</u>	<u>DL</u>	<u>Units</u>
Benzene	0.200U	0.400	0.120	ug/L
Ethylbenzene	0.500U	1.00	0.310	ug/L
o-Xylene	0.500U	1.00	0.310	ug/L
P & M -Xylene	1.00U	2.00	0.620	ug/L
Toluene	0.500U	1.00	0.310	ug/L
Xylenes (total)	1.50U	3.00	1.00	ug/L

## Surrogates

1,2-Dichloroethane-D4 (surr)	106	81-118	%
4-Bromofluorobenzene (surr)	98	85-114	%
Toluene-d8 (surr)	98.9	89-112	%

## Batch Information

Analytical Batch: VMS19267  
Analytical Method: EPA 602/624  
Instrument: VPA 780/5975 GC/MS  
Analyst: CMC  
Analytical Date/Time: 8/5/2019 2:31:00PM

Prep Batch: VXX34588  
Prep Method: SW5030B  
Prep Date/Time: 8/5/2019 6:00:00AM  
Prep Initial Wt./Vol.: 5 mL  
Prep Extract Vol: 5 mL

## Blank Spike Summary

Blank Spike ID: LCS for HBN 1194285 [VXX34588]  
 Blank Spike Lab ID: 1523665  
 Date Analyzed: 08/05/2019 14:46

Spike Duplicate ID: LCSD for HBN 1194285  
 [VXX34588]  
 Spike Duplicate Lab ID: 1523666  
 Matrix: Water (Surface, Eff., Ground)

QC for Samples: 1194285001, 1194285011

## Results by EPA 602/624

Parameter	Blank Spike (ug/L)			Spike Duplicate (ug/L)			CL	RPD (%)	RPD CL
	Spike	Result	Rec (%)	Spike	Result	Rec (%)			
Benzene	30	26.6	89	30	26.2	87	( 79-120 )	1.60	(< 20 )
Ethylbenzene	30	26.8	89	30	26.6	89	( 79-121 )	0.75	(< 20 )
o-Xylene	30	26.6	89	30	26.7	89	( 78-122 )	0.23	(< 20 )
P & M -Xylene	60	53.4	89	60	53.4	89	( 80-121 )	0.02	(< 20 )
Toluene	30	25.7	86	30	25.2	84	( 80-121 )	2.10	(< 20 )
Xylenes (total)	90	80.0	89	90	80.0	89	( 79-121 )	0.06	(< 20 )
<b>Surrogates</b>									
1,2-Dichloroethane-D4 (surr)	30	101	101	30	102	102	( 81-118 )	0.49	
4-Bromofluorobenzene (surr)	30	99	99	30	98.9	99	( 85-114 )	0.10	
Toluene-d8 (surr)	30	98.7	99	30	98.9	99	( 89-112 )	0.24	

## Batch Information

Analytical Batch: VMS19267  
 Analytical Method: EPA 602/624  
 Instrument: VPA 780/5975 GC/MS  
 Analyst: CMC

Prep Batch: VXX34588  
 Prep Method: SW5030B  
 Prep Date/Time: 08/05/2019 06:00  
 Spike Init Wt./Vol.: 30 ug/L Extract Vol: 5 mL  
 Dupe Init Wt./Vol.: 30 ug/L Extract Vol: 5 mL

Print Date: 08/20/2019 9:46:10AM

## Method Blank

Blank ID: MB for HBN 1797709 (WFI/2832)  
Blank Lab ID: 1524804

Matrix: Water (Surface, Eff., Ground)

QC for Samples:  
1194285001, 1194285002, 1194285003, 1194285004, 1194285005

## Results by SM21 4500NO3-F

<u>Parameter</u>	<u>Results</u>	<u>LOQ/CL</u>	<u>DL</u>	<u>Units</u>
Nitrate-N	0.100U	0.200	0.0500	mg/L
Nitrite-N	0.100U	0.200	0.0500	mg/L
Total Nitrate/Nitrite-N	0.100U	0.200	0.0500	mg/L

## Batch Information

Analytical Batch: WFI2832  
Analytical Method: SM21 4500NO3-F  
Instrument: Astoria segmented flow  
Analyst: EWW  
Analytical Date/Time: 8/9/2019 2:28:09PM

Print Date: 08/20/2019 9:46:12AM



## Method Blank

Blank ID: MB for HBN 1797709 (WFI/2832)  
Blank Lab ID: 1524806

Matrix: Water (Surface, Eff., Ground)

QC for Samples:  
1194285001, 1194285002, 1194285003, 1194285004, 1194285005

## Results by SM21 4500NO3-F

<u>Parameter</u>	<u>Results</u>	<u>LOQ/CL</u>	<u>DL</u>	<u>Units</u>
Nitrate-N	0.100U	0.200	0.0500	mg/L
Nitrite-N	0.100U	0.200	0.0500	mg/L
Total Nitrate/Nitrite-N	0.100U	0.200	0.0500	mg/L

## Batch Information

Analytical Batch: WFI2832  
Analytical Method: SM21 4500NO3-F  
Instrument: Astoria segmented flow  
Analyst: EWW  
Analytical Date/Time: 8/9/2019 3:34:39PM

Print Date: 08/20/2019 9:46:12AM

## Method Blank

Blank ID: MB for HBN 1797709 (WFI/2832)

Blank Lab ID: 1524808

QC for Samples:

Matrix: Water (Surface, Eff., Ground)

## Results by SM21 4500NO3-F

<u>Parameter</u>	<u>Results</u>	<u>LOQ/CL</u>	<u>DL</u>	<u>Units</u>
Nitrate-N	0.100U	0.200	0.0500	mg/L
Nitrite-N	0.100U	0.200	0.0500	mg/L
Total Nitrate/Nitrite-N	0.100U	0.200	0.0500	mg/L

## Batch Information

Analytical Batch: WFI2832

Analytical Method: SM21 4500NO3-F

Instrument: Astoria segmented flow

Analyst: EWW

Analytical Date/Time: 8/9/2019 4:20:09PM

Print Date: 08/20/2019 9:46:12AM

## Blank Spike Summary

Blank Spike ID: LCS for HBN 1194285 [WFI2832]

Blank Spike Lab ID: 1524803

Date Analyzed: 08/09/2019 14:26

Matrix: Water (Surface, Eff., Ground)

QC for Samples: 1194285001, 1194285002, 1194285003, 1194285004, 1194285005

## Results by SM21 4500NO3-F

Blank Spike (mg/L)				
Parameter	Spike	Result	Rec (%)	CL
Nitrate-N	2.5	2.52	101	( 70-130 )
Nitrite-N	2.5	2.58	103	( 90-110 )
Total Nitrate/Nitrite-N	5	5.10	102	( 90-110 )

## Batch Information

Analytical Batch: WFI2832

Analytical Method: SM21 4500NO3-F

Instrument: Astoria segmented flow

Analyst: EWW

## Blank Spike Summary

Blank Spike ID: LCS for HBN 1194285 [WFI2832]

Blank Spike Lab ID: 1524805

Date Analyzed: 08/09/2019 15:32

Matrix: Water (Surface, Eff., Ground)

QC for Samples: 1194285001, 1194285002, 1194285003, 1194285004, 1194285005

## Results by SM21 4500NO3-F

Blank Spike (mg/L)

Parameter	Spike	Result	Rec (%)	CL
Nitrate-N	2.5	2.52	101	( 70-130 )
Nitrite-N	2.5	2.67	107	( 90-110 )
Total Nitrate/Nitrite-N	5	5.18	104	( 90-110 )

## Batch Information

Analytical Batch: WFI2832

Analytical Method: SM21 4500NO3-F

Instrument: Astoria segmented flow

Analyst: EWW

Print Date: 08/20/2019 9:46:13AM

## Blank Spike Summary

Blank Spike ID: LCS for HBN 1194285 [WFI2832]

Blank Spike Lab ID: 1524807

Date Analyzed: 08/09/2019 16:18

Matrix: Water (Surface, Eff., Ground)

QC for Samples:

## Results by SM21 4500NO3-F

Blank Spike (mg/L)

Parameter	Spike	Result	Rec (%)	CL
Nitrate-N	2.5	2.47	99	( 70-130 )
Nitrite-N	2.5	2.61	105	( 90-110 )
Total Nitrate/Nitrite-N	5	5.09	102	( 90-110 )

## Batch Information

Analytical Batch: WFI2832

Analytical Method: SM21 4500NO3-F

Instrument: Astoria segmented flow

Analyst: EWW

Print Date: 08/20/2019 9:46:13AM

## Matrix Spike Summary

Original Sample ID: 1194337009  
MS Sample ID: 1524592 MS  
MSD Sample ID: 1524593 MSD

Analysis Date: 08/09/2019 15:48  
Analysis Date: 08/09/2019 15:50  
Analysis Date: 08/09/2019 15:52  
Matrix: Water (Surface, Eff., Ground)

QC for Samples: 1194285001, 1194285002, 1194285003, 1194285004, 1194285005

## Results by SM21 4500NO3-F

Parameter	Sample	Matrix Spike (mg/L)			Spike Duplicate (mg/L)			CL	RPD (%)	RPD CL
		Spike	Result	Rec (%)	Spike	Result	Rec (%)			
Total Nitrate/Nitrite-N	0.140J	5.00	5.51	107	5.00	5.60	109	90-110	1.70	(< 25 )

## Batch Information

Analytical Batch: WFI2832  
Analytical Method: SM21 4500NO3-F  
Instrument: Astoria segmented flow  
Analyst: EWW  
Analytical Date/Time: 8/9/2019 3:50:24PM

Print Date: 08/20/2019 9:46:14AM

## Matrix Spike Summary

Original Sample ID: 1194397001  
MS Sample ID: 1524594 MS  
MSD Sample ID: 1524595 MSD

Analysis Date: 08/09/2019 15:08  
Analysis Date: 08/09/2019 15:10  
Analysis Date: 08/09/2019 15:11  
Matrix: Drinking Water

QC for Samples: 1194285001, 1194285002, 1194285003, 1194285004, 1194285005

## Results by SM21 4500NO3-F

Parameter	Sample	Matrix Spike (mg/L)			Spike Duplicate (mg/L)			CL	RPD (%)	RPD CL
		Spike	Result	Rec (%)	Spike	Result	Rec (%)			
Total Nitrate/Nitrite-N	8.14	20.0	29.3	106	20.0	30.5	112 *	90-110	3.90	(< 25 )

## Batch Information

Analytical Batch: WFI2832  
Analytical Method: SM21 4500NO3-F  
Instrument: Astoria segmented flow  
Analyst: EWW  
Analytical Date/Time: 8/9/2019 3:10:09PM

Print Date: 08/20/2019 9:46:14AM

## Method Blank

Blank ID: MB for HBN 1798080 [WXX/12974]  
Blank Lab ID: 1526063

Matrix: Water (Surface, Eff., Ground)

QC for Samples:  
1194285001, 1194285002, 1194285003, 1194285004, 1194285005

## Results by SM21 4500P-B,E

<u>Parameter</u>	<u>Results</u>	<u>LOQ/CL</u>	<u>DL</u>	<u>Units</u>
Total Phosphorus	0.0100U	0.0200	0.00500	mg/L

## Batch Information

Analytical Batch: WDA4629  
Analytical Method: SM21 4500P-B,E  
Instrument: Discrete Analyzer 2  
Analyst: DMM  
Analytical Date/Time: 8/15/2019 11:44:38AM

Prep Batch: WXX12974  
Prep Method: SM21 4500P-B,E  
Prep Date/Time: 8/14/2019 9:30:00PM  
Prep Initial Wt./Vol.: 25 mL  
Prep Extract Vol: 25 mL

Print Date: 08/20/2019 9:46:15AM



## Blank Spike Summary

Blank Spike ID: LCS for HBN 1194285 [WXX12974]  
 Blank Spike Lab ID: 1526064  
 Date Analyzed: 08/15/2019 11:45

Spike Duplicate ID: LCSD for HBN 1194285 [WXX12974]  
 Spike Duplicate Lab ID: 1526065  
 Matrix: Water (Surface, Eff., Ground)

QC for Samples: 1194285001, 1194285002, 1194285003, 1194285004, 1194285005

## Results by SM21 4500P-B,E

Parameter	Blank Spike (mg/L)			Spike Duplicate (mg/L)			CL	RPD (%)	RPD CL
	Spike	Result	Rec (%)	Spike	Result	Rec (%)			
Total Phosphorus	0.2	0.189	94	0.2	0.170	85	( 75-125 )	10.50	(< 25 )

## Batch Information

Analytical Batch: WDA4629  
 Analytical Method: SM21 4500P-B,E  
 Instrument: Discrete Analyzer 2  
 Analyst: DMM

Prep Batch: WXX12974  
 Prep Method: SM21 4500P-B,E  
 Prep Date/Time: 08/14/2019 21:30  
 Spike Init Wt./Vol.: 0.2 mg/L Extract Vol: 25 mL  
 Dupe Init Wt./Vol.: 0.2 mg/L Extract Vol: 25 mL

Print Date: 08/20/2019 9:46:16AM

## Matrix Spike Summary

Original Sample ID: 1194284001  
MS Sample ID: 1526066 MS  
MSD Sample ID: 1526067 MSD

Analysis Date: 08/15/2019 11:47  
Analysis Date: 08/15/2019 11:48  
Analysis Date: 08/15/2019 11:49  
Matrix: Water (Surface, Eff., Ground)

QC for Samples: 1194285001, 1194285002, 1194285003, 1194285004, 1194285005

## Results by SM21 4500P-B,E

Parameter	Sample	Matrix Spike (mg/L)			Spike Duplicate (mg/L)			CL	RPD (%)	RPD CL
		Spike	Result	Rec (%)	Spike	Result	Rec (%)			
Total Phosphorus	0.0277	0.200	.207	90	0.200	0.210	91	75-125	1.80	(< 25 )

## Batch Information

Analytical Batch: WDA4629  
Analytical Method: SM21 4500P-B,E  
Instrument: Discrete Analyzer 2  
Analyst: DMM  
Analytical Date/Time: 8/15/2019 11:48:32AM

Prep Batch: WXX12974  
Prep Method: Total Phosphorus (W) Ext.  
Prep Date/Time: 8/14/2019 9:30:00PM  
Prep Initial Wt./Vol.: 25.00mL  
Prep Extract Vol: 25.00mL

Print Date: 08/20/2019 9:46:17AM

**1194285**



## Locations Nationwide

Alaska

Maryland

## New Jersey

New York

## North Carolina

## Florida

Profile #362575 CN

[www.us.sqs.com](http://www.us.sqs.com)

[illegible]



## e-Sample Receipt Form

SGS Workorder #:

1194285



1 1 9 4 2 8 5

Review Criteria		Condition (Yes, No, N/A)	Exceptions Noted below	
<b>Chain of Custody / Temperature Requirements</b>			N/A	Exemption permitted if sampler hand carries/delivers.
Were Custody Seals intact? Note # & location	Yes	2 front		
COC accompanied samples?	Yes			
DOD: Were samples received in COC corresponding coolers?	N/A			
N/A **Exemption permitted if chilled & collected <8 hours ago, or for samples where chilling is not required				
Temperature blank compliant* (i.e., 0-6 °C after CF)?	Yes	Cooler ID: 1	@ 5.1 °C	Therm. ID: D58
If samples received without a temperature blank, the "cooler temperature" will be documented instead & "COOLER TEMP" will be noted to the right. "ambient" or "chilled" will be noted if neither is available.		Cooler ID:	@	°C Therm. ID:
		Cooler ID:	@	°C Therm. ID:
		Cooler ID:	@	°C Therm. ID:
		Cooler ID:	@	°C Therm. ID:
*If >6°C, were samples collected <8 hours ago?		N/A		
If <0°C, were sample containers ice free?		N/A		
Note: Identify containers received at non-compliant temperature . Use form FS-0029 if more space is needed.				
<b>Holding Time / Documentation / Sample Condition Requirements</b>		Note: Refer to form F-083 "Sample Guide" for specific holding times.		
Were samples received within holding time?	Yes			
Do samples match COC** (i.e., sample IDs, dates/times collected)?	Yes			
**Note: If times differ <1hr, record details & login per COC.				
***Note: If sample information on containers differs from COC, SGS will default to COC information				
Were analytical requests clear? (i.e., method is specified for analyses with multiple option for analysis (Ex: BTEX, Metals))	Yes			
Were proper containers (type/mass/volume/preservative***) used?		Yes	N/A	***Exemption permitted for metals (e.g. 200.8/6020A).
<b>Volatile / LL-Hg Requirements</b>				
Were Trip Blanks (i.e., VOAs, LL-Hg) in cooler with samples?	N/A			
Were all water VOA vials free of headspace (i.e., bubbles ≤ 6mm)?	N/A			
Were all soil VOAs field extracted with MeOH+BFB?	N/A			
Note to Client: Any "No", answer above indicates non-compliance with standard procedures and may impact data quality.				
Additional notes (if applicable):				

## Sample Containers and Preservatives

<u>Container Id</u>	<u>Preservative</u>	<u>Container Condition</u>	<u>Container Id</u>	<u>Preservative</u>	<u>Container Condition</u>
1194285001-A	H2SO4 to pH < 2	OK			
1194285001-B	HNO3 to pH < 2	OK			
1194285001-C	HCL to pH < 2	OK			
1194285001-D	HCL to pH < 2	OK			
1194285001-E	HCL to pH < 2	OK			
1194285002-A	H2SO4 to pH < 2	OK			
1194285002-B	HNO3 to pH < 2	OK			
1194285003-A	H2SO4 to pH < 2	OK			
1194285003-B	HNO3 to pH < 2	OK			
1194285004-A	H2SO4 to pH < 2	OK			
1194285004-B	HNO3 to pH < 2	OK			
1194285005-A	H2SO4 to pH < 2	OK			
1194285005-B	HNO3 to pH < 2	OK			
1194285006-A	HNO3 to pH < 2	OK			
1194285007-A	HNO3 to pH < 2	OK			
1194285008-A	HNO3 to pH < 2	OK			
1194285009-A	HNO3 to pH < 2	OK			
1194285010-A	HNO3 to pH < 2	OK			
1194285011-A	HCL to pH < 2	OK			
1194285011-B	HCL to pH < 2	OK			
1194285011-C	HCL to pH < 2	OK			

### Container Condition Glossary

Containers for bacteriological, low level mercury and VOA vials are not opened prior to analysis and will be assigned condition code OK unless evidence indicates than an inappropriate container was submitted.

OK - The container was received at an acceptable pH for the analysis requested.

BU - The container was received with headspace greater than 6mm.

DM - The container was received damaged.

FR - The container was received frozen and not usable for Bacteria or BOD analyses.

IC - The container provided for microbiology analysis was not a laboratory-supplied, pre-sterilized container and therefore was not suitable for analysis.

NC- The container provided was not preserved or was under-preserved. The method does not allow for additional preservative added after collection.

PA - The container was received outside of the acceptable pH for the analysis requested. Preservative was added upon receipt and the container is now at the correct pH. See the Sample Receipt Form for details on the amount and lot # of the preservative added.

PH - The container was received outside of the acceptable pH for the analysis requested. Preservative was added upon receipt, but was insufficient to bring the container to the correct pH for the analysis requested. See the Sample Receipt Form for details on the amount and lot # of the preservative added.



August 15, 2019

Service Request No:K1907050

Julie Shumway  
SGS North America, Inc.  
200 West Potter Drive  
Anchorage, AK 99518

**Laboratory Results for: 1194285**

Dear Julie,

Enclosed are the results of the sample(s) submitted to our laboratory August 02, 2019  
For your reference, these analyses have been assigned our service request number **K1907050**.

Analyses were performed according to our laboratory's NELAP-approved quality assurance program. The test results meet requirements of the current NELAP standards, where applicable, and except as noted in the laboratory case narrative provided. For a specific list of NELAP-accredited analytes, refer to the certifications section at [www.alsglobal.com](http://www.alsglobal.com). All results are intended to be considered in their entirety, and ALS Group USA Corp. dba ALS Environmental (ALS) is not responsible for use of less than the complete report. Results apply only to the items submitted to the laboratory for analysis and individual items (samples) analyzed, as listed in the report.

Please contact me if you have any questions. My extension is 3364. You may also contact me via email at [howard.holmes@alsglobal.com](mailto:howard.holmes@alsglobal.com).

Respectfully submitted,

**ALS Group USA, Corp. dba ALS Environmental**

Howard Holmes  
Project Manager

ADDRESS 1317 S. 13th Avenue, Kelso, WA 98626  
PHONE +1 360 577 7222 | FAX +1 360 636 1068  
ALS Group USA, Corp.  
dba ALS Environmental



## Narrative Documents

**ALS Environmental—Kelso Laboratory**  
1317 South 13th Avenue, Kelso, WA 98626  
Phone (360) 577-7222 Fax (360) 425-9096  
[www.alsglobal.com](http://www.alsglobal.com)

**Client:** SGS North America, Inc. (SGS Environmental)  
**Project:** 1194285  
**Sample Matrix:** Water

**Service Request:** K1907050  
**Date Received:** 08/02/2019

### CASE NARRATIVE

All analyses were performed consistent with the quality assurance program of ALS Environmental. This report contains analytical results for samples for the Tier II level requested by the client.

### Sample Receipt:

Five water samples were received for analysis at ALS Environmental on 08/02/2019. Any discrepancies upon initial sample inspection are annotated on the sample receipt and preservation form included within this report. The samples were stored at minimum in accordance with the analytical method requirements.

### Metals:

No significant anomalies were noted with this analysis.

Approved by



Date

08/15/2019



### SAMPLE DETECTION SUMMARY

<b>CLIENT ID: USFWS-Cunningham Park</b>	<b>Lab ID: K1907050-001</b>
---	-----------------------------

Analyte	Results	Flag	MDL	MRL	Units	Method
Calcium	11.6		0.003	0.021	mg/L	200.7
Iron	0.638		0.008	0.021	mg/L	200.7
Magnesium	1.16		0.0004	0.0053	mg/L	200.7

<b>CLIENT ID: USFWS-Beaver Creek</b>	<b>Lab ID: K1907050-002</b>
--------------------------------------	-----------------------------

Analyte	Results	Flag	MDL	MRL	Units	Method
Calcium	21.2		0.003	0.021	mg/L	200.7
Iron	1.89		0.008	0.021	mg/L	200.7
Magnesium	5.03		0.0004	0.0053	mg/L	200.7

<b>CLIENT ID: USFWS-Kenai River</b>	<b>Lab ID: K1907050-003</b>
-------------------------------------	-----------------------------

Analyte	Results	Flag	MDL	MRL	Units	Method
Calcium	11.6		0.003	0.021	mg/L	200.7
Iron	0.580		0.008	0.021	mg/L	200.7
Magnesium	1.13		0.0004	0.0053	mg/L	200.7

<b>CLIENT ID: USFWS-Pillars</b>	<b>Lab ID: K1907050-004</b>
---------------------------------	-----------------------------

Analyte	Results	Flag	MDL	MRL	Units	Method
Calcium	11.6		0.003	0.021	mg/L	200.7
Iron	0.594		0.008	0.021	mg/L	200.7
Magnesium	1.12		0.0004	0.0053	mg/L	200.7

<b>CLIENT ID: USFWS-Poachers Cove</b>	<b>Lab ID: K1907050-005</b>
---------------------------------------	-----------------------------

Analyte	Results	Flag	MDL	MRL	Units	Method
Calcium	11.7		0.003	0.021	mg/L	200.7
Iron	0.588		0.008	0.021	mg/L	200.7
Magnesium	1.13		0.0004	0.0053	mg/L	200.7



## Sample Receipt Information

**ALS Environmental—Kelso Laboratory**  
1317 South 13th Avenue, Kelso, WA 98626  
Phone (360) 577-7222 Fax (360) 425-9096  
[www.alsglobal.com](http://www.alsglobal.com)

SGS North America Inc.  
CHAIN OF CUSTODY RECORD



Locations Nationwide

Alaska Florida  
New Jersey Colorado  
Texas North Carolina  
Virginia Louisiana

www.us.sgs.com

K1907050

CLIENT: SGS North America Inc. - Alaska Division					SGS Reference: <b>ALS Kelso</b>										Page 1 of 1	
CONTACT: Julie Shumway PHONE NO: (907) 562-2343					Additional Comments: All soils report out in dry weight unless											
PROJECT NAME: 1194285 PWSID#: NPDL#:					# C O N T A I N E R S  PRESERVATIVE USED: TYPE C = COMP G = GRAB MI = Multi Incremental Soils	HNO3	HNO3	HNO3								
REPORTS TO: Julie Shumway E-MAIL: Julie.Shumway@sgs.com Env.Alaska.RefLabTeam@sgs.com						Calcium by 200.7	Iron by 200.7	Magnesium by 200.7								
INVOICE TO: SGS - Alaska QUOTE #: P.O. #: 1194285																
RESERVED for lab use	SAMPLE IDENTIFICATION	DATE mm/dd/yy	TIME HHMM	MATRIX/ MATRIX CODE												
	USFWS - Cunningham Park	7/30/2019	09:11:00	Water	1	G =	X	X	X						1194285006	Location ID
	USFWS - Beaver Creek	7/30/2019	09:54:00	Water	1	G =	X	X	X						1194285007	
	USFWS - Kenai River	7/30/2019	10:27:00	Water	1	G =	X	X	X						1194285008	
	USFWS - Pillars	7/30/2019	10:44:00	Water	1	G =	X	X	X						1194285009	
	USFWS - Poachers Cove	7/30/2019	11:04:00	Water	1	G =	X	X	X						1194285010	
Relinquished By: (1)		Date	Time	Received By:		DOD Project?		No		Data Deliverable Requirements:						
		8/2/19	0925			Report to DL (J Flags)?		Yes		Level II						
						Cooler ID:										
						Requested Turnaround Time and-or Special Instructions:										
Relinquished By: (3)		Date	Time	Received By:		Temp Blank °C:										
						Chain of Custody Seal: (Circle)										
Relinquished By: (4)		Date	Time	Received For Laboratory By:		or Ambient [ ] INTACT BROKEN ABSENT										
		8/1/19	0943													

[ X ] 209 W. Potter Drive Anchorage, AK 99518 Tel: (907) 562-2343 Fax: (907) 561-5301  
 [ ] 5500 Business Drive Wilmington, NC 28405 Tel: (910) 350-1903 Fax: (910) 350-1557

[http://www.sgs.com/terms\\_and\\_conditions.htm](http://www.sgs.com/terms_and_conditions.htm)

PC H2

## Cooler Receipt and Preservation Form

Client SGS Service Request K19 07050  
 Received: 8/2/19 Opened: 8/2/19 By: A Unloaded: 8/2/19 By: D

1. Samples were received via? USPS Fed Ex UPS DHL PDX Courier Hand Delivered  
 2. Samples were received in: (circle) Cooler Box Envelope Other NA  
 3. Were custody seals on coolers? NA Y N If yes, how many and where? 2 sides  
 If present, were custody seals intact? Y N If present, were they signed and dated? Y N

Raw Cooler Temp	Corrected Cooler Temp	Raw Temp Blank	Corrected Temp Blank	Corr. Factor	Thermometer ID	Cooler/COC ID	Tracking Number	NA	Filed
3.5	3.7	2.6	2.8	+0.2	371	NA	12 A8619W016833		
							2744		

4. Packing material: Inserts Baggies Bubble Wrap Gel Packs Wet Ice Dry Ice Sleeves  
 5. Were custody papers properly filled out (ink, signed, etc.)? NA Y N  
 6. Were samples received in good condition (temperature, unbroken)? Indicate in the table below. NA Y N  
 If applicable, tissue samples were received: Frozen Partially Thawed Thawed  
 7. Were all sample labels complete (i.e analysis, preservation, etc.)? NA Y N  
 8. Did all sample labels and tags agree with custody papers? Indicate major discrepancies in the table on page 2. NA Y N  
 9. Were appropriate bottles/containers and volumes received for the tests indicated? NA Y N  
 10. Were the pH-preserved bottles (see SMO GEN SOP) received at the appropriate pH? Indicate in the table below NA Y N  
 11. Were VOA vials received without headspace? Indicate in the table below. NA Y N  
 12. Was C12/Res negative? NA Y N

Sample ID on Bottle	Sample ID on COC	Identified by:

Sample ID	Bottle Count Bottle Type	Out of Temp	Head- space	Broke	pH	Reagent	Volume added	Reagent Lot Number	Initials	Time

Notes, Discrepancies, &amp; Resolutions:



## Miscellaneous Forms

**ALS Environmental—Kelso Laboratory**  
1317 South 13th Avenue, Kelso, WA 98626  
Phone (360) 577-7222 Fax (360) 425-9096  
[www.alsglobal.com](http://www.alsglobal.com)

### **Inorganic Data Qualifiers**

- \* The result is an outlier. See case narrative.
- # The control limit criteria is not applicable. See case narrative.
- B The analyte was found in the associated method blank at a level that is significant relative to the sample result as defined by the DOD or NELAC standards.
- E The result is an estimate amount because the value exceeded the instrument calibration range.
- J The result is an estimated value.
- U The analyte was analyzed for, but was not detected ("Non-detect") at or above the MRL/MDL.  
*DOD-QSM 4.2 definition* : Analyte was not detected and is reported as less than the LOD or as defined by the project. The detection limit is adjusted for dilution.
- i The MRL/MDL or LOQ/LOD is elevated due to a matrix interference.
- X See case narrative.
- Q See case narrative. One or more quality control criteria was outside the limits.
- H The holding time for this test is immediately following sample collection. The samples were analyzed as soon as possible after receipt by the laboratory.

### **Metals Data Qualifiers**

- # The control limit criteria is not applicable. See case narrative.
- J The result is an estimated value.
- E The percent difference for the serial dilution was greater than 10%, indicating a possible matrix interference in the sample.
- M The duplicate injection precision was not met.
- N The Matrix Spike sample recovery is not within control limits. See case narrative.
- S The reported value was determined by the Method of Standard Additions (MSA).
- U The analyte was analyzed for, but was not detected ("Non-detect") at or above the MRL/MDL.  
*DOD-QSM 4.2 definition* : Analyte was not detected and is reported as less than the LOD or as defined by the project. The detection limit is adjusted for dilution.
- W The post-digestion spike for furnace AA analysis is out of control limits, while sample absorbance is less than 50% of spike absorbance.
- i The MRL/MDL or LOQ/LOD is elevated due to a matrix interference.
- X See case narrative.
- + The correlation coefficient for the MSA is less than 0.995.
- Q See case narrative. One or more quality control criteria was outside the limits.

### **Organic Data Qualifiers**

- \* The result is an outlier. See case narrative.
- # The control limit criteria is not applicable. See case narrative.
- A A tentatively identified compound, a suspected aldol-condensation product.
- B The analyte was found in the associated method blank at a level that is significant relative to the sample result as defined by the DOD or NELAC standards.
- C The analyte was qualitatively confirmed using GC/MS techniques, pattern recognition, or by comparing to historical data.
- D The reported result is from a dilution.
- E The result is an estimated value.
- J The result is an estimated value.
- N The result is presumptive. The analyte was tentatively identified, but a confirmation analysis was not performed.
- P The GC or HPLC confirmation criteria was exceeded. The relative percent difference is greater than 40% between the two analytical results.
- U The analyte was analyzed for, but was not detected ("Non-detect") at or above the MRL/MDL.  
*DOD-QSM 4.2 definition* : Analyte was not detected and is reported as less than the LOD or as defined by the project. The detection limit is adjusted for dilution.
- i The MRL/MDL or LOQ/LOD is elevated due to a chromatographic interference.
- X See case narrative.
- Q See case narrative. One or more quality control criteria was outside the limits.

### **Additional Petroleum Hydrocarbon Specific Qualifiers**

- F The chromatographic fingerprint of the sample matches the elution pattern of the calibration standard.
- L The chromatographic fingerprint of the sample resembles a petroleum product, but the elution pattern indicates the presence of a greater amount of lighter molecular weight constituents than the calibration standard.
- H The chromatographic fingerprint of the sample resembles a petroleum product, but the elution pattern indicates the presence of a greater amount of heavier molecular weight constituents than the calibration standard.
- O The chromatographic fingerprint of the sample resembles an oil, but does not match the calibration standard.
- Y The chromatographic fingerprint of the sample resembles a petroleum product eluting in approximately the correct carbon range, but the elution pattern does not match the calibration standard.
- Z The chromatographic fingerprint does not resemble a petroleum product.

**ALS Group USA Corp. dba ALS Environmental (ALS) - Kelso**  
**State Certifications, Accreditations, and Licenses**

<b>Agency</b>	<b>Web Site</b>	<b>Number</b>
Alaska DEH	<a href="http://dec.alaska.gov/eh/lab/cs/csapproval.htm">http://dec.alaska.gov/eh/lab/cs/csapproval.htm</a>	UST-040
Arizona DHS	<a href="http://www.azdhs.gov/lab/license/env.htm">http://www.azdhs.gov/lab/license/env.htm</a>	AZ0339
Arkansas - DEQ	<a href="http://www.adeq.state.ar.us/techsvs/labcert.htm">http://www.adeq.state.ar.us/techsvs/labcert.htm</a>	88-0637
California DHS (ELAP)	<a href="http://www.cdph.ca.gov/certlic/labs/Pages/ELAP.aspx">http://www.cdph.ca.gov/certlic/labs/Pages/ELAP.aspx</a>	2795
DOD ELAP	<a href="http://www.denix.osd.mil/edqw/Accreditation/AccreditedLabs.cfm">http://www.denix.osd.mil/edqw/Accreditation/AccreditedLabs.cfm</a>	L16-58-R4
Florida DOH	<a href="http://www.doh.state.fl.us/lab/EnvLabCert/WaterCert.htm">http://www.doh.state.fl.us/lab/EnvLabCert/WaterCert.htm</a>	E87412
Hawaii DOH	<a href="http://health.hawaii.gov/">http://health.hawaii.gov/</a>	-
ISO 17025	<a href="http://www.pjllabs.com/">http://www.pjllabs.com/</a>	L16-57
Louisiana DEQ	<a href="http://www.deq.louisiana.gov/page/la-lab-accreditation">http://www.deq.louisiana.gov/page/la-lab-accreditation</a>	03016
Maine DHS	<a href="http://www.maine.gov/dhhs/">http://www.maine.gov/dhhs/</a>	WA01276
Minnesota DOH	<a href="http://www.health.state.mn.us/accreditation">http://www.health.state.mn.us/accreditation</a>	053-999-457
Nevada DEP	<a href="http://ndep.nv.gov/bsdwlabservice.htm">http://ndep.nv.gov/bsdwlabservice.htm</a>	WA01276
New Jersey DEP	<a href="http://www.nj.gov/dep/enforcement/oqa.html">http://www.nj.gov/dep/enforcement/oqa.html</a>	WA005
New York - DOH	<a href="https://www.wadsworth.org/regulatory/elap">https://www.wadsworth.org/regulatory/elap</a>	12060
North Carolina DEQ	<a href="https://deq.nc.gov/about/divisions/water-resources/water-resources-data/water-sciences-home-page/laboratory-certification-branch/non-field-lab-certification">https://deq.nc.gov/about/divisions/water-resources/water-resources-data/water-sciences-home-page/laboratory-certification-branch/non-field-lab-certification</a>	605
Oklahoma DEQ	<a href="http://www.deq.state.ok.us/CSDnew/labcert.htm">http://www.deq.state.ok.us/CSDnew/labcert.htm</a>	9801
Oregon – DEQ (NELAP)	<a href="http://public.health.oregon.gov/LaboratoryServices/EnvironmentalLaboratoryAccreditation/Pages/index.aspx">http://public.health.oregon.gov/LaboratoryServices/EnvironmentalLaboratoryAccreditation/Pages/index.aspx</a>	WA100010
South Carolina DHEC	<a href="http://www.scdhec.gov/environment/EnvironmentalLabCertification/">http://www.scdhec.gov/environment/EnvironmentalLabCertification/</a>	61002
Texas CEQ	<a href="http://www.tceq.texas.gov/field/qa/env_lab_accreditation.html">http://www.tceq.texas.gov/field/qa/env_lab_accreditation.html</a>	T104704427
Washington DOE	<a href="http://www.ecy.wa.gov/programs/eap/labs/lab-accreditation.html">http://www.ecy.wa.gov/programs/eap/labs/lab-accreditation.html</a>	C544
Wyoming (EPA Region 8)	<a href="https://www.epa.gov/region8-waterops/epa-region-8-certified-drinking-water">https://www.epa.gov/region8-waterops/epa-region-8-certified-drinking-water</a>	-
Kelso Laboratory Website	<a href="http://www.alsglobal.com">www.alsglobal.com</a>	NA

Analyses were performed according to our laboratory's NELAP-approved quality assurance program. A complete listing of specific NELAP-certified analytes, can be found in the certification section at [www.ALSGlobal.com](http://www.ALSGlobal.com) or at the accreditation bodies web site.

Please refer to the certification and/or accreditation body's web site if samples are submitted for compliance purposes. The states highlighted above, require the analysis be listed on the state certification if used for compliance purposes and if the method/analyte is offered by that state.

## Acronyms

ASTM	American Society for Testing and Materials
A2LA	American Association for Laboratory Accreditation
CARB	California Air Resources Board
CAS Number	Chemical Abstract Service registry Number
CFC	Chlorofluorocarbon
CFU	Colony-Forming Unit
DEC	Department of Environmental Conservation
DEQ	Department of Environmental Quality
DHS	Department of Health Services
DOE	Department of Ecology
DOH	Department of Health
EPA	U. S. Environmental Protection Agency
ELAP	Environmental Laboratory Accreditation Program
GC	Gas Chromatography
GC/MS	Gas Chromatography/Mass Spectrometry
LOD	Limit of Detection
LOQ	Limit of Quantitation
LUFT	Leaking Underground Fuel Tank
M	Modified
MCL	Maximum Contaminant Level is the highest permissible concentration of a substance allowed in drinking water as established by the USEPA.
MDL	Method Detection Limit
MPN	Most Probable Number
MRL	Method Reporting Limit
NA	Not Applicable
NC	Not Calculated
NCASI	National Council of the Paper Industry for Air and Stream Improvement
ND	Not Detected
NIOSH	National Institute for Occupational Safety and Health
PQL	Practical Quantitation Limit
RCRA	Resource Conservation and Recovery Act
SIM	Selected Ion Monitoring
TPH	Total Petroleum Hydrocarbons
tr	Trace level is the concentration of an analyte that is less than the PQL but greater than or equal to the MDL.



**ALS Group USA, Corp.**

dba ALS Environmental

## Analyst Summary report

**Client:** SGS North America, Inc. (SGS Environmental)  
**Project:** 1194285/

**Service Request:** K1907050

**Sample Name:** USFWS-Cunningham Park  
**Lab Code:** K1907050-001  
**Sample Matrix:** Water

**Date Collected:** 07/30/19**Date Received:** 08/2/19

**Analysis Method**  
200.7

**Extracted/Digested By**  
YZOOK

**Analyzed By**  
JCHAN

**Sample Name:** USFWS-Beaver Creek  
**Lab Code:** K1907050-002  
**Sample Matrix:** Water

**Date Collected:** 07/30/19**Date Received:** 08/2/19

**Analysis Method**  
200.7

**Extracted/Digested By**  
YZOOK

**Analyzed By**  
JCHAN

**Sample Name:** USFWS-Kenai River  
**Lab Code:** K1907050-003  
**Sample Matrix:** Water

**Date Collected:** 07/30/19**Date Received:** 08/2/19

**Analysis Method**  
200.7

**Extracted/Digested By**  
YZOOK

**Analyzed By**  
JCHAN

**Sample Name:** USFWS-Pillars  
**Lab Code:** K1907050-004  
**Sample Matrix:** Water

**Date Collected:** 07/30/19**Date Received:** 08/2/19

**Analysis Method**  
200.7

**Extracted/Digested By**  
YZOOK

**Analyzed By**  
JCHAN

**Sample Name:** USFWS-Poachers Cove  
**Lab Code:** K1907050-005  
**Sample Matrix:** Water

**Date Collected:** 07/30/19**Date Received:** 08/2/19

**Analysis Method**  
200.7

**Extracted/Digested By**  
YZOOK

**Analyzed By**  
JCHAN



## Sample Results

**ALS Environmental—Kelso Laboratory**  
1317 South 13th Avenue, Kelso, WA 98626  
Phone (360) 577-7222 Fax (360) 425-9096  
[www.alsglobal.com](http://www.alsglobal.com)



## Metals

**ALS Environmental—Kelso Laboratory**  
1317 South 13th Avenue, Kelso, WA 98626  
Phone (360) 577-7222 Fax (360) 425-9096  
[www.alsglobal.com](http://www.alsglobal.com)

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** SGS North America, Inc. (SGS Environmental)  
**Project:** 1194285  
**Sample Matrix:** Water  
**Sample Name:** USFWS-Cunningham Park  
**Lab Code:** K1907050-001

**Service Request:** K1907050  
**Date Collected:** 07/30/19 09:11  
**Date Received:** 08/02/19 09:25  
**Basis:** NA

Total Metals

Analyte Name	Analysis Method	Result	Units	MRL	MDL	Dil.	Date Analyzed	Date Extracted	Q
Calcium	200.7	11.6	mg/L	0.021	0.003	1	08/07/19 11:14	08/06/19	
Iron	200.7	0.638	mg/L	0.021	0.008	1	08/07/19 11:14	08/06/19	
Magnesium	200.7	1.16	mg/L	0.0053	0.0004	1	08/07/19 11:14	08/06/19	

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** SGS North America, Inc. (SGS Environmental)  
**Project:** 1194285  
**Sample Matrix:** Water  
**Sample Name:** USFWS-Beaver Creek  
**Lab Code:** K1907050-002

**Service Request:** K1907050  
**Date Collected:** 07/30/19 09:54  
**Date Received:** 08/02/19 09:25

**Basis:** NA

Total Metals

Analyte Name	Analysis Method	Result	Units	MRL	MDL	Dil.	Date Analyzed	Date Extracted	Q
Calcium	200.7	21.2	mg/L	0.021	0.003	1	08/07/19 11:17	08/06/19	
Iron	200.7	1.89	mg/L	0.021	0.008	1	08/07/19 11:17	08/06/19	
Magnesium	200.7	5.03	mg/L	0.0053	0.0004	1	08/07/19 11:17	08/06/19	

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** SGS North America, Inc. (SGS Environmental)  
**Project:** 1194285  
**Sample Matrix:** Water  
**Sample Name:** USFWS-Kenai River  
**Lab Code:** K1907050-003

**Service Request:** K1907050  
**Date Collected:** 07/30/19 10:27  
**Date Received:** 08/02/19 09:25  
**Basis:** NA

Total Metals

Analyte Name	Analysis Method	Result	Units	MRL	MDL	Dil.	Date Analyzed	Date Extracted	Q
Calcium	200.7	11.6	mg/L	0.021	0.003	1	08/07/19 11:19	08/06/19	
Iron	200.7	0.580	mg/L	0.021	0.008	1	08/07/19 11:19	08/06/19	
Magnesium	200.7	1.13	mg/L	0.0053	0.0004	1	08/07/19 11:19	08/06/19	

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** SGS North America, Inc. (SGS Environmental)  
**Project:** 1194285  
**Sample Matrix:** Water  
**Sample Name:** USFWS-Pillars  
**Lab Code:** K1907050-004

**Service Request:** K1907050  
**Date Collected:** 07/30/19 10:44  
**Date Received:** 08/02/19 09:25

**Basis:** NA

Total Metals

Analyte Name	Analysis Method	Result	Units	MRL	MDL	Dil.	Date Analyzed	Date Extracted	Q
Calcium	200.7	11.6	mg/L	0.021	0.003	1	08/07/19 11:22	08/06/19	
Iron	200.7	0.594	mg/L	0.021	0.008	1	08/07/19 11:22	08/06/19	
Magnesium	200.7	1.12	mg/L	0.0053	0.0004	1	08/07/19 11:22	08/06/19	

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** SGS North America, Inc. (SGS Environmental)  
**Project:** 1194285  
**Sample Matrix:** Water  
**Sample Name:** USFWS-Poachers Cove  
**Lab Code:** K1907050-005

**Service Request:** K1907050  
**Date Collected:** 07/30/19 11:04  
**Date Received:** 08/02/19 09:25  
**Basis:** NA

Total Metals

Analyte Name	Analysis Method	Result	Units	MRL	MDL	Dil.	Date Analyzed	Date Extracted	Q
Calcium	200.7	11.7	mg/L	0.021	0.003	1	08/07/19 11:24	08/06/19	
Iron	200.7	0.588	mg/L	0.021	0.008	1	08/07/19 11:24	08/06/19	
Magnesium	200.7	1.13	mg/L	0.0053	0.0004	1	08/07/19 11:24	08/06/19	





## QC Summary Forms

**ALS Environmental—Kelso Laboratory**  
1317 South 13th Avenue, Kelso, WA 98626  
Phone (360) 577-7222 Fax (360) 425-9096  
[www.alsglobal.com](http://www.alsglobal.com)



## Metals

**ALS Environmental—Kelso Laboratory**  
1317 South 13th Avenue, Kelso, WA 98626  
Phone (360) 577-7222 Fax (360) 425-9096  
[www.alsglobal.com](http://www.alsglobal.com)

ALS Group USA, Corp.  
dba ALS Environmental

Analytical Report

**Client:** SGS North America, Inc. (SGS Environmental)  
**Project:** 1194285  
**Sample Matrix:** Water  
**Sample Name:** Method Blank  
**Lab Code:** KQ1910864-05

**Service Request:** K1907050  
**Date Collected:** NA  
**Date Received:** NA  
**Basis:** NA

Total Metals

Analyte Name	Analysis Method	Result	Units	MRL	MDL	Dil.	Date Analyzed	Date Extracted	Q
Calcium	200.7	ND U	mg/L	0.021	0.003	1	08/07/19 10:45	08/06/19	
Iron	200.7	ND U	mg/L	0.021	0.008	1	08/07/19 10:45	08/06/19	
Magnesium	200.7	ND U	mg/L	0.0053	0.0004	1	08/07/19 10:45	08/06/19	

ALS Group USA, Corp.  
dba ALS Environmental

QA/QC Report

**Client:** SGS North America, Inc. (SGS Environmental)  
**Project:** 1194285  
**Sample Matrix:** Water

**Service Request:** K1907050  
**Date Analyzed:** 08/07/19

**Lab Control Sample Summary**  
**Total Metals**

**Units:**mg/L  
**Basis:**NA

**Lab Control Sample**  
KQ1910864-06

Analyte Name	Analytical Method	Result	Spike Amount	% Rec	% Rec Limits
Calcium	200.7	12.1	12.5	97	85-115
Iron	200.7	2.42	2.50	97	85-115
Magnesium	200.7	12.1	12.5	97	85-115