

#	Ref Point	Tol (%)	CH1 Measured	CH1 Status	CH2 Measured	CH2 Status	Row Result
#1	150 V	1%	150.91 V	PASS	150.65 V	PASS	PASS
#2	112.5 V	1%	111.78 V	PASS	111.95 V	PASS	PASS
#3	75 V	1%	74.48 V	PASS	74.86 V	PASS	PASS
#4	37.5 V	1%	37.18 V	PASS	37.39 V	PASS	PASS

#	Ref Point	Tol (%)	CH1 Measured	CH1 Status	CH2 Measured	CH2 Status	Row Result
#1	100 V	1%	100.12 V	PASS	99.88 V	PASS	PASS
#2	200 V	1%	200.05 V	PASS	199.94 V	PASS	PASS

