Pulsed Electro Magnetic Field (PEMF)

SAFETY

* No car keys, (one-foot) cell phones, hearing aids, pacemakers, credit cards... etc.... can be within 3 feet of a running treatment coil.
* The intensity dials MUST be turned to a soft stop all the way counter-clockwise before plugging anything in + turning on.

Instructions

1. Plug power cord in.
2. Plug in the coil attachment: the left side dial controls the left plug, the Right controls the right. Both sides can be run at the same time + are independent.
3. ~~SLOWLY~~ ~~turn~~ Turn on the GREEN switch. When it lights up you have it running. SLOWLY turn the dial clockwise. The intensity can be adjusted up or down during a treatment.
4. VERY IMPORTANT: At the end of a treatment, ALWAYS turn the dial all the way counter-clockwise to a soft stop.

General Treatment times: 10-15 mn. per body part
More intensity = deeper treatment.

$3/minute for the first 60. $2/mn. after.

LEFT DIAL	RIGHT DIAL
154.1	142.0