

校正証明書
ISO 9847

製造者：		校正証明書番号：	S23082257-EP24-126
製品名：	全天日射計	発行日：	2024/8/26
モデル名：	MS-40S	校正日：	2024/5/27
ISOクラス：	Class C	校正方法：	LM-10
シリアルNo.：	S23082257		
付属品：	ケーブル		

校正時の環境条件

温度	25±3	[°C]
照射強度	1000±50	[W/m ²]
照射強度の安定性	0.3	[%]

校正値

感度定数	12.93	[μV/W·m ⁻²]
不確かさ	0.83	[%] (包含係数 $k = 1.96$)

校正品はISO/IEC9847:2023 屋内校正("ランプから直接照射する装置での校正", type A1) にて、トレーサビリティの取れた標準器と比較校正されています。不確かさは「the Guide to the Expression of Uncertainty in Measurement (GUM)」に基づいて計算されています。

使用標準器	モデル名	シリアルNo.	有効期限
日射計	MS-40	S19122278	2024/7/25
デジタルマルチメータ	34401A	3146A24500	2024/8/21

品質保証部 部長

S-series Setting Report

Issue Date:

26-Aug-24

Factory setting data	
Sensor type	Pyranometer
Model name	MS-40S
Serial number (S*****)	23082257
Manufacturing date	May 28, 2024
Calibration date	May 27, 2024
Analog sensitivity [$\mu\text{V/W/m}^2$]	12.93
Hardware version	7
Firmware version (*.***)	7003

Current output range	Current output
Enable/Disable	Enable (Default)
Output range corresponding to 4-20 [mA].	0 - 1600 [W/m ²]

Current output range	Current output
Enable/Disable	Disable (Default)
Output range corresponding to 0-10 [mA].	0 - 1600 [W/m ²]

Digital Communication	Setting
Modbus RTU	Enable
Address	57
Baudrate	19200
Parity	EVEN
Reg. type	EKO-S