

NDIS High Cost Assistive Technology Assessment Report Bioness L300 Go Functional Electrical Stimulation System

Executive Summary

This report supports funding for the Bioness L300 Go Functional Electrical Stimulation system for a 35-year-old participant with Multiple Sclerosis. The participant completed an in-clinic trial, an 8-week community trial and a 3-week washout period. Objective outcome measures, participant logbooks, participant feedback, carer feedback and neurologist support demonstrated clinically meaningful improvements in mobility, safety, endurance, confidence and participation. Withdrawal of the device resulted in measurable deterioration, confirming the functional benefit of the device.

Participant Background

Diagnosis: Multiple Sclerosis. Primary impairments include right foot drop, right ankle inversion during gait, right knee hyperextension, severe fatigue and reduced community participation.

Clinical Assessment Findings

Right ankle dorsiflexion 0/5, plantarflexion 3/5, eversion 0/5, inversion 1/5. Moderate gastrocnemius and soleus spasticity. Independent indoors with a four-wheeled walker.

Outcome Measures

TUG improved from 42.57 seconds to 33.11 seconds. 6MWT improved from 57m to 90m. DGI improved from 11/24 to 14/24. Several measures declined following washout.

Community Trial Findings

Improved foot clearance, confidence, endurance, community participation and safety. No falls were reported throughout the trial.

Washout Findings

Increased fatigue, stiffness, toe scuffing, tripping, reduced confidence and reduced participation were reported without the device.

Recommendation

Long-term provision of the Bioness L300 Go is strongly recommended under the NDIS as a reasonable and necessary support.