



General Details

Assessor	Samuel Prieto Ayllon
Assessment Date	24/06/2025
Assigned Reviewer	Samuel Prieto Ayllon
Next Review Date	24/06/2026

Signed Off By	Oraib AlKheetanRojas
Signed Off On	19/05/2026
Sign Off Comments	

Operation Assessed	Accoustic Imager - Fluke SV600 - General Risk Assessment
--------------------	---

Associated with specific area

-  Basement 2
-  B2 029 (Robotics Lab)
-  All

Description of work area and/or activity assessed

This RA (Risk Assessment) pertains to the general usage and operation of the Fluke SV600 (PoE powered) accoustic imager. The sensor will be operated mounted on a bench/tripod or as a payload for one of the ground robots, used to capture, detect, and classify accoustic signals for research. The activities can take place in both indoor and outdoor scenarios, with the risks identified in this RA appying to both of them.

Reason for Risk Assessment

To document and control any health, safety, and equipment-damage risks introduced by actively using the SV-600 in active research (both stand-alone and robot-mounted).

Overall Current Risk

Low Risk



Average Number of Persons Affected

2	 Employees
2	 Students
2	 Visitors

Current Rating	Hazard Information	Task Status
Low Risk	 Electricity Contact with or damage to the Power over Ethernet (PoE)-powered lead (less than 57 VDC) could cause a mild electric shock, short-circuit, or spark during installation, relocation, or maintenance of the SV600	0 0 0
	Measures Currently in place to prevent risk of injury	
	PoE supply limited to Safety Extra-Low Voltage (SELV) via 802.3at injector/switch Switch port is de-energised until the unit is fully connected ("power-on last" rule) All cables are properly routed when in operation Visual inspection of connectors and insulation before usage Operator trained in the proper use of the equipment	

Potential Rating	Additional Controls Required
Low Risk	No Remedial Actions Entered If there are any reasonably practicable actions which can be taken to reduce the risk associated with this hazard please use the form above to enter them.