

1 Native phantom images

Planning CT + original/corrected MRI



2 Detect and index grid nodes

DoG peaks + subvoxel localization; no report input



3 Rigid alignment, separately per MRI

91 central nodes; nominal radius ≤ 30 mm



4 Evaluate identical outer nodes

757 held-out nodes; paired protocol distributions

