Technical note on MRU Re-calibration and validation

This note applies to

- MRU types: Compact, Marine, Subsea and Ex
- MRU models: 3000, 6000 and 9000

The Norwegian Subsea MRU calibration certificate is valid for 4 years, and full product specification is guaranteed in this period under normal operating conditions.

Notes:

1. The specified accuracy in roll & pitch is usually met over the MRU's lifetime under normal operating conditions as Norwegian Subsea uses only validated sensors with exceptional stability.

2. The specified heave accuracy of 5 cm / 5% is maintained over the MRU's lifetime.

3. The lifetime static accuracy in roll & pitch, without re-calibration and under normal operating conditions, does not exceed

    - 0.1 degrees for the 3000 model
    - 0.05 degrees for the 6000 model
    - 0.02 degrees for the 9000 model

4. The computed MTBF (mean time between failure) for a NORSUB MRU Marine is 100 000 hours, but actual lifetime depends on use, vibrations, and temperature. An MRU lifetime of 10 years or more under normal continuous operation can be expected.

Fredrik Dukan

CEO