

INNOVATIVE MEDICAL SOLUTIONS INC.

Quality Manual

QM-RM-002: Risk Management for Medical Devices

Document ID: QM-RM-002

Revision: 1.5

Effective Date: February 5, 2024

1. PURPOSE AND SCOPE

This Quality Manual section establishes the risk management process for medical devices in accordance with ISO 14971:2019 and FDA guidance documents.

2. RISK MANAGEMENT PROCESS

2.1 Risk Analysis

Risk analysis shall be conducted systematically to:

- Identify intended use and reasonably foreseeable misuse
- Identify hazards and hazardous situations
- Estimate risk for each hazardous situation

2.2 Risk Evaluation

- Risk acceptability criteria shall be established
- Risks shall be evaluated against acceptance criteria
- Risk-benefit analysis shall be performed where applicable

2.3 Risk Control

Risk control measures shall be implemented in order:

precautions in labeling

2.4 Residual Risk Assessment

- Residual risk shall be evaluated after risk control
- Overall residual risk shall be acceptable
- Benefits shall outweigh residual risks