

ACME MEDICAL DEVICES, INC.
Standard Operating Procedure
SOP-DE-001: Design Controls for Medical Devices

Document ID: SOP-DE-001

Revision: 2.1

Effective Date: January 15, 2024

1. PURPOSE AND SCOPE

This Standard Operating Procedure establishes the design controls process for medical device development in accordance with FDA 21 CFR 820.30 and ISO 13485 requirements.

2. DESIGN AND DEVELOPMENT PLANNING

2.1 Design planning shall be documented and shall identify:

- Design and development activities
- Personnel responsible for each design activity
- Verification and validation requirements
- Design transfer criteria

3. DESIGN INPUTS

3.1 Design inputs shall be documented and shall include:

- Intended use and indications for use
- User needs and user requirements
- Functional requirements and performance requirements
- Safety requirements including risk analysis
- Applicable statutory and regulatory requirements
- Applicable standards and guidance documents

4. DESIGN OUTPUTS

4.1 Design outputs shall be documented and shall:

- Meet the design input requirements
- Contain specifications for purchasing, production, and servicing
- Define acceptance criteria for the device
- Identify characteristics critical to device safety and performance

5. DESIGN REVIEW