

Linearity is a fundamental characteristic of many analytic measurement methods, whereby there is a straight-line relationship between “true” analyte concentrations and measured concentrations. In this context, linearity refers to the relationship between the predicted and observed measurement results and not to the relationship between instrument signal output and analyte concentration. For most assays, this relationship is linear within the AMR.

AMR VERIFICATION

Laboratories are required to verify that the appropriate relationship is maintained over the AMR. Laboratories may verify and use an AMR that is narrower than the range defined by the manufacturer. This may be appropriate when materials available for method validation and/or AMR verification are not available to verify the full range claimed by the manufacturer, or reporting values across the full range defined by the manufacturer is not clinically relevant. For many assays, results beyond the AMR can be reported through dilution or concentration studies (see CBG.12500 & CBG.12600). AMR verification is not required for calculated test results (refer to the Definition of Terms in the All Common Checklist) as long as the individual results contributing to the calculation have AMR verification.

Minimum requirements for AMR verification can be met by using matrix appropriate materials, which include low, mid and high concentration or activity range of the AMR with recovery of results that fall within a defined range of the target value. Records of AMR verification must be available.

CLOSENESS OF SAMPLE CONCENTRATIONS OR ACTIVITIES TO THE UPPER AND LOWER LIMITS OF THE AMR

When verifying the AMR, it is required that materials used are near the upper and lower limits of the AMR. Factors to consider in verifying the AMR are the expected analytic imprecision near the limits, the clinical impact of errors near the limits, and the availability of test specimens near the limits. It may be difficult to obtain specimens with values near the limits for some analytes. In such cases, reasonable procedures should be adopted based on available specimen materials. The closeness of sample concentrations or activities to the upper and lower limits of the AMR are defined at the laboratory director's discretion. The method manufacturer's instructions for verifying the AMR must be followed, when available. The laboratory director must define limits for accepting or rejecting verification tests of the AMR.

CBG.11700 Calibration Procedure

Phase II



The laboratory calibrates each test system as defined and reviews the calibration records for acceptability.

NOTE: Calibration of FDA-cleared/approved methods must be performed following the manufacturer's instructions, at minimum, including the number, type, and concentration of calibration materials, frequency of calibration, and criteria for acceptable performance. Calibration procedures are typically specified in the manufacturer's instructions but may also be established by the laboratory.

The calibration procedure must define the limits of acceptable variation, eg, +/- 20% of the expected value. These limits should be applied to all standard and control samples run after the calibration is performed. The procedure should also specify the actions to be taken if a control or standard sample falls outside the defined range. For some analytes (eg, enzymes) calibration is limited to the product of the reaction, rather than the enzyme concentration or activity itself.

REFERENCES

- 1) Department of Health and Human Services, Centers for Medicare & Medicaid Services. Clinical laboratory improvement amendments of 1988; final rule. *Fed Register*. 1992(Feb 28):7165 [42CFR493.1217]
- 2) Department of Health and Human Services, Centers for Medicare & Medicaid Services. Medicare, Medicaid and CLIA Programs; Laboratory Requirements Relating to Quality Systems and Certain Personnel Qualifications; final rule. *Fed Register*. 2003(Jan 24):3707 [42CFR493.1255]
- 3) Clinical and Laboratory Standards Institute (CLSI). *Evaluation of Matrix Effects*. 4th ed. CLSI document EP14. Clinical and Laboratory Standards Institute, Wayne, PA; 2022.
- 4) Miller WG. Quality control. In: Henry's Clinical Diagnostic and Management by Laboratory Methods, 21st Edition, ed. McPherson RA, Pincus MR. Saunders Elsevier, 2007,p 99-111

CBG.11800 Calibration and Calibration Verification Materials

Phase II