

RD.10.2 The radiology department implements corrective actions accordingly.

Standard Intent:

The safety of diagnostic imaging equipment is ensured through performing the necessary tests that done minimally annual with the corrective actions accordingly, those test are: Automatic Exposure Control (AEC) test, Kvp reproducibility and repeatability, Half Value Layer test, Alignment of collimator and x-ray field, Mean glandular dose test (for mammography).
