

## PHYSICAL FACILITIES

*Utilities should be adequate for the overall workload of the cytogenetics section, and must meet all safety requirements.*

### Inspector Instructions:

OBSERVE



- Temperature and humidity are sufficient

**CYG.61400 Climate Control**

**Phase I**

**Ambient temperature and humidity are maintained within a defined and acceptable range to facilitate optimal chromosome spreading.**

**Evidence of Compliance:**

- ✓ Temperature and humidity records in the slide preparation area