

**Evidence of Compliance:**

- ✓ Records of temperature checks

**REFERENCES**

- 1) Hanzlick RL, et al. Autopsy Facility Design. In: Collins KA, ed. *Autopsy Pathology and Reporting*. 3rd ed. Northfield, IL: College of American Pathologists: 2017; chap 9.

**ANP.32450 Scale/Balance****Phase I**

**A scale and/or balance are provided for reliable weighing of organs.**

*NOTE: If infants or fetuses are autopsied at the institution, accuracy of balances to 1.0 gm for infants and 0.1 gm for fetuses must be verified by periodic calibration.*

**Evidence of Compliance:**

- ✓ Record of scale calibration checks and scale in use is appropriate for the types of cases performed

**REFERENCES**

- 1) Hanzlick RL, et al. Autopsy Facility Design. In: Collins KA, ed. *Autopsy Pathology and Reporting*. 3rd ed. Northfield, IL: College of American Pathologists: 2017; chap 9.
- 2) Tan CD, et al. Autopsy Performance. In: Collins KA, ed. *Autopsy Pathology and Reporting*. 3rd ed. Northfield, IL: College of American Pathologists: 2017; chap 15.
- 3) Conran RM, et al. The Pediatric Autopsy. In: Collins KA, ed. *Autopsy Pathology and Reporting*. 3rd ed. Northfield, IL: College of American Pathologists: 2017; chap 17.

**ANP.32500 Temperature and Ventilation****Phase I**

**Ambient temperature and ventilation control are adequate.**

*NOTE: Airborne infectious agent control requires appropriate ventilation.*

**REFERENCES**

- 1) Hazlett SO. Perspectives in pathology. The newly designed morgue. *Advance/Lab*. 2000;9(1):10-11
- 2) Hanzlick RL, et al. Autopsy Facility Design. In: Collins KA, ed. *Autopsy Pathology and Reporting*. 3rd ed. Northfield, IL: College of American Pathologists: 2017; chap 9.

**ANP.32550 Photographic Equipment****Phase I**

**Photographic equipment is available, convenient, and functional.**

**REFERENCES**

- 1) Belanger AJ, et al. Implementation of a practical digital imaging system for routine gross photography in an autopsy environment. *Arch Pathol Lab Med*. 2000;124:160-165
- 2) Hanzlick RL, et al. Autopsy Facility Design. In: Collins KA, ed. *Autopsy Pathology and Reporting*. 3rd ed. Northfield, IL: College of American Pathologists: 2017; chap 9.
- 3) Oliver WR. Considerations for Gross Autopsy Photography. In: Collins KA, ed. *Autopsy Pathology and Reporting*. 3rd ed. Northfield, IL: College of American Pathologists: 2017; chap 30.
- 4) Schoppe C. Photomicrography. In: Collins KA, ed. *Autopsy Pathology and Reporting*. 3rd ed. Northfield, IL: College of American Pathologists: 2017; chap 31.

## AUTOPSY PERFORMANCE AND RECORDS

### Inspector Instructions:



- Sampling of policies and procedures for autopsy performance, intra- and extra-departmental consultations, reporting, and record retention
- Sampling of records of case review/pre-autopsy discussion
- Specimen collection records (as applicable)
- Sampling of final autopsy reports for completeness and pathology review