FRESENIUS KABI
caring for life

Procedure Report - Aurora

Chicago

PROCEDURE REPORT

Donation ID	0520439880	**Instrument Type**	Aurora
Procedure Type	Plasma Collection, Saline	**Procedure Status**	Procedure Complete
Phlebotomist Start ID	BB18B7	**Region**	Midwest
Phlebotomist End ID	BB18B7	**Site**	Chicago
Operator Start ID		**Collection Start Time**	1/4/2021 6:06 PM
Operator End ID	BB18B7	**Collection End Time**	1/4/2021 7:56 PM

DONOR INFORMATION

Donor ID		**Donor Hemoglobin**	
Donor Gender		**Donor Weight**	
Donor Height			

PROCEDURE DATA

Procedure Key	20210104180652_BC08441_PR	**Blood Sample Taken**	
Donation Setup ID		**Number Of Cycles**	9
Plasma Removed Time	1/4/2021 7:56 PM	**WB Processed**	2379 mL
AC Solution		**Total RBC Loss**	
AC Ratio	6:100	**Saline to Return**	500 mL
AC Used	155 mL	**Saline Used**	459 mL

DONATION TIMES

Type	Start Time	End Time	Duration
Needle Duration	1/4/2021 7:11 PM	1/4/2021 7:54 PM	00:43
Collection	1/4/2021 7:11 PM	1/4/2021 7:53 PM	00:42
Disconnect Time	1/4/2021 7:53 PM	1/4/2021 7:54 PM	00:01
Post Processing	1/4/2021 7:54 PM	1/4/2021 7:56 PM	00:02

SOFT GOODS

No soft goods exist for this procedure

PRODUCTS

Sub Components

Primary	Type	Target Volume	Actual Volume
Yes	Anticoagulated Plasma	825	825

Exceptions

No exceptions exist for this product

DEVICE INFORMATION

Instrument Serial Number	BC08441	**Instrument Name**	104
Admin Configuration	03845A4645745D38	**Software Version**	2.0
Language Pack Version	0.0	**Hardware Controller SW Version**	10
MPU SW Version	2.0.18.109	**Safety Controller SW Version**	1

EVENTS

Operator ID	Date	Description	Event Category	Alert Alarm
	1/4/2021 6:06 PM	Changed: Target Saline Infusion Volume->500	Operator Tasks	
	1/4/2021 6:06 PM	Changed: Target Return Rate->150	Operator Tasks	
	1/4/2021 6:06 PM	Changed: Procedure ID->	Operator Tasks	
	1/4/2021 6:06 PM	Changed: Donation Setup ID->	Operator Tasks	
	1/4/2021 6:06 PM	Changed: Donor ID->	Operator Tasks	
	1/4/2021 6:06 PM	Changed: Cuff Pressure Draw Setting->52.00000	Operator Tasks	
	1/4/2021 6:06 PM	Changed: Target Draw Rate->120	Operator Tasks	
	1/4/2021 6:06 PM	Changed: Squeeze Threshold->-110	Operator Tasks	

FRESENIUS KABI
caring for life

Procedure Report - Aurora

Chicago

Operator ID	Date	Description	Event Category	Alert Alarm
BB18B7	1/4/2021 7:04 PM	Changed: Operator ID->BDBZS	Operator Tasks	
BB18B7	1/4/2021 7:04 PM	Resume From Stand by Selected	Operator Tasks	
BB18B7	1/4/2021 7:07 PM	Install Disposable Set Task Done	Operator Tasks	
BB18B7	1/4/2021 7:08 PM	Connect Solutions Task Done	Operator Tasks	
BB18B7	1/4/2021 7:10 PM	Changed: Procedure ID->0520439880	Operator Tasks	
BB18B7	1/4/2021 7:10 PM	Changed: Target Plasma Collection Volume->825	Operator Tasks	
BB18B7	1/4/2021 7:10 PM	Check Plasma Container Connection Task Done	Operator Tasks	
BB18B7	1/4/2021 7:11 PM	Venipuncture Task Done, Ready to Start	Operator Tasks	
BB18B7	1/4/2021 7:11 PM	Begin Collection	Operator Tasks	
BB18B7	1/4/2021 7:48 PM	Plasma Volume Reached	Instrument Events	
BB18B7	1/4/2021 7:54 PM	Disconnect Donor Task Done	Operator Tasks	
BB18B7	1/4/2021 7:56 PM	Remove Plasma Collection Container Task Done	Operator Tasks	