



GSOP-1405.1 v3.0

Change Control

PURPOSE

This procedure describes the end-to-end process for Change Control (CC), including documentation within electronic Quality Management System (eQMS), governance, and oversight expectations. This procedure provides a consistent, global approach to the management of Changes and is aligned to the requirements set out in [LQS-1400.1 Change Control](#) and [HSE-STD-100 HSE Management System](#).

SCOPE

This procedure applies to all GxP and Health, Safety, and Environmental (HSE) relevant CCs entered within the eQMS.

This procedure applies to activities related to product and data quality, and product safety and efficacy information, for pharmaceuticals and medical devices for which Lilly shares or is responsible from a legal or regulatory level, including but not limited to:

- Associated data and information
- Good Laboratory Practice (GLP)
- Good Quality Practice (GQP)
- Good Promotional Practices (GPP)
- Good Vigilance Practice (GVP)
- Good Clinical Practice (GCP)
- Good Distribution Practice (GDP)
- Good Manufacturing Practice (GMP)
- Regulatory authorization

The execution of a change within the Global Service Management Tool is out of scope of this document.



GSOP-1405.1 v3.0

Change Control

TABLE OF CONTENTS

PROCESS FLOWCHART	3
INSTRUCTIONS	4
1.0 Change Control Governance	4
1.1 Change Control Local Governance	5
1.2 Change Control Global Governance.....	6
2.0 Change Pre-Planning.....	6
3.0 Change Initiation	6
4.0 Change Classification.....	8
5.0 Impact Assessments.....	9
6.0 Risk Assessment.....	11
7.0 Change Plan Definition	12
8.0 Change Plan Review and Approval	14
9.0 Change Implementation.....	15
10.0 Effectiveness Check	19
11.0 Extension Requests.....	19
12.0 Amendment.....	19
TERMS & DEFINITIONS	20
REFERENCED DOCUMENTS.....	20
REVISION HISTORY	21
APPENDIX A: Workflow Driven by Change Classification	22



GSOP-1405.1 v3.0

Change Control

PROCESS FLOWCHART

Change Control (CC) is a structured and risk-based approach used to transition individuals, teams, and sites/functions from a current state to a desired future state.




Figure 1: Change Control Process Overview



GSOP-1405.1 v3.0

Change Control

INSTRUCTIONS

1.0 Change Control Governance

Table 1 lists the roles and responsibilities for users executing Change Control records. Users of eQMS must be suitably trained for the tasks they are performing.

Table 1: Roles and Responsibilities

Change Control Role	Responsibilities
Change Owner	Manages Change Control from identification through completion. This includes documenting the proposed Change, facilitating, and summarizing impact assessment(s), and assigning the actions necessary to implement a Change. NOTE: Change Owner cannot be assigned to any approval roles in the Change Control.
Impact Assessment Owner	Subject Matter Expert (SME) in the respective area, responsible for identifying potential impact of a proposed Change.
External Collaborator	Complete assigned tasks as a user external to Lilly by using an external link allowing temporary access to the assigned task for completion. Tasks within the change control process that can be assigned to and executed by an external collaborator include impact assessments and change actions.
Risk Assessment Owner	Assesses known and potential risks and identifies mitigations and/or the residual risk of a proposed Change.
Change Plan Reviewer(s) (Moderate or Major Changes Only)	Performs a review of a proposed Change to ensure the content is appropriate.
Change Review Board (CRB) (Major Changes Only)	The Change Review Board (CRB) consists of a cross-functional, cross-site, and/or multi-site team established to ensure consistent and robust review of impact, risks, resources, and timelines for, at minimum, major Changes. Changes that impact sites/functions globally should be escalated to a global CRB. Verdicts of the CRB must be captured within the record by a representative with use of the Change Plan Reviewer role.
Quality Plan Approver(s)	Approves the proposed Change, prior to implementation, in the event there is quality impact. Must be fulfilled by a QA representative.
HSE Approver(s)	Approves the proposed Change, prior to and after implementation, in the event there is HSE impact. Must be fulfilled by an HSE representative.



GSOP-1405.1 v3.0

Change Control

Change Control Role	Responsibilities
Additional Approver(s)	Approves the proposed Change prior to implementation. At minimum, it must include a representative of the functional organization(s) impacted, to approve resources and timelines associated with the change. Additional approvers may also include regulatory stakeholders if a regulatory impact is identified.
QA/Batch Approver (Batch Impacting Changes Only)	Verifies completion of batch release impacting change actions. Must be fulfilled by a GMP QA Representative.
Change Action Owner	Performs and documents the completion of change actions created in accordance with the change plan.
Change Action Reviewer(s) (Moderate or Major Changes Only)	Reviews the completion of change actions at the end of the pre-implementation and implementation phases, for moderate and major Changes. This task can be assigned to an SME and is not limited to QA personnel.
Quality Implementation Approver(s)	Reviews the entirety of the executed Change Control to ensure that the actions taken meet the intent of the pre-approved Change Plan. They are responsible for completing final approval of the change. They are also responsible for approving abandonment of a change control, if requested. Must be fulfilled by a QA representative.
EC Owner	Executes assigned effectiveness checks (ECs) and provides results in accordance with the instructions and acceptance criteria provided within the effectiveness check record. Assigns a verdict of the EC based on the results.
EC Verifier	Reviews and approves the execution and associated outcomes of an EC. Must be fulfilled by Quality or HSE (for Quality or HSE ECs, respectively).
Third Party	Complete assigned tasks as a user external to Lilly by using an external link allowing temporary access to the assigned task for completion. Tasks within the change control process that can be assigned to and executed by an external collaborator include impact assessments and change actions.

1.1 Change Control Local Governance

Local sites and areas must apply all requirements from this procedure. It is the functional site and area's responsibility to establish processes for the Change Review Board (CRB), monitor their own changes, and report through the Management Review process.



GSOP-1405.1 v3.0

Change Control

1.2 Change Control Global Governance

The Change Global Process Owner (GPO) is responsible for the Change Control process, associated performance, and the global CRB. The GPO ensures that changes to the eQMS are managed appropriately and manages supporting tasks such as global document reviews/updates. The GPO monitors the overall performance of the Change Control system and escalates to appropriate levels of management when issues or opportunities are identified, and actions need to take place.

HSE retains full governance and ownership of all HSE-Only records, including ensuring compliance and timely execution across all HSE records.

2.0 Change Pre-Planning

Prior to initiating a planned proposed GxP change, a Change Owner must refer to [GSOP-1405.1-TL.1 Change Pre-Planning](#) to determine if a Change Control is necessary and in which system it must be initiated.

Changes originating from the Supplier Change Notification (SCN) process are managed in accordance with [GSOP-1301.1-TL.11 Supplier Change Notification](#).

3.0 Change Initiation

When initiating a Change Control, it is the responsibility of the Change Owner to include information that adequately describes the proposed Change in accordance with the steps below.

Responsible Role for this Section: Change Owner

Step	Action
3.1	Initiate a CC documenting the "What" and "Why" behind the proposed Change inclusive of the following details: • Change Title: Brief descriptive title for the change • Description: What is changing, Current state vs. future state • Justification: Why the Change should proceed, inclusive of the benefits of the change • Source: Select the source of the change (known driver behind the change) • Area(s) of Impact: Indicate the type of impact associated with this change • Due Date: The date when all change actions are complete and approval of the implementation summary by Quality Implementation Approver (or HSE Approver - for HSE only changes) has been granted.



GSOP-1405.1 v3.0

Change Control

Step	Action	
3.2	Within the CC, document if any of the following scenarios are applicable: - Emergency Change – A change proposal which if not immediately implemented may have: - Adverse impact to Quality of a batch or utility - Interruption, discontinuation, or restriction of commercial product supply that impacts the patient or end user - Adverse impact to employee health, safety, or the environment Emergency Changes allow for the Change to be performed prior to receipt of formal Change Plan Approval within eQMS. - Temporary Change – A CC that will not become a permanent practice, with a defined time period and scope of use. These changes must have actions to both implement and reverse the change, as applicable. If a Temporary Change becomes a permanent Change, a new Change Control must be initiated referring to the temporary change. - Parent/Child Change – Document if the Change Control will necessitate a hierarchical (parent-child) relationship to other Change Controls. This may often be used for more complex global, multi-site, or multi-function changes. Refer to GSOP-1405.1-TL.9 Parent-Child Change Controls for instructions. - Batch Release Impacting Change – If aspects of the Change must be completed prior to performing batch release, refer to GSOP-1405.1-TL.7 Batch Impacting Changes for instructions.	
3.3	IF	THEN
	Change Control is required to be initiated as an Emergency Change Control	Document the following details as soon as possible after receiving a verbal agreement: - Description and justification for the use of an emergency change - A preliminary risk assessment - Description of emergency steps implemented, and date performed - Quality and Business representative (minimum of an M3, or equivalent, for each role) who provided agreement to proceed with emergency change. <i>For emergency changes with HSE impact:</i> HSE who provided agreement to proceed (minimum site HSE leader, or equivalent representative). During the change plan approval, include the HSE representative as a "Business" representative. Once agreement is obtained, the Change may be implemented. Include individual(s) who provided agreement as Change Plan approvers and perform remaining steps within this procedure.
	Else	Proceed to step 3.4.



GSOP-1405.1 v3.0

Change Control

Step	Action	
3.4	Document the following details pertaining to scope, as applicable: • Related products • Related materials or batches • Related assets (example, equipment, computer systems) • Related records (example, Change Controls, CAPAs, regulatory events) • Impacted organizations, regions, or third parties • Impacted processes or documents • Impacted studies • Other impact(s)	
3.5	IF	THEN
	Change Control is identified as batch impacting	At least one batch and/or material must be associated to the CC. When identified, batch and/or material information will be made available to batch disposition personnel. Information will also be available to batch disposition personnel once the batch impacting Change actions are completed. Refer to GSOP-1405.1-TL.7 Batch Impacting Changes for further guidance.
	Else	Proceed to step 4.0 .

4.0 Change Classification

Change classification is determined via a structured process that supports risk-based decision-making, evaluated through completion of the Change Classification checklist.

Based on results of the checklist, Changes may be classified as Minor, Moderate, or Major which will drive workflow requirements in accordance with **Appendix A: Workflow Driven by Change Classification**.

Within eQMS, there is a Quality and HSE classification. HSE classification is applicable when the Change impacts health, safety, or environmental aspects. In all other cases, a Quality Classification is required. When there is impact to both Quality and HSE, both classifications are independently determined and the highest of the two will drive workflow, with Quality involvement, in accordance with **Appendix A: Workflow Driven by Change Classification**.

Refer to [GSOP-1405.1-TL.2 Quality Change Classification](#) and [GSOP-1405.1-TL.3 HSE Change Classification](#) for more information on the scoring matrix and criteria used to combine severity and complexity ratings into the final classification, along with the detailed questions and guidance.



GSOP-1405.1 v3.0 Change Control

Responsible Role for this Section: Change Owner

Step	Action								
4.1	Consult with Quality and/or HSE, as needed, and complete the Change Classification checklist to assign a Quality and/or HSE classification in accordance with the table below: <table> <tr> <th>IF</th><th>THEN</th></tr> <tr> <td>Quality Impact Only</td><td>Perform classification in accordance with GSOP-1405.1-TL.2 Quality Change Classification.</td></tr> <tr> <td>HSE Impact Only</td><td>Perform classification in accordance with GSOP-1405.1-TL.3 HSE Change Classification.</td></tr> <tr> <td>Quality and HSE Impact</td><td>Classify in accordance with both supplements above. The highest resulting classification will drive the workflow and corresponding approval levels.</td></tr> </table>	IF	THEN	Quality Impact Only	Perform classification in accordance with GSOP-1405.1-TL.2 Quality Change Classification .	HSE Impact Only	Perform classification in accordance with GSOP-1405.1-TL.3 HSE Change Classification .	Quality and HSE Impact	Classify in accordance with both supplements above. The highest resulting classification will drive the workflow and corresponding approval levels.
IF	THEN								
Quality Impact Only	Perform classification in accordance with GSOP-1405.1-TL.2 Quality Change Classification .								
HSE Impact Only	Perform classification in accordance with GSOP-1405.1-TL.3 HSE Change Classification .								
Quality and HSE Impact	Classify in accordance with both supplements above. The highest resulting classification will drive the workflow and corresponding approval levels.								
4.2	Document rationale for the resulting Quality and/or HSE classification including support for answers to questions within the classification checklist.								

5.0 Impact Assessments

Step	Action	Responsible
5.1	Impact assessments must consider evaluating impact across the following areas, at minimum: • Data integrity • Product quality • Pharmaceutical quality systems • Documentation • Validation • Regulatory/legal compliance • Patient or human research subject safety • Quality and integrity of safety/efficacy information • Calibration • Maintenance • Computer Systems • Quality Control • Any other ongoing changes, of similar scope	Change Owner



GSOP-1405.1 v3.0

Change Control

Step	Action		Responsible
	IF	THEN	
	Moderate or Major Change	Create and assign Impact Assessments to appropriate SME(s) based on CC details and scope. Consider use of Impact Assessment Templates to guide the direction for Impact Assessment Owner.	
	Minor Change	Only an Impact Assessment Summary is required. This is to be performed by the Change Owner independently or on behalf of a group of functional representatives (for example, SMEs, review panel, cross-functional team). Once completed, proceed to Section 7.0 Change Plan Definition	
	Area of Impact includes GMP/ Manufacturing Site	Refer to GSOP-1405.1-TL.10 Proposed Changes Requiring Regulatory Affairs Evaluation to confirm and document whether a CMC Regulatory Review is required. If required, relate and assign an Impact Assessment to Regulatory.	
	Impact Assessments are to be performed by a Supplier	Assign impact assessments in accordance with GSOP-1301.1-TL.11 Supplier Change Notification .	
5.2	Perform impact assessments in accordance with GSOP-1405.1-TL.4 Change Impact Assessment . An Impact Assessment Owner may perform this assessment independently or on behalf of a group of functional representatives (for example, SMEs, review panel, cross-functional team). If a Change Description or Change Justification changes during or after Impact Assessments have been completed, prior to Change Plan approval, the Change Owner is responsible for ensuring that Impact Assessment Owners are informed and can re-perform their assessments, as needed.		Impact Assessment Owner
5.3	Utilize the AI Impact Assessment Summary Tool to document a summary inclusive of: • How many impact assessments were completed • Types of impact assessments completed		Change Owner



GSOP-1405.1 v3.0 Change Control

Step	Action		Responsible
	Summary of identified impacts to any of the details within the scope section Review and edit (as needed) the summary generated by the tool for accuracy or to address any cases where there is misalignment between impact assessments. Once sufficient, acknowledge the results to return the summary to the Change record. NOTE: Change Owner may manually create a summary inclusive of the details above in lieu of tool usage.		
5.4	IF	THEN	Change Owner
	CMC Regulatory Impact Assessment is performed	Document outcome of the CMC Regulatory Impact Assessment within the Regulatory Results. If <i>Do Then Tell</i> or <i>Tell/Wait/Do</i> , relate the associated RIM Event to the Change Control and include CMC Regulatory representative as an Additional Approver, prior to routing the CC for review and approval.	
	Else	Proceed to section 6.0 .	

6.0 Risk Assessment

The purpose of a Risk Assessment is to identify and evaluate known and potential risks associated with a proposed change, focusing on the likelihood and severity of adverse effects. For Moderate and Major CCs, Risk Assessments are required to be assigned. For Minor CCs, additional Risk Assessments are not needed beyond the understood minimal risk identified in the Change Classification checklist.

Step	Action	Responsible
6.1	Create and assign Risk Assessments to the appropriate Risk Assessment Owner(s) based on the CC details, scope, classification, and impact(s) identified during impact assessment. Refer to GSOP-1405.1-TL.5 Risk Assessment in Change Control, for guidance on the required level of formality. NOTE: A relation to a pre-existing risk evaluation may be utilized in lieu of creating a new one, if it appropriately encompasses the risk associated with the CC.	Change Owner
6.2	Perform Risk Assessments in accordance with GSOP-1405.1-TL.5 Risk Assessment in Change Control .	Risk Assessment Owner



GSOP-1405.1 v3.0

Change Control

7.0 Change Plan Definition

Responsible Role for this Section: Change Owner

Step	Action
7.1	Once all impact and risk assessments have been completed, review the classification to ensure it still accurately reflects the appropriate severity and complexity associated with the CC.
7.2	Document the change plan inclusive of the following details for the implementation of the change: • Summary of the objectives, milestones, and general timelines including dependencies • High-level summary of phased implementations, including actions to mitigate risk associated with the implementation of the change • The strategy behind the use of Batch Impacting Change actions in accordance with GSOP-1405.1-TL7 Batch Impacting Changes, if applicable • Scope and systems involved in execution • Functions or resources needed
7.3	Create and assign Change Actions required to implement the Change Plan to any of the below action step phases: • Pre-Implementation • Implementation • Post-Implementation When creating change actions, ensure that change actions: • Describe the action to be performed • Contain specific criteria in which success will be measured against • Are identified as batch impacting, if applicable, in accordance with GSOP-1405.1-TL7 Batch Impacting Changes. If the below scenarios are applicable, ensure the following: • For batch impacting CCs, at least one batch impacting change action is created. • For temporary CCs, actions necessary for both the implementation of the change and the return to the original state are included Consider and utilize applicable change action templates, as needed. Refer to GSOP-1405.1-RS.1 Change Action Phasing for additional details. Refer to GSOP-1301.1-TL.11 Supplier Change Notification for assignment of change actions to suppliers.



GSOP-1405.1 v3.0

Change Control

Step	Action	
	NOTE: - All CCs require at least one change action, and it can be any step phase. The use of all action step phases is not mandatory. - The use of the 'Other' action step phase is to be used during the Change Update process, if needed. - Change action due dates are targets to support timely completion of the CC.	
7.4	IF	THEN
	Moderate OR Major Change	Create an EC and document an EC plan in accordance with GSOP-1405.1-TL.11 Effectiveness Check to: - Verify implementation of the CC meets its predefined goals. - Detect any issues or unintended consequences such as the realization of the most prevalent residual risks, adverse impacts, or deviations related to the CC. - Ensure the change is sustainable over the monitoring period of the effectiveness check, as applicable. NOTE: - The performance of an effectiveness check on a Moderate or Major parent change control may be used in lieu of distinct effectiveness checks on child change controls if the measurement of effectiveness is mutually applicable to both the parent and child change control. - For Minor changes, the Implementation Summary serves as the evaluation of effectiveness.
	Else	Proceed to step 7.5 .
7.5	Route the CC for review and approval. Approval, at minimum, consists of the following functions: - An approver from Quality and/or HSE based on the area of classification. If no impact to one of Quality or HSE, an approver is not required from that respective area. - An additional approver representing the impacted site/function(s) to approve resources and timelines associated with the change Additional approvers from impacted functions, including Regulatory, may be added if identified during the impact assessment process. The table below reflects the minimum levels of reviewers and approvers based on change classification:	



GSOP-1405.1 v3.0 Change Control

Step	Action		
	Change Classification	Change Plan Reviewer	Approvers (Quality and/or HSE, or Additional Approvers)
	Minor	N/A	Quality Approver: Quality Unit Representative HSE Approver: HSE Unit Representative Additional Approver(s): Business/Organization Unit Representative
	Moderate	SME(s)	M2/P4/R4, or higher
	Major	Change Review Board (Identify CRB representative)	M3/P5/R6, or higher

8.0 Change Plan Review and Approval

Step	Action		Responsible
8.1	IF	THEN	Change Control Reviewer (Moderate)
	Moderate Change	Review the CC to confirm the following: - CC properly documents description, justification, and classification - Accuracy and sufficient detail within the scope (example, batch numbers) - Impacts and risks have been accurately assessed and mitigated within the change plan, as needed - Change actions are adequately defined, inclusive of batch impacting designation - Effectiveness check criteria are appropriate and measurable for the nature of the change	
	Else	Proceed to step 8.2 .	
8.2	IF	THEN	Change Review Board (CRB)
	Major Change	Major changes are required to be presented and approved by the Change Review Board (CRB). Documentation of the CRB review is performed by a CRB member in the Review step of the Major change. CRB reviews the CC to confirm the following:	



GSOP-1405.1 v3.0

Change Control

Step	Action	Responsible
	- CC properly documents description, justification, and classification - Accuracy and sufficient detail within the scope (example, batch numbers) - Impacts and risks have been accurately assessed and mitigated within the change plan, as needed - Change actions are adequately defined, inclusive of batch impacting designation - Effectiveness check criteria are appropriate and measurable for the nature of the change	
8.3	Approval signifies that the Approvers have reviewed the record for accuracy and completeness, with respect to the Approver role and that there is an agreement that the risk(s) have been appropriately identified, mitigated and the Change is authorized for implementation. NOTE: Any modification of the Change Plan must be approved by all original Approvers, or equivalent. Modifications to implementation plan actions or success criteria must have appropriate quality oversight.	Quality Plan Approver/ Additional Approver/ HSE Approver

9.0 Change Implementation

Step	Action	Responsible						
9.1	Perform and document evidence of actions taken in accordance with the completion criteria established within the change action(s).	Change Action Owner						
	For situations in which execution varies from the completion criteria or Change Action instructions, Change Action Owner performs as follows:							
	<table><tr><th>IF</th><th>THEN</th></tr><tr><td>The intent of the Change Action is still met</td><td>Document the variation within the change.</td></tr><tr><td>The intent of the Change Action cannot be met</td><td>Contact the Change Owner and Quality and/or HSE to discuss, and document justification prior to progressing the record forward.</td></tr></table>		IF	THEN	The intent of the Change Action is still met	Document the variation within the change.	The intent of the Change Action cannot be met	Contact the Change Owner and Quality and/or HSE to discuss, and document justification prior to progressing the record forward.
	IF		THEN					
The intent of the Change Action is still met	Document the variation within the change.							
The intent of the Change Action cannot be met	Contact the Change Owner and Quality and/or HSE to discuss, and document justification prior to progressing the record forward.							
9.2	<table><tr><th>IF</th><th>THEN</th></tr><tr><td>Change is: Moderate or Major </td><td>Review all executed Change Actions within the phase for evidence that they have been completed in accordance with the completion</td></tr></table>	IF	THEN	Change is: Moderate or Major	Review all executed Change Actions within the phase for evidence that they have been completed in accordance with the completion	Change Action Reviewer		
	IF	THEN						
Change is: Moderate or Major	Review all executed Change Actions within the phase for evidence that they have been completed in accordance with the completion							



GSOP-1405.1 v3.0

Change Control

Step	Action		Responsible
	Changes, and in Pre-Implementation or Implementation phases	criteria, or sufficient rationale is documented for how the intent of the Change Action was met. If not, return the record with comments for edits or actions required for the change actions to obtain approval. NOTE: Actions in the prior phase must be reviewed before actions in the following phase can begin. Refer to GSOP-1405.1-RS.1 Change Action Phasing .	
	Else	Proceed to step 9.3 .	
9.3	IF	THEN	QA Batch Approver
	Change Action is Batch Impacting	Verify Batch Impacting Change Actions in accordance with GSOP-1405.1-TL.7 Batch Impacting Changes .	
	Else	Proceed to step 9.4 .	
9.4	IF	THEN	Change Action Owner
	Change Action is overdue	Modify the change action due date to the date in which it is expected to be completed. If a Change Action due date is required to be modified beyond the CC due date, an extension request will be required on the CC. Contact the Change Owner to have them perform an Extension Request in accordance with GSOP-1405.1-TL.6, Extension Requests .	
	Else	Proceed to step 9.5 .	



GSOP-1405.1 v3.0

Change Control

Step	Action		Responsible
9.5	IF	THEN	Change Owner
	- The scope of the CC has altered, OR - Any actions are required to be added or updated	Utilize Change Updates for the following: - Correct and clarify the scope of a Change Plan or action - Revise the extent of actions to reflect a Change in available resources Change Updates are not to be utilized to modify the intent of the Change Plan. In those cases, abandonment and creation of a new Change Control may be required. Identify that a Change Update is required and perform the following: - Review the Change Plan and update necessary sections. - Add or update any incomplete Change Actions. For each update, include the rationale behind the change. - Once the update is complete, provide a holistic description of the Change Update with justification. - Route changes back to original (or equivalent) Change Plan Approvers. Approver(s) must perform the same level of approval as initially required per section 9.0 step 2. NOTE: - Once a Change Update is initiated, Change Actions cannot be progressed until obtaining Change Update Approval. - Once a given implementation phase has been initiated or completed, actions can no longer be added within that phase. In those cases, utilize 'Other' actions for creation of any additional actions.	
	Else	Proceed to step 9.6 .	



GSOP-1405.1 v3.0

Change Control

Step	Action		Responsible
9.6	IF	THEN	Change Owner
	CC can and/or should no longer be implemented or changes to the change plan significantly alter the original impact and/or risk assessments	Abandon the CC by performing the following: - Using Change Update, create and perform new Change actions to undo completed Change actions within the CC record, as necessary when the completed actions impact the original state. - Cancel all outstanding Change actions Once complete, document the rationale for not moving forward with the change and route for approval of the abandonment of the CC. Note: Abandonment must only be performed prior to the completion of all Change actions. It is not to be used in the event of failed implementations.	
	Else	Proceed to step 9.7 .	
9.7	Once CC implementation has been completed, utilize the AI Implementation Summary Tool to generate a summary inclusive of: - A statement as to whether implementation progressed as planned - Whether actions taken were performed in accordance with the description and completion criteria of Change Actions. If not, document any recorded justifications for the departure - A holistic statement as to whether the implementation of all Change Actions met the intent of the Change Plan. If not, document any recorded justification for the departure Review and edit (as needed) the summary generated by the tool. In the event of misalignment with the change plan, next steps (example, new CC, deviation, etc.) to be taken within summary. Once sufficient, acknowledge the results to return the summary to the CC record. In addition to the implementation summary, record the actual implementation date describing when the CC was operationally effective. Once complete, route the CC for approval. NOTE: Change Owner may manually create a summary inclusive of the details above in lieu of tool usage.		Change Owner



GSOP-1405.1 v3.0

Change Control

Step	Action	Responsible								
9.8	Review and approve the CC if the following criteria are met: - Implementation summary is accurate - Completed change actions and implementation summary meet the objectives of the change plan In the event there is misalignment between the implementation, consider the following options:	Quality Implementation Approver/ HSE Approver								
	<table><tr><th>IF</th><th>THEN</th></tr><tr><td>Documentation within the implementation summary is insufficient</td><td>Reject the CC to the Change Owner for corrections</td></tr><tr><td>Documentation of completed change actions are insufficient</td><td>Return the record to implementation to ensure appropriate details are captured</td></tr><tr><td>Implementation was not performed in accordance with the intent of the Change Plan</td><td>Ensure appropriate next steps are initiated (example, relate new CC, deviation, etc.) prior to approving the CC</td></tr></table>		IF	THEN	Documentation within the implementation summary is insufficient	Reject the CC to the Change Owner for corrections	Documentation of completed change actions are insufficient	Return the record to implementation to ensure appropriate details are captured	Implementation was not performed in accordance with the intent of the Change Plan	Ensure appropriate next steps are initiated (example, relate new CC, deviation, etc.) prior to approving the CC
	IF		THEN							
	Documentation within the implementation summary is insufficient		Reject the CC to the Change Owner for corrections							
	Documentation of completed change actions are insufficient		Return the record to implementation to ensure appropriate details are captured							
Implementation was not performed in accordance with the intent of the Change Plan	Ensure appropriate next steps are initiated (example, relate new CC, deviation, etc.) prior to approving the CC									

10.0 Effectiveness Check

For Moderate and/or Major changes, perform effectiveness checks in accordance with [GSOP-1405.1-TL.11 Effectiveness Check](#).

11.0 Extension Requests

Overdue change controls must be extended prior to routing for final approval. Extension requests are to be initiated by the Change Owner as soon as it is known that the change control will not be completed by its assigned due date in accordance with [GSOP-1405.1-TL.6 Extension Request](#).

12.0 Amendment

The amendment process is only to be utilized to add supporting information, not impacting original scope or intent, to a closed CC. If an amendment is required, request initiation of the amendment process in accordance with the steps below.

Step	Action	Responsible
12.1	Contact and provide rationale to QA to initiate the Amendment process, regardless of Quality/HSE change classification.	Change Owner



GSOP-1405.1 v3.0 Change Control

Step	Action	Responsible						
12.2	Ensure the proposed amendment: • has sufficient rationale to explain the need for additional supporting information • will not alter the original scope or intent of the CC If so, with the use of the Quality Implementation Approver role, initiate the amendment workflow.	Quality Implementation Approver						
12.3	Add the amended information. Once complete, route the Amendment to the original (or equivalent) Change Plan Approvers for approval.	Change Owner						
12.4	Evaluate the completed amendment did not alter the original scope or intent of the CC, and proceed to the verdict: <table><tr><th>IF</th><th>THEN</th></tr><tr><td>Acceptable</td><td>Approve the amendment. Upon approval, the Change Control record will be closed.</td></tr><tr><td>Not Acceptable</td><td>Reject the amendment, inclusive of rationale. The record will be returned to the amendment Owner for revision.</td></tr></table>	IF	THEN	Acceptable	Approve the amendment. Upon approval, the Change Control record will be closed.	Not Acceptable	Reject the amendment, inclusive of rationale. The record will be returned to the amendment Owner for revision.	Quality Implementation Approver
IF	THEN							
Acceptable	Approve the amendment. Upon approval, the Change Control record will be closed.							
Not Acceptable	Reject the amendment, inclusive of rationale. The record will be returned to the amendment Owner for revision.							

TERMS & DEFINITIONS

The following are terms and definitions used within this document.

Term	Definition
N/A	N/A

REFERENCED DOCUMENTS

Document Number	Description
GSOP-1405.1-TL.1	Change Pre-Planning
GSOP-1405.1-TL.2	Quality Change Classification Supplement
GSOP-1405.1-TL.3	HSE Change Classification Supplement
GSOP-1405.1-TL.4	Change Impact Assessment



GSOP-1405.1 v3.0 Change Control

Document Number	Description
GSOP-1405.1-TL.5	Risk Assessment
GSOP-1405.1-TL.6	Extension Requests
GSOP-1405.1-TL.7	Batch Impacting Changes
GSOP-1405.1-TL.9	Parent-Child Change Controls
GSOP-1405.1-TL.10	Proposed Changes Requiring Regulatory Affairs Evaluation
GSOP-1405.1-TL.11	Effectiveness Check
GSOP-1405.1-RS.1	Change Action Phasing
GSOP-1301.1-TL.11	Supplier Change Notification
HSE-STD-100	HSE Management System
LQS-1400.1	Change Control

REVISION HISTORY

List of Changes
Version 3: Administrative change. Updated document number for Supplier Change Notification.
Version 2: Updated Section 3.1 to add a description for Due Date that includes "Due Date: The date when all change actions are complete and approval of the implementation summary by Quality Implementation Approver (or HSE Approver - for HSE only changes) has been granted." Updated Section 5.4 from "Do/Wait/Tell" to "Do Then Tell" to align with system terminology. Updated Section 7.5 to change the Quality/HSE and Additional Approver(s) requirement from "M1/P3/R3, or higher" to Quality Approver: Quality Unit Representative, HSE Approver: HSE Unit Representative, Additional Approver(s): Business/ Organization Unit Representative. Updated section 11.0 from "Overdue change controls must be extended prior to routing for final M2." To "Overdue change controls must be extended prior to routing for final approval.
Version 1: New document developed as part of QMS I3 to create Process based QMS.



GSOP-1405.1 v3.0

Change Control

APPENDIX A: WORKFLOW DRIVEN BY CHANGE CLASSIFICATION

The classification level assigned to a change determines the depth and structure of the workflow that follows. This ensures that the level of control applied to a change is commensurate with its risk and complexity, avoiding unnecessary burden for low-risk changes while maintaining robust oversight for high-impact ones.

The main workflow distinction lies in how Minor changes are handled. These changes are intentionally streamlined to reduce the administrative burden while maintaining essential controls. For Minor changes, the resulting classification is an assessment of the potential adverse impact associated with the change, thus reducing the need for further workflow elements focused on mitigation of associated risks. This allows for faster execution and simplified documentation, appropriate for lower risk and complexity.

In contrast, Moderate and Major changes follow a similar workflow structure, both requiring formal evaluations, impact and risk assessments, and effectiveness checks. The depth and formality of these steps may vary, but the overall process remains consistent across these two higher classification levels.

Table 2: Workflow Element Overview for Minor Change vs Moderate and Major Change

Workflow Element	Minor Change	Moderate/Major Change
Impact Assessment (IA)	Required (IA Summary only, IA task(s) not required)	Required (both IA task(s) and IA Summary)
Risk Assessment (RA)	Not required	Required (General or FMEA)
Child Changes	Not allowed	Allowed
Change Plan Review	Not allowed	Allowed
Post-Phase Review of Change Action	Not allowed	Required
Effectiveness Check (EC) Creation	Not allowed	Required

Document Approval Signatures

Approved Date: 17 Jul 2026

Approval Verdict: Approve	Thomas Davey (L108497@lilly.com) Quality Approver 24-Jun-2026 11:39:48 GMT+0000
Approval Verdict: Approve	Allison Griffin (L111479@lilly.com) Approver 17-Jul-2026 21:02:18 GMT+0000
Reviewer / Approver Additional Details	N/A