THE UNIVERSITY OF TEXAS
MD Anderson Cancer Center
Making Cancer History®

Radiation Dosimetry Services
8060 El Rio St, ERD 1.200
Houston, TX 77054-4186
Main: (713) 745-8999 Fax: (713) 794-1371
Email: RDS@MDAnderson.org
Website: https://rds.mdanderson.org

RESULTS OF LITHIUM FLUORIDE DOSIMETER IRRADIATION FOR MONITORING OUTPUT OF RADIATION THERAPY MACHINES

Date of irradiation: August 29, 2019

Institution: 3566 - Cameroon Oncology Center
Douala, Littoral, Cameroon

Person irradiating capsules: Imandi Paulin Harmoniel, M.Sc

Radiation machine: Varian 21EX Serial number: 1813

Target-surface distance: 100 cm

Electron energy: 06 MeV TLD plug ID: G598

TLD at d_{max} :

Dose stated by institution:	300.0 cGy
Dose measured by MDACC:	302 cGy water
Ratio of MDACC / institution:	1.01

TLD at depth:

Depth of TLD:	2.0 cm
MDACC measured % depth dose (%dd):	76.8
Institution's depth of this %dd:	1.9 cm
MDACC depth - Institution depth:	0.1 cm

This information should be used only as a check of machine operation, not as a machine calibration, nor should it be used as an alternative to frequent calibration by a qualified physicist. Agreement to within ± 5% is considered a satisfactory check for the dose at d_{max} . For the TLD at depth, agreement to within ± 0.5 cm is considered an acceptable check. Under no circumstances should the calibration of a radiation machine be altered on the basis of a TLD check. The MDACC calculation of dose is based on the TG-51 protocol [Med Phys 26(9), Sept 1999].

Date of report: September 26, 2019
Date of TLD readings: September 25, 2019
TLD readings by: Jennifer Larson
Calculations by: Jennifer Larson
Checked by: Francisco Aguirre

Rebecca M. Howell

Rebecca M. Howell, PhD, DABR
Director, Radiation Dosimetry Services

THE UNIVERSITY OF TEXAS
MD Anderson Cancer Center
Making Cancer History®

Radiation Dosimetry Services
8060 El Rio St, ERD 1.200
Houston, TX 77054-4186
Main: (713) 745-8999 Fax: (713) 794-1371
Email: RDS@MDAnderson.org
Website: https://rds.mdanderson.org

RESULTS OF LITHIUM FLUORIDE DOSIMETER IRRADIATION FOR MONITORING OUTPUT OF RADIATION THERAPY MACHINES

Date of irradiation: August 29, 2019

Institution: 3566 - Cameroon Oncology Center
Douala, Littoral, Cameroon

Person irradiating capsules: Imandi Paulin Harmoniel, M.Sc

Radiation machine: Varian 21EX Serial number: 1813

Target-surface distance: 100 cm

Electron energy: 09 MeV TLD plug ID: H470

TLD at d_{max} :

Dose stated by institution:	300.0 cGy
Dose measured by MDACC:	295 cGy water
Ratio of MDACC / institution:	0.98

TLD at depth:

Depth of TLD:	3.3 cm
MDACC measured % depth dose (%dd):	62.9
Institution's depth of this %dd:	3.4 cm
MDACC depth - Institution depth:	-0.1 cm

This information should be used only as a check of machine operation, not as a machine calibration, nor should it be used as an alternative to frequent calibration by a qualified physicist. Agreement to within ± 5% is considered a satisfactory check for the dose at d_{max} . For the TLD at depth, agreement to within ± 0.5 cm is considered an acceptable check. Under no circumstances should the calibration of a radiation machine be altered on the basis of a TLD check. The MDACC calculation of dose is based on the TG-51 protocol [Med Phys 26(9), Sept 1999].

Date of report: September 26, 2019
Date of TLD readings: September 25, 2019
TLD readings by: Jennifer Larson
Calculations by: Jennifer Larson
Checked by: Francisco Aguirre

Rebecca M. Howell

Rebecca M. Howell, PhD, DABR
Director, Radiation Dosimetry Services

THE UNIVERSITY OF TEXAS
MD Anderson Cancer Center
Making Cancer History®

Radiation Dosimetry Services
8060 El Rio St, ERD 1.200
Houston, TX 77054-4186
Main: (713) 745-8999 Fax: (713) 794-1371
Email: RDS@MDAnderson.org
Website: https://rds.mdanderson.org

RESULTS OF LITHIUM FLUORIDE DOSIMETER IRRADIATION FOR MONITORING OUTPUT OF RADIATION THERAPY MACHINES

Date of irradiation: August 29, 2019

Institution: 3566 - Cameroon Oncology Center
Douala, Littoral, Cameroon

Person irradiating capsules: Imandi Paulin Harmoniel, M.Sc

Radiation machine: Varian 21EX Serial number: 1813

Target-surface distance: 100 cm

Electron energy: 12 MeV TLD plug ID: W335

TLD at d_{max} :

Dose stated by institution:	300.0 cGy
Dose measured by MDACC:	296 cGy water
Ratio of MDACC / institution:	0.99

TLD at depth:

Depth of TLD:	4.7 cm
MDACC measured % depth dose (%dd):	64.6
Institution's depth of this %dd:	4.7 cm
MDACC depth - Institution depth:	0.0 cm

This information should be used only as a check of machine operation, not as a machine calibration, nor should it be used as an alternative to frequent calibration by a qualified physicist. Agreement to within ± 5% is considered a satisfactory check for the dose at d_{max} . For the TLD at depth, agreement to within ± 0.5 cm is considered an acceptable check. Under no circumstances should the calibration of a radiation machine be altered on the basis of a TLD check. The MDACC calculation of dose is based on the TG-51 protocol [Med Phys 26(9), Sept 1999].

Date of report: September 26, 2019
Date of TLD readings: September 25, 2019
TLD readings by: Jennifer Larson
Calculations by: Jennifer Larson
Checked by: Francisco Aguirre

Rebecca M. Howell

Rebecca M. Howell, PhD, DABR
Director, Radiation Dosimetry Services