MATERIAL SPECIFICATION
Specification: raw material RM-20005
Document No.: SPEC-RM-8005
Revision: 01
Effective date: 2025-01-06
Approved by: Grace Okafor

1. SCOPE
Acceptance specification for this raw material at all sites.

2. LIMITS

Parameter	Limit	Unit	Method
Microbial count	0 - 100	cfu/g	Plate count
Assay	98.0 - 101.0	%	HPLC
Viscosity	1200 - 1600	mPa.s	Brookfield
pH	5.0 - 7.5	-	Potentiometric

3. SAMPLING
Sample every lot before release and test against the limits below.