MATERIAL SPECIFICATION
Specification: Purified water
Document No.: SPEC-RM-0003
Revision: 02
Effective date: 2024-02-27
Material: RM-10003
Site: Northfield
Approved by: Maya Moreau

1. SCOPE
Acceptance specification for Purified water (RM-10003) received at all sites.

2. LIMITS

Parameter	Limit	Unit	Method
Conductivity	0.00 - 1.17	uS/cm	In-house
Total organic carbon	0 - 500	ppb	In-house
Microbial count	0 - 100	cfu/g	Ph. Eur.

3. SAMPLING
Sample per SOP-QA-0003. Test every lot before release.