


# IC Calibration

Arthur Adriaens

March 19, 2024

## 0.1 Setup

## 0.2 Procedure



The connectors normally connecting to the directional coupler were connected to the VNA, after which power ramps for different frequencies were performed, for the above shown picture this was a ramp from -20dbm to -2dbm which covers the range of 100W to 6310W as the directional coupler induces an attenuation of -70db (this was measured, all S-parameters and original calibration files are available on the TOMAS github).

I then proceeded to fit a straight line to one of the ramps, this results in a relation between time and voltage

$$V = at + b \quad (1)$$

For this specific ramp we chose, the time is related to the input power as:

$$P = \omega t + \rho \quad (2)$$

As the ramp takes 1 second and the power difference is 18dbm, we have that

$$\omega = P_{fin} - P_{init} = 18 \quad (3)$$

Now using the end time of the ramp  $t_{fin}$ , we can find  $\rho$ :

$$\rho = P_{fin} - (P_{fin} - P_{init})t_{fin} = -2 - 18t_{fin} \quad (4)$$


Now we can fill these into the relation with the voltage:

$$V = \frac{a}{\omega}P - \frac{\rho a}{\omega} + b := a'P + b' \quad (5)$$

I.e the relation we need is:

$$P = \frac{V - b'}{a'} \quad (6)$$

Old setup



New setup




Figure 1: The new setup splits the signals of the forward and reflected in two, using one side to measure power and the other side to measure the phase

## 0.3 Results

### 0.3.1 Old setup

This part is useful for the people who did measurements whilst the new box was installed prior to 13/03/24. Note that some of the accuracy of  $b$  is lost when transforming to  $b'$  due to the inaccuracy of  $t_{fin}$  even though this is the case,  $a'$  is still as accurate as  $a$ , and as such I try to keep as many significant digits as possible.

Signal	a	error on a (%)	b	error on b (%)	$\omega$	$t_{fin}$
ICF	0.457843	0.17	0.713159	0.32	18	3.348
ICR	0.442106	0.27	0.890641	0.36	18	3.198

So the power we get from the voltage is:

$$P_{ICF}(\text{dBm}) = \frac{V_{ICF} - 2.296065}{0.02543572} \quad (7)$$

$$P_{ICR}(\text{dBm}) = \frac{V_{ICR} - 2.1.856293}{0.02456144} \quad (8)$$

### 0.3.2 New setup

This setup is in effect as of the 13th of march 2024.

25MHz						
Signal	a	error on a (%)	b	error on b (%)	$\omega$	$t_{fin}$
ICF	0.4636	0.037	0.910352	0.040	18	2.673
ICR	0.453835	0.083	0.791354	0.126	18	3.144

So the power we get from the voltage is:

$$P_{ICF}(\text{dBm}) = \frac{V_{ICF} - 2.200913}{0.02575556} \quad (9)$$

$$P_{ICR}(\text{dBm}) = \frac{V_{ICR} - 2.268637}{0.02521306} \quad (10)$$

35MHz						
Signal	a	error on a (%)	b	error on b (%)	$\omega$	$t_{fin}$
ICF	0.464247	0.016	0.982976	0.031	18	2.503
ICR	0.447934	0.09703	0.667036	0.1896	18	3.431

$$P_{ICF}(\text{dBm}) = \frac{V_{ICF} - 2.196569}{0.0257915} \quad (11)$$

$$P_{ICR}(\text{dBm}) = \frac{V_{ICR} - 2.253668}{0.02488522} \quad (12)$$

$$(13)$$

45MHz						
Signal	a	error on a (%)	b	error on b (%)	$\omega$	$t_{fin}$
ICF	0.459813	0.039	0.595329	0.051	18	3.353
ICR	0.446204	0.09459	0.935934	0.1042	18	2.822

$$P_{ICF}(\text{dBm}) = \frac{V_{ICF} - 2.188172}{0.02554517} \quad (14)$$

$$P_{ICR}(\text{dBm}) = \frac{V_{ICR} - 2.2447}{0.02478911} \quad (15)$$

### 0.3.3 ICV0-3

Signal	a	error on a (%)	b	error on b (%)	$\omega$	$t_{fin}$
ICV3	0.707133	0.04609	0.328061	0.2369	28	2.764
ICV2	0.706039	0.02707	-0.138149	0.4025	28	3.4346
ICV1	0.702135	0.02917	0.576769	0.06987	28	2.4675
ICV0	0.694397	0.03	0.123464	0.4494	28	3.1335