# Full Paper

# Analysis of Residual Explosives by Swab Touch Spray Ionization Mass Spectrometry

Ryan M. Bain+,[a] Patrick W. Fedick+,[a] Jonathan M. Dilger,[b] and R. Graham Cooks*[a]

**Abstract:** Swab touch spray ionization mass spectrometry, an ambient ionization technique, has been applied to the analysis of six explosives from various surfaces including glass, metal, Teflon, plastic, human hands and three types of gloves (nitrile, vinyl and latex). A swab, attached to a metallic handle, was used to sample explosive residues and acted as the ion source. The explosives, 1,3,5-trinitro-1,3,5-triazinane (RDX), 1,3,5,7-tetranitro-1,3,5,7-tetrazocane (HMX), and 2,2-bis[(nitrooxy)methyl]propane-1,3-diyl dinitrate (PETN) had an absolute limit of detection of 10 ng from all the surfaces except for PETN from the nitrile gloves (limit of detection 100 ng). Sodium perchlorate, 2-methyl-1,3,5-trinitrobenzene (TNT) and tetra-butylammonium perchlorate had limits of detection of 100 pg, 10 pg, and 1 pg, respectively from all surfaces. This study demonstrates the feasibility of swab touch spray ionization mass spectrometry for detection of a wide array of explosives from a variety of forensically applicable surfaces with disposable, commercial, tamperproof and individually-wrapped conductive swabs without complicated/lengthy sample preparations or extractions.

**Keywords:** Mass Spectrometry · Ambient Ionization · Explosives

## 1 Introduction

Residual and trace explosives analysis by mass spectrometry typically involves the use of a swab or wipe followed by a lengthier sample preparation procedure [1]. Liquid chromatography or gas chromatography coupled to mass spectrometry [2] and ion mobility [3] represent the current state of the art methods in explosives detection. Colorimetric detection methods have been developed but typically suffer from low specificity [4]. Raman spectroscopy methodologies look promising as a good alternative; however, there can be issues with fluorescence and background mitigation [5]. Bulk detection of explosives is commonly performed by x-ray diffraction as highlighted in a recent review [6]. Radioactive sources (typically [63]Ni), are also commonly used, but they represent a major cost burden due to additional administrative oversight to track the sources and routine testing for radiation leakages [7]. A more rapid and simpler test for explosives would be useful in a variety of applications, including airport security and forensic crime scenes. The relatively new field of ambient ionization mass spectrometry pertains to ionization under atmospheric pressure with the stipulation of little to no sample preparation (unlike most electrospray ionization or atmospheric pressure chemical ionization experiments) [8]. Desorption electrospray ionization [9] (DESI) and direct analysis in real-time [10] (DART) were the first ambient ionization sources to be reported and many other have followed including low temperature plasma ionization [11] (LTP), paper spray ionization [12] (PS) and swab touch spray ionization [13]. As rapid analysis of explosive residues is desirable, both DESI [14] and DART [15]

have been employed for this application [16]. These methods have proven especially useful in the analysis on skin, clothing, and other surfaces; however, both methodologies require specialized equipment. Examples include the DART plasma generator and the DESI spray setup, of which neither is readily coupled to a portable instrument which would be desirable for in-situ analysis [17]. A recent review has detailed the current advances of ambient ionization to the trace detection of explosive residues [18].

Low temperature plasma ionization has been applied extensively to explosive residues [19]. The low temperature characteristic of the plasma, as well as being safe to the touch, enables residues to be sampled from a variety of surfaces, including human skin without physical damage or sample degradation [11]. This technique has also been used with portable mass spectrometers, where a backpack MS showed promise for in-situ measurements for homeland security and defense applications [20]. While LTP is promising, it does require that a specialized ionization source be used and it utilizes high AC potentials which have an extra level

[a] R. M. Bain,+ P. W. Fedick,+ R. G. Cooks
Department of Chemistry, Purdue University, West Lafayette, Indiana 47907, United States
*e-mail: cooks@purdue.edu
[b] J. M. Dilger
Naval Surface Warfare Center Crane Division, Crane, Indiana 47522, United States
[+] These authors contributed equally

of safety concern [11]. By contrast, one of the simplest ambient ionization techniques, PS, uses only a low current DC potential to ionize analytes from a sharply pointed piece of paper wetted with solvent [21]. As shown previously, this allows detection of 2-methyl-1,3,5-trinitrobenzene (TNT), 2,2-bis[(nitrooxy)methyl]propane-1,3-diyl dinitrate (PETN), 1,3,5,7-tetranitro-1,3,5,7-tetrazocane (HMX) and 1,3,5-trinitro-1,3,5-triazinane (RDX) [22]. PS has a low barrier of entrance into a forensic setting as the paper acts as both as the sampling device and the ionization source. While PS shows great promise, the stability of the tip is essential for successful ionization, and this can make swabbing with the paper substrate difficult [23]. If the tip of the paper substrate is damaged during swabbing, subsequent analysis will be adversely affected, although proper training can alleviate this problem [24].

Swab touch spray ionization, as performed on this experiment, utilizes a rayon swab or an alternative swabbing substrate attached to a conductive handle [25]. Surface analytes are attached to the swab surface by moving the swab over the surface of interest. Solvent is continually applied to the swab's tip after sampling the surface, and a high voltage is applied to the metallic handle to generate a Taylor cone which then leads to ionization of the sample [25]. Similar to paper spray ionization, the swab in swab touch spray ionization acts both as the sampling device and the ionization source. Swab touch spray ionization has the added advantage over PS of being able to perform analysis from an individually-packed, tamperproof swab, rather than a fragile hand-cut paper triangle [26]. Previously we have demonstrated the field-ability of swab touch spray ionization with portable instrumentation for the analysis of organic gunshot residues 1,3-diethyl-1,3-diphenylurea (ethyl centralite) and 1,3-dimethyl-1,3-diphenylurea (methyl centralite) [27]. Chemical warfare agent simulants have also been analyzed by this methodology [28]. Herein, we demonstrate the feasibility of swab touch spray ionization mass spectrometry for the analysis of six residual explosives from various forensically applicable surfaces.

## 2 Experimental

### 2.1 Chemicals and Surfaces Preparation

The explosives TNT, PETN, HMX and RDX were purchased from Cerilliant (Round Rock, Texas), while sodium perchlorate and tetrabutylammonium perchlorate were purchased from Sigma-Aldrich (St. Louis, MO). HPLC grade acetonitrile was purchased from Fisher Scientific (Hampton, NH). Samples were prepared from 1.0 mg/mL stock solutions in acetonitrile. Direct analysis of explosives was performed by depositing 5.0 µL of solution on the swab using a pipette. Lower concentrations for limit of detection studies were prepared by serial dilutions (orders of magnitude) from the stock solutions to provide conservative limits of detection

from the surface. This method created a conservative approximate limit of detection that can be expected in the field and is not determined by extrapolation. Surfaces bearing trace explosives were prepared by spotting 1.0 µL of reagent onto each surface. Surfaces included black nitrile exam gloves (Ammex, Seattle, WA), diamond grip latex gloves (Microflex, Reno, NV), anti-static vinyl gloves (OAK Technical, Matteson, IL), glass microscope slides (Gold Seal, Portsmouth, NH), polytetrafluoroethylene (PTFE) plugs (Swagelok, Indianapolis, IN), stainless-steel plugs (Swagelok, Indianapolis, IN), blue polyethylene flat-top screw cap (Fisher Scientific, Hampton, NH) and human hands (those of the authors). Once the sample had been spotted on the surface, it was allowed to dry prior to swabbing with a swab pre-wetted with 20 µL of acetonitrile.

### 2.2 Swab Touch Spray Ionization

Ionization was performed from rayon medical grade sterile swabs (Copan Diagnostics, Murritea, CA). Each swab was opened by breaking the tamper-proof seal and used only once. Prior to swabbing, the swabs were pre-wetted with 20 µL of acetonitrile. This volume was selected as the swabs remained saturated with solvent during swabbing. To collect samples, the surface was sampled by swabbing using a circular motion (Figure 1). Then swabs were positioned 5–10 mm above the ion transfer capillary of the mass spectrometer. The position was an important variable. If the swab was too close to the inlet, discharge would occur. Conversely, if the swab was too far away, a stable cone would not be produced resulting in an unstable signal or no signal at all. Acetonitrile was added to the positioned swab using a syringe pump (Harvard Apparatus standard infusion only PHD 22/2000 syringe pump, Holliston, MA) connected to a 0.5 mL gastight syringe (Hamilton, Reno, NV) at a flow rate 10–30 mL/min. This flow was varied to maintain a steady spray without over wetting and unnecessarily diluting the sample. A high voltage of $+/-5.5$ kV was applied to the aluminum handle, and the generation of a spray could be observed by eye. The flow rate as well as the posi-



**Figure 1.** Pictorial representation of swab touch spray ionization source demonstrating sampling (left) and ionization using a commercial ion trap mass spectrometer (right).

tion were both optimized in order to achieve a steady signal. Efforts have been made previously to achieve robust signal with 3D printed parts and a camera to aid in alignment [26]. Similarly, a rail system has been used to aid in alignment for non-expert users and in-field applications with miniature instrumentation [29].

## 2.3 Mass Spectrometry

All mass spectral analyses were performed using a Thermo LTQ Orbitrap XL Hybrid Ion Trap-Orbitrap mass spectrometer (San Jose, CA). Spectra were collected in negative ion mode, except when analyzing tetrabutylammonium perchlorate, which was collected in both positive and negative ion modes. The instrument also served as the supply of the high voltage applied to the metallic shaft of the swab. The entire time from swabbing through the collection of the spectra is ca 2 minutes.

## 3 Results and Discussion

Direct analysis of 5 µg of TNT using the product ion scan (Figure 2) gave a spectrum rich in fragment ions. The fragmentation of TNT has been previously studied by DESI [14a] and LTP [17] and both the radical anion ($m/z$ 227) and deprotonated ($m/z$ 226) TNT were observed. Using an isolation width of 3 Th on the LTQ and a peak centered at $m/z$ 226.5, both the radical anion and deprotonated TNT were isolated and fragmented. Product ions at $m/z$ 210, $m/z$ 197, and $m/z$ 196 correspond to the loss of OH, the loss of NO, and the loss of HNO, respectively. The product ion mass spectrum recorded for 10 pg of TNT swabbed from a glass surface (Figure S1) gave this as the limit of detection.

Full scan mass spectra recorded for the direct analysis of 5.0 µg of RDX, HMX, and PETN from a swab are shown in Figure 3. The full scan mass spectra at the limit of detection of 10 ng of RDX, HMX, and PETN swabbed off a glass slide are recorded Figure S2. In agreement with the literature [19b, 30], RDX was identified by four ions that form adducts with RDX; $[RDX+Cl_{35/37}]^-$ at $m/z$ 257 and $m/z$ 259, $[RDX+NO_2-H]^-$ at $m/z$ 267 and $[RDX+NO_3]^-$ at $m/z$ 284. Similarly, HMX and PETN were identified by the corresponding four adducts. HMX was identified by $[HMX+Cl_{35/37}]^-$ at $m/z$ 331 and $m/z$ 333, $[HMX+NO_2-H]^-$ at $m/z$ 341 and $[HMX+NO_3]^-$ at $m/z$ 358. PETN was identified by $[PETN+Cl_{35/37}]^-$ at $m/z$ 351 and $m/z$ 353, $[PETN+NO_2-H]^-$ at $m/z$ 361 and $[PETN+NO_3]^-$ at $m/z$ 378. PETN had a limit of detection of 10 ng on all surfaces except for the nitrile glove surface, which had a measured limit of 100 ng. These three explosives were identified by their adducts rather than their fragmentation profiles as no measurable fragments are detected on the ion trap mass spectrometer, as is consistent with previous results [14d].



**Figure 3.** A) Swab touch spray ionization mass spectrum of 5 µg of RDX in negative ion mode. The ions at $m/z$ 257 and $m/z$ 259 correspond to $[RDX+Cl_{35/37}]^-$ at $m/z$ 267 to $[RDX+NO_2-H]^-$ and at $m/z$ 284 to $[RDX+NO_3]^-$. B) Mass spectrum of 5 µg of HMX in negative ion mode. The ions at $m/z$ 331 and $m/z$ 333 correspond to $[HMX+Cl_{35/37}]^-$ at $m/z$ 341 to $[HMX+NO_2-H]^-$ and at $m/z$ 358 to $[HMX+NO_3]^-$. C): Mass spectrum of 5 µg of PETN in negative ion mode. The ions at $m/z$ 351 and $m/z$ 353 correspond to $[PETN+Cl_{35/37}]^-$ at $m/z$ 361 to $[PETN+NO_2-H]^-$ and at $m/z$ 378 to $[PETN+NO_3]^-$.



**Figure 2.** MS/MS product ion scan of TNT using swab touch spray ionization. The ion at $m/z$ 226 corresponds to the $[M-H]^-$ species and the ion at $m/z$ 227 corresponds to the $[M]^-\cdot$ as expected. Product ions at $m/z$ 210, $m/z$ 197, and $m/z$ 196 correspond to the loss of OH, the loss of NO, and the loss of HNO, respectively.

Explosive perchlorate compounds were also of interest in this study because of their widespread use [31]. The direct analysis of 5.0 µg of tetrabutylammonium perchlorate was analyzed in full scan negative ion mode to look for perchlorate ions, as well as in the positive ion mode to measure tetrabutylammonium (Figure 4). Tetrabutylammonium at $m/z$ 242 produced fragments at $m/z$ 186, $m/z$ 184, $m/z$ 142, $m/z$ 130 and $m/z$ 100 corresponding to tribu-

R. M. Bain, P. W. Fedick, J. M. Dilger, R. G. Cooks



**Figure 4.** A) Swab touch spray ionization mass spectrometry product ion scan mass spectrum of tetrabutylammonium perchlorate in positive ion mode shows tetrabutylammonium at *m/z* 242 and characteristic fragment ions including those of *m/z* 186, 184, 142, 130, and 100. B) Swab touch spray ionization mass spectrometry full scan mass spectrum in negative ion mode of tetrabutylammonium perchlorate shows the perchlorate ions at *m/z* 99 and *m/z* 101.

**Table 2.** Approximate limits of detection for explosives from gloves and human hands by swab touch spray ionization mass spectrometry.

| Explosive | Vinyl glove | Nitrile glove | Latex glove | Exposed hand |
|---|---|---|---|---|
| TNT | 10 pg | 10 pg | 10 pg | 10 pg |
| RDX | 10 ng | 10 ng | 10 ng | 10 ng |
| HMX | 10 ng | 10 ng | 10 ng | 10 ng |
| PETN | 10 ng | 100 ng | 10 ng | 10 ng |
| Sodium perchlorate | 100 pg | 100 pg | 100 pg | 100 pg |
| Tetrabutylammonium perchlorate | 1 pg | 1 pg | 1 pg | 1 pg |

tylammonium, dehydrogenated tributyl ammonium, dehydrogenated methyl dibutylammonium, dibutylammonium and loss of ethane from dehydrated dibutyl ammonium cation. The product ion scan mass spectrum establishing the limit of detection as 1 pg of tetrabutylammonium perchlorate swabbed from a glass surface can be found in Figure S3. Sodium perchlorate was also analyzed (Figure S4) and the characteristic perchlorate peaks were used for identification.

In addition to characterizing explosives applied directly to the swab (Figure 2–4) and trace residues on glass (S1–S4), a variety of forensically applicable surfaces were also screened for explosive residues. Identical to the experiments using glass, a known amount was spotted onto the substrate and the sample was allowed to dry. The swab, pre-wetted with 20 μL of acetonitrile, was rastered over the surface in a circular motion to collect the explosive residue. Surfaces and their corresponding limits of detection can be found in Table 1 for each solid surface. Additionally, for screening and forensic purposes, gloves and human hands were swabbed for all explosives and limits of detection for each explosive can be found in Table 2.

Analysis from common surfaces for residual amounts of explosives is applicable to forensic and security analyses, from airports to crime scenes [14d, 20a]. The ability to screen people by analyzing human hands is also of great interest in situations such as airport screenings. Similarly, finding gloves near a crime scene that have residual explosives can be traced back to a perpetrator using fingerprints [32] inside the glove or using cameras between the crime scene and the location where the gloves were discovered. Additionally, the successful swabbing of a variety of surfaces shows the robustness of this technique and suggests feasibility across a variety of other surfaces as well. The surfaces studied all gave similar results except that the limit of detection for PETN was an order of magnitude higher for nitrile gloves than other surfaces or other explosives. This is likely a result of chemical selectivity specifically related to PETN and nitrile.

## 4 Conclusion

To recapitulate, the analysis of six different explosive residues from a variety of surfaces (glass, stainless steel, polyethylene, PTFE, human hands, nitrile gloves, vinyl gloves and latex gloves) was performed with low absolute limits of detection without the need for complex and time-consuming extraction or sample preparation steps. Limits of detection for RDX, HMX, and PETN were 10 ng from all surfaces, with the exception of PETN on the nitrile gloves. Sodium perchlorate, TNT and tetra-butyl ammonium perchlorate had limits of detection of 100 pg, 10 pg, and 1 pg, respectively from all surfaces examined. The wide breadth of surfaces demonstrates the applicability of swab touch spray ionization for forensic, military and homeland security applications. In addition, as the explosives were detectable from human skin, airport security could employ this technology in conjunction with the current ion mobility systems. With swab touch spray ionization previously being demonstrated using portable mass spectrometers [27, 29],

**Table 1.** Approximate limits of detection for explosives from various surfaces by swab touch spray ionization mass spectrometry.

| Explosive | Glass | PTFE | Stainless steel | Polyethylene |
|---|---|---|---|---|
| TNT | 10 pg | 10 pg | 10 pg | 10 pg |
| RDX | 10 ng | 10 ng | 10 ng | 10 ng |
| HMX | 10 ng | 10 ng | 10 ng | 10 ng |
| PETN | 10 ng | 10 ng | 10 ng | 10 ng |
| Sodium perchlorate | 100 pg | 100 pg | 100 pg | 100 pg |
| Tetrabutylammonium perchlorate | 1 pg | 1 pg | 1 pg | 1 pg |

this technique can be useful for *in-situ* analysis for real time detection of explosive residues.

The advantages of swab touch spray ionization are the ease of use, contained commercial packaging, low entrance barrier, and the fact that your sampling device is also your ionization source. When utilizing PS the paper triangle is the sampling and ionization source; however, swab touch spray does not require the analyst to carefully handle the device to preserve the tip as is needed in PS. Swab touch spray has detection limits comparable to PS [22] while providing a forensically feasible platform. While the detection limits have been shown to be lower for LTP, swab touch spray does not require additional parts such as the plasma generation and AC power supply [19a].

## Acknowledgements

## References

[1] a) O. Dalby, D. Butler, W. Birkett Jason, Analysis of gunshot residue and associated materials-A Review *J. Forensic Sci.* **2010**, *55*, 924–943; b) D. S. Moore, Instrumentation for trace detection of high explosives *Rev. Sci. Instrum.* **2004**, *75*, 2499–2512; c) L. Barron, E. Gilchrist, Ion chromatography-mass spectrometry: A review of recent technologies and applications in forensic and environmental explosives analysis *Anal. Chim. Acta* **2014**, *806*, 27–54.

[2] a) J. M. Perr, K. G. Furton, J. R. Almirall, Gas chromatography positive chemical ionization and tandem mass spectrometry for the analysis of organic high explosives *Talanta* **2005**, *67*, 430–436; b) M. Palit, D. Pardasani, A. K. Gupta, D. K. Dubey, Application of single drop microextraction for analysis of chemical warfare agents and related compounds in water by gas chromatography/mass spectrometry *Anal. Chem.* **2005**, *77*, 711–717; c) X. Xu, M. Koeberg, C.-J. Kuijpers, E. Kok, Development and validation of highly selective screening and confirmatory methods for the qualitative forensic analysis of organic explosive compounds with high performance liquid chromatography coupled with (photodiode array and) LTQ ion trap/Orbitrap mass spectrometric detections (HPLC-(PDA)-LTQOrbitrap) *Sci. Justice* **2014**, *54*, 3–21.

[3] a) J. Lee, S. Park, S. G. Cho, E. M. Goh, S. Lee, S.-S. Koh, J. Kim, Analysis of explosives using corona discharge ionization combined with ion mobility spectrometry-mass spectrometry *Talanta* **2014**, *120*, 64–70; b) Z. Du, T. Sun, J. Zhao, D. Wang, Z. Zhang, W. Yu, Development of a plug-type IMS-MS instrument and its applications in resolving problems existing in in-situ detection of illicit drugs and explosives by IMS *Talanta* **2018**, *184*, 65–72.

[4] J. Almog, S. Zitrin, in *Aspects of Explosives Detection* (Eds.: M. Marshall, J. C. Oxley), Elsevier, Amsterdam, **2009**, pp. 41–58.

[5] D. S. Moore, R. J. Scharff, Portable Raman explosives detection *Anal. Bioanal. Chem.* **2009**, *393*, 1571–1578.

[6] K. Wells, D. A. Bradley, A review of X-ray explosives detection techniques for checked baggage *Appl. Radiat. Isot.* **2012**, *70*, 1729–1746.

[7] R. G. Ewing, D. A. Atkinson, G. A. Eiceman, G. J. Ewing, A critical review of ion mobility spectrometry for the detection of explosives and explosive related compounds *Talanta* **2001**, *54*, 515–529.

[8] D. N. Correa, J. M. Santos, L. S. Eberlin, M. N. Eberlin, S. F. Teunissen, Forensic chemistry and ambient mass spectrometry: A perfect couple destined for a happy marriage? *Anal. Chem.* **2016**, *88*, 2515–2526.

[9] Z. Takáts, J. M. Wiseman, B. Gologan, R. G. Cooks, Mass spectrometry sampling under ambient conditions with desorption electrospray ionization *Science* **2004**, *306*, 471–473.

[10] R. B. Cody, J. A. Laramée, H. D. Durst, Versatile new ion source for the analysis of materials in open air under ambient conditions *Anal. Chem.* **2005**, *77*, 2297–2302.

[11] J. D. Harper, N. A. Charipar, C. C. Mulligan, X. Zhang, R. G. Cooks, Z. Ouyang, Low-temperature plasma probe for ambient desorption ionization *Anal. Chem.* **2008**, *80*, 9097–9104.

[12] H. Wang, J. Liu, R. G. Cooks, Z. Ouyang, Paper spray for direct analysis of complex mixtures using mass spectrometry *Angew. Chem. Int. Ed.* **2010**, *49*, 877–880.

[13] A. K. Jarmusch, V. Pirro, K. S. Kerian, R. G. Cooks, Detection of strep throat causing bacterium directly from medical swabs by touch spray-mass spectrometry *Analyst* **2014**, *139*, 4785–4789.

[14] a) Z. Takats, I. Cotte-Rodriguez, N. Talaty, H. Chen, R. G. Cooks, Direct, trace level detection of explosives on ambient surfaces by desorption electrospray ionization mass spectrometry *Chem. Commun.* **2005**, 1950–1952; b) N. Talaty, C. C. Mulligan, D. R. Justes, A. U. Jackson, R. J. Noll, R. G. Cooks, Fabric analysis by ambient mass spectrometry for explosives and drugs *Analyst* **2008**, *133*, 1532–1540; c) D. R. Justes, N. Talaty, I. Cotte-Rodriguez, R. G. Cooks, Detection of explosives on skin using ambient ionization mass spectrometry *Chem. Commun.* **2007**, 2142–2144; d) I. Cotte-Rodríguez, Z. Takáts, N. Talaty, H. Chen, R. G. Cooks, Desorption electrospray ionization of explosives on surfaces: Sensitivity and selectivity enhancement by reactive desorption electrospray ionization *Anal. Chem.* **2005**, *77*, 6755–6764.

[15] a) F. Rowell, J. Seviour, A. Y. Lim, C. G. Elumbaring-Salazar, J. Loke, J. Ma, Detection of nitro-organic and peroxide explosives in latent fingermarks by DART- and SALDI-TOF-mass spectrometry *Forensic Sci. Int.* **2012**, *221*, 84–91; b) E. Sisco, J. Dake, C. Bridge, Screening for trace explosives by AccuTOF™-DART®: An in-depth validation study *Forensic Sci. Int.* **2013**, *232*, 160–168; c) J. M. Nilles, T. R. Connell, S. T. Stokes, D. H. Dupont, Explosives detection using direct analysis in real time (DART) mass spectrometry *Propellants Explos. Pyrotech.* **2010**, *35*, 446–451.

[16] T. P. Forbes, E. Sisco, Recent advances in ambient mass spectrometry of trace explosives *Analyst* **2018**, *143*, 1948–1969.

[17] N. L. Sanders, S. Kothari, G. Huang, G. Salazar, R. G. Cooks, Detection of explosives as negative ions directly from surfaces using a miniature mass spectrometer *Anal. Chem.* **2010**, *82*, 5313–5316.

[18] T. P. Forbes, E. Sisco, Recent advances in ambient mass spectrometry of trace explosives *Analyst* **2018**, *143*, 1948–1969.

[19] a) Y. Zhang, X. Ma, S. Zhang, C. Yang, Z. Ouyang, X. Zhang, Direct detection of explosives on solid surfaces by low temperature plasma desorption mass spectrometry *Analyst* **2009**, *134*, 176–181; b) W. Chen, K. Hou, X. Xiong, Y. Jiang, W. Zhao, L. Hua,

P. Chen, Y. Xie, Z. Wang, H. Li, Non-contact halogen lamp heating assisted LTP ionization miniature rectilinear ion trap: a platform for rapid, on-site explosives analysis *Analyst* **2013**, *138*, 5068–5073; c) J. F. Garcia-Reyes, J. D. Harper, G. A. Salazar, N. A. Charipar, Z. Ouyang, R. G. Cooks, Detection of explosives and related compounds by low-temperature plasma ambient ionization mass spectrometry *Anal. Chem.* **2011**, *83*, 1084–1092.

[20] a) P. I. Hendricks, J. K. Dalgleish, J. T. Shelley, M. A. Kirleis, M. T. McNicholas, L. Li, T.-C. Chen, C.-H. Chen, J. S. Duncan, F. Boudreau, R. J. Noll, J. P. Denton, T. A. Roach, Z. Ouyang, R. G. Cooks, Autonomous in situ analysis and real-time chemical detection using a backpack miniature mass spectrometer: Concept, instrumentation development, and performance *Anal. Chem.* **2014**, *86*, 2900–2908; b) J. K. Dalgleish, K. Hou, Z. Ouyang, R. G. Cooks, In situ explosive detection using a miniature plasma ion source and a portable mass spectrometer *Anal. Lett.* **2012**, *45*, 1440–1446.

[21] S. F. Teunissen, P. W. Fedick, B. J. A. Berendsen, M. W. F. Nielen, M. N. Eberlin, R. Graham Cooks, A. C. van Asten, Novel selectivity-based forensic toxicological validation of a paper spray mass spectrometry method for the quantitative determination of eight amphetamines in whole blood *J. Am. Soc. Mass Spectrom.* **2017**, *28*, 2665–2676.

[22] C. W. Tsai, C. A. Tipple, R. A. Yost, Application of paper spray ionization for explosives analysis *Rapid Commun. Mass Spectrom.* **2017**, *31*, 1565–1572.

[23] R. D. Espy, A. R. Muliadi, Z. Ouyang, R. G. Cooks, Spray mechanism in paper spray ionization *Int. J. Mass Spectrom.* **2012**, *325–327*, 167–171.

[24] Z. E. Lawton, A. Traub, W. L. Fatigante, J. Mancias, A. E. O'Leary, S. E. Hall, J. R. Wieland, H. Oberacher, M. C. Gizzi, C. C. Mulligan, Analytical validation of a portable mass spectrometer featuring interchangeable, ambient ionization sources for high throughput forensic evidence screening *J. Am. Soc. Mass Spectrom.* **2017**, *28*, 1048–1059.

[25] V. Pirro, A. K. Jarmusch, M. Vincenti, R. G. Cooks, Direct drug analysis from oral fluid using medical swab touch spray mass spectrometry *Anal. Chim. Acta* **2015**, *861*, 47–54.

[26] A. K. Jarmusch, V. Pirro, D. L. Logsdon, R. G. Cooks, Direct ion generation from swabs *Talanta* **2018**, *184*, 356–363.

[27] P. W. Fedick, R. M. Bain, Swab touch spray mass spectrometry for rapid analysis of organic gunshot residue from human hand and various surfaces using commercial and fieldable mass spectrometry systems *Forensic Chemistry* **2017**, *5*, 53–57.

[28] D. T. Snyder, L. J. Szalwinski, R. L. Schrader, V. Pirro, R. Hilger, R. G. Cooks, Precursor and neutral loss scans in an RF scanning linear quadrupole ion trap *J. Am. Soc. Mass Spectrom.* **2018**, *29*, 1345–1354.

[29] P. W. Fedick, W. L. Fatigante, Z. E. Lawton, A. E. O'Leary, S. E. Hall, R. M. Bain, S. T. Ayrton, J. A. Ludwig, C. C. Mulligan, A low-cost, simplified platform of interchangeable, ambient ionization sources for rapid, forensic evidence screening on portable mass spectrometric instrumentation *Instruments* **2018**, *2*(2), 5.

[30] I. Cotte-Rodriguez, R. G. Cooks, Non-proximate detection of explosives and chemical warfare agent simulants by desorption electrospray ionization mass spectrometry *Chem. Commun.* **2006**, 2968–2970.

[31] E. Sokol, A. U. Jackson, R. G. Cooks, Trace detection of inorganic oxidants using desorption electrospray ionization (DESI) mass spectrometry *Cent. Eur. J. Chem.* **2011**, *9*, 790–797.

[32] D. R. Ifa, N. E. Manicke, A. L. Dill, R. G. Cooks, Latent fingerprint chemical imaging by mass spectrometry" *Science* **2008**, *321*, 805.