

SOP Compliance Review Report

21 CFR Part 820 — Quality System Regulation

Date Generated: May 11, 2026 23:17 UTC
SOP File: sample_sop.txt
Device Class: Class II
SOP Type: Cleaning

Overall GMP Readiness

NEEDS REVISION

Rationale: The SOP covers most operational steps adequately but is missing the Revision History Table (21 CFR 820.40(b)) and the Non-Conformance Reporting Procedure (21 CFR 820.90), both mandatory for Class II device compliance.

Summary Scorecard

Total Items	PRESENT	INCOMPLETE	MISSING	Pass Rate
15	12	1	2	80%

Detailed Compliance Findings

ID	Checklist Item	Reg. Section	Status	Evidence / Gap	Recommendation
U01	Purpose Statement	21 CFR 820.40	PRESENT	This procedure establishes the steps required to clean manufacturing equipment...	—
U02	Scope	21 CFR 820.40	PRESENT	This SOP applies to all production equipment...Class 10,000 (ISO 7) cleanroom suite on the second floor of Building A.	—
U03	Responsibilities	21 CFR 820.20(b)	PRESENT	Manufacturing Technician...Line Supervisor...QA Manager...	—
U04	References	21 CFR 820.40	PRESENT	21 CFR Part 820, Subpart G; ISO 14644-1; SOP-ENV-002; SOP-EQ-011	—
U05	Definitions / Abbreviations	21 CFR 820.3	PRESENT	PPE, ISO, QA, SOP, WFI, IPA, Bioburden — all defined in Section 5.	—
U06	Materials and Equipment List	21 CFR 820.70(g)	PRESENT	Section 6 enumerates cleaning agents, wipes, bucket, scraper, pH strips, and datalogger.	—
U07	Safety Precautions / PPE Requirements	21 CFR 820.70(e)	PRESENT	Section 7 specifies nitrile gloves, safety glasses, cleanroom gown, N95 respirator, and IPA hazard controls.	—
U08	Step-by-Step Procedure	21 CFR 820.40	PRESENT	Eight numbered steps with action-verb format from STEP 1 (Notify and Isolate) to STEP 8 (Release).	—
U09	Documentation / Forms / Records	21 CFR 820.180	PRESENT	Section 9 specifies Cleaning Log Form QA-CL-001 with 2-year retention per 21 CFR 820.180.	—
U10	Revision History Table	21 CFR 820.40(b)	MISSING	Not found in SOP	Add a Revision History Table including columns for Version, Revision Date, Author, and Change Summary. Per 21 CFR 820.40(b), all document changes must be identified and retained.
U11	Approval Signatures Block	21 CFR 820.40(a)	PRESENT	Author, Reviewed, and Approved signature lines present with named titles for Jane Smith, Mark Rivera, and Dr. Lisa Chen.	—
CL01	Cleaning Agent Specifications	21 CFR 820.70(e)	PRESENT	Section 6.1 specifies 70% IPA USP Grade, WFI, and CLN-DET-103 at 1:50 dilution with approved supplier reference.	—

ID	Checklist Item	Reg. Section	Status	Evidence / Gap	Recommendation
CL02	Rinse / Residue Verification	21 CFR 820.70(e)	INCOMPLETE	Section 4.1 requires WFI wipe rinse twice but does not specify a verification method (e.g., pH or conductivity test).	Add a quantitative residue verification step — for example, a pH or conductivity measurement of the final rinse wipe with a defined acceptance criterion (e.g., pH 6.5–7.5).
CL03	Non-Conformance Reporting Procedure	21 CFR 820.90	MISSING	Not found in SOP	Add a dedicated section describing actions to take when cleaning fails (e.g., failed pH, visible residue after repeated cleaning). Specify disposition of affected equipment, CAPA initiation trigger, and notification chain per 21 CFR 820.90.
CL04	Cleaning Frequency / Schedule	21 CFR 820.70(f)	PRESENT	Cleaning Frequency Schedule table specifies: between product lots, end of shift (daily minimum), after equipment maintenance, and after any spill.	—