

	DISPOSAFE HEALTH AND LIFE CARE LIMITED Plot No.37A, Sector-06, Ballabgarh (Faridabad) Haryana India 121006							Doc. No	: F/ASSEM/03:00
								Supersedes	: None
								Effective Date	: 12/7/2024
BALLOON & VALVE INTEGRITY REPORT							Page: 1 of 1		

BALLOON IPSCTION REPORT - P17										
DATE	CHECKED QTY	ACCEPT QTY	HOLD QTY	REJ. QTY	REJ. %	STRUCK BALLOON	BALLOOM BRUST	LEAKAGE	OTHERS	Remarks
01/04/2025	9612	9612		15	0.16	11	1	3		
02/04/2025	9907	9858		49	0.49	27	3	19		
03/04/2025	10746	10683		63	0.59	46	10	7		
04/04/2025	9573	9555		18	0.19	12	2	4		
05/04/2025	7038	7018		20	0.28	8	1	11		
07/04/2025	4553	4498		55	1.21	27	12	16		
08/04/2025	9400	9326		71	0.76	28	17	26		
09/04/2025	8912	8886		26	0.29	10		16		
10/04/2025	10350	10279		71	0.69	16	8	47		
11/04/2025	12214	12114		100	0.82	61	2	37		
12/04/2025	7250	7214		36	0.50	5	3	28		
14/04/2025	6182	6088		94	1.52	69	4	21		
15/04/2025	11557	11479		78	0.67	39	10	29		
16/04/2025	8339	8267		57	0.68	32	7	18		
17/04/2025	10209	10104		105	1.03	65	6	34		
18/04/2025	8453	8390		63	0.75	35	1	27		
19/04/2025	10889	10794		95	0.87	59	2	34		
21/04/2025					#DIV/0!					
22/04/2025	7281	7244		37	0.51	23	1	13		
23/04/2025	7370	7220		150	2.04	108	2	40		
24/04/2025	6878	6755		123	1.79	95	4	24		
25/04/2025	8614	8485		129	1.50	115	4	10		
26/04/2025	5986	5937		49	0.82	44	2	3		
28/04/2025	8183	7955		228	2.79	50	17	161		
29/04/2025	8788	8723		65	0.74	22	16	27		
30/04/2025	7811	7698		113	1.45	57	6	50		
TOTAL	216095	214182		1910		1064	141	705		
TOTAL %		99.11		0.88		0.49	0.07	0.33		

VALVE INTEGRITY										
CHECKED QTY	ACCEPT QTY	HOLD QTY	REJ. QTY	REJ. %	LEAKAGE	90/10	BUBBLE	THIN SPOD	OTHERS	Remarks
9612	9483		129	1.34	21	64		28	16	
9858	9767		91	0.92	55	12		22	2	
10683	10632		51	0.48	27	4		20		
9555	9481		74	0.77	33	19		22		
7018	6963	1	55	0.78	27	17		11		
4498	4443		55	1.22	22	19		14		
9326	9100		226	2.42	140	75		11		
8886	8820		66	0.74	61			5		
10279	10159		120	1.17	102	4		14		
12114	1208		56	0.46	20	15		21		
7214	7148		66	0.91	50			16		
6088	5967		121	1.99	22	52		47		
11479	11376		103	0.90	56	6		41		
8267	7477		790	9.56	56			734		
10104	8874		1230	12.17	90			1140		
8390	6557	1794	39	0.46	39					
10794	9967	684	143	1.32	85			58		
				#DIV/0!						
7244	7177	1867	67	0.92	55	12				
7220	7143	1100	77	1.07	76	1				
6755	6694	1855	61	0.90	61					
8485	8425	331	60	0.71	34			26		
5937	5887	985	58	0.98	36	5		16		
7955	7640	481	315	3.96	237			78		
8723	8597	126		0.00	73			53		
7698	7570	612	128	1.66	88			40		
214182	196555	9836	4181		1566	305	0	2417	18	
	91.77	4.59	1.95		0.73	0.14	0.00		0.01	



COMMULATIVE DETAILS IN APRIL 25

BALLOON INS. REPORT

STRUCK BA	1064
BALLOON B	141
LEAKAGE	705
REJ QTY	1910

VALV INTEGRITY

LEAKAGE	1566
90/10	305
BUBBLE	0
THIN SPOD	2417
OTHERS	18
REJ QTY	4306





DISPOSAFE HEALTH AND LIFE CARE LIMITED

Plot No.37A, Sector-06, Ballabgarh (Faridabad) Haryana India 121006

Doc. No : F/ASSEM/03:
Supersedes : None
Effective Date : 12/7/2024
Page: 1 of 1

DAILY BALLOON & VALVE INTEGRITY REPORT

BALLOON ISPECTION REPORT - P17

DATE	CHECKED QTY	ACCEPT QTY	HOLD QTY	REJ. QTY	REJ. %	STRUCK BALLOON	BALLOON BRUST	LEAKAGE	OTHERS	Remarks
01/05/2025	6229	6123		106	1.70	78	2	26		
02/05/2025	4208	4188		20	0.48	15	1	4		
03/05/2025	6642	6591		51	0.77	35	8	8		
04/05/2025	SUNDAY									
05/05/2025	5847	5698		149	2.55	79	24	46		
06/05/2025	6490	6397		93	1.43	52	18	23		
07/05/2025	7268	7164		104	1.43	70	2	32		
08/05/2025	7808	7747		61	0.78	39	12	10		
09/05/2025	7976	7917		59	0.74	43	10	6		
10/05/2025	9000	8994		6	0.07	5		1		
12/05/2025	8485	8432		53	0.62	29	15	9		
13/05/2025	10218	10134		84	0.82	46	32	6		
14/05/2025	10056	9930		126	1.25	101	9	16		
15/05/2025	11144	11047		97	0.87	68	4	25		
16/05/2025	10936	10791		145	1.33	108	2	35		
17/05/2025	11780	11702		78	0.66	38	26	16		
19/05/2025	9892	9889		3	0.03	2		1		
20/05/2025	11914	11810		104	0.87	69	10	25		
21/05/2025	10380	10328		52	0.50	20	29	3		
22/05/2025	10200	10080		120	1.18	82	34	4		
23/05/2025	10669	10618		51	0.48	32	7	12		
24/05/2025	10423	10360		63	0.60	50	10	3		
25/05/2025	10141	10076		65	0.64	34	17	14		
26/05/2025	10141	10076		65	0.64	34	17	14		
27/05/2025	11162	10925		237	2.12	146	91			
28/05/2025	9310	9305		5	0.05	2		3		
29/05/2025	9675	9570		105	1.09	78	21	6		
30/05/2025	10710	10631		79	0.74	39	28	12		
31/05/2025					#DIV/0!					
TOTAL	248704	246523		2181		1394	429	360		
TOTAL %		99.12		0.88		0.56	0.17	0.14		

VALVE INTEGRITY




CHECKED QTY	ACCEPT QTY	HOLD QTY	REJ. QTY	REJ. %	LEAKAGE	90/10	BUBBLE	THIN SPOD	OTHERS	Remarks
6125	6000	1633	123	2.01	95	11		17		
4188	4138	18	22	0.53	9			13		
6591	6570	968	21	0.32	9			12		
5698	5555		143	2.51	96			47		
6397	6243		154	2.41	78			61	15	
7164	7025		79	1.10	55			24		
7747	7666		81	1.05	60			21		
7917	7879		38	0.48	24	4		19		
8994	8992		2	0.02	2					
8432	8382		50	0.59				50		
10134	10076		58	0.57	51			7		
9930	9786		144	1.45	56			88		
11047	10955		92	0.83	39			53		
10791	10651		140	1.30	27			113		
11702	11558		144	1.23	69			75		
9889	9883		6	0.06	4			2		
11810	11737		73	0.62	32			41		
10328	10287		41	0.40	14			27		
10080	10022		58	0.58	21			37		
10618	10548		70	0.66	20			50		
10360	10277		83	0.80	36			47		
10076	9979		97	0.96	61			36		
10076	9979		97	0.96	61			36		
10309	10275		34	0.33	20	3		11		
11243	11079		164	1.46	63			101		
11354	11211		143	1.26	64			66	13	
10524	10393		131	1.24	42			31	58	
				#DIV/0!						
TOTAL	249524	247146	2619	2288		1108	18	0	1085	86
TOTAL %		99.05	1.05	0.92		0.44	0.01	0.00		0.03

COMMULATIVE DETAILS IN MAY 25



BALLOON INS. REPORT	
STRUCK BALLOON	1394
BALLOON BRUST	429
LEAKAGE	360
REJ QTY	2183

VALV INTEGRITY	
LEAKAGE	1108
90/10	18
BUBBLE	0
THIN SPOD	1085
OTHERS	86
REJ QTY	2297





DISPOSAFE HEALTH AND LIFE CARE LIMITED

Plot No.37A, Sector-06, Ballabgarh (Faridabad) Haryana India 121006

Doc. No : F/ASSEM/03:00
Supersedes : None
Effective Date : 12/7/2024
Page: 1 of 1

DAILY BALLOON & VALVE INTEGRITY REPORT

BALLOON INSPECTION REPORT - P17

DATE	CHECKED QTY	ACCEPT QTY	HOLD QTY	REJ. QTY	REJ. %	STRUCK BALLOON	BALLOON M BRUST	LEAKAGE	OTHERS	Remarks
01/06/2025										
02/06/2025	10047	9713		334	3.32	58	126	150		
03/06/2025	9465	9361		104	1.10	73	16	12		
04/06/2025	10582	10541		41	0.39	22	15	4		
05/06/2025	9649	9603		46	0.48	29	19			
06/06/2025	10226	10169		67	0.66	39	16	2		
07/06/2025	10346	10290		56	0.54	36	15	5		
08/06/2025										
09/06/2025	10892	10805		87	0.80	6	8	73		
10/06/2025	10481	10065		416	3.97	10	25	381		
11/06/2025	10274	10168		106	1.03	29	4	73		
12/06/2025	10118	10071		47	0.46	24	8	15		
13/06/2025	10789	10715		74	0.69	22	31	21		
14/06/2025	10431	10349		82	0.79	48	24	10		
15/06/2025										
16/06/2025	10851	10808		43	0.40	23	13	7		
17/06/2025	9506	9446		60	0.63	35	11	14		
18/06/2025	10137	10073		64	0.63	40	16	8		
19/06/2025	12684	12635		49	0.39	27	8	14		
20/06/2025	11697	11592		105	0.90	53	46	6		
21/06/2025	12405	12340		65	0.52	47	13	5		
22/06/2025	10383	10358		25	0.24	17	3	5		
23/06/2025	11300	11240		60	0.53	60				
24/06/2025	10261	10218		43	0.42	7	18	18		
25/06/2025	10758	10730		28	0.26	19	6	3		
26/06/2025	9518	9497		21	0.22	12	6	3		
27/06/2025	13746	13720		26	0.19	12	9	5		
28/06/2025	11974	11940		34	0.28	24		10		
29/06/2025										
30/06/2025	12690	12672		18	0.14	13	1	4		
TOTAL	281210	279119		2101		785	457	848		
TOTAL %		99.26		0.75		0.28	0.16	0.30		

VALVE INTEGRITY

CHECKED QTY	ACCEPT QTY	HOLD QTY	REJ. QTY	REJ. %	LEAKAGE	90/10	BUBBLE	THIN SPOD	OTHERS	Remarks
10226	10004		222	2.17	74			117	31	
10626	10479		149	1.40	69			54	26	
9907	9459		448	4.52	72			331	45	
12699	11918		781	6.15	54			586	141	
11489	11272		217	1.89	64			75	78	
10290	9979		311	3.02	31			45	235	
10805	10679		126	1.17	55			34	37	
10065	9453		612	6.08	313			257	42	
10168	9882		286	2.81	76			99	111	
10071	9696		329	3.27	102	7		11	201	
10715	10386		329	3.07	204		2	35	88	
10349	9878		471	4.55	164			229	78	
10808	10613		285	2.64	130			35	120	
9446	9179		267	2.83	71			82	114	
10073	9846		227	2.25	120			18	89	
12635	12284		351	2.78	140			16	145	
11592	11114		478	4.12	151			189	138	
12340	12066		274	2.22	46			56	172	
8140	7625		515	6.33	27	2		104	382	
12218	11602		616	5.04	98			197	321	
10458	10236		222	2.12	112			30	80	
10745	10624		121	1.13	43			17	61	
10491	10367		124	1.18	50			9	65	
12720	12609		111	0.87	80			24	7	
11940	11809		131	1.10	113			18		
12888	12790		98	0.76	81			17		
283904	275849	0	8101		2540	9	2	2685	2807	
TOTAL %	97.16	0.00	2.85		0.89	0.00	0.00		0.99	

COMMULATIVE DETAILS IN MAY 25




BALLOON INS. REPORT

STRUCK BALLOON	785
BALLOON BRUST	457
LEAKAGE	848
REJ QTY	2090


VALV INTEGRITY

LEAKAGE	2540
90/10	9
BUBBLE	2
THIN SPOD	2685
OTHERS	2807
REJ QTY	8043


BALLOON REJECTION IN MAY 25



VALVE INTEGRITY REJECTION IN MAY 25






REJ TREND



BALLOON ISPECTION REPORT - P17											VALVE INTEGRITY										
DATE	CHECKED QTY	ACCEPT QTY	HOLD QTY	REJ. QTY	REJ. %	STRUCK BALLOON	BALLOOM BRUST	LEAKAGE	OTHERS	Remarks	CHECKED QTY	ACCEPT QTY	HOLD QTY	REJ. QTY	REJ. %	LEAKAGE	90/10	BUBBLE	THIN SPOD	OTHERS	Remarks
01/07/2025	11448	11387		61	0.53	49	1	11			12387	12234		153	1.24	54			94	5	
02/07/2025	12083	12049		34	0.28	13	9	12			12049	11972		77	0.64	66	1		8	3	
03/07/2025	9565	9498		67	0.70	10	17	40			9498	9362		136	1.43	58	14		52	12	
04/07/2025	11248	11193		55	0.49	27	2	26			11193	11002		191	1.71	91			65	35	
05/07/2025	12619	12578		41	0.32	16	4	21			12578	12501		77	0.61	56	4		8	9	
WEEKLY	56963	56705	0	258	0.45	115	33	110	0	0	57705	57071	0	634	1.10	325	19	0	227	64	0
06/07/2025															#DIV/0!						
07/07/2025	11752	11737		15	0.13	12		3			10644	10578		66	0.62	53			5	8	
08/07/2025	11153	11121		32	0.29	26	2	4			10945	10902		43	0.39	36			5	2	
09/07/2025	11394	11364		30	0.26	19	2	9			10366	10319		47	0.45	28			11	8	
10/07/2025	9652	9611		41	0.42	27	14				9611	9509		102	1.06	64			21	17	
11/07/2025	10694	10661		33	0.31	16	10	7			10661	10592		69	0.65	44			17	8	
12/07/2025	11488	11447		41	0.36	33	6	2			11447	11367		80	0.70	69			9	2	
WEEKLY 1	66133	65941	0	192	0.29	133	34	25	0	0	63674.00	63267	0	407	0.64	294	0	0	68	45	0
13/07/2025	SUNDAY																				
14/07/2025	9998	9971		27	0.27	27					9971	9948		23	0.23	17			4	2	
15/07/2025	10458	10422		36	0.34	36					10422	10371		51	0.49	44	1		3	3	
16/07/2025	10839	10766		73	0.67	69		4			10766	10686		80	0.74	44	7		8	21	
17/07/2025	12473	12456		17	0.14	17					12456	12381		75	0.60	52			4	19	
18/07/2025	11421	11396		25	0.22	22		3			11396	11378		18	0.16	9	1	1	5	2	
19/07/2025	10557	10520		37	0.35	37					10520	10500		20	0.19	16		1	3		
WEEKLY 2	65746	65531	0	215	0.33	208	0	7	0	0	65531	65264	0	267	0.41	182	9	2	27	47	0
20/07/2025	SUNDAY																				
21/07/2025	11070	11041		29	0.26	25	4				11041	11012		29	0.26	29					
22/07/2025	13366	13340		17	0.13	13	4				13340	13260		89	0.67	72			15	2	
23/07/2025	13505	13479		26	0.19	26					13479	13469		10	0.07	6			3	1	
24/07/2025	11893	11869		24	0.20	21	2	1			11869	11860		9	0.08	5			4		
25/07/2025	13520	13488		32	0.24	2					13488	13473		15	0.11	7			8		
26/07/2025	12451	12410		41	0.33	41					12410	12400		10	0.08	4			1	5	
WEEKLY 3	75805	75627	0	169	0.22	128	10	1	0	0	75627	75474	0	162	0.21	123	0	0	31	8	0
27/07/2025	SUNDAY																				
28/07/2025	11082	11068		14	0.13	13	1				11074	11061		13	0.12	10			3		
29/07/2025	13451	13422		29	0.22	29					13416	13369		47	0.35	46			1		
30/07/2025	9594	9571		23	0.24	20	2	1			9571	9532		39	0.41	7			25	7	
31/07/2025	14970	14945		25	0.17	22	3				14945	14939		6	0.04	2			4		
WEEKLY 4	49097	49006	0	91	0.19	84	6	1	0	0	49006	48901	0	105	0.21	65	0	0	33	7	0
TOTAL	313744	312810	0	925	0.29	668	83	144	0		262537	309977	0	1575	0.60	989	28	2	386	171	
PERCENTAGE OF TOTAL	99.70			0.29		0.21	0.03	0.05			PERCENTAGE OF TOTAL	118.07	0.00	0.60		0.38	0.01	0.00	0.15	0.07	

BALLOON INS. REPORT	
STRUCK BALLOON	668
BALLOON BRUST	83
LEAKAGE	144
REJ QTY	895

VALV INTEGRITY	
LEAKAGE	989
90/10	28
BUBBLE	2
THIN SPOD	386
OTHERS	171
REJ QTY	1576





DISPOSAFE HEALTH AND LIFE CARE LIMITED

Plot No.37A, Sector-06, Ballabgarh (Faridabad) Haryana India 121006

Doc. No : F/ASSEM/03:00
Supersedes : 00
Effective Date : 11/11/2024
Page: 1 of 1

DAILY BALLOON & VALVE INTEGRITY REPORT

BALLOON ISPECTION REPORT - P17

VALVE INTEGRITY




BALLOON INSPECTION REPORT - P17											VALVE INTEGRITY										
DATE	CHECKED QTY	ACCEPT QTY	HOLD QTY	REJ. QTY	REJ. %	STRUCK BALLOON	BALLOOM BRUST	LEAKAGE	OTHERS	Remarks	CHECKED QTY	ACCEPT QTY	HOLD QTY	REJ. QTY	REJ. %	LEAKAGE	90/10	BUBBLE	THIN SPOD	OTHERS	Remarks
01/08/2025	11911	11886		25	0.21	24	1				11886	11845		41	0.34	30			6	5	
02/08/2025	11330	11154		10	0.09	4	6				11149	11109		40	0.36	36			4		
WEEKLY 1	23241	23040	0	35	0.15	28	7	0	0	0	23035	22954	0	81	0.35	66	0	0	10	5	0
03/08/2025					#DIV/0!										#DIV/0!						
04/08/2025	11108	11103		5	0.05	5					8103	8080		23	0.28	7			16		
05/08/2025	11104	1105		54	0.49	12	7	35			9136	8717		419	4.59	80			235	104	
06/08/2025	9696	9682		14	0.14	14					13059	13036		23	0.18	10	1	2	2	8	
07/08/2025	10891	10881		10	0.09	10					12418	12396		22	0.18	20			2		
08/08/2025	8078	8049		9	0.11	9					8049	8028		21	0.26	14			7		
09/08/2025	HOLIDAY																				
WEEKLY 2	50877	40820	0	92	0.18	50	7	35	0	0	50765	50257	0	508	1.00	131	1	2	262	112	0
10/08/2025					#DIV/0!										#DIV/0!						
11/08/2025	9570	9558		12	0.13	12					7217	7181		36	0.50	29		6	1		
12/08/2025	7334	7255		79	1.08	50		29			9069	9032		37	0.41	34		1		2	
13/08/2025	8183	8080		49	0.60						8048	8032		16	0.20	15				1	
14/08/2025	8455	8408		47	0.56	46	1				8887	8869		18	0.20	17			1		
15/08/2025	HOLIDAY																				
16/08/2025	HOLIDAY																				
WEEKLY 3	33542	33301	0	187	0.56	108	1	29	0	0	33221	33114	0	107	0.32	95	0	7	2	3	0
17/08/2025	8240	8209		33	0.40	33					5929	5886		23	0.39	21			2		
18/08/2025	7808	7776		32	0.41	32					100561	10037		19	0.02	19					
19/08/2025	13121	13098		23	0.18	22	1				7690	703		67	0.87	14			53		
20/08/2025	12865	12822		38	0.30	38					10848	10704		144	1.33	9			135		
21/08/2025	11422	11393		29	0.25	3	26				10788	10715		73	0.68	9			62	2	
22/08/2025	10571	10544		27	0.26	27					11557	11140		417	3.61	59			339	19	
23/08/2025	13789	13776		13	0.09	10	3				13944	13830		114	0.82	50			46	18	
WEEKLY 4	77816	77618	0	195	0.25	165	30	0	0	0	161317	63015	0	857	0.53	181	0	0	637	39	0
24/08/2025					#DIV/0!										#DIV/0!						
25/08/2025	10091	10078		13	0.13	13					11115	10524		861	7.75	12			833	16	
26/08/2025	11695	11666		29	0.25	21	6	2			13991	13724		267	1.91	25		1	175	66	
27/08/2025	9025	8972		53	0.59	43	10				12422	12162		260	2.09	28			202	30	
28/08/2025	11884	11864		20	0.17		20				11864	11833		31	0.26	5			25	1	
29/08/2025	12461	12426		35	0.28	31	4				12426	12108		318	2.56	21			291	6	
30/08/2025	12947	12922		25	0.19	9	16				11439	11387		52	0.45	3			45	4	
WEEKLY 5	68103	67928	0	175	0.26	117	56	2	0	0	73257	71738	0	1789	2.44	94	0	1	1571	123	0
31/08/2025					#DIV/0!										#DIV/0!						
TOTAL	253579	242707	0	684	0.27	468	101	66	0	0	341595	241078	0	3342	0.98	567	1	10	2482	282	0
TOTAL %		95.71	#DIV/0!	0.27		0.18	0.04	0.03			TOTAL %	70.57	0.00	0.98		0.17	0.00	0.00	0.73	0.08	

COMMULATIVE DETAILS IN MAY 25



BALLOON INS. REPORT	
STRUCK BALLOON	468
BALLOON BRUST	101
LEAKAGE	66
REJ QTY	635

VALV INTEGRITY	
LEAKAGE	567
90/10	1
BUBBLE	10
THIN SPOD	2482
OTHERS	282
REJ QTY	3342





DISPOSAFE HEALTH AND LIFE CARE LIMITED
Plot No.37A, Sector-06, Ballabgarh (Faridabad) Haryana India 121006

Doc. No : F/ASSEM/03:00
Supersedes : 00
Effective Date : 11/11/2024
Page: 1 of 1




DAILY BALLOON & VALVE INTEGRITY REPORT

BALLOON ISPECTION REPORT - P17											VALVE INTEGRITY										
DATE	CHECKED QTY	ACCEPT QTY	HOLD QTY	REJ. QTY	REJ. %	STRUCK BALLOON	BALLOOM BRUST	LEAKAGE	OTHERS	Remarks	CHECKED QTY	ACCEPT QTY	HOLD QTY	REJ. QTY	REJ. %	LEAKAGE	90/10	BUBBLE	THIN SPOD	OTHERS	Remarks
01/09/2025	10967	10930		37	0.34	17	9	11			11477	11397		80	0.70	9			67	4	
02/09/2025	11858	11817		46	0.39	38	5	3			12748	12495		253	1.98	132			83	38	
03/09/2025	11554	11529		25	0.22	17	8				11529	11363		166	1.44	25			95	46	
04/09/2025	14519	14508		11	0.08	10	1				8798	8791		7	0.08				3	4	
05/09/2025	12736	12730		6	0.05	6					13192	13180		12	0.09	2			5	5	
06/09/2025	11786	11782		4	0.03	4					13029	12997		32	0.25	3			21	8	
WEEKLY 1	73420	73296	0	129	0.18	92	23	14	0	0	70773	70223	0	550	0.78	171	0	0	274	105	0
07/09/2025					#DIV/0!										#DIV/0!						
08/09/2025	13650	13642		8	0.06	7	1				14640	14625		15	0.10	1			12	2	
09/09/2025	12705	12675		30	0.24	28	2				11915	11905		10	0.08	6			4		
10/09/2025	13817	13779		38	0.28	37	1				13545	13494		51	0.38	3			41	7	
11/09/2025	10444	10426		18	0.17	14	4				12423	12364		59	0.47	6		1	41	11	
12/09/2025	11328	11296		32	0.28	20	12				11012	10651		361	3.28	8			312	41	
13/09/2025	9586	9529		57	0.59	57					11816	11634		182	1.54	42			55	85	
WEEKLY 2	71530	71347	0	183	0.26	163	20	0	0	0	75351	74673	0	678	0.90	66	0	1	465	146	0
14/09/2025					#DIV/0!										#DIV/0!						
15/09/2025	11571	11502		64	0.55	62	2				10753	10604		149	1.39	15			113	21	
16/09/2025	10972	10941		31	0.28	29	2				11715	11684		31	0.26	6			25		
17/09/2025	10508	10459		49	0.47	44	5				10459	10399		60	0.57	15			37	8	
18/09/2025	16601	16554		47	0.28	47					16554	16481		73	0.44	28			43	2	
19/09/2025	9729	9640		89	0.91	66	23				9640	9332		308	3.20	44			186	78	
20/09/2025	14316	14292		24	0.17	22	2				13443	13371		72	0.54	4			66	2	
WEEKLY 3	73697	73388	0	304	0.41	270	34	0	0	0	72564	71871	0	693	0.96	112	0	0	470	111	0
21/09/2025					#DIV/0!										#DIV/0!						
22/09/2025	10523	10486		37	0.35	28	9				1335	11168		167	12.51	10			144	13	
23/09/2025	14259	14228		31	0.22	27	4				14225	14127		101	0.71	7	3	1	90		
24/09/2025	12251	12238		31	0.25	11	1	1			11453	11290		163	1.42	12			151		
25/09/2025	12536	12500		36	0.29	22	14				13285	13202		83	0.62	2			74	7	
26/09/2025	11079	11030		49	0.44	29	8	12			9564	8271		1293	13.52	12	899		370	12	
27/09/2025	11094	10971		123	1.11	87	21	15			12434	11232		1202	9.67	109	658	1	412	22	
WEEKLY 4	71742	71453	0	307	0.43	204	57	28	0	0	62296	69290	0	3009	4.83	152	1560	2	1241	54	0
28/09/2025					#DIV/0!										#DIV/0!						
29/09/2025	12630	12591		38	0.30	36	3				10021	9883		138	1.38	9			129		
30/09/2025	8894	8876		18	0.20	18					8914	8592		322	3.61	5			254	63	
WEEKLY 5	21524	21467	0	56	0.26	54	3	0	0	0	18935	18475	0	460	2.43	14	0	0	383	63	0
TOTAL	311913	310951	0	979	0.31	783	137	42	0	0	299919	304532	0	5390	1.80	515	1560	3	2833	479	0
TOTAL %	99.69	#REF!	0.31			0.25	0.04	0.01			TOTAL %	101.54	0.00	1.80		0.17	0.52	0.00	0.94	0.16	

COMMULATIVE DETAILS IN MAY 25

BALLOON INS. REPORT
STRUCK BALLOON
BALLOON BRUST
LEAKAGE
REJ QTY

VALV INTEGRITY
LEAKAGE
90/10
BUBBLE
THIN SPOD
OTHERS
REJ QTY





DISPOSAFE HEALTH AND LIFE CARE LIMITED

Plot No.37A, Sector-06, Ballabgarh (Faridabad) Haryana India 121006

Doc. No : F/ASSEM/03:00
Supersedes : 00
Effective Date : 11/11/2024
Page: 1 of 1

DAILY BALLOON & VALVE INTEGRITY REPORT

BALLOON ISPECTION REPORT - P17

DATE	CHECKED QTY	ACCEPT QTY	HOLD QTY	REJ. QTY	REJ. %	STRUCK BALLOON	BALLOOM BRUST	LEAKAGE	OTHERS	Remarks
01/10/2025	9602	9427		175	1.82	55	100	20		
02/10/2025										
03/10/2025	11083	11011		72	0.65	70	2			
04/10/2025	10522	10377		145	1.38	142	1			
WEEKLY 1	31207	30815	0	392	1.26	267	103	20	0	0
05/10/2025					#DIV/0!					
06/10/2025	10913	10724		189	1.73	182	7			
07/10/2025	9571	9230		69	0.72	67	2			
08/10/2025	7944	7841		103	1.30	78	23	2		
09/10/2025	7894	7688		206	2.61	85	93	28		
10/10/2025	7181	7065		116	1.62	114		2		
11/10/2025	7578	7479		99	1.31	86	8	5		
WEEKLY 2	51081	50027	0	782	1.53	612	133	37	0	0
12/10/2025					#DIV/0!					
13/10/2025	8993	8884		109	1.21	97	12			
14/10/2025	7687	7546		141	1.83	111	28	2		
15/10/2025	7324	7272		52	0.71	46	6			
16/10/2025	6783	6749		34	0.50	31	3			
17/10/2025	6974	6909		65	0.93	35	14	16		
18/10/2025	4728	4679		49	1.04	14	35			
WEEKLY 3	42489	42039	0	450	1.06	334	98	18	0	0
19/10/2025	3100	3092		8	0.26	4		4		
20/10/2025										
21/10/2025										
22/10/2025										
23/10/2025										
24/10/2025					#DIV/0!					
25/10/2025					#DIV/0!					
WEEKLY 4	3100	3092	0	8	0.26	4	0	4	0	0
26/10/2025					#DIV/0!					
27/10/2025					#DIV/0!					
28/10/2025					#DIV/0!					
29/10/2025					#DIV/0!					
30/10/2025					#DIV/0!					
31/10/2025					#DIV/0!					
WEEKLY 5	0	0	0	0	#DIV/0!	0	0	0	0	0
TOTAL	127877	125973	0	1632	1.28	1217	334	79	0	0
TOTAL %		98.51	#REF!	1.28		0.95	0.26	0.06		

VALVE INTEGRITY

CHECKED QTY	ACCEPT QTY	HOLD QTY	REJ. QTY	REJ. %	LEAKAGE	90/10	BUBBLE	THIN SPOD	OTHERS	Remarks
12019	11157		62	0.52	6			55	1	
VIJAYA DASHMI										
11019	10857		162	1.47	7	53		102		
10377	10345		32	0.31		2		30		
33415	32359	0	256	0.77	13	55	0	187	1	0
				#DIV/0!						
10724	10677		47	0.44	18	4	2	23		
9230	9217		13	0.14				12	1	
7841	7723		118	1.50	56		1	59	2	
7707	7394		313	4.06	191			122		
7046	6996		50	0.71	18			31	1	
7479	7361		118	1.58				74	44	
50027	49368	0	659	1.32	283	4	3	321	48	0
				#DIV/0!						
8506	8402		104	1.22	2			102		
7924	7812		112	1.41	24			88		
7298	6897		401	5.49	45			356		
6749	6709		40	0.59				40		
6909	6867		42	0.61	9			33		
4679	4630		40	0.85				40		
42065	41317	0	739	1.76	80	0	0	659	0	0
3092	3069		23	0.74		2		21		
HAPPY DIWALI										
				#DIV/0!						
				#DIV/0!						
3092	3069	0	23	0.74	0	2	0	21	0	0
				#DIV/0!						
				#DIV/0!						
				#DIV/0!						
				#DIV/0!						
				#DIV/0!						
				#DIV/0!						
0	0	0	0	#DIV/0!	0	0	0	0	0	0
128599	126113	0	1677	1.30	376	61	3	1188	49	0
TOTAL %	98.07	0.00	1.30		0.29	0.05	0.00	0.92	0.04	



COMMULATIVE DETAILS

BALLOON INS. REPORT

STRUCK BALLOON
BALLOON BRUST
LEAKAGE
REJ QTY

VALV INTEGRITY

LEAKAGE
90/10
BUBBLE
THIN SPOD
OTHERS
REJ QTY





DISPOSAFE HEALTH AND LIFE CARE LIMITED

Plot No.37A, Sector-06, Ballabgarh (Faridabad) Haryana India 121006

Doc. No :
Supersedes :
Effective Date :
Page: 1 of 1

DAILY BALLOON & VALVE INTEGRITY REPORT

BALLOON ISPECTION REPORT - P17

DATE	CHECKED QTY	ACCEPT QTY	HOLD QTY	REJ. QTY	REJ. %	STRUCK BALLOON	BALLOOM BRUST	LEAKAGE	OTHERS	Remarks
01/11/2025	3064	3043		21	0.69	16	5			
WEEKLY 1	3064	3043	0	21	0.69	16	5	0	0	0
02/11/2025					#DIV/0!					
03/11/2025	6681	6670		11	0.16	10	1			
04/11/2025	10796	10751		43	0.40	2				
05/11/2025	8738	8712		26	0.30	26				
06/11/2025	8862	8843		19	0.21	16	1	2		
07/11/2025	11854	11833		21	0.18	19	2			
08/11/2025	7646	7554		92	1.20	61	30	1		
WEEKLY 2	54577	54363	0	212	0.39	134	34	3	0	0
09/11/2025					#DIV/0!					
10/11/2025	6315	6063		252	3.99	195	57			
11/11/2025	5600	5095		500	8.93	7	107	386		
12/11/2025	7442	7438		4	0.05	1	3			
13/11/2025	7475	7466		9	0.12		9			
14/11/2025	8216	7965		251	3.06		15	21	215	
15/11/2025	7661	9637		24	0.31		11	13		
WEEKLY 3	42709	43664	0	1040	2.44	203	202	420	215	0
16/11/2025					#DIV/0!					
17/11/2025	10482	10434		48	0.46		2	9	37	
18/11/2025	10449	10437		12	0.11		2	8		
19/11/2025	11093	11074		19	0.17	5	2	12		
20/11/2025	10233	10226		7	0.07		2	5		
21/11/2025	8584	8566		18	0.21	7	1	10		
22/11/2025	4382	4367		15	0.34	2	9	4		
WEEKLY 4	55223	55104	0	119	0.22	14	18	48	37	0
23/11/2025					#DIV/0!					
24/11/2025	4457	4431		26	0.58	4	2	20		
25/11/2025	3159	3144		15	0.47			15		
26/11/2025	4012	3987		25	0.62	12		3		
27/11/2025	7900	7834		66	0.84	61		5		
28/11/2025	8234	7899		335	4.07	314	5	16		
29/11/2025	9878	9557		21	0.21	17	3	1		
30/11/2025					#DIV/0!					
WEEKLY 5	37640	36852	0	488	1.30	408	10	60	0	0
TOTAL	193213	193026	0	1880	0.97	775	269	531	252	0
TOTAL %	99.90	0.00	0.97			0.40	0.14	0.27		

VALVE INTEGRITY

	CHECKED QTY	ACCEPT QTY	HOLD QTY	REJ. QTY	REJ. %	LEAKAGE	90/10	BUBBLE	THIN SPOD	OTHERS	Remarks
	3043	3024		19	0.62				19		
	3043	3024	0	19	0.62	0	0	0	19	0	0
				#DIV/0!							
	6670	6650		20	0.30	2			18		
	10751	10709		42	0.39	2			38	2	
	8712	8702		10	0.11				9	1	
	8843	8820		23	0.26				23		
	11833	11800		33	0.28	1			32		
	7554	7463		91	1.20	10			76	5	
	54363	54144	0	219	0.40	15	0	0	196	8	0
				#DIV/0!							
	6063	5885		178	2.94	126			50	2	
	5095	4836		311	6.10	244			67		
	7438	7360		78	1.05	15			63		
	7466	7338		128	1.71	71			57		
	7965	7709		256	3.21	161			95		
	9637	9533		104	1.08	29			75		
	43664	42661	0	1055	2.42	646	0	0	407	2	0
				#DIV/0!							
	10434	10337		97	0.93	30			67		
	10437	10390		47	0.45	26			21		
	11074	11027		47	0.42	10			37		
	10233	10208		25	0.24	8			17		
	8566	8503		63	0.74	11			52		
	4367	4262		105	2.40	50			55		
	55111	54727	0	384	0.70	135	0	0	249	0	0
				#DIV/0!							
	4431	4396		35	0.79	12			23		
	3144	3117		27	0.86	18			9		
	3987	3941		46	1.15	11			35		
	7834	7782		52	0.66	10			42		
	6643	6603		40	0.60	9			31		
	10814	10704		110	1.02	80			30		
				#DIV/0!							
	36853	36543	0	310	0.84	140	0	0	170	0	0
	193034	191099	0	1987	1.03	936	0	0	1041	10	0
	TOTAL %	99.00	0.00	1.03		0.48	0.00	0.00	0.54	0.01	




COMMULATIVE DETAILS

BALLOON INS. REPORT

STRUCK BALLOON
BALLOON BRUST
LEAKAGE
REJ QTY

VALV INTEGRITY

LEAKAGE
90/10
BUBBLE
THIN SPOD
OTHERS
REJ QTY





DISPOSAFE HEALTH AND LIFE CARE LIMITED

Plot No.37A, Sector-06, Ballabgarh (Faridabad) Haryana India 121006

Doc. No :
Supersedes :
Effective Date :
Page: 1 of 1

DAILY BALLOON & VALVE INTEGRITY REPORT

BALLOON ISPECTION REPORT - P17

DATE	CHECKED QTY	ACCEPT QTY	HOLD QTY	REJ. QTY	REJ. %	STRUCK BALLOON	BALLOOM BRUST	LEAKAGE	OTHERS	Remarks
01/12/2025	5518	5274		244	4.42	3	70	171		
02/12/2025	4946	4842		104	2.10	4	30	70		
03/12/2025	0	0	0	0	#DIV/0!	0	0	0	0	0
04/12/2025	5882	5805		77	1.31	59	10	8		
05/12/2025	6903	6887		16	0.23	10	6			
06/12/2025	7635	7608		27	0.35	26	1			
WEEKLY 1	30884	30416	0	468	1.52	102	117	249	0	0
07/12/2025					#DIV/0!					
08/12/2025	10297	10292		5	0.05	5				
09/12/2025	9562	9533		29	0.30	22	7			
10/12/2025	8019	8012		7	0.09	2	5			
11/12/2025	6369	6300		69	1.08	64	5			
12/12/2025	10454	10440		14	0.13	14				
13/12/2025	11183	11161		22	0.20	21	1			
WEEKLY 2	55884	55738	0	146	0.26	128	18	0	0	0
14/12/2025					#DIV/0!					
15/12/2025	9323	9304		19	0.20	14	3	2		
16/12/2025	10355	10334		21	0.20	20	1			
17/12/2025	12450	12418		32	0.26	30	2			
18/12/2025	12431	12419		12	0.10	1	2			
19/12/2025	12453	12413		40	0.32	35	3	2		
20/12/2025	14969	14950		19	0.13	18	1			
WEEKLY 3	71981	71838	0	143	0.20	118	12	4	0	0
21/12/2025					#DIV/0!					
22/12/2025	12172	12165		7	0.06	6	1			
23/12/2025	15112	15107		5	0.03	4	1			
24/12/2025	14559	14529		30	0.21	27	3			
25/12/2025	14878	14841		37	0.25	32	5			
26/12/2025	16825	16772		53	0.32	52	1			
27/12/2025	10873	10857		16	0.15	15	1			
WEEKLY 4	84419	84271	0	148	0.18	136	12	0	0	0
28/12/2025					#DIV/0!					
29/12/2025	16602	16565		37	0.22	28	9			
30/12/2025	11537	11521		16	0.14	9	7			
31/12/2025	13441	13411		30	0.22	26	4			
WEEKLY 5	41580	41497	0	83	0.20	63	20	0	0	0
TOTAL	284748	283760	0	988	0.35	547	179	253	0	0
TOTAL %	99.65	0.00	0.35		0.19	0.06	0.09			

VALVE INTEGRITY

CHECKED QTY	ACCEPT QTY	HOLD QTY	REJ. QTY	REJ. %	LEAKAGE	90/10	BUBBLE	THIN SPOD	OTHERS	Remarks
5276	5178		99	1.88	66			33		
6488	6358		130	2.00	24			106		
0	0	0	0	#DIV/0!	0	0	0	0	0	0
5805	5745		60	1.03	11			49		
6899	6812		87	1.26	6			81		
7608	7573		35	0.46	2			33		
32076	31666	0	411	1.28	109	0	0	302	0	0
				#DIV/0!						
10292	10240		52	0.51	10			42		
9549	9488		61	0.64	6			55		
8012	7822		190	2.37	4			186		
6122	4797	700	625	10.21				625		
10618	10049	300	269	2.53	14			255		
11161	11026	52	83	0.74	17			66		
55754	53422	1052	1280	2.30	51	0	0	1229	0	0
				#DIV/0!						
8547	8318	95	193	2.26	19			115		
11844	11680	26	138	1.17	15			123		
8419	8032	273	114	1.35	8			106		
13447	12232		1215	9.04	1	919		295		
10466	8715		175	1.67	3	1705		43		
15537	13575		1962	12.63	2	1960				
68260	62552	394	3797	5.56	48	4584	0	682	0	0
				#DIV/0!						
10000	8407		1593	15.93	3	1538		52		
15615	13123		2492	15.96	17	2458		17		
15790	11969		3821	24.20	28	3595		198		
18731	11798		1933	10.32	3	1901		19		
18108	16531		1577	8.71	11	1537		29		
13065	11830		1235	9.45	32	1183		20		
91309	73658	0	12651	13.86	94	12212	0	335	0	0
				#DIV/0!						
12940	10353		2587	19.99	23	2553		11		
15410	10383		5027	32.62	33	4994				
8516	5327		3189	37.45	5	3184				
36866	26063	0	10803	29.30	61	10731	0	11	0	0
284265	247361	1446	28942	10.18	363	27527	0	2559	0	0
TOTAL %	87.02	0.51	10.18		0.13	9.68	0.00	0.90	0.00	




COMMULATIVE DETAILS

BALLOON INS. REPORT

STRUCK BALLOON
BALLOON BRUST
LEAKAGE
REJ QTY

VALV INTEGRITY

LEAKAGE
90/10
BUBBLE
THIN SPOD
OTHERS
REJ QTY





DISPOSAFE HEALTH AND LIFE CARE LIMITED

Plot No.37A, Sector-06, Ballabgarh (Faridabad) Haryana India 121006

DAILY BALLOON & VALVE INTEGRITY REPORT

BALLOON INSPECTION REPORT - P17

DATE	CHECKED QTY	ACCEPT QTY	HOLD QTY	REJ. QTY	REJ. %	STRUCK BALLOON	BALLOOM BRUST	LEAKAGE	OTHERS	Remarks
01/01/2026	7690	7659		31	0.40	20	11			
02/01/2026	10000	10001		8	0.08	6	2			
03/01/2026	8304	8293		11	0.13	10	1			
WEEKLY 1	25994	25953	0	50	0.19	36	14	0	0	0
04/01/2026					#DIV/0!					
05/01/2026	10585	10535		50	0.47	32	16	2		
06/01/2026	13037	13032		5	0.04	5				
07/01/2026	10869	10804		65	0.60	16	49			
08/01/2026	10114	10097		17	0.17	17				
09/01/2026	14014	13976		38	0.27	29	9			
10/01/2026	12916	12889		27	0.21	16	10	1		
WEEKLY 2	71535	71333	0	202	0.28	115	84	3	0	0
11/01/2026					#DIV/0!					
12/01/2026	15439	15391		48	0.31	40	7	1		
13/01/2026	13217	13207		10	0.08	7	3			
14/01/2026	12899	12866		33	0.26	26	6			
15/01/2026	12155	12107		48	0.39	40	8			
16/01/2026	15338	15310		28	0.18	13	7	8		
17/01/2026	11494	11469		25	0.22	14	11			
WEEKLY 3	80542	80350	0	192	0.24	140	42	9	0	0
18/01/2026					#DIV/0!					
19/01/2026	12313	12282		31	0.25	11	20			
20/01/2026	11317	11289		28	0.25	15	12	1		
21/01/2026	14909	14836		73	0.49	41	32			
22/01/2026	14777	14721		56	0.38	32	24			
23/01/2026	10267	10236		31	0.30	17	13	1		
24/01/2026	13468	13428		40	0.30	29	11			
WEEKLY 4	77051	76792	0	259	0.34	145	112	2	0	0
25/01/2026					#DIV/0!					
26/01/2026					#DIV/0!					
27/01/2026	17410	17351		59	0.34	44	15			
28/01/2026					#DIV/0!					
29/01/2026					#DIV/0!					
30/01/2026					#DIV/0!					
31/01/2026					#DIV/0!					
WEEKLY 5	17410	17351	0	59	0.34	44	15	0	0	0
TOTAL	272532	271779	0	762	0.28	480	267	14	0	0
TOTAL %	99.72	0.00	0.28		0.18	0.10	0.01			

VALVE INTEGRITY




CHECKED QTY	ACCEPT QTY	HOLD QTY	REJ. QTY	REJ. %	LEAKAGE	90/10	BUBBLE	THIN SPOD	OTHERS	Remarks
8000	6686		1314	16.43	8	1300		6		
9659	7942	992	725	7.51	1	724				
10010	8311	1578	121	1.21		119		2		
27669	22939	2570	2160	7.81	9	2143	0	8	0	0
				#DIV/0!						
9618	4806	3267	1545	16.06	4	1541				
10993	10993	1992	584	5.31	10	574				
11685	9120	1774	312	2.67	11	301			479	
15156	12105	1700	826	5.45	6	816	4	525		
16954	14813	1547	418	2.47	2	415		1	176	
12892	12187	290	309	2.40	9	300			106	
77298	64024	10570	3994	5.17	42	3947	4	526	761	0
				#DIV/0!						
11300	10497	305	332	2.94	7	325			166	
14709	14472		131	0.89	4	122		5	106	
12802	11768	720	306	2.39	14	263			9	
13185	12186	301	298	2.26	11	217			70	
11062	9656	1196	210	1.90	2	207			1	
10262	10222		40	0.39	8	32				
73320	68801	2522	1317	1.80	46	1166	0	5	352	0
				#DIV/0!						
11724	11691		33	0.28	2	25		5	1	
12358	11337	802	219	1.77	4	200		5	10	
15391	14687	356	348	2.26	13	331		3	1	
12586	12497		89	0.71	13	54		20	2	
11557	11021	354	182	1.57	21	147	1	13		
15327	15158		169	1.10	16	102	27	23	1	
78943	76391	1512	1040	1.32	69	859	28	69	15	0
				#DIV/0!						
				#DIV/0!						
15596	15435		161	1.03	11	103		47		
				#DIV/0!						
				#DIV/0!						
				#DIV/0!						
				#DIV/0!						
15596	15435	0	161	1.03	11	103	0	47	0	0
272826	247590	17174	8672	3.18	177	8218	32	655	1128	0
TOTAL %	90.75	6.29	3.18		0.06	3.01	0.01	0.24	0.41	

COMMULATIVE DETAILS



BALLOON INS. REPORT	
STRUCK BALLOON	
BALLOON BRUST	
LEAKAGE	
REJ QTY	

VALV INTEGRITY	
LEAKAGE	
90/10	
BUBBLE	
THIN SPOD	
OTHERS	
REJ QTY	





DISPOSAFE HEALTH AND LIFE CARE LIMITED
Plot No.37A, Sector-06, Ballabgarh (Faridabad) Haryana India 121006

Doc. No : F/ASSEM/03:00
Supersedes : None
Effective Date: 12/7/2024
Page: 1 of 1

DAILY BALLOON & VALVE INTEGRITY REPORT YEAR 2025-26

BALLOON ISPECTION REPORT - P17											VALVE INTEGRITY										
MONTH	CHECKED QTY	ACCEPT QTY	HOLD QTY	REJ. QTY	REJ. %	STRUCK BALLOON	BALLOOM BRUST	LEAKAGE	OTHERS	Remarks	CHECKED QTY	ACCEPT QTY	HOLD QTY	REJ. QTY	REJ. %	LEAKAGE	90/10	BUBBLE	THIN SPOD	OTHERS	Remarks
APRIL 25	216095	214182	0	1910	0.88	1064	141	705	0		214182	196555	9836	4181	1.95	1566	305	0	2417	18	
MAY 25	248704	246523	0	2181	0.88	1394	429	360	0		249524	247146	2619	2288	0.92	1108	18	0	1085	86	
JUNE 25	281210	279119	0	2101	0.75	785	457	848	0		283904	275849	0	8101	2.85	2540	9	2	2685	2807	
JULY 25	313744	312810	0	925	0.29	668	83	144	0		262537	309977	0	1575	0.60	989	28	2	386	171	
AUGUST 25	253579	242707	0	684	0.27	468	101	66	0		341595	241078	0	3342	0.98	567	1	10	2482	282	
SEPTEMBER 25	311913	310951	0	979	0.31	783	137	42	0		299919	304532	0	5390	1.80	515	1560	3	2833	479	
OCTOBER 25	127877	125973	0	1632	1.28	1217	334	79	0		128599	126113	0	1677	1.30	376	61	3	1188	49	
NOVEMBER 25					#DIV/0!										#DIV/0!						
DECEMBER 25					#DIV/0!										#DIV/0!						
JANUARY 26					#DIV/0!										#DIV/0!						
FEBRUARY 26					#DIV/0!										#DIV/0!						
MARCH 26					#DIV/0!										#DIV/0!						
TOTAL	1753122	1732265		10412		6379	1682	2244	0		1780260	1701250	12455	26554		7661	1982	20	13076	3892	
TOTAL %		98.81		0.59		0.36	0.10	0.13				95.56	0.70	1.49		0.43	0.11	0.00		0.22	

COMMULATIVE DETAILS YEAR 2025-26

BALLOON INS. REPORT

STRUCK BALL	6379
BALLOON BRU	1682
LEAKAGE	2244
REJ QTY	10305

VALV INTEGRITY

LEAKAGE	7661
90/10	1982
BUBBLE	20
THIN SPOD	13076
OTHERS	3892
REJ QTY	26631

