



The Hong Kong University of Science and Technology

Health, Safety and Environment Office

Project Safety Review Form

Reference Number: 17048

Date Received from Department: 2017-07-19

Target Review Completion Date: 2017-08-02

Department: CBE

Anticipated Location (Room numbers): 7109

Extension: N/A

Name of Researcher: Siu Yip KUNG

Name of Supervisor: Prof Zhengtang LUO

Project Title: Design of Adsorption Devices for Blood Purification

Area	Comment	Department Response
Occupational Hygiene	Name Signature Date Pak IP	
	-Conduct chemical preparation work in fume hood. -Follow established procedures strictly when adding KMnO ₄ to sulphuric acid. -Wear proper gloves and chemical resistant goggles. -Store incompatible chemicals separately. -Ensure proper balancing of centrifuge rotor. -Follow operating procedures of various analytical equipment. -Caution other users in the lab for any unattended experiment.	
Safety Engineering	--	
Environmental Protection	--	
Health Physics	--	
Peer Review	--	