



The Hong Kong University of Science and Technology

Health, Safety and Environment Office

Project Safety Review Form

Reference Number: 17047

Date Received from Department: 2017-07-14

Target Review Completion Date: 2017-07-28

Department: CBE

Anticipated Location (Room numbers): 6121,6104,6107,6251

Extension: N/A

Name of Researcher: Yin Yung CHAN

Name of Supervisor: Prof Fei SUN

Project Title: Genetically encoded

Area	Comment	Department Response
Occupational Hygiene	Name Signature Date Pak IP	
	-Ensure proper balancing of centrifuge rotor. -Adopt aseptic technique when handling bacterial culture. -Follow established procedures when operating confocal microscope. -Autoclave all biological materials before disposal.	
Safety Engineering	--	
Environmental Protection	--	
Health Physics	--	
Peer Review	--	