



# The Hong Kong University of Science and Technology

## Health, Safety and Environment Office

### Project Safety Review Form

**Reference Number:** 17026

**Date Received from Department:** 2017-07-19

**Target Review Completion Date:** 2017-07-28

**Department:** CBE

**Anticipated Location (Room numbers):** 7102

**Extension:** N/A

**Name of Researcher:** Tiehuai LI

**Name of Supervisor:** Minhua SHAO

**Project Title:** A New Class of Electrocatalysts for Hydrogen Evolution Reaction

Area	Comment	Department Response
Occupational Hygiene	Name                      Signature                      Date  Pak IP	
	-Ensure proper balancing of centrifuge rotor. -Conduct sample preparation work in fume hood. -Avoid skin contact with hot surface. -Insulate all electrical contact.	
Safety Engineering	--	
Environmental Protection	--	
Health Physics	--	
Peer Review	--	