

## PREMIER MAGNESIA - GILES CHEMICAL

**COMPANY FORM** 

Title: USP Analytical Balance Calibration Check Log Stephen

Ballew

Page : 1 of 2

Revision : 00 Effective Date : 2/21/12

Form Number: QA-LAB-43-F1



Balance Modle: XS105DU Serial Number: B139Z9Z316

Reference Weight: Certificate No.: Nominal Value Warning Limit +/- Control Limit +/-

 No. 1
 MT0021716
 100 g
 0.0002 g
 0.0003 g

 No. 2
 MT0021716
 5 g
 0.00001 g
 0.00003 g

Date	Result Ref. Weight No. 1	Result Ref. Weight No. 2	Within Limits (Yes/No) Warning	Within Limits (Pass/Fail) Control	Name	Initials	Retest (yes/no)



## PREMIER MAGNESIA - GILES CHEMICAL

**COMPANY FORM** 

Title: USP Analytical Balance
Calibration Check Log
Author: Stephen
Ballew

Page : 2 of 2

Form Number: QA-LAB-43-F1

Revision: 00

Effective Date : 2/21/12



Revision Number	Revision Date	Effective Date	Revision Author	Quality Approval	Production Approval	Revision Description
00	02/21/2012	02/21/2012	Stephen Ballew	Deborah Durbin	Jason Bumgarner	New Document