	GILES CHEMICAL ~ PREMIER MAGNESIA	
	Company Form	
	Title: USP Analytical Balance Calibration Check Log	Number: L12-PR-100-F012
	Owner: Hunter Douglas	Revision: 02
	Effective Date: 07/20/2015	Page: 1 of 1

Balance Model: XS105DU **Serial Number: B139292316**

Weight Set No.	Serial No.	Nominal Value	Warning Limit +/–	Control Limit +/–
No. 1	6432410037	100g/5 g	0.0002 g/0.00001 g	0.0003 g/0.00003 g
No. 2	B417533411			

Date	Weight Set No.	Result 100g Weight	Result 5g Weight	Within Limits (Yes/No) Warning	Within Limits (Pass/Fail) Control	Performed By	Reviewed By Initials	Retest (yes/no)

Page ___ of ___

Controlled Document

Only those quality documents viewed through the Giles Chemical electronic Documentation System are officially controlled. All other copies, whether viewed through another computer program or a printed version, are not controlled and, therefore, the Quality Unit at Giles assumes no responsibility for accuracy of the document